

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005757.D
 Acq On : 08 Jun 2021 01:14
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Quant Time: Jun 08 03:23:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	126009	20.000	ng	0.00
21) Naphthalene-d8	10.775	136	502662	20.000	ng	-0.01
39) Acenaphthene-d10	14.592	164	335328	20.000	ng	-0.01
64) Phenanthrene-d10	17.339	188	692454	20.000	ng	0.00
76) Chrysene-d12	21.404	240	786606	20.000	ng	#-0.01
86) Perylene-d12	23.827	264	929113	20.000	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	635798	78.446	ng	0.00
7) Phenol-d6	7.134	99	799855	72.392	ng	0.00
23) Nitrobenzene-d5	9.140	82	755011	83.533	ng	0.00
42) 2,4,6-Tribromophenol	16.080	330	447572	104.089	ng	-0.01
45) 2-Fluorobiphenyl	13.228	172	1737521	68.574	ng	0.00
79) Terphenyl-d14	19.880	244	2955429	70.502	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.399	88	133898	36.097	ng	92
3) Pyridine	3.810	79	357306	38.911	ng	# 93
4) n-Nitrosodimethylamine	3.716	42	160017	39.127	ng	# 24
6) Aniline	7.293	93	445552	34.496	ng	94
8) 2-Chlorophenol	7.534	128	349241	38.483	ng	96
9) Benzaldehyde	7.104	77	179032	37.425	ng	92
10) Phenol	7.163	94	399970	35.455	ng	96
11) bis(2-Chloroethyl)ether	7.393	93	296371	34.027	ng	97
12) 1,3-Dichlorobenzene	7.857	146	378941	36.786	ng	99
13) 1,4-Dichlorobenzene	8.004	146	386032	36.976	ng	97
14) 1,2-Dichlorobenzene	8.328	146	374393	37.497	ng	98
15) Benzyl Alcohol	8.210	79	311297	39.278	ng	# 82
16) 2,2'-oxybis(1-Chloropr...	8.504	45	287627	28.848	ng	89
17) 2-Methylphenol	8.416	107	281799	38.373	ng	97
18) Hexachloroethane	9.051	117	140745	37.990	ng	92
19) n-Nitroso-di-n-propyla...	8.787	70	252642	35.947	ng	98
20) 3+4-Methylphenols	8.745	107	386537	39.509	ng	95
22) Acetophenone	8.798	105	480538	35.840	ng	# 98
24) Nitrobenzene	9.175	77	391710	39.016	ng	92
25) Isophorone	9.710	82	715985	35.680	ng	97
26) 2-Nitrophenol	9.892	139	197731	51.208	ng	# 80
27) 2,4-Dimethylphenol	9.951	122	292104	37.925	ng	99
28) bis(2-Chloroethoxy)met...	10.187	93	390032	36.115	ng	98
29) 2,4-Dichlorophenol	10.428	162	364576	44.010	ng	99
30) 1,2,4-Trichlorobenzene	10.639	180	391620	39.913	ng	98
31) Naphthalene	10.828	128	1108783	37.041	ng	97
32) Benzoic acid	10.092	122	106406	27.460	ng	94
33) 4-Chloroaniline	10.934	127	475655	38.162	ng	# 91
34) Hexachlorobutadiene	11.116	225	255491	41.690	ng	98
35) Caprolactam	11.739	113	115564	50.941	ng	# 73
36) 4-Chloro-3-methylphenol	12.057	107	380544	43.642	ng	95
37) 2-Methylnaphthalene	12.433	142	743883	35.072	ng	# 88
38) 1-Methylnaphthalene	12.651	142	714454	35.503	ng	# 97
40) 1,2,4,5-Tetrachloroben...	12.798	216	400948	35.844	ng	97
41) Hexachlorocyclopentadiene	12.775	237	179460	28.817	ng	97
43) 2,4,6-Trichlorophenol	13.039	196	292336	43.088	ng	94

mohammad

6/8/2021 3:37:26 PM

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005757.D
 Acq On : 08 Jun 2021 01:14
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Quant Time: Jun 08 03:23:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

mohammad

6/8/2021 3:37:26 PM

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.110	196	311823	42.501	ng	# 92
46) 1,1'-Biphenyl	13.433	154	930357	34.169	ng	98
47) 2-Chloronaphthalene	13.475	162	770416	34.809	ng	99
48) 2-Nitroaniline	13.680	65	228197	43.861	ng	# 69
49) Acenaphthylene	14.316	152	1280787	37.126	ng	99
50) Dimethylphthalate	14.057	163	1008102	37.796	ng	99
51) 2,6-Dinitrotoluene	14.175	165	239867	41.485	ng	# 65
52) Acenaphthene	14.657	154	768048	34.249	ng	96
53) 3-Nitroaniline	14.498	138	253304	43.815	ng	80
54) 2,4-Dinitrophenol	14.710	184	29643	17.678	ng	98
55) Dibenzofuran	14.992	168	1267943	36.887	ng	99
56) 4-Nitrophenol	14.810	139	204388	42.903	ng	# 59
57) 2,4-Dinitrotoluene	14.957	165	339768	46.745	ng	# 67
58) Fluorene	15.639	166	978389	35.601	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.216	232	301362m	47.475	ng	
60) Diethylphthalate	15.410	149	1049599	39.221	ng	98
61) 4-Chlorophenyl-phenyle...	15.627	204	492583	35.851	ng	96
62) 4-Nitroaniline	15.657	138	274551	45.184	ng	# 53
63) Azobenzene	15.922	77	993773	35.461	ng	95
65) 4,6-Dinitro-2-methylph...	15.722	198	89297	30.457	ng	91
66) n-Nitrosodiphenylamine	15.845	169	913887	38.732	ng	# 92
67) 4-Bromophenyl-phenylether	16.522	248	320440	38.414	ng	95
68) Hexachlorobenzene	16.645	284	383296	38.590	ng	# 94
69) Atrazine	16.798	200	289076	42.870	ng	96
70) Pentachlorophenol	16.986	266	213260	42.810	ng	93
71) Phenanthrene	17.380	178	1515127	35.121	ng	100
72) Anthracene	17.469	178	1509634	35.441	ng	99
73) Carbazole	17.739	167	1490355	35.550	ng	99
74) Di-n-butylphthalate	18.280	149	1670712	37.215	ng	# 92
75) Fluoranthene	19.333	202	1922565	37.125	ng	98
77) Benzidine	19.504	184	686330	45.943	ng	99
78) Pyrene	19.686	202	1974726	35.578	ng	99
80) Butylbenzylphthalate	20.545	149	842119	38.412	ng	# 91
81) Benzo(a)anthracene	21.386	228	2104198	36.450	ng	99
82) 3,3'-Dichlorobenzidine	21.315	252	776455	41.641	ng	# 97
83) Chrysene	21.439	228	2031599	36.343	ng	99
84) Bis(2-ethylhexyl)phtha...	21.304	149	1174262	37.451	ng	# 96
85) Di-n-octyl phthalate	22.227	149	2197960	41.709	ng	98
87) Indeno(1,2,3-cd)pyrene	26.392	276	2663475	33.694	ng	# 94
88) Benzo(b)fluoranthene	23.086	252	2379856	35.303	ng	# 97
89) Benzo(k)fluoranthene	23.139	252	2281048	35.709	ng	# 98
90) Benzo(a)pyrene	23.727	252	2228013	36.330	ng	# 97
91) Dibenzo(a,h)anthracene	26.409	278	2310709	33.683	ng	# 95
92) Benzo(g,h,i)perylene	27.180	276	2196092	33.190	ng	# 93


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
Data File : BP005757.D
Acq On : 08 Jun 2021 01:14
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampled :
SSTDCCC040

Manual Integrations
APPROVED

Quant Time: Jun 08 03:23:13 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 07 12:13:53 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.969 min Scan# 830
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	152	Resp:	126009
Ion	Ratio	Lower	Upper
152	100		
150	153.5	119.8	179.8
115	58.4	49.0	73.4




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED



#2
1,4-Dioxane
Concen: 36.097 ng
RT: 3.399 min Scan# 53
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14


Tgt Ion:	88	Resp:	133898
Ion	Ratio	Lower	Upper
88	100		
58	60.7	42.5	63.7
43	26.2	19.8	29.8







#3
Pyridine
Concen: 38.911 ng
RT: 3.810 min Scan# 123
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion: 79 Resp: 357306
Ion Ratio Lower Upper
79 100
52 53.6 40.3 60.5
51 30.6 20.0 30.0#

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#4
n-Nitrosodimethylamine
Concen: 39.127 ng
RT: 3.716 min Scan# 107
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion: 42 Resp: 160017
Ion Ratio Lower Upper
42 100
74 130.2 122.6 183.8
44 5.6 146.9 220.3#








#5
2-Fluorophenol
Concen: 78.446 ng
RT: 5.528 min Scan# 415
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 112 Resp: 635798
Ion Ratio Lower Upper
112 100
64 51.2 33.6 50.4#
63 29.0 18.6 27.8#


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#6
Aniline
Concen: 34.496 ng
RT: 7.293 min Scan# 715
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 93 Resp: 445552
Ion Ratio Lower Upper
93 100
66 38.3 27.0 40.4
65 18.7 14.4 21.6







#7
Phenol-d6
Concen: 72.392 ng
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 799855



Ion	Ratio	Lower	Upper
99	100		
42	17.0	12.9	19.3
71	38.6	26.7	40.1

Manual Integrations
APPROVED



mohammad




6/8/2021 3:37:26 PM




#8
2-Chlorophenol
Concen: 38.483 ng
RT: 7.534 min Scan# 756
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 128 Resp: 349241


Ion	Ratio	Lower	Upper
128	100		
130	32.0	14.1	54.1
64	39.4	17.2	57.2







#9
Benzaldehyde
Concen: 37.425 ng
RT: 7.104 min Scan# 683
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 179032
Ion Ratio Lower Upper
77 100
105 98.5 89.4 129.4
106 94.9 79.9 119.9

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#10
Phenol
Concen: 35.455 ng
RT: 7.163 min Scan# 693
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion: 94 Resp: 399970
Ion Ratio Lower Upper
94 100
65 28.6 8.0 48.0
66 40.7 17.3 57.3








#11
bis(2-Chloroethyl)ether
Concen: 34.027 ng
RT: 7.393 min Scan# 732
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 93	Resp: 296371
Ion Ratio	Lower Upper
93 100	
63 66.9	49.6 89.6
95 32.3	14.4 54.4


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#12
1,3-Dichlorobenzene
Concen: 36.786 ng
RT: 7.857 min Scan# 811
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:146	Resp: 378941		
Ion Ratio	Lower	Upper	
146 100			
148 64.2	51.4	77.0	
75 31.8	23.9	35.9	







#13
1,4-Dichlorobenzene
Concen: 36.976 ng
RT: 8.004 min Scan# 836
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:146 Resp: 386032
Ion Ratio Lower Upper
146 100
148 64.5 52.2 78.4
111 39.4 34.2 51.4



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#14
1,2-Dichlorobenzene
Concen: 37.497 ng
RT: 8.328 min Scan# 891
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:146 Resp: 374393
Ion Ratio Lower Upper
146 100
148 65.8 51.5 77.3
111 40.9 34.0 51.0







#15
Benzyl Alcohol
Concen: 39.278 ng
RT: 8.210 min Scan# 871
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 311297
Ion Ratio Lower Upper
79 100
108 74.3 64.0 96.0
77 64.3 72.7 109.1#

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#16
2,2'-oxybis(1-Chloropropane)
Concen: 28.848 ng
RT: 8.504 min Scan# 921
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14


Tgt Ion: 45 Resp: 287627
Ion Ratio Lower Upper
45 100
77 20.7 7.0 47.0
79 16.5 1.0 41.0







#17
2-Methylphenol
Concen: 38.373 ng
RT: 8.416 min Scan# 906
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	107	Resp:	281799
Ion Ratio	Lower	Upper	
107	100		
108	110.8	85.8	128.8
77	49.6	41.4	62.0
79	47.9	38.6	57.8

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#18
Hexachloroethane
Concen: 37.990 ng
RT: 9.051 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:	117	Resp:	140745
Ion Ratio	Lower	Upper	
117	100		
119	97.8	80.1	120.1
201	109.4	98.8	148.2







#19
n-Nitroso-di-n-propylamine
Concen: 35.947 ng
RT: 8.787 min Scan# 969
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 70 Resp: 252642

Ion	Ratio	Lower	Upper
70	100		
42	49.9	39.7	59.5
101	10.6	7.4	11.0
130	24.4	17.2	25.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED



mohammad




6/8/2021 3:37:26 PM




#20
3+4-Methylphenols
Concen: 39.509 ng
RT: 8.745 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 107 Resp: 386537

Ion	Ratio	Lower	Upper
107	100		
108	90.0	69.5	109.5
77	34.8	21.5	61.5
79	30.7	16.8	56.8







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.775 min Scan# 1307
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	136	Resp:	502662
Ion Ratio	Lower	Upper	
136	100		
137	10.7	8.9	13.3
54	5.7	5.0	7.4
68	5.1	3.7	5.5



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#22
Acetophenone
Concen: 35.840 ng
RT: 8.798 min Scan# 971
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	105	Resp:	480538
Ion Ratio	Lower	Upper	
105	100		
71	3.1	0.9	1.3
51	22.6	18.2	27.2
120	22.9	17.0	25.6







#23
Nitrobenzene-d5
Concen: 83.533 ng
RT: 9.140 min Scan# 1029
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 82 Resp: 755011

Ion	Ratio	Lower	Upper
82	100		
128	44.7	30.5	45.7
54	41.7	33.9	50.9



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED



mohammad




6/8/2021 3:37:26 PM




#24
Nitrobenzene
Concen: 39.016 ng
RT: 9.175 min Scan# 1035
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 77 Resp: 391710

Ion	Ratio	Lower	Upper
77	100		
123	45.6	31.7	47.5
65	13.1	11.5	17.3







#25
Isophorone
Concen: 35.680 ng
RT: 9.710 min Scan# 1126
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	82	Resp:	715985
Ion	Ratio	Lower	Upper
82	100		
95	8.4	7.5	11.3
138	19.5	14.6	22.0



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#26
2-Nitrophenol
Concen: 51.208 ng
RT: 9.892 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	139	Resp:	197731
Ion	Ratio	Lower	Upper
139	100		
109	35.2	40.7	61.1#
65	45.7	47.8	71.6#







#27
2,4-Dimethylphenol
Concen: 37.925 ng
RT: 9.951 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	122	Resp:	292104
Ion	Ratio	Lower	Upper
122	100		
107	116.5	93.8	140.8
121	56.4	45.6	68.4



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#28
bis(2-Chloroethoxy)methane
Concen: 36.115 ng
RT: 10.187 min Scan# 1207
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	93	Resp:	390032
Ion	Ratio	Lower	Upper
93	100		
95	32.2	24.9	37.3
123	12.1	9.5	14.3







#29
2,4-Dichlorophenol
Concen: 44.010 ng
RT: 10.428 min Scan# 1248
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:162 Resp: 364576
Ion Ratio Lower Upper
162 100
164 64.3 44.1 84.1
98 33.8 12.6 52.6



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#30
1,2,4-Trichlorobenzene
Concen: 39.913 ng
RT: 10.639 min Scan# 1284
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:180 Resp: 391620
Ion Ratio Lower Upper
180 100
182 94.9 77.2 115.8
145 28.7 21.7 32.5







#31
Naphthalene
Concen: 37.041 ng
RT: 10.828 min Scan# 1316
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:128 Resp: 1108783
Ion Ratio Lower Upper
128 100
129 11.0 8.0 12.0
127 12.9 11.1 16.7



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#32
Benzoic acid
Concen: 27.460 ng
RT: 10.092 min Scan# 1191
Delta R.T. -0.018 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:122 Resp: 106406
Ion Ratio Lower Upper
122 100
105 118.9 105.1 145.1
77 95.8 69.6 109.6







#33
4-Chloroaniline
Concen: 38.162 ng
RT: 10.934 min Scan# 1334
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	127	Resp:	475655
Ion Ratio	Lower	Upper	
127	100		
129	32.4	25.5	38.3
65	29.5	30.0	45.0#
92	19.3	20.2	30.2#



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#34
Hexachlorobutadiene
Concen: 41.690 ng
RT: 11.116 min Scan# 1365
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14


Tgt Ion:	225	Resp:	255491
Ion Ratio	Lower	Upper	
225	100		
223	61.8	51.1	76.7
227	64.3	50.2	75.2







#35
Caprolactam
Concen: 50.941 ng
RT: 11.739 min Scan# 1471
Delta R.T. 0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:113 Resp: 115564
Ion Ratio Lower Upper
113 100
55 126.8 72.0 112.0#
56 101.0 65.1 105.1

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#36
4-Chloro-3-methylphenol
Concen: 43.642 ng
RT: 12.057 min Scan# 1525
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:107 Resp: 380544
Ion Ratio Lower Upper
107 100
144 26.9 20.1 30.1
142 82.6 62.2 93.4







#37
2-Methylnaphthalene
Concen: 35.072 ng
RT: 12.433 min Scan# 1589
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:142 Resp: 743883
Ion Ratio Lower Upper
142 100
141 87.9 74.7 112.1
115 38.6 43.8 65.8#



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#38
1-Methylnaphthalene
Concen: 35.503 ng
RT: 12.651 min Scan# 1626
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:142 Resp: 714454
Ion Ratio Lower Upper
142 100
141 93.8 77.0 115.6
116 4.2 5.0 7.6#







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.592 min Scan# 1956
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:164 Resp: 335328
Ion Ratio Lower Upper
164 100
162 101.3 73.5 110.3
160 44.4 40.6 61.0



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#40
1,2,4,5-Tetrachlorobenzene
Concen: 35.844 ng
RT: 12.798 min Scan# 1651
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:216 Resp: 400948
Ion Ratio Lower Upper
216 100
214 78.0 60.6 91.0
179 20.8 15.3 22.9
108 17.8 13.8 20.6







#41
Hexachlorocyclopentadiene
Concen: 28.817 ng
RT: 12.775 min Scan# 1647
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	237	Resp:	179460
Ion	Ratio	Lower	Upper
237	100		
235	62.8	40.9	80.9
272	13.4	0.0	34.6



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#42
2,4,6-Tribromophenol
Concen: 104.089 ng
RT: 16.080 min Scan# 2209
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	330	Resp:	447572
Ion	Ratio	Lower	Upper
330	100		
332	97.4	76.3	114.5
141	30.4	22.8	34.2







#43
2,4,6-Trichlorophenol
Concen: 43.088 ng
RT: 13.039 min Scan# 1692
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:196	Resp:	292336	
Ion	Ratio	Lower	Upper
196	100		
198	95.6	81.4	122.2
200	31.4	28.2	42.2



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#44
2,4,5-Trichlorophenol
Concen: 42.501 ng
RT: 13.110 min Scan# 1704
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt	Ion:196	Resp:	311823
Ion	Ratio	Lower	Upper
196	100		
198	97.7	71.0	106.4
97	48.7	38.5	57.7
132	26.5	15.6	23.4#







#45
2-Fluorobiphenyl
Concen: 68.574 ng
RT: 13.228 min Scan# 1724
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:172 Resp: 1737521
Ion Ratio Lower Upper
172 100
171 34.8 28.1 42.1
170 23.3 19.6 29.4



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#46
1,1'-Biphenyl
Concen: 34.169 ng
RT: 13.433 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:154 Resp: 930357
Ion Ratio Lower Upper
154 100
153 40.1 18.9 58.9
76 11.1 0.0 32.1







#47
2-Chloronaphthalene
Concen: 34.809 ng
RT: 13.475 min Scan# 1766
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	162	Resp:	770416
Ion	Ratio	Lower	Upper
162	100		
127	38.5	31.4	47.0
164	32.2	26.6	39.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#48
2-Nitroaniline
Concen: 43.861 ng
RT: 13.680 min Scan# 1801
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	65	Resp:	228197
Ion	Ratio	Lower	Upper
65	100		
92	74.8	48.8	73.2#
138	123.2	69.5	104.3#







#49
Acenaphthylene
Concen: 37.126 ng
RT: 14.316 min Scan# 1909
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:152 Resp: 1280787
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.6
153 13.2 10.5 15.7



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#50
Dimethylphthalate
Concen: 37.796 ng
RT: 14.057 min Scan# 1865
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:163 Resp: 1008102
Ion Ratio Lower Upper
163 100
194 4.4 3.9 5.9
164 10.2 8.1 12.1









#51
2,6-Dinitrotoluene
Concen: 41.485 ng
RT: 14.175 min Scan# 1885
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:165 Resp: 239867
Ion Ratio Lower Upper
165 100
63 53.0 67.4 101.0#
89 58.6 73.8 110.6#




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED



#52
Acenaphthene
Concen: 34.249 ng
RT: 14.657 min Scan# 1967
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14


Tgt Ion:154 Resp: 768048
Ion Ratio Lower Upper
154 100
153 111.6 90.9 136.3
152 53.5 48.5 72.7







#53
3-Nitroaniline
Concen: 43.815 ng
RT: 14.498 min Scan# 1940
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:138 Resp: 253304
Ion Ratio Lower Upper
138 100
108 10.5 10.2 15.4
92 110.1 108.1 162.1

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#54
2,4-Dinitrophenol
Concen: 17.678 ng
RT: 14.710 min Scan# 1976
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Tgt Ion:184 Resp: 29643
Ion Ratio Lower Upper
184 100
63 67.6 51.2 76.8
154 69.5 55.8 83.8







#55
Dibenzenofuran
Concen: 36.887 ng
RT: 14.992 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:168 Resp: 1267943
Ion Ratio Lower Upper
168 100
139 39.5 31.4 47.0
169 13.6 11.4 17.0

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#56
4-Nitrophenol
Concen: 42.903 ng
RT: 14.810 min Scan# 1993
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Tgt Ion:139 Resp: 204388
Ion Ratio Lower Upper
139 100
109 84.0 122.8 162.8#
65 89.7 110.0 150.0#






#57
2,4-Dinitrotoluene
Concen: 46.745 ng
RT: 14.957 min Scan# 2018
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:165 Resp: 339768
Ion Ratio Lower Upper
165 100
63 42.0 56.7 85.1#
89 72.3 85.8 128.6#
182 2.7 3.0 4.6#





Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

mohammad



6/8/2021 3:37:26 PM




#58
Fluorene
Concen: 35.601 ng
RT: 15.639 min Scan# 2134
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:166 Resp: 978389
Ion Ratio Lower Upper
166 100
165 98.4 78.8 118.2
167 13.3 10.9 16.3







#59
2,3,4,6-Tetrachlorophenol
Concen: 47.475 ng m
RT: 15.216 min Scan# 2062
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	232	Resp:	301362
Ion Ratio	Lower	Upper	
232	100		
131	38.7	26.5	39.7
130	2.6	2.1	3.1
166	23.8	19.4	29.0



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#60
Diethylphthalate
Concen: 39.221 ng
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	149	Resp:	1049599
Ion Ratio	Lower	Upper	
149	100		
177	22.5	19.2	28.8
150	12.8	9.9	14.9







#61
4-Chlorophenyl-phenylether
Concen: 35.851 ng
RT: 15.627 min Scan# 2132
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	204	Resp:	492583
Ion Ratio	Lower	Upper	
204	100		
206	33.0	26.0	39.0
141	59.9	44.2	66.2



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#62
4-Nitroaniline
Concen: 45.184 ng
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	138	Resp:	274551
Ion Ratio	Lower	Upper	
138	100		
92	51.5	50.9	90.9
108	95.3	154.7	194.7#







#63
Azobenzene
Concen: 35.461 ng
RT: 15.922 min Scan# 2182
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	77	Resp:	993773
Ion	Ratio	Lower	Upper
77	100		
182	27.2	5.5	45.5
105	20.0	0.0	36.5
51	28.1	10.3	50.3



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.339 min Scan# 2423
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	188	Resp:	692454
Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.0	7.6
80	7.6	5.9	8.9







#65
4,6-Dinitro-2-methylphenol
Concen: 30.457 ng
RT: 15.722 min Scan# 2148
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	198	Resp:	89297
Ion Ratio	Lower	Upper	
198	100		
51	33.5	10.3	50.3
105	39.9	13.6	53.6



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#66
n-Nitrosodiphenylamine
Concen: 38.732 ng
RT: 15.845 min Scan# 2169
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	169	Resp:	913887
Ion Ratio	Lower	Upper	
169	100		
168	66.1	55.2	82.8
167	35.9	36.6	54.8#







#67
4-Bromophenyl-phenylether
Concen: 38.414 ng
RT: 16.522 min Scan# 2284
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	248	Resp:	320440
Ion Ratio	Lower	Upper	
248	100		
250	96.8	77.5	116.3
141	72.5	50.1	75.1



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#68
Hexachlorobenzene
Concen: 38.590 ng
RT: 16.645 min Scan# 2305
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	284	Resp:	383296
Ion Ratio	Lower	Upper	
284	100		
142	31.9	20.8	31.2#
249	30.5	23.7	35.5







#69
Atrazine
Concen: 42.870 ng
RT: 16.798 min Scan# 2331
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:200 Resp: 289076
Ion Ratio Lower Upper
200 100
173 25.2 0.7 40.7
215 49.8 29.0 69.0



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#70
Pentachlorophenol
Concen: 42.810 ng
RT: 16.986 min Scan# 2363
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:266 Resp: 213260
Ion Ratio Lower Upper
266 100
268 63.7 49.4 74.0
264 59.0 53.7 80.5







#71
Phenanthrene
Concen: 35.121 ng
RT: 17.380 min Scan# 2430
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:178 Resp: 1515127
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.4 12.3 18.5



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#72
Anthracene
Concen: 35.441 ng
RT: 17.469 min Scan# 2445
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14


Tgt Ion:178 Resp: 1509634
Ion Ratio Lower Upper
178 100
176 18.8 15.2 22.8
179 15.5 12.6 19.0







#73
 Carbazole
 Concen: 35.550 ng
 RT: 17.739 min Scan# 2491
 Delta R.T. -0.006 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040




Tgt Ion:167 Resp: 1490355
 Ion Ratio Lower Upper
 167 100
 166 20.9 17.4 26.0
 139 13.1 10.4 15.6

Manual Integrations
 APPROVED




mohammad



6/8/2021 3:37:26 PM




#74
 Di-n-butylphthalate
 Concen: 37.215 ng
 RT: 18.280 min Scan# 2583
 Delta R.T. -0.012 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14




Tgt Ion:149 Resp: 1670712
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.6 11.4
 104 5.5 8.9 13.3#







#75
Fluoranthene
Concen: 37.125 ng
RT: 19.333 min Scan# 2762
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:202 Resp: 1922565
Ion Ratio Lower Upper
202 100
101 8.3 0.0 26.1
203 17.7 0.0 37.2

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.404 min Scan# 3114
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:240 Resp: 786606
Ion Ratio Lower Upper
240 100
120 8.2 4.2 6.2#
236 26.1 20.2 30.2







#77
Benzidine
Concen: 45.943 ng
RT: 19.504 min Scan# 2791
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:184 Resp: 686330
Ion Ratio Lower Upper
184 100
185 14.0 11.1 16.7
183 11.8 9.8 14.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#78
Pyrene
Concen: 35.578 ng
RT: 19.686 min Scan# 2822
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14


Tgt Ion:202 Resp: 1974726
Ion Ratio Lower Upper
202 100
200 20.9 16.9 25.3
203 17.6 13.8 20.6







#79
Terphenyl-d14
Concen: 70.502 ng
RT: 19.880 min Scan# 2855
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:244 Resp: 2955429
Ion Ratio Lower Upper
244 100
212 7.6 5.8 8.6
122 8.1 5.2 7.8#

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#80
Butylbenzylphthalate
Concen: 38.412 ng
RT: 20.545 min Scan# 2968
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Tgt Ion:149 Resp: 842119
Ion Ratio Lower Upper
149 100
91 69.0 60.2 90.4
206 22.2 23.9 35.9#






#81
Benzo(a)anthracene
Concen: 36.450 ng
RT: 21.386 min Scan# 3111
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:228 Resp: 2104198
Ion Ratio Lower Upper
228 100
226 27.4 21.2 31.8
229 19.8 15.6 23.4




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

mohammad



6/8/2021 3:37:26 PM




#82
3,3'-Dichlorobenzidine
Concen: 41.641 ng
RT: 21.315 min Scan# 3099
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:252 Resp: 776455
Ion Ratio Lower Upper
252 100
254 64.3 49.9 74.9
126 9.3 5.5 8.3#









#83
Chrysene
Concen: 36.343 ng
RT: 21.439 min Scan# 3120
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:228 Resp: 2031599
Ion Ratio Lower Upper
228 100
226 30.3 24.8 37.2
229 19.8 15.4 23.0




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED



#84
Bis(2-ethylhexyl)phthalate
Concen: 37.451 ng
RT: 21.304 min Scan# 3097
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14


Tgt Ion:149 Resp: 1174262
Ion Ratio Lower Upper
149 100
167 27.5 23.8 35.8
279 5.4 5.7 8.5#







#85
Di-n-octyl phthalate
Concen: 41.709 ng
RT: 22.227 min Scan# 3254
Delta R.T. -0.018 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:149 Resp: 2197960
Ion Ratio Lower Upper
149 100
167 1.6 1.5 2.3
43 7.1 5.1 7.7

Manual Integrations
APPROVED




mohammad



6/8/2021 3:37:26 PM




#86
Perylene-d12
Concen: 20.000 ng
RT: 23.827 min Scan# 3526
Delta R.T. -0.024 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:264 Resp: 929113
Ion Ratio Lower Upper
264 100
260 24.3 20.6 30.8
265 21.4 19.0 28.6







#87
 Indeno(1,2,3-cd)pyrene
 Concen: 33.694 ng
 RT: 26.392 min Scan# 3962
 Delta R.T. -0.036 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14

Tgt Ion: 276 Resp: 2663475
 Ion Ratio Lower Upper
 276 100
 138 18.4 8.4 12.6#
 277 25.6 20.6 31.0



Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED






mohammad


6/8/2021 3:37:26 PM



#88
 Benzo(b)fluoranthene
 Concen: 35.303 ng
 RT: 23.086 min Scan# 3400
 Delta R.T. -0.024 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14

Tgt Ion: 252 Resp: 2379856
 Ion Ratio Lower Upper
 252 100
 253 21.9 18.2 27.2
 125 7.4 3.4 5.0#








#89
Benzo(k)fluoranthene
Concen: 35.709 ng
RT: 23.139 min Scan# 3409
Delta R.T. -0.018 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:252 Resp: 2281048
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 7.2 3.8 5.6#


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED






mohammad


6/8/2021 3:37:26 PM



#90
Benzo(a)pyrene
Concen: 36.330 ng
RT: 23.727 min Scan# 3509
Delta R.T. -0.018 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14


Tgt Ion:252 Resp: 2228013
Ion Ratio Lower Upper
252 100
253 21.9 17.1 25.7
125 7.9 3.8 5.6#







#91
Dibenzo(a,h)anthracene
Concen: 33.683 ng
RT: 26.409 min Scan# 3965
Delta R.T. -0.030 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:278 Resp: 2310709
Ion Ratio Lower Upper
278 100
139 11.9 5.0 7.6#
279 24.2 18.6 28.0

Manual Integrations
APPROVED




mohammad

6/8/2021 3:37:26 PM



#92
Benzo(g,h,i)perylene
Concen: 33.190 ng
RT: 27.180 min Scan# 4096
Delta R.T. -0.036 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14



Tgt Ion:276 Resp: 2196092
Ion Ratio Lower Upper
276 100
277 23.5 19.8 29.8
138 15.9 7.7 11.5#

