

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060724\
 Data File : BP020554.D
 Acq On : 07 Jun 2024 10:16
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 07 11:52:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	39.861	0.3	97	0.00
3	Pyridine	40.000	38.419	4.0	91	0.00
4	n-Nitrosodimethylamine	40.000	34.954	12.6	83	0.00
5 S	2-Fluorophenol	80.000	81.260	-1.6	96	0.00
6	Aniline	40.000	35.596	11.0	83	-0.01
7 S	Phenol-d6	80.000	75.341	5.8	89	0.00
8	2-Chlorophenol	40.000	39.433	1.4	93	0.00
9	Benzaldehyde	40.000	34.283	14.3	93	0.00
10 C	Phenol	40.000	37.345	6.6	88	0.00
11	bis(2-Chloroethyl)ether	40.000	36.702	8.2	87	-0.01
12	1,3-Dichlorobenzene	40.000	39.759	0.6	96	0.00
13 C	1,4-Dichlorobenzene	40.000	39.486	1.3	94	0.00
14	1,2-Dichlorobenzene	40.000	39.035	2.4	93	0.00
15	Benzyl Alcohol	40.000	35.751	10.6	83	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	32.056	19.9	77	-0.01
17	2-Methylphenol	40.000	37.370	6.6	87	0.00
18	Hexachloroethane	40.000	39.133	2.2	93	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	33.926	15.2	79	-0.01
20	3+4-Methylphenols	40.000	36.938	7.7	86	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	85	0.00
22	Acetophenone	40.000	39.489	1.3	84	-0.01
23 S	Nitrobenzene-d5	80.000	81.682	-2.1	85	-0.01
24	Nitrobenzene	40.000	39.962	0.1	85	-0.01
25	Isophorone	40.000	36.771	8.1	77	0.00
26 C	2-Nitrophenol	40.000	47.075	-17.7	105	-0.01
27	2,4-Dimethylphenol	40.000	39.228	1.9	81	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.055	7.4	79	0.00
29 C	2,4-Dichlorophenol	40.000	42.250	-5.6	87	0.00
30	1,2,4-Trichlorobenzene	40.000	41.312	-3.3	88	-0.01
31	Naphthalene	40.000	40.020	-0.1	86	0.00
32	Benzoic acid	40.000	44.064	-10.2	103	0.02
33	4-Chloroaniline	40.000	36.621	8.4	77	-0.01
34 C	Hexachlorobutadiene	40.000	43.616	-9.0	94	-0.01
35	Caprolactam	40.000	35.208	12.0	72	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.718	5.7	79	-0.01
37	2-Methylnaphthalene	40.000	37.319	6.7	80	-0.02
38	1-Methylnaphthalene	40.000	37.231	6.9	81	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	76	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.062	-10.2	85	-0.01
41 P	Hexachlorocyclopentadiene	40.000	50.863	-27.2#	94	-0.02
42 S	2,4,6-Tribromophenol	80.000	93.119	-16.4	84	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.144	-15.4	85	0.00
44	2,4,5-Trichlorophenol	40.000	44.371	-10.9	81	0.00
45 S	2-Fluorobiphenyl	80.000	83.591	-4.5	82	0.00
46	1,1'-Biphenyl	40.000	40.887	-2.2	78	0.00
47	2-Chloronaphthalene	40.000	40.911	-2.3	78	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060724\
 Data File : BP020554.D
 Acq On : 07 Jun 2024 10:16
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 07 11:52:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.640	-4.1	75	0.00
49	Acenaphthylene	40.000	40.546	-1.4	78	0.00
50	Dimethylphthalate	40.000	37.248	6.9	72	0.00
51	2,6-Dinitrotoluene	40.000	40.849	-2.1	76	-0.01
52 C	Acenaphthene	40.000	39.179	2.1	77	0.00
53	3-Nitroaniline	40.000	38.524	3.7	71	0.00
54 P	2,4-Dinitrophenol	40.000	49.265	-23.2	107	0.01
55	Dibenzofuran	40.000	38.626	3.4	75	0.00
56 P	4-Nitrophenol	40.000	38.412	4.0	71	0.01
57	2,4-Dinitrotoluene	40.000	40.711	-1.8	74	0.00
58	Fluorene	40.000	37.853	5.4	74	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	42.899	-7.2	78	0.00
60	Diethylphthalate	40.000	35.191	12.0	68	0.00
61	4-Chlorophenyl-phenylether	40.000	38.457	3.9	75	0.00
62	4-Nitroaniline	40.000	36.134	9.7	69	0.00
63	Azobenzene	40.000	34.932	12.7	68	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	68	0.00
65	4,6-Dinitro-2-methylphenol	40.000	49.691	-24.2	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.656	-4.1	72	0.00
67	4-Bromophenyl-phenylether	40.000	43.987	-10.0	76	0.00
68	Hexachlorobenzene	40.000	43.271	-8.2	75	0.00
69	Atrazine	40.000	35.961	10.1	61	-0.01
70 C	Pentachlorophenol	40.000	45.749	-14.4	77	0.01
71	Phenanthrene	40.000	39.727	0.7	68	0.00
72	Anthracene	40.000	40.226	-0.6	68	-0.01
73	Carbazole	40.000	38.509	3.7	66	0.00
74	Di-n-butylphthalate	40.000	36.655	8.4	60	0.00
75 C	Fluoranthene	40.000	38.200	4.5	65	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	72	-0.01
77	Benzydine	40.000	24.527	38.7#	37	0.00
78	Pyrene	40.000	37.205	7.0	65	-0.01
79 S	Terphenyl-d14	80.000	75.010	6.2	64	-0.02
80	Butylbenzylphthalate	40.000	34.813	13.0	62	0.00
81	Benzo(a)anthracene	40.000	39.910	0.2	72	-0.02
82	3,3'-Dichlorobenzidine	40.000	46.893	-17.2	80	-0.01
83	Chrysene	40.000	39.979	0.1	72	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	34.918	12.7	58	-0.02
85 c	Di-n-octyl phthalate	40.000	45.193	-13.0	81	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.853	-14.6	116	-0.01
88	Benzo(b)fluoranthene	40.000	36.079	9.8	95	0.00
89	Benzo(k)fluoranthene	40.000	35.404	11.5	94	0.00
90 C	Benzo(a)pyrene	40.000	39.603	1.0	103	0.02
91	Dibenzo(a,h)anthracene	40.000	45.294	-13.2	115	0.01
92	Benzo(g,h,i)perylene	40.000	46.035	-15.1	117	0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060724\
Data File : BP020554.D
Acq On : 07 Jun 2024 10:16
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 07 11:52:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0