

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005780.D
 Acq On : 08 Jun 2021 18:31
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Jun 09 01:21:55 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 01:17:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	74329	20.00	ng	0.00
21) Naphthalene-d8	10.763	136	278183	20.00	ng	0.00
39) Acenaphthene-d10	14.581	164	177760	20.00	ng	0.00
64) Phenanthrene-d10	17.322	188	382592	20.00	ng	0.00
76) Chrysene-d12	21.392	240	434411	20.00	ng	0.00
86) Perylene-d12	23.810	264	503027	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	93789	19.62	ng	0.00
7) Phenol-d6	7.116	99	119584	18.35	ng	0.00
23) Nitrobenzene-d5	9.122	82	109341	21.86	ng	0.00
42) 2,4,6-Tribromophenol	16.069	330	59531	26.12	ng	0.00
45) 2-Fluorobiphenyl	13.210	172	271049	20.18	ng	0.00
79) Terphenyl-d14	19.869	244	474351	20.49	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.387	88	21481	9.82	ng	97
3) Pyridine	3.817	79	46659	8.61	ng	98
4) n-Nitrosodimethylamine	3.711	42	22976	9.52	ng	# 97
6) Aniline	7.281	93	66112	8.68	ng	98
8) 2-Chlorophenol	7.522	128	53528	10.00	ng	98
9) Benzaldehyde	7.099	77	26886	9.53	ng	97
10) Phenol	7.146	94	61339	9.22	ng	98
11) bis(2-Chloroethyl)ether	7.381	93	45910	8.94	ng	97
12) 1,3-Dichlorobenzene	7.846	146	60173	9.90	ng	99
13) 1,4-Dichlorobenzene	7.993	146	60912	9.89	ng	100
14) 1,2-Dichlorobenzene	8.310	146	59272	10.06	ng	98
15) Benzyl Alcohol	8.199	79	44604	9.54	ng	92
16) 2,2'-oxybis(1-Chloropr...	8.493	45	41882	7.12	ng	98
17) 2-Methylphenol	8.405	107	41041	9.47	ng	97
18) Hexachloroethane	9.040	117	21828	9.99	ng	91
19) n-Nitroso-di-n-propyla...	8.763	70	37294	9.00	ng	96
20) 3+4-Methylphenols	8.728	107	55155	9.56	ng	92
22) Acetophenone	8.787	105	71207	9.60	ng	# 98
24) Nitrobenzene	9.163	77	57903	10.42	ng	98
25) Isophorone	9.693	82	101451	9.14	ng	99
26) 2-Nitrophenol	9.875	139	25683	14.39	ng	94
27) 2,4-Dimethylphenol	9.934	122	40384	9.47	ng	97
28) bis(2-Chloroethoxy)met...	10.175	93	54463	9.11	ng	98
29) 2,4-Dichlorophenol	10.416	162	48779	10.64	ng	98
30) 1,2,4-Trichlorobenzene	10.628	180	56579	10.42	ng	98
31) Naphthalene	10.816	128	158308	9.56	ng	99
32) Benzoic acid	10.034	122	16887	9.10	ng	93
33) 4-Chloroaniline	10.928	127	63296	9.18	ng	98
34) Hexachlorobutadiene	11.099	225	37858	11.16	ng	98
35) Caprolactam	11.704	113	13264	10.56	ng	90
36) 4-Chloro-3-methylphenol	12.046	107	50522	10.47	ng	99
37) 2-Methylnaphthalene	12.422	142	113791	9.69	ng	97
38) 1-Methylnaphthalene	12.640	142	108110	9.71	ng	97
40) 1,2,4,5-Tetrachloroben...	12.787	216	60596	10.22	ng	99
41) Hexachlorocyclopentadiene	12.763	237	19209	5.82	ng	96
43) 2,4,6-Trichlorophenol	13.022	196	40722	11.32	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005780.D
 Acq On : 08 Jun 2021 18:31
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Jun 09 01:21:55 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 01:17:24 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.098	196	43765	11.25	ng	98
46) 1,1'-Biphenyl	13.422	154	139462	9.66	ng	100
47) 2-Chloronaphthalene	13.463	162	115084	9.81	ng	97
48) 2-Nitroaniline	13.663	65	28163	12.32	ng	99
49) Acenaphthylene	14.304	152	175069	9.57	ng	98
50) Dimethylphthalate	14.040	163	145473	10.29	ng	100
51) 2,6-Dinitrotoluene	14.157	165	32048	12.13	ng	96
52) Acenaphthene	14.645	154	107574	9.05	ng	99
53) 3-Nitroaniline	14.487	138	30673	11.78	ng	94
54) 2,4-Dinitrophenol	14.698	184	12910	12.33	ng	98
55) Dibenzofuran	14.981	168	181708	9.97	ng	99
56) 4-Nitrophenol	14.792	139	23782	11.10	ng	97
57) 2,4-Dinitrotoluene	14.945	165	42435	13.43	ng	96
58) Fluorene	15.628	166	141221	9.69	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.204	232	39290	11.68	ng	94
60) Diethylphthalate	15.398	149	143290	10.10	ng	98
61) 4-Chlorophenyl-phenyle...	15.622	204	74399	10.21	ng	97
62) 4-Nitroaniline	15.645	138	32049	11.73	ng	94
63) Azobenzene	15.910	77	129749	8.73	ng	97
65) 4,6-Dinitro-2-methylph...	15.704	198	22278	12.22	ng	97
66) n-Nitrosodiphenylamine	15.834	169	125591	9.63	ng	98
67) 4-Bromophenyl-phenylether	16.516	248	47678	10.34	ng	98
68) Hexachlorobenzene	16.634	284	59158	10.78	ng	95
69) Atrazine	16.781	200	39159	10.51	ng	98
70) Pentachlorophenol	16.981	266	28157	10.23	ng	99
71) Phenanthrene	17.369	178	229648	9.63	ng	98
72) Anthracene	17.457	178	225129	9.57	ng	100
73) Carbazole	17.728	167	216025	9.33	ng	99
74) Di-n-butylphthalate	18.275	149	242345	9.77	ng	99
75) Fluoranthene	19.327	202	281719	9.85	ng	99
77) Benzidine	19.504	184	80377	9.74	ng	98
78) Pyrene	19.675	202	293865	9.59	ng	99
80) Butylbenzylphthalate	20.539	149	114562	10.45	ng	99
81) Benzo(a)anthracene	21.374	228	316649	9.93	ng	99
82) 3,3'-Dichlorobenzidine	21.304	252	111311	10.81	ng	97
83) Chrysene	21.427	228	306499	9.93	ng	99
84) Bis(2-ethylhexyl)phtha...	21.298	149	175567	10.14	ng	99
85) Di-n-octyl phthalate	22.221	149	307691	10.57	ng	99
87) Indeno(1,2,3-cd)pyrene	26.351	276	432265	10.10	ng	99
88) Benzo(b)fluoranthene	23.068	252	341452	9.36	ng	99
89) Benzo(k)fluoranthene	23.121	252	334635	9.68	ng	99
90) Benzo(a)pyrene	23.704	252	320710	9.66	ng	99
91) Dibenzo(a,h)anthracene	26.368	278	370971	9.99	ng	99
92) Benzo(g,h,i)perylene	27.139	276	362445	10.12	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005780.D
 Acq On : 08 Jun 2021 18:31
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010

Quant Time: Jun 09 01:21:55 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 01:17:24 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.957 min Scan# 828
Delta R.T. -0.001 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:152 Resp: 74329
Ion Ratio Lower Upper
152 100
150 155.7 125.6 188.4
115 61.1 48.5 72.7



#2
1,4-Dioxane
Concen: 9.82 ng
RT: 3.387 min Scan# 51
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 88 Resp: 21481
Ion Ratio Lower Upper
88 100
58 61.0 46.7 70.1
43 27.0 20.3 30.5









#3
Pyridine
Concen: 8.61 ng
RT: 3.817 min Scan# 124
Delta R.T. 0.012 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 79 Resp: 46659
Ion Ratio Lower Upper
79 100
52 52.4 41.7 62.5
51 28.2 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 9.52 ng
RT: 3.711 min Scan# 106
Delta R.T. 0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 42 Resp: 22976
Ion Ratio Lower Upper
42 100
74 135.1 105.4 158.0
44 7.9 4.2 6.4#









#5
2-Fluorophenol
Concen: 19.62 ng
RT: 5.516 min Scan# 413
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:112 Resp: 93789
Ion Ratio Lower Upper
112 100
64 54.2 41.4 62.2
63 31.0 23.8 35.8



#6
Aniline
Concen: 8.68 ng
RT: 7.281 min Scan# 713
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 93 Resp: 66112
Ion Ratio Lower Upper
93 100
66 35.5 29.7 44.5
65 18.5 15.3 22.9









#7
Phenol-d6
Concen: 18.35 ng
RT: 7.116 min Scan# 685
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 99 Resp: 119584
Ion Ratio Lower Upper
99 100
42 17.6 13.4 20.0
71 38.1 30.6 46.0



#8
2-Chlorophenol
Concen: 10.00 ng
RT: 7.522 min Scan# 754
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 128 Resp: 53528
Ion Ratio Lower Upper
128 100
130 32.5 12.4 52.4
64 38.3 20.3 60.3









#9
Benzaldehyde
Concen: 9.53 ng
RT: 7.099 min Scan# 682
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 77 Resp: 26886
Ion Ratio Lower Upper
77 100
105 100.6 76.7 116.7
106 97.7 76.2 116.2



#10
Phenol
Concen: 9.22 ng
RT: 7.146 min Scan# 690
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 94 Resp: 61339
Ion Ratio Lower Upper
94 100
65 27.8 8.8 48.8
66 42.2 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 8.94 ng
RT: 7.381 min Scan# 730
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	93	Resp:	45910
Ion	Ratio	Lower	Upper
93	100		
63	69.1	46.2	86.2
95	31.7	12.2	52.2



#12
1,3-Dichlorobenzene
Concen: 9.90 ng
RT: 7.846 min Scan# 809
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	146	Resp:	60173
Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.5	77.3
75	33.8	25.7	38.5









#13
1,4-Dichlorobenzene
Concen: 9.89 ng
RT: 7.993 min Scan# 834
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:146 Resp: 60912
Ion Ratio Lower Upper
146 100
148 63.7 51.3 76.9
111 41.4 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 10.06 ng
RT: 8.310 min Scan# 888
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:146 Resp: 59272
Ion Ratio Lower Upper
146 100
148 65.1 50.3 75.5
111 42.1 33.7 50.5









#15
Benzyl Alcohol
Concen: 9.54 ng
RT: 8.199 min Scan# 869
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	79	Resp:	44604
Ion	Ratio	Lower	Upper
79	100		
108	71.6	62.5	93.7
77	60.0	52.9	79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 7.12 ng
RT: 8.493 min Scan# 919
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	45	Resp:	41882
Ion	Ratio	Lower	Upper
45	100		
77	21.5	2.0	42.0
79	17.7	0.0	36.7









#17
2-Methylphenol
Concen: 9.47 ng
RT: 8.405 min Scan# 904
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	107	Resp:	41041
Ion	Ratio	Lower	Upper
107	100		
108	113.4	87.7	131.5
77	48.4	40.2	60.2
79	50.1	39.0	58.4



#18
Hexachloroethane
Concen: 9.99 ng
RT: 9.040 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	117	Resp:	21828
Ion	Ratio	Lower	Upper
117	100		
119	101.7	76.2	114.2
201	115.7	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 9.00 ng
RT: 8.763 min Scan# 965
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 70	Resp:	37294
Ion	Ratio	Lower Upper
70	100	
42	47.0	40.8 61.2
101	10.7	8.8 13.2
130	22.2	18.9 28.3



#20
3+4-Methylphenols
Concen: 9.56 ng
RT: 8.728 min Scan# 959
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 107	Resp:	55155
Ion	Ratio	Lower Upper
107	100	
108	81.6	71.0 111.0
77	34.2	16.6 56.6
79	29.3	12.7 52.7









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.763 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	136	Resp:	278183
Ion Ratio	Lower	Upper	
136	100		
137	11.2	8.6	12.8
54	5.2	4.6	6.8
68	4.7	3.6	5.4



#22
Acetophenone
Concen: 9.60 ng
RT: 8.787 min Scan# 969
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	105	Resp:	71207
Ion Ratio	Lower	Upper	
105	100		
71	1.5	1.9	2.9#
51	21.2	18.4	27.6
120	21.9	17.9	26.9









#23
Nitrobenzene-d5
Concen: 21.86 ng
RT: 9.122 min Scan# 1026
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 82 Resp: 109341
Ion Ratio Lower Upper
82 100
128 45.1 35.7 53.5
54 42.1 35.0 52.4



#24
Nitrobenzene
Concen: 10.42 ng
RT: 9.163 min Scan# 1033
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 77 Resp: 57903
Ion Ratio Lower Upper
77 100
123 46.2 38.2 57.2
65 13.3 11.4 17.2









#25
Isophorone
Concen: 9.14 ng
RT: 9.693 min Scan# 1123
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	82	Resp:	101451
Ion	Ratio	Lower	Upper
82	100		
95	8.7	6.5	9.7
138	18.9	15.0	22.4



#26
2-Nitrophenol
Concen: 14.39 ng
RT: 9.875 min Scan# 1154
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	139	Resp:	25683
Ion	Ratio	Lower	Upper
139	100		
109	34.0	28.8	43.2
65	45.4	41.0	61.6









#27
2,4-Dimethylphenol
Concen: 9.47 ng
RT: 9.934 min Scan# 1164
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:122 Resp: 40384
Ion Ratio Lower Upper
122 100
107 113.5 94.5 141.7
121 57.9 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 9.11 ng
RT: 10.175 min Scan# 1205
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 93 Resp: 54463
Ion Ratio Lower Upper
93 100
95 31.7 25.9 38.9
123 11.7 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 10.64 ng
RT: 10.416 min Scan# 1246
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	162	Resp:	48779
Ion Ratio	Lower	Upper	
162	100		
164	63.2	44.9	84.9
98	32.8	14.3	54.3



#30
1,2,4-Trichlorobenzene
Concen: 10.42 ng
RT: 10.628 min Scan# 1282
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	180	Resp:	56579
Ion Ratio	Lower	Upper	
180	100		
182	95.5	77.9	116.9
145	29.7	24.7	37.1









#31
Naphthalene
Concen: 9.56 ng
RT: 10.816 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:128 Resp: 158308
Ion Ratio Lower Upper
128 100
129 10.3 9.0 13.6
127 13.6 10.7 16.1



#32
Benzoic acid
Concen: 9.10 ng
RT: 10.034 min Scan# 1181
Delta R.T. -0.053 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:122 Resp: 16887
Ion Ratio Lower Upper
122 100
105 136.5 103.2 143.2
77 96.7 76.2 116.2









#33
4-Chloroaniline
Concen: 9.18 ng
RT: 10.928 min Scan# 1333
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	127	Resp:	63296
Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.1	39.1
65	30.9	22.5	33.7
92	18.7	15.0	22.6



#34
Hexachlorobutadiene
Concen: 11.16 ng
RT: 11.099 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	225	Resp:	37858
Ion	Ratio	Lower	Upper
225	100		
223	65.6	51.0	76.6
227	65.5	51.4	77.2









#35
Caprolactam
Concen: 10.56 ng
RT: 11.704 min Scan# 1465
Delta R.T. -0.012 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	113	Resp:	13264
Ion	Ratio	Lower	Upper
113	100		
55	105.2	98.9	138.9
56	85.5	71.7	111.7



#36
4-Chloro-3-methylphenol
Concen: 10.47 ng
RT: 12.046 min Scan# 1523
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	107	Resp:	50522
Ion	Ratio	Lower	Upper
107	100		
144	26.7	21.1	31.7
142	83.7	66.0	99.0









#37
2-Methylnaphthalene
Concen: 9.69 ng
RT: 12.422 min Scan# 1587
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:142 Resp: 113791
Ion Ratio Lower Upper
142 100
141 87.0 71.8 107.6
115 37.7 31.6 47.4



#38
1-Methylnaphthalene
Concen: 9.71 ng
RT: 12.640 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:142 Resp: 108110
Ion Ratio Lower Upper
142 100
141 94.8 73.3 109.9
116 4.4 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.581 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:164 Resp: 177760
Ion Ratio Lower Upper
164 100
162 101.1 80.2 120.2
160 44.9 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.22 ng
RT: 12.787 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:216 Resp: 60596
Ion Ratio Lower Upper
216 100
214 78.1 61.9 92.9
179 20.9 16.8 25.2
108 17.5 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 5.82 ng
RT: 12.763 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	237	Resp:	19209
Ion	Ratio	Lower	Upper
237	100		
235	61.9	44.5	84.5
272	15.1	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 26.12 ng
RT: 16.069 min Scan# 2207
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	330	Resp:	59531
Ion	Ratio	Lower	Upper
330	100		
332	98.2	77.8	116.8
141	31.9	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 11.32 ng
RT: 13.022 min Scan# 1689
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:	196	Resp:	40722
Ion Ratio	Lower	Upper	
196	100		
198	96.4	79.5	119.3
200	30.1	25.0	37.6



#44
2,4,5-Trichlorophenol
Concen: 11.25 ng
RT: 13.098 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	196	Resp:	43765
Ion Ratio	Lower	Upper	
196	100		
198	96.4	74.7	112.1
97	47.0	37.4	56.0
132	25.3	20.0	30.0









#45
2-Fluorobiphenyl
Concen: 20.18 ng
RT: 13.210 min Scan# 1721
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:172	Resp:	271049
Ion Ratio	Lower	Upper
172	100	
171	35.2	27.9 41.9
170	23.0	18.6 27.8



#46
1,1'-Biphenyl
Concen: 9.66 ng
RT: 13.422 min Scan# 1757
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:154	Resp:	139462
Ion Ratio	Lower	Upper
154	100	
153	40.6	20.4 60.4
76	10.5	0.0 30.8









#47
2-Chloronaphthalene
Concen: 9.81 ng
RT: 13.463 min Scan# 1764
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:162 Resp: 115084
Ion Ratio Lower Upper
162 100
127 37.0 31.4 47.0
164 32.0 26.6 39.8



#48
2-Nitroaniline
Concen: 12.32 ng
RT: 13.663 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion: 65 Resp: 28163
Ion Ratio Lower Upper
65 100
92 70.7 57.7 86.5
138 115.3 92.1 138.1









#49
Acenaphthylene
Concen: 9.57 ng
RT: 14.304 min Scan# 1907
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:152	Resp:	175069
Ion Ratio	Lower	Upper
152	100	
151	20.9	16.2 24.2
153	13.9	10.6 16.0



#50
Dimethylphthalate
Concen: 10.29 ng
RT: 14.040 min Scan# 1862
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:163	Resp:	145473
Ion Ratio	Lower	Upper
163	100	
194	4.3	3.4 5.2
164	10.1	8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 12.13 ng
RT: 14.157 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:165 Resp: 32048
Ion Ratio Lower Upper
165 100
63 54.0 41.3 61.9
89 59.4 44.8 67.2



#52
Acenaphthene
Concen: 9.05 ng
RT: 14.645 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:154 Resp: 107574
Ion Ratio Lower Upper
154 100
153 111.4 89.4 134.2
152 52.3 43.0 64.4









#53
3-Nitroaniline
Concen: 11.78 ng
RT: 14.487 min Scan# 1938
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:138 Resp: 30673
Ion Ratio Lower Upper
138 100
108 10.6 8.4 12.6
92 117.2 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 12.33 ng
RT: 14.698 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:184 Resp: 12910
Ion Ratio Lower Upper
184 100
63 57.8 47.0 70.6
154 68.0 52.3 78.5









#55
Dibenzenofuran
Concen: 9.97 ng
RT: 14.981 min Scan# 2022
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:168 Resp: 181708
Ion Ratio Lower Upper
168 100
139 38.2 31.0 46.6
169 13.2 10.8 16.2



#56
4-Nitrophenol
Concen: 11.10 ng
RT: 14.792 min Scan# 1990
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:139 Resp: 23782
Ion Ratio Lower Upper
139 100
109 87.8 72.3 112.3
65 90.6 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 13.43 ng
RT: 14.945 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	165	Resp:	42435
Ion Ratio	Lower	Upper	
165	100		
63	37.6	32.3	48.5
89	67.3	55.9	83.9
182	2.5	2.2	3.4



#58
Fluorene
Concen: 9.69 ng
RT: 15.628 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	166	Resp:	141221
Ion Ratio	Lower	Upper	
166	100		
165	99.2	78.0	117.0
167	13.6	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 11.68 ng
RT: 15.204 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	232	Resp:	39290
Ion Ratio	Lower	Upper	
232	100		
131	43.1	32.2	48.4
130	2.6	2.0	3.0
166	28.6	19.4	29.2



#60
Diethylphthalate
Concen: 10.10 ng
RT: 15.398 min Scan# 2093
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	149	Resp:	143290
Ion Ratio	Lower	Upper	
149	100		
177	23.0	17.5	26.3
150	11.8	9.9	14.9









#61
4-Chlorophenyl-phenylether
Concen: 10.21 ng
RT: 15.622 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:204 Resp: 74399
Ion Ratio Lower Upper
204 100
206 32.2 26.6 40.0
141 58.2 48.6 73.0



#62
4-Nitroaniline
Concen: 11.73 ng
RT: 15.645 min Scan# 2135
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:138 Resp: 32049
Ion Ratio Lower Upper
138 100
92 48.5 31.2 71.2
108 91.7 78.6 118.6









#63
Azobenzene
Concen: 8.73 ng
RT: 15.910 min Scan# 2180
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion: 77	Resp: 129749
Ion Ratio	Lower Upper
77 100	
182 28.3	6.0 46.0
105 20.2	0.0 38.8
51 29.0	8.8 48.8



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:188	Resp: 382592
Ion Ratio	Lower Upper
188 100	
94 7.3	5.6 8.4
80 7.5	6.3 9.5









#65
4,6-Dinitro-2-methylphenol
Concen: 12.22 ng
RT: 15.704 min Scan# 2145
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:198 Resp: 22278
Ion Ratio Lower Upper
198 100
51 34.4 14.0 54.0
105 42.8 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 9.63 ng
RT: 15.834 min Scan# 2167
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:169 Resp: 125591
Ion Ratio Lower Upper
169 100
168 66.9 54.4 81.6
167 37.6 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 10.34 ng
RT: 16.516 min Scan# 2283
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:248 Resp: 47678
Ion Ratio Lower Upper
248 100
250 97.9 79.8 119.8
141 70.1 57.8 86.6



#68
Hexachlorobenzene
Concen: 10.78 ng
RT: 16.634 min Scan# 2303
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:284 Resp: 59158
Ion Ratio Lower Upper
284 100
142 31.2 26.9 40.3
249 27.9 24.5 36.7









#69
Atrazine
Concen: 10.51 ng
RT: 16.781 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDICC010


Tgt Ion:200 Resp: 39159
Ion Ratio Lower Upper
200 100
173 25.2 4.1 44.1
215 45.9 27.7 67.7



#70
Pentachlorophenol
Concen: 10.23 ng
RT: 16.981 min Scan# 2362
Delta R.T. 0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:266 Resp: 28157
Ion Ratio Lower Upper
266 100
268 64.6 51.6 77.4
264 61.5 50.3 75.5









#71
Phenanthrene
Concen: 9.63 ng
RT: 17.369 min Scan# 2428
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:178 Resp: 229648
Ion Ratio Lower Upper
178 100
176 18.3 15.1 22.7
179 14.5 12.4 18.6



#72
Anthracene
Concen: 9.57 ng
RT: 17.457 min Scan# 2443
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:178 Resp: 225129
Ion Ratio Lower Upper
178 100
176 18.4 14.9 22.3
179 15.3 12.2 18.2









#73
 Carbazole
 Concen: 9.33 ng
 RT: 17.728 min Scan# 2489
 Delta R.T. -0.000 min
 Lab File: BP005780.D
 Acq: 08 Jun 2021 18:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Tgt Ion:	167	Resp:	216025
Ion Ratio	Lower	Upper	
167	100		
166	21.8	17.1	25.7
139	13.2	10.7	16.1



#74
 Di-n-butylphthalate
 Concen: 9.77 ng
 RT: 18.275 min Scan# 2582
 Delta R.T. -0.000 min
 Lab File: BP005780.D
 Acq: 08 Jun 2021 18:31

Tgt Ion:	149	Resp:	242345
Ion Ratio	Lower	Upper	
149	100		
150	9.0	7.3	10.9
104	5.3	4.5	6.7









#75
Fluoranthene
Concen: 9.85 ng
RT: 19.327 min Scan# 2761
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:202 Resp: 281719
Ion Ratio Lower Upper
202 100
101 7.5 0.0 27.6
203 17.3 0.0 37.8



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.392 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:240 Resp: 434411
Ion Ratio Lower Upper
240 100
120 7.8 6.5 9.7
236 25.8 20.6 31.0









#77
Benzidine
Concen: 9.74 ng
RT: 19.504 min Scan# 2791
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:184	Resp:	80377
Ion Ratio	Lower	Upper
184	100	
185	14.6	11.0 16.4
183	11.4	9.5 14.3



#78
Pyrene
Concen: 9.59 ng
RT: 19.675 min Scan# 2820
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:202	Resp:	293865
Ion Ratio	Lower	Upper
202	100	
200	21.1	16.3 24.5
203	17.2	14.2 21.2









#79
 Terphenyl-d14
 Concen: 20.49 ng
 RT: 19.869 min Scan# 2853
 Delta R.T. -0.000 min
 Lab File: BP005780.D
 Acq: 08 Jun 2021 18:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC010


Tgt Ion:244	Resp: 474351
Ion Ratio	Lower Upper
244 100	
212 7.0	6.0 9.0
122 8.3	7.1 10.7



#80
 Butylbenzylphthalate
 Concen: 10.45 ng
 RT: 20.539 min Scan# 2967
 Delta R.T. -0.000 min
 Lab File: BP005780.D
 Acq: 08 Jun 2021 18:31

Tgt Ion:149	Resp: 114562
Ion Ratio	Lower Upper
149 100	
91 67.9	54.6 81.8
206 22.4	16.7 25.1









#81
Benzo(a)anthracene
Concen: 9.93 ng
RT: 21.374 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:228 Resp: 316649
Ion Ratio Lower Upper
228 100
226 27.4 21.6 32.4
229 20.0 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 10.81 ng
RT: 21.304 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:252 Resp: 111311
Ion Ratio Lower Upper
252 100
254 65.9 50.6 76.0
126 10.0 7.0 10.6









#83
Chrysene
Concen: 9.93 ng
RT: 21.427 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:	228	Resp:	306499
Ion Ratio	Lower	Upper	
228	100		
226	29.3	23.8	35.6
229	20.5	15.4	23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 10.14 ng
RT: 21.298 min Scan# 3096
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:	149	Resp:	175567
Ion Ratio	Lower	Upper	
149	100		
167	28.0	21.9	32.9
279	5.5	4.3	6.5









#85
Di-n-octyl phthalate
Concen: 10.57 ng
RT: 22.221 min Scan# 3253
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:149	Resp:	307691
Ion Ratio	Lower	Upper
149	100	
167	1.5	1.4 2.0
43	7.3	5.6 8.4



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.810 min Scan# 3523
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:264	Resp:	503027
Ion Ratio	Lower	Upper
264	100	
260	23.1	18.6 28.0
265	22.4	17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 10.10 ng
RT: 26.351 min Scan# 3955
Delta R.T. -0.018 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:276 Resp: 432265
Ion Ratio Lower Upper
276 100
138 18.7 15.3 22.9
277 25.0 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 9.36 ng
RT: 23.068 min Scan# 3397
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:252 Resp: 341452
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 7.3 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 9.68 ng
RT: 23.121 min Scan# 3406
Delta R.T. -0.000 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31




Instrument :
BNA_P
ClientSampleId :
SSTDIC010


Tgt Ion:252 Resp: 334635
Ion Ratio Lower Upper
252 100
253 22.6 17.6 26.4
125 8.2 6.0 9.0



#90
Benzo(a)pyrene
Concen: 9.66 ng
RT: 23.704 min Scan# 3505
Delta R.T. -0.006 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:252 Resp: 320710
Ion Ratio Lower Upper
252 100
253 22.6 17.9 26.9
125 8.8 6.5 9.7









#91
Dibenzo(a,h)anthracene
Concen: 9.99 ng
RT: 26.368 min Scan# 3958
Delta R.T. -0.018 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Instrument :
BNA_P
ClientSampleId :
SSTDIC010

Tgt Ion:278 Resp: 370971
Ion Ratio Lower Upper
278 100
139 12.1 9.3 13.9
279 23.9 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 10.12 ng
RT: 27.139 min Scan# 4089
Delta R.T. -0.018 min
Lab File: BP005780.D
Acq: 08 Jun 2021 18:31

Tgt Ion:276 Resp: 362445
Ion Ratio Lower Upper
276 100
277 23.8 18.9 28.3
138 17.0 12.6 18.8

