

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005782.D
 Acq On : 08 Jun 2021 19:40
 Operator : CG/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICCC040

Quant Time: Jun 09 01:22:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 01:17:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	89064	20.00	ng	0.00
21) Naphthalene-d8	10.763	136	336042	20.00	ng	0.00
39) Acenaphthene-d10	14.581	164	209912	20.00	ng	0.00
64) Phenanthrene-d10	17.328	188	444246	20.00	ng	0.00
76) Chrysene-d12	21.392	240	474006	20.00	ng	0.00
86) Perylene-d12	23.816	264	542200	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.517	112	424401	74.08	ng	0.00
7) Phenol-d6	7.122	99	553673	70.90	ng	0.00
23) Nitrobenzene-d5	9.122	82	522278	86.44	ng	0.00
42) 2,4,6-Tribromophenol	16.069	330	273749	101.70	ng	0.00
45) 2-Fluorobiphenyl	13.216	172	1195722	75.39	ng	0.00
79) Terphenyl-d14	19.869	244	1875156	74.23	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.387	88	95133	36.28	ng	100
3) Pyridine	3.805	79	206933	31.88	ng	100
4) n-Nitrosodimethylamine	3.705	42	104003	35.98	ng	100
6) Aniline	7.281	93	305927	33.51	ng	100
8) 2-Chlorophenol	7.522	128	239626	37.36	ng	100
9) Benzaldehyde	7.099	77	128395	37.97	ng	100
10) Phenol	7.152	94	271332	34.03	ng	100
11) bis(2-Chloroethyl)ether	7.381	93	204795	33.27	ng	100
12) 1,3-Dichlorobenzene	7.846	146	267512	36.74	ng	100
13) 1,4-Dichlorobenzene	7.993	146	271373	36.78	ng	100
14) 1,2-Dichlorobenzene	8.310	146	261430	37.04	ng	100
15) Benzyl Alcohol	8.199	79	209574	37.41	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.493	45	191654	27.20	ng	100
17) 2-Methylphenol	8.405	107	186882	36.00	ng	100
18) Hexachloroethane	9.040	117	100104	38.23	ng	100
19) n-Nitroso-di-n-propyla...	8.769	70	169103	34.04	ng	100
20) 3+4-Methylphenols	8.734	107	253721	36.69	ng	100
22) Acetophenone	8.787	105	337398	37.64	ng	100
24) Nitrobenzene	9.163	77	264239	39.37	ng	100
25) Isophorone	9.693	82	473525	35.30	ng	100
26) 2-Nitrophenol	9.875	139	127508	59.14	ng	100
27) 2,4-Dimethylphenol	9.940	122	187551	36.42	ng	100
28) bis(2-Chloroethoxy)met...	10.175	93	241942	33.51	ng	100
29) 2,4-Dichlorophenol	10.416	162	224582	40.55	ng	100
30) 1,2,4-Trichlorobenzene	10.628	180	248152	37.83	ng	100
31) Naphthalene	10.816	128	709497	35.45	ng	100
32) Benzoic acid	10.087	122	141236	62.98	ng	100
33) 4-Chloroaniline	10.928	127	289810	34.78	ng	100
34) Hexachlorobutadiene	11.099	225	167461	40.87	ng	100
35) Caprolactam	11.716	113	69375	45.74	ng	100
36) 4-Chloro-3-methylphenol	12.046	107	227795	39.08	ng	100
37) 2-Methylnaphthalene	12.422	142	503110	35.48	ng	100
38) 1-Methylnaphthalene	12.640	142	479933	35.67	ng	100
40) 1,2,4,5-Tetrachloroben...	12.787	216	279308	39.89	ng	100
41) Hexachlorocyclopentadiene	12.763	237	123784	31.75	ng	100
43) 2,4,6-Trichlorophenol	13.022	196	190548	44.86	ng	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005782.D
 Acq On : 08 Jun 2021 19:40
 Operator : CG/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICCC040

Quant Time: Jun 09 01:22:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 01:17:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.098	196	206179	44.89	ng	100
46) 1,1'-Biphenyl	13.422	154	643908	37.78	ng	100
47) 2-Chloronaphthalene	13.463	162	503492	36.34	ng	100
48) 2-Nitroaniline	13.663	65	136881	50.69	ng	100
49) Acenaphthylene	14.304	152	787608	36.47	ng	100
50) Dimethylphthalate	14.045	163	628909	37.67	ng	100
51) 2,6-Dinitrotoluene	14.163	165	146042	46.81	ng	100
52) Acenaphthene	14.645	154	469046	33.41	ng	100
53) 3-Nitroaniline	14.487	138	142833	46.44	ng	100
54) 2,4-Dinitrophenol	14.698	184	81760	66.15	ng	100
55) Dibenzofuran	14.981	168	772193	35.89	ng	100
56) 4-Nitrophenol	14.792	139	118648	46.88	ng	100
57) 2,4-Dinitrotoluene	14.945	165	199740	53.55	ng	100
58) Fluorene	15.628	166	604122	35.12	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.122	232	183031	46.06	ng	99
60) Diethylphthalate	15.398	149	638552	38.12	ng	100
61) 4-Chlorophenyl-phenyle...	15.622	204	312464	36.33	ng	100
62) 4-Nitroaniline	15.645	138	146662	45.46	ng	100
63) Azobenzene	15.910	77	583906	33.28	ng	100
65) 4,6-Dinitro-2-methylph...	15.710	198	122339	57.79	ng	100
66) n-Nitrosodiphenylamine	15.834	169	535918	35.40	ng	100
67) 4-Bromophenyl-phenylether	16.516	248	208824	39.02	ng	100
68) Hexachlorobenzene	16.634	284	247583	38.85	ng	100
69) Atrazine	16.787	200	195929	45.29	ng	100
70) Pentachlorophenol	16.975	266	152188	47.62	ng	100
71) Phenanthrene	17.369	178	974096	35.20	ng	100
72) Anthracene	17.463	178	964757	35.30	ng	100
73) Carbazole	17.728	167	927051	34.47	ng	100
74) Di-n-butylphthalate	18.275	149	1079800	37.49	ng	100
75) Fluoranthene	19.328	202	1176747	35.42	ng	100
77) Benzidine	19.504	184	450440	50.04	ng	100
78) Pyrene	19.675	202	1210421	36.19	ng	100
80) Butylbenzylphthalate	20.539	149	508208	42.49	ng	100
81) Benzo(a)anthracene	21.380	228	1245818	35.81	ng	100
82) 3,3'-Dichlorobenzidine	21.310	252	424593	37.79	ng	100
83) Chrysene	21.433	228	1203819	35.74	ng	100
84) Bis(2-ethylhexyl)phtha...	21.298	149	738129	39.07	ng	100
85) Di-n-octyl phthalate	22.222	149	1321852	41.63	ng	100
87) Indeno(1,2,3-cd)pyrene	26.368	276	1685819	36.54	ng	100
88) Benzo(b)fluoranthene	23.074	252	1364681	34.69	ng	100
89) Benzo(k)fluoranthene	23.121	252	1316583	35.32	ng	100
90) Benzo(a)pyrene	23.710	252	1276378	35.66	ng	100
91) Dibenzo(a,h)anthracene	26.386	278	1456125	36.37	ng	100
92) Benzo(g,h,i)perylene	27.157	276	1429345	37.02	ng	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.958 min Scan# 828
 Delta R.T. -0.000 min
 Lab File: BP005782.D
 Acq: 08 Jun 2021 19:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion:152 Resp: 89064
 Ion Ratio Lower Upper
 152 100
 150 157.0 125.6 188.4
 115 60.6 48.5 72.7



#2
 1,4-Dioxane
 Concen: 36.28 ng
 RT: 3.387 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: BP005782.D
 Acq: 08 Jun 2021 19:40

Tgt Ion: 88 Resp: 95133
 Ion Ratio Lower Upper
 88 100
 58 58.4 46.7 70.1
 43 25.4 20.3 30.5









#3
Pyridine
Concen: 31.88 ng
RT: 3.805 min Scan# 122
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion: 79 Resp: 206933
Ion Ratio Lower Upper
79 100
52 52.1 41.7 62.5
51 30.1 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 35.98 ng
RT: 3.705 min Scan# 105
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion: 42 Resp: 104003
Ion Ratio Lower Upper
42 100
74 131.7 105.4 158.0
44 5.3 4.2 6.4









#5
2-Fluorophenol
Concen: 74.08 ng
RT: 5.517 min Scan# 413
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:112 Resp: 424401
Ion Ratio Lower Upper
112 100
64 51.8 41.4 62.2
63 29.8 23.8 35.8



#6
Aniline
Concen: 33.51 ng
RT: 7.281 min Scan# 713
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion: 93 Resp: 305927
Ion Ratio Lower Upper
93 100
66 37.1 29.7 44.5
65 19.1 15.3 22.9









#7
Phenol-d6
Concen: 70.90 ng
RT: 7.122 min Scan# 686
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion: 99 Resp: 553673
Ion Ratio Lower Upper
99 100
42 16.7 13.4 20.0
71 38.3 30.6 46.0



#8
2-Chlorophenol
Concen: 37.36 ng
RT: 7.522 min Scan# 754
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion: 128 Resp: 239626
Ion Ratio Lower Upper
128 100
130 32.4 12.4 52.4
64 40.3 20.3 60.3









#9
Benzaldehyde
Concen: 37.97 ng
RT: 7.099 min Scan# 682
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion: 77 Resp: 128395
Ion Ratio Lower Upper
77 100
105 96.7 76.7 116.7
106 96.2 76.2 116.2



#10
Phenol
Concen: 34.03 ng
RT: 7.152 min Scan# 691
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion: 94 Resp: 271332
Ion Ratio Lower Upper
94 100
65 28.8 8.8 48.8
66 40.7 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 33.27 ng
RT: 7.381 min Scan# 730
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion: 93 Resp: 204795
Ion Ratio Lower Upper
93 100
63 66.2 46.2 86.2
95 32.2 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 36.74 ng
RT: 7.846 min Scan# 809
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:146 Resp: 267512
Ion Ratio Lower Upper
146 100
148 64.4 51.5 77.3
75 32.1 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 36.78 ng
RT: 7.993 min Scan# 834
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:146 Resp: 271373
Ion Ratio Lower Upper
146 100
148 64.1 51.3 76.9
111 41.3 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 37.04 ng
RT: 8.310 min Scan# 888
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:146 Resp: 261430
Ion Ratio Lower Upper
146 100
148 62.9 50.3 75.5
111 42.1 33.7 50.5









#15
Benzyl Alcohol
Concen: 37.41 ng
RT: 8.199 min Scan# 869
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion: 79 Resp: 209574
Ion Ratio Lower Upper
79 100
108 78.1 62.5 93.7
77 66.1 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 27.20 ng
RT: 8.493 min Scan# 919
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion: 45 Resp: 191654
Ion Ratio Lower Upper
45 100
77 22.0 2.0 42.0
79 16.7 0.0 36.7









#17
2-Methylphenol
Concen: 36.00 ng
RT: 8.405 min Scan# 904
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	107	Resp:	186882
Ion	Ratio	Lower	Upper
107	100		
108	109.6	87.7	131.5
77	50.2	40.2	60.2
79	48.7	39.0	58.4



#18
Hexachloroethane
Concen: 38.23 ng
RT: 9.040 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	117	Resp:	100104
Ion	Ratio	Lower	Upper
117	100		
119	95.2	76.2	114.2
201	104.2	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 34.04 ng
RT: 8.769 min Scan# 966
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	70	Resp:	169103
Ion	Ratio	Lower	Upper
70	100		
42	51.0	40.8	61.2
101	11.0	8.8	13.2
130	23.6	18.9	28.3



#20
3+4-Methylphenols
Concen: 36.69 ng
RT: 8.734 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	107	Resp:	253721
Ion	Ratio	Lower	Upper
107	100		
108	91.0	71.0	111.0
77	36.6	16.6	56.6
79	32.7	12.7	52.7









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.763 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	136	Resp:	336042
Ion Ratio	Lower	Upper	
136	100		
137	10.7	8.6	12.8
54	5.7	4.6	6.8
68	4.5	3.6	5.4



#22
Acetophenone
Concen: 37.64 ng
RT: 8.787 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	105	Resp:	337398
Ion Ratio	Lower	Upper	
105	100		
71	2.4	1.9	2.9
51	23.0	18.4	27.6
120	22.4	17.9	26.9









#23
Nitrobenzene-d5
Concen: 86.44 ng
RT: 9.122 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion: 82 Resp: 522278
Ion Ratio Lower Upper
82 100
128 44.6 35.7 53.5
54 43.7 35.0 52.4



#24
Nitrobenzene
Concen: 39.37 ng
RT: 9.163 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion: 77 Resp: 264239
Ion Ratio Lower Upper
77 100
123 47.7 38.2 57.2
65 14.3 11.4 17.2









#25
Isophorone
Concen: 35.30 ng
RT: 9.693 min Scan# 1123
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion: 82 Resp: 473525
Ion Ratio Lower Upper
82 100
95 8.1 6.5 9.7
138 18.7 15.0 22.4



#26
2-Nitrophenol
Concen: 59.14 ng
RT: 9.875 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:139 Resp: 127508
Ion Ratio Lower Upper
139 100
109 36.0 28.8 43.2
65 51.3 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 36.42 ng
RT: 9.940 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:122 Resp: 187551
Ion Ratio Lower Upper
122 100
107 118.1 94.5 141.7
121 59.3 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 33.51 ng
RT: 10.175 min Scan# 1205
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion: 93 Resp: 241942
Ion Ratio Lower Upper
93 100
95 32.4 25.9 38.9
123 12.7 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 40.55 ng
RT: 10.416 min Scan# 1246
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:162 Resp: 224582
Ion Ratio Lower Upper
162 100
164 64.9 44.9 84.9
98 34.3 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 37.83 ng
RT: 10.628 min Scan# 1282
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:180 Resp: 248152
Ion Ratio Lower Upper
180 100
182 97.4 77.9 116.9
145 30.9 24.7 37.1









#31
Naphthalene
Concen: 35.45 ng
RT: 10.816 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:128 Resp: 709497
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.6
127 13.4 10.7 16.1



#32
Benzoic acid
Concen: 62.98 ng
RT: 10.087 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:122 Resp: 141236
Ion Ratio Lower Upper
122 100
105 123.2 103.2 143.2
77 96.2 76.2 116.2









#33
4-Chloroaniline
Concen: 34.78 ng
RT: 10.928 min Scan# 1333
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	127	Resp:	289810
Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.1	39.1
65	28.1	22.5	33.7
92	18.8	15.0	22.6



#34
Hexachlorobutadiene
Concen: 40.87 ng
RT: 11.099 min Scan# 1362
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	225	Resp:	167461
Ion	Ratio	Lower	Upper
225	100		
223	63.8	51.0	76.6
227	64.3	51.4	77.2









#35
Caprolactam
Concen: 45.74 ng
RT: 11.716 min Scan# 1467
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	113	Resp:	69375
Ion	Ratio	Lower	Upper
113	100		
55	118.9	98.9	138.9
56	91.7	71.7	111.7



#36
4-Chloro-3-methylphenol
Concen: 39.08 ng
RT: 12.046 min Scan# 1523
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	107	Resp:	227795
Ion	Ratio	Lower	Upper
107	100		
144	26.4	21.1	31.7
142	82.5	66.0	99.0









#37
2-Methylnaphthalene
Concen: 35.48 ng
RT: 12.422 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	142	Resp:	503110
Ion	Ratio	Lower	Upper
142	100		
141	89.7	71.8	107.6
115	39.5	31.6	47.4



#38
1-Methylnaphthalene
Concen: 35.67 ng
RT: 12.640 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	142	Resp:	479933
Ion	Ratio	Lower	Upper
142	100		
141	91.6	73.3	109.9
116	4.4	3.5	5.3









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.581 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:164 Resp: 209912
Ion Ratio Lower Upper
164 100
162 100.2 80.2 120.2
160 43.1 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.89 ng
RT: 12.787 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:216 Resp: 279308
Ion Ratio Lower Upper
216 100
214 77.4 61.9 92.9
179 21.0 16.8 25.2
108 18.6 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 31.75 ng
RT: 12.763 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	237	Resp:	123784
Ion Ratio	Lower	Upper	
237	100		
235	64.5	44.5	84.5
272	12.5	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 101.70 ng
RT: 16.069 min Scan# 2207
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	330	Resp:	273749
Ion Ratio	Lower	Upper	
330	100		
332	97.3	77.8	116.8
141	31.8	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 44.86 ng
RT: 13.022 min Scan# 1689
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:196 Resp: 190548
Ion Ratio Lower Upper
196 100
198 99.4 79.5 119.3
200 31.3 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 44.89 ng
RT: 13.098 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:196 Resp: 206179
Ion Ratio Lower Upper
196 100
198 93.4 74.7 112.1
97 46.7 37.4 56.0
132 25.0 20.0 30.0









#45
2-Fluorobiphenyl
Concen: 75.39 ng
RT: 13.216 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:172 Resp: 1195722
Ion Ratio Lower Upper
172 100
171 34.9 27.9 41.9
170 23.2 18.6 27.8



#46
1,1'-Biphenyl
Concen: 37.78 ng
RT: 13.422 min Scan# 1757
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:154 Resp: 643908
Ion Ratio Lower Upper
154 100
153 40.4 20.4 60.4
76 10.8 0.0 30.8









#47
2-Chloronaphthalene
Concen: 36.34 ng
RT: 13.463 min Scan# 1764
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:162 Resp: 503492
Ion Ratio Lower Upper
162 100
127 39.2 31.4 47.0
164 33.2 26.6 39.8



#48
2-Nitroaniline
Concen: 50.69 ng
RT: 13.663 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion: 65 Resp: 136881
Ion Ratio Lower Upper
65 100
92 72.1 57.7 86.5
138 115.1 92.1 138.1









#49
Acenaphthylene
Concen: 36.47 ng
RT: 14.304 min Scan# 1907
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	152	Resp:	787608
Ion Ratio	Lower	Upper	
152	100		
151	20.2	16.2	24.2
153	13.3	10.6	16.0



#50
Dimethylphthalate
Concen: 37.67 ng
RT: 14.045 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	163	Resp:	628909
Ion Ratio	Lower	Upper	
163	100		
194	4.3	3.4	5.2
164	10.2	8.2	12.2









#51
2,6-Dinitrotoluene
Concen: 46.81 ng
RT: 14.163 min Scan# 1883
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:165 Resp: 146042
Ion Ratio Lower Upper
165 100
63 51.6 41.3 61.9
89 56.0 44.8 67.2



#52
Acenaphthene
Concen: 33.41 ng
RT: 14.645 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:154 Resp: 469046
Ion Ratio Lower Upper
154 100
153 111.8 89.4 134.2
152 53.7 43.0 64.4









#53
3-Nitroaniline
Concen: 46.44 ng
RT: 14.487 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:138 Resp: 142833
Ion Ratio Lower Upper
138 100
108 10.5 8.4 12.6
92 109.8 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 66.15 ng
RT: 14.698 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:184 Resp: 81760
Ion Ratio Lower Upper
184 100
63 58.8 47.0 70.6
154 65.4 52.3 78.5









#55
Dibenzofuran
Concen: 35.89 ng
RT: 14.981 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	168	Resp:	772193
Ion Ratio	Lower	Upper	
168	100		
139	38.8	31.0	46.6
169	13.5	10.8	16.2



#56
4-Nitrophenol
Concen: 46.88 ng
RT: 14.792 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	139	Resp:	118648
Ion Ratio	Lower	Upper	
139	100		
109	92.3	72.3	112.3
65	92.5	72.5	112.5









#57
2,4-Dinitrotoluene
Concen: 53.55 ng
RT: 14.945 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:	165	Resp:	199740
Ion Ratio	Lower	Upper	
165	100		
63	40.4	32.3	48.5
89	69.9	55.9	83.9
182	2.8	2.2	3.4



#58
Fluorene
Concen: 35.12 ng
RT: 15.628 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:	166	Resp:	604122
Ion Ratio	Lower	Upper	
166	100		
165	97.5	78.0	117.0
167	13.1	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 46.06 ng
RT: 15.122 min Scan# 2046
Delta R.T. -0.082 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:232	Resp: 183031
Ion Ratio	Lower Upper
232 100	
131 39.7	32.2 48.4
130 2.5	2.0 3.0
166 24.0	19.4 29.2



#60
Diethylphthalate
Concen: 38.12 ng
RT: 15.398 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:149	Resp: 638552
Ion Ratio	Lower Upper
149 100	
177 21.9	17.5 26.3
150 12.4	9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 36.33 ng
RT: 15.622 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:204 Resp: 312464
Ion Ratio Lower Upper
204 100
206 33.3 26.6 40.0
141 60.8 48.6 73.0



#62
4-Nitroaniline
Concen: 45.46 ng
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:138 Resp: 146662
Ion Ratio Lower Upper
138 100
92 51.2 31.2 71.2
108 98.6 78.6 118.6









#63
Azobenzene
Concen: 33.28 ng
RT: 15.910 min Scan# 2180
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion: 77	Resp: 583906
Ion Ratio	Lower Upper
77 100	
182 26.0	6.0 46.0
105 18.8	0.0 38.8
51 28.8	8.8 48.8



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.328 min Scan# 2421
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:188	Resp: 444246
Ion Ratio	Lower Upper
188 100	
94 7.0	5.6 8.4
80 7.9	6.3 9.5









#65
4,6-Dinitro-2-methylphenol
Concen: 57.79 ng
RT: 15.710 min Scan# 2146
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:198 Resp: 122339
Ion Ratio Lower Upper
198 100
51 34.0 14.0 54.0
105 39.5 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 35.40 ng
RT: 15.834 min Scan# 2167
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:169 Resp: 535918
Ion Ratio Lower Upper
169 100
168 68.0 54.4 81.6
167 36.3 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 39.02 ng
RT: 16.516 min Scan# 2283
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:248 Resp: 208824
Ion Ratio Lower Upper
248 100
250 99.8 79.8 119.8
141 72.2 57.8 86.6



#68
Hexachlorobenzene
Concen: 38.85 ng
RT: 16.634 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:284 Resp: 247583
Ion Ratio Lower Upper
284 100
142 33.6 26.9 40.3
249 30.6 24.5 36.7









#69
Atrazine
Concen: 45.29 ng
RT: 16.787 min Scan# 2329
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:200 Resp: 195929
Ion Ratio Lower Upper
200 100
173 24.1 4.1 44.1
215 47.7 27.7 67.7



#70
Pentachlorophenol
Concen: 47.62 ng
RT: 16.975 min Scan# 2361
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:266 Resp: 152188
Ion Ratio Lower Upper
266 100
268 64.5 51.6 77.4
264 62.9 50.3 75.5









#71
Phenanthrene
Concen: 35.20 ng
RT: 17.369 min Scan# 2428
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:178 Resp: 974096
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.5 12.4 18.6



#72
Anthracene
Concen: 35.30 ng
RT: 17.463 min Scan# 2444
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:178 Resp: 964757
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.2 12.2 18.2









#73
 Carbazole
 Concen: 34.47 ng
 RT: 17.728 min Scan# 2489
 Delta R.T. 0.000 min
 Lab File: BP005782.D
 Acq: 08 Jun 2021 19:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC040


Tgt Ion:	167	Resp:	927051
Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.1	25.7
139	13.4	10.7	16.1



#74
 Di-n-butylphthalate
 Concen: 37.49 ng
 RT: 18.275 min Scan# 2582
 Delta R.T. 0.000 min
 Lab File: BP005782.D
 Acq: 08 Jun 2021 19:40

Tgt Ion:	149	Resp:	1079800
Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	5.6	4.5	6.7









#75
Fluoranthene
Concen: 35.42 ng
RT: 19.328 min Scan# 2761
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:202 Resp: 1176747
Ion Ratio Lower Upper
202 100
101 7.6 0.0 27.6
203 17.8 0.0 37.8



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.392 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:240 Resp: 474006
Ion Ratio Lower Upper
240 100
120 8.1 6.5 9.7
236 25.8 20.6 31.0









#77
Benzidine
Concen: 50.04 ng
RT: 19.504 min Scan# 2791
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:184 Resp: 450440
Ion Ratio Lower Upper
184 100
185 13.7 11.0 16.4
183 11.9 9.5 14.3



#78
Pyrene
Concen: 36.19 ng
RT: 19.675 min Scan# 2820
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:202 Resp: 1210421
Ion Ratio Lower Upper
202 100
200 20.4 16.3 24.5
203 17.7 14.2 21.2









#79
Terphenyl-d14
Concen: 74.23 ng
RT: 19.869 min Scan# 2853
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:244 Resp: 1875156
Ion Ratio Lower Upper
244 100
212 7.5 6.0 9.0
122 8.9 7.1 10.7



#80
Butylbenzylphthalate
Concen: 42.49 ng
RT: 20.539 min Scan# 2967
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:149 Resp: 508208
Ion Ratio Lower Upper
149 100
91 68.2 54.6 81.8
206 20.9 16.7 25.1









#81
Benzo(a)anthracene
Concen: 35.81 ng
RT: 21.380 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:228 Resp: 1245818
Ion Ratio Lower Upper
228 100
226 27.0 21.6 32.4
229 19.7 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 37.79 ng
RT: 21.310 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:252 Resp: 424593
Ion Ratio Lower Upper
252 100
254 63.3 50.6 76.0
126 8.8 7.0 10.6





#83

Chrysene

Concen: 35.74 ng

RT: 21.433 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BP005782.D

Acq: 08 Jun 2021 19:40

Tgt Ion:228 Resp: 1203819





Ion	Ratio	Lower	Upper
228	100		
226	29.7	23.8	35.6
229	19.3	15.4	23.2

Instrument :

BNA_P

ClientSampleId :

SSTDICCC040



#84

Bis(2-ethylhexyl)phthalate

Concen: 39.07 ng

RT: 21.298 min Scan# 3096




Delta R.T. 0.000 min


Lab File: BP005782.D

Acq: 08 Jun 2021 19:40

Tgt Ion:149 Resp: 738129

Ion	Ratio	Lower	Upper
149	100		
167	27.4	21.9	32.9
279	5.4	4.3	6.5









#85
Di-n-octyl phthalate
Concen: 41.63 ng
RT: 22.222 min Scan# 3253
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:149 Resp: 1321852
Ion Ratio Lower Upper
149 100
167 1.7 1.4 2.0
43 7.0 5.6 8.4



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.816 min Scan# 3524
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:264 Resp: 542200
Ion Ratio Lower Upper
264 100
260 23.3 18.6 28.0
265 22.0 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 36.54 ng
RT: 26.368 min Scan# 3958
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:276 Resp: 1685819
Ion Ratio Lower Upper
276 100
138 19.1 15.3 22.9
277 25.5 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 34.69 ng
RT: 23.074 min Scan# 3398
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:252 Resp: 1364681
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 7.8 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 35.32 ng
RT: 23.121 min Scan# 3406
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40




Instrument :
BNA_P
ClientSampleId :
SSTDICCC040


Tgt Ion:252 Resp: 1316583
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 7.5 6.0 9.0



#90
Benzo(a)pyrene
Concen: 35.66 ng
RT: 23.710 min Scan# 3506
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:252 Resp: 1276378
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 8.1 6.5 9.7









#91
Dibenzo(a,h)anthracene
Concen: 36.37 ng
RT: 26.386 min Scan# 3961
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:278 Resp: 1456125
Ion Ratio Lower Upper
278 100
139 11.6 9.3 13.9
279 23.5 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 37.02 ng
RT: 27.157 min Scan# 4092
Delta R.T. 0.000 min
Lab File: BP005782.D
Acq: 08 Jun 2021 19:40

Tgt Ion:276 Resp: 1429345
Ion Ratio Lower Upper
276 100
277 23.6 18.9 28.3
138 15.7 12.6 18.8

