

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005801.D
 Acq On : 09 Jun 2021 08:09
 Operator : CG/JU
 Sample : M2589-26
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B8(4-6)

Quant Time: Jun 09 09:35:15 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 02:47:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	75825	20.000	ng	0.00
21) Naphthalene-d8	10.763	136	275754	20.000	ng	0.00
39) Acenaphthene-d10	14.581	164	162125	20.000	ng	0.00
64) Phenanthrene-d10	17.328	188	338401	20.000	ng	0.00
76) Chrysene-d12	21.398	240	384708	20.000	ng	0.00
86) Perylene-d12	23.821	264	459847	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	616612	130.489	ng	0.00
7) Phenol-d6	7.122	99	651636	106.393	ng	0.00
23) Nitrobenzene-d5	9.122	82	493145	87.678	ng	0.00
42) 2,4,6-Tribromophenol	16.069	330	369013	132.634	ng	0.00
45) 2-Fluorobiphenyl	13.210	172	1077859	87.301	ng	0.00
79) Terphenyl-d14	19.874	244	1798268	87.527	ng	0.00
Target Compounds						
31) Naphthalene	10.810	128	41993	2.632	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
Data File : BP005801.D
Acq On : 09 Jun 2021 08:09
Operator : CG/JU
Sample : M2589-26
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
18-B8(4-6)

Quant Time: Jun 09 09:35:15 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 02:47:38 2021
Response via : Initial Calibration

