

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005786.D
 Acq On : 08 Jun 2021 21:56
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP060821

Quant Time: Jun 09 11:59:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	37.157	7.1	97	0.00
3	Pyridine	40.000	38.557	3.6	96	0.00
4	n-Nitrosodimethylamine	40.000	36.141	9.6	94	0.00
5 S	2-Fluorophenol	80.000	74.746	6.6	97	0.00
6	Aniline	40.000	37.261	6.8	97	0.00
7 S	Phenol-d6	80.000	75.249	5.9	97	0.00
8	2-Chlorophenol	40.000	36.822	7.9	97	0.00
9	Benzaldehyde	40.000	41.273	-3.2	98	0.00
10 C	Phenol	40.000	37.057	7.4	97	0.00
11	bis(2-Chloroethyl)ether	40.000	36.882	7.8	97	0.00
12	1,3-Dichlorobenzene	40.000	36.703	8.2	97	0.00
13 C	1,4-Dichlorobenzene	40.000	36.413	9.0	97	0.00
14	1,2-Dichlorobenzene	40.000	36.530	8.7	96	0.00
15	Benzyl Alcohol	40.000	37.145	7.1	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.255	9.4	94	0.00
17	2-Methylphenol	40.000	37.246	6.9	97	0.00
18	Hexachloroethane	40.000	37.057	7.4	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.546	8.6	95	0.00
20	3+4-Methylphenols	40.000	37.084	7.3	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	38.406	4.0	96	0.00
23 S	Nitrobenzene-d5	80.000	77.044	3.7	96	0.00
24	Nitrobenzene	40.000	37.509	6.2	96	0.00
25	Isophorone	40.000	37.086	7.3	94	0.00
26 C	2-Nitrophenol	40.000	38.373	4.1	97	0.00
27	2,4-Dimethylphenol	40.000	37.536	6.2	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.577	8.6	95	0.00
29 C	2,4-Dichlorophenol	40.000	38.034	4.9	98	0.00
30	1,2,4-Trichlorobenzene	40.000	37.100	7.2	97	0.00
31	Naphthalene	40.000	37.104	7.2	96	0.00
32	Benzoic acid	40.000	37.951	5.1	93	0.00
33	4-Chloroaniline	40.000	37.304	6.7	96	0.00
34 C	Hexachlorobutadiene	40.000	36.927	7.7	97	0.00
35	Caprolactam	40.000	40.544	-1.4	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.357	6.6	96	0.00
37	2-Methylnaphthalene	40.000	37.430	6.4	98	0.00
38	1-Methylnaphthalene	40.000	37.006	7.5	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.413	6.5	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.727	8.2	96	0.00
42 S	2,4,6-Tribromophenol	80.000	75.938	5.1	95	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.460	6.3	95	0.00
44	2,4,5-Trichlorophenol	40.000	37.154	7.1	94	0.00
45 S	2-Fluorobiphenyl	80.000	74.268	7.2	95	0.00
46	1,1'-Biphenyl	40.000	37.753	5.6	95	0.00
47	2-Chloronaphthalene	40.000	36.334	9.2	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005786.D
 Acq On : 08 Jun 2021 21:56
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP060821

Quant Time: Jun 09 11:59:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.751	5.6	96	0.00
49	Acenaphthylene	40.000	37.118	7.2	95	0.00
50	Dimethylphthalate	40.000	36.509	8.7	96	0.00
51	2,6-Dinitrotoluene	40.000	37.205	7.0	95	0.00
52 C	Acenaphthene	40.000	36.608	8.5	96	0.00
53	3-Nitroaniline	40.000	37.489	6.3	95	0.00
54 P	2,4-Dinitrophenol	40.000	37.254	6.9	94	0.00
55	Dibenzofuran	40.000	36.338	9.2	95	0.00
56 P	4-Nitrophenol	40.000	37.712	5.7	94	0.00
57	2,4-Dinitrotoluene	40.000	37.986	5.0	95	0.00
58	Fluorene	40.000	36.398	9.0	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.893	5.3	97	0.00
60	Diethylphthalate	40.000	36.937	7.7	96	0.00
61	4-Chlorophenyl-phenylether	40.000	36.241	9.4	95	0.00
62	4-Nitroaniline	40.000	36.865	7.8	94	0.00
63	Azobenzene	40.000	36.559	8.6	94	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.858	5.4	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.334	9.2	96	0.00
67	4-Bromophenyl-phenylether	40.000	36.385	9.0	95	0.00
68	Hexachlorobenzene	40.000	35.566	11.1	94	0.00
69	Atrazine	40.000	42.797	-7.0	96	0.00
70 C	Pentachlorophenol	40.000	39.367	1.6	94	0.00
71	Phenanthrene	40.000	35.841	10.4	94	0.00
72	Anthracene	40.000	36.285	9.3	95	0.00
73	Carbazole	40.000	36.200	9.5	95	0.00
74	Di-n-butylphthalate	40.000	36.643	8.4	95	0.00
75 C	Fluoranthene	40.000	36.273	9.3	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzydine	40.000	40.420	-1.1	85	0.00
78	Pyrene	40.000	36.656	8.4	96	0.00
79 S	Terphenyl-d14	80.000	74.062	7.4	96	0.00
80	Butylbenzylphthalate	40.000	37.508	6.2	96	0.00
81	Benzo(a)anthracene	40.000	36.266	9.3	96	0.00
82	3,3'-Dichlorobenzidine	40.000	35.163	12.1	94	0.00
83	Chrysene	40.000	36.776	8.1	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.430	6.4	97	0.00
85 c	Di-n-octyl phthalate	40.000	37.631	5.9	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.597	8.5	99	0.00
88	Benzo(b)fluoranthene	40.000	37.329	6.7	100	0.00
89	Benzo(k)fluoranthene	40.000	35.074	12.3	95	0.00
90 C	Benzo(a)pyrene	40.000	36.425	8.9	98	0.00
91	Dibenzo(a,h)anthracene	40.000	36.631	8.4	99	0.00
92	Benzo(g,h,i)perylene	40.000	36.944	7.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
Data File : BP005786.D
Acq On : 08 Jun 2021 21:56
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP060821

Quant Time: Jun 09 11:59:14 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 11:54:55 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0