

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005790.D
 Acq On : 09 Jun 2021 01:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 11:59:56 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	0.00
2	1,4-Dioxane	40.000	36.751	8.1	108	0.00
3	Pyridine	40.000	38.733	3.2	108	0.00
4	n-Nitrosodimethylamine	40.000	36.469	8.8	107	0.00
5 S	2-Fluorophenol	80.000	74.912	6.4	109	0.00
6	Aniline	40.000	37.184	7.0	109	0.00
7 S	Phenol-d6	80.000	75.184	6.0	109	0.00
8	2-Chlorophenol	40.000	37.112	7.2	110	0.00
9	Benzaldehyde	40.000	41.497	-3.7	111	0.00
10 C	Phenol	40.000	36.856	7.9	109	0.00
11	bis(2-Chloroethyl)ether	40.000	36.525	8.7	108	0.00
12	1,3-Dichlorobenzene	40.000	36.917	7.7	110	0.00
13 C	1,4-Dichlorobenzene	40.000	36.540	8.7	109	0.00
14	1,2-Dichlorobenzene	40.000	36.362	9.1	108	0.00
15	Benzyl Alcohol	40.000	38.225	4.4	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.748	10.6	105	0.00
17	2-Methylphenol	40.000	37.286	6.8	109	0.00
18	Hexachloroethane	40.000	37.169	7.1	109	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.282	6.8	109	0.00
20	3+4-Methylphenols	40.000	37.892	5.3	110	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	38.069	4.8	109	0.00
23 S	Nitrobenzene-d5	80.000	77.051	3.7	110	0.00
24	Nitrobenzene	40.000	37.385	6.5	110	0.00
25	Isophorone	40.000	37.290	6.8	109	0.00
26 C	2-Nitrophenol	40.000	37.984	5.0	110	0.00
27	2,4-Dimethylphenol	40.000	37.121	7.2	109	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.443	8.9	108	0.00
29 C	2,4-Dichlorophenol	40.000	37.218	7.0	110	0.00
30	1,2,4-Trichlorobenzene	40.000	35.892	10.3	108	0.00
31	Naphthalene	40.000	36.569	8.6	109	0.00
32	Benzoic acid	40.000	37.701	5.7	106	0.00
33	4-Chloroaniline	40.000	36.572	8.6	108	0.00
34 C	Hexachlorobutadiene	40.000	36.096	9.8	108	0.00
35	Caprolactam	40.000	39.238	1.9	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.356	6.6	110	0.00
37	2-Methylnaphthalene	40.000	36.495	8.8	109	0.00
38	1-Methylnaphthalene	40.000	36.342	9.1	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.141	7.1	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.437	6.4	111	0.00
42 S	2,4,6-Tribromophenol	80.000	74.378	7.0	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.383	6.5	107	0.00
44	2,4,5-Trichlorophenol	40.000	37.484	6.3	107	0.00
45 S	2-Fluorobiphenyl	80.000	73.985	7.5	106	0.00
46	1,1'-Biphenyl	40.000	37.772	5.6	107	0.00
47	2-Chloronaphthalene	40.000	36.368	9.1	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
 Data File : BP005790.D
 Acq On : 09 Jun 2021 01:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 11:59:56 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.271	4.3	109	0.00
49	Acenaphthylene	40.000	36.797	8.0	107	0.00
50	Dimethylphthalate	40.000	36.751	8.1	108	0.00
51	2,6-Dinitrotoluene	40.000	36.895	7.8	107	0.00
52 C	Acenaphthene	40.000	36.108	9.7	107	0.00
53	3-Nitroaniline	40.000	37.553	6.1	107	0.00
54 P	2,4-Dinitrophenol	40.000	37.240	6.9	106	0.00
55	Dibenzofuran	40.000	36.049	9.9	106	0.00
56 P	4-Nitrophenol	40.000	38.074	4.8	106	0.00
57	2,4-Dinitrotoluene	40.000	38.207	4.5	108	0.00
58	Fluorene	40.000	36.005	10.0	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.433	6.4	107	0.00
60	Diethylphthalate	40.000	37.017	7.5	108	0.00
61	4-Chlorophenyl-phenylether	40.000	35.562	11.1	105	0.00
62	4-Nitroaniline	40.000	37.126	7.2	107	0.00
63	Azobenzene	40.000	37.247	6.9	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.260	4.4	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.363	9.1	107	0.00
67	4-Bromophenyl-phenylether	40.000	35.892	10.3	105	0.00
68	Hexachlorobenzene	40.000	35.286	11.8	104	0.00
69	Atrazine	40.000	42.684	-6.7	107	0.00
70 C	Pentachlorophenol	40.000	39.256	1.9	105	0.00
71	Phenanthrene	40.000	36.173	9.6	106	0.00
72	Anthracene	40.000	36.397	9.0	106	0.00
73	Carbazole	40.000	36.557	8.6	106	0.00
74	Di-n-butylphthalate	40.000	37.315	6.7	108	0.00
75 C	Fluoranthene	40.000	36.308	9.2	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzydine	40.000	45.050	-12.6	104	0.00
78	Pyrene	40.000	36.699	8.3	106	0.00
79 S	Terphenyl-d14	80.000	72.748	9.1	104	0.00
80	Butylbenzylphthalate	40.000	38.081	4.8	108	0.00
81	Benzo(a)anthracene	40.000	36.036	9.9	105	0.00
82	3,3'-Dichlorobenzidine	40.000	35.217	12.0	104	0.00
83	Chrysene	40.000	36.534	8.7	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.846	5.4	108	0.00
85 c	Di-n-octyl phthalate	40.000	38.226	4.4	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	110	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.740	8.1	111	0.01
88	Benzo(b)fluoranthene	40.000	36.574	8.6	109	0.00
89	Benzo(k)fluoranthene	40.000	35.858	10.4	108	0.00
90 C	Benzo(a)pyrene	40.000	36.346	9.1	109	0.00
91	Dibenzo(a,h)anthracene	40.000	36.641	8.4	110	0.00
92	Benzo(g,h,i)perylene	40.000	37.044	7.4	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060821\
Data File : BP005790.D
Acq On : 09 Jun 2021 01:21
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 09 11:59:56 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 11:54:55 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0