

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060822\
 Data File : BP010720.D
 Acq On : 08 Jun 2022 10:43
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 08 12:31:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 07 19:29:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	40.475	-1.2	102	0.00
3	Pyridine	40.000	41.307	-3.3	101	0.00
4	n-Nitrosodimethylamine	40.000	43.310	-8.3	105	0.00
5 S	2-Fluorophenol	80.000	83.155	-3.9	103	0.00
6	Aniline	40.000	42.139	-5.3	104	0.00
7 S	Phenol-d6	80.000	84.515	-5.6	103	0.00
8	2-Chlorophenol	40.000	41.446	-3.6	103	0.00
9	Benzaldehyde	40.000	37.352	6.6	105	0.00
10 C	Phenol	40.000	41.813	-4.5	102	0.00
11	bis(2-Chloroethyl)ether	40.000	41.479	-3.7	105	0.00
12	1,3-Dichlorobenzene	40.000	40.966	-2.4	104	0.00
13 C	1,4-Dichlorobenzene	40.000	41.234	-3.1	102	0.00
14	1,2-Dichlorobenzene	40.000	41.093	-2.7	104	0.00
15	Benzyl Alcohol	40.000	42.791	-7.0	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.046	-7.6	108	0.00
17	2-Methylphenol	40.000	42.349	-5.9	105	0.00
18	Hexachloroethane	40.000	41.492	-3.7	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.150	-5.4	105	0.00
20	3+4-Methylphenols	40.000	42.639	-6.6	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	41.493	-3.7	105	0.00
23 S	Nitrobenzene-d5	80.000	83.701	-4.6	105	0.00
24	Nitrobenzene	40.000	41.737	-4.3	105	0.00
25	Isophorone	40.000	42.175	-5.4	106	0.00
26 C	2-Nitrophenol	40.000	43.237	-8.1	106	0.00
27	2,4-Dimethylphenol	40.000	41.728	-4.3	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.694	-4.2	107	0.00
29 C	2,4-Dichlorophenol	40.000	41.952	-4.9	104	0.00
30	1,2,4-Trichlorobenzene	40.000	41.569	-3.9	106	0.00
31	Naphthalene	40.000	41.367	-3.4	105	0.00
32	Benzoic acid	40.000	39.899	0.3	102	0.00
33	4-Chloroaniline	40.000	42.237	-5.6	104	0.00
34 C	Hexachlorobutadiene	40.000	41.047	-2.6	103	0.00
35	Caprolactam	40.000	42.948	-7.4	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.512	-6.3	106	0.00
37	2-Methylnaphthalene	40.000	41.504	-3.8	105	0.00
38	1-Methylnaphthalene	40.000	41.518	-3.8	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.325	-0.8	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.255	-0.6	108	0.00
42 S	2,4,6-Tribromophenol	80.000	81.574	-2.0	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.683	-4.2	105	0.00
44	2,4,5-Trichlorophenol	40.000	42.396	-6.0	106	0.00
45 S	2-Fluorobiphenyl	80.000	81.955	-2.4	107	0.00
46	1,1'-Biphenyl	40.000	40.714	-1.8	106	0.00
47	2-Chloronaphthalene	40.000	40.650	-1.6	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060822\
 Data File : BP010720.D
 Acq On : 08 Jun 2022 10:43
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 08 12:31:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 07 19:29:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.395	-6.0	107	0.00
49	Acenaphthylene	40.000	41.448	-3.6	107	0.00
50	Dimethylphthalate	40.000	40.923	-2.3	107	0.00
51	2,6-Dinitrotoluene	40.000	42.004	-5.0	108	0.00
52 C	Acenaphthene	40.000	41.103	-2.8	107	0.00
53	3-Nitroaniline	40.000	42.547	-6.4	105	0.00
54 P	2,4-Dinitrophenol	40.000	39.882	0.3	109	0.00
55	Dibenzofuran	40.000	41.015	-2.5	107	0.00
56 P	4-Nitrophenol	40.000	41.672	-4.2	104	0.00
57	2,4-Dinitrotoluene	40.000	42.245	-5.6	106	0.00
58	Fluorene	40.000	40.982	-2.5	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.767	-4.4	106	0.00
60	Diethylphthalate	40.000	41.137	-2.8	108	0.00
61	4-Chlorophenyl-phenylether	40.000	40.504	-1.3	105	0.00
62	4-Nitroaniline	40.000	43.887	-9.7	108	0.00
63	Azobenzene	40.000	42.116	-5.3	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.090	-5.2	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.264	-3.2	107	0.00
67	4-Bromophenyl-phenylether	40.000	40.459	-1.1	105	0.00
68	Hexachlorobenzene	40.000	40.040	-0.1	107	0.00
69	Atrazine	40.000	40.981	-2.5	108	0.00
70 C	Pentachlorophenol	40.000	42.330	-5.8	107	0.00
71	Phenanthrene	40.000	40.520	-1.3	106	0.00
72	Anthracene	40.000	40.960	-2.4	107	0.00
73	Carbazole	40.000	41.643	-4.1	107	0.00
74	Di-n-butylphthalate	40.000	41.149	-2.9	108	0.00
75 C	Fluoranthene	40.000	40.577	-1.4	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	49.177	-22.9	104	0.00
78	Pyrene	40.000	40.545	-1.4	106	0.00
79 S	Terphenyl-d14	80.000	81.958	-2.4	107	0.00
80	Butylbenzylphthalate	40.000	41.534	-3.8	107	0.00
81	Benzo(a)anthracene	40.000	40.732	-1.8	104	0.00
82	3,3'-Dichlorobenzidine	40.000	41.577	-3.9	105	0.00
83	Chrysene	40.000	40.705	-1.8	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.234	-3.1	107	0.00
85 c	Di-n-octyl phthalate	40.000	41.278	-3.2	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.498	-3.7	104	0.00
88	Benzo(b)fluoranthene	40.000	41.429	-3.6	106	0.00
89	Benzo(k)fluoranthene	40.000	41.320	-3.3	103	0.00
90 C	Benzo(a)pyrene	40.000	41.978	-4.9	105	0.00
91	Dibenzo(a,h)anthracene	40.000	41.944	-4.9	105	0.00
92	Benzo(g,h,i)perylene	40.000	41.279	-3.2	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060822\
Data File : BP010720.D
Acq On : 08 Jun 2022 10:43
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 08 12:31:46 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 07 19:29:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0