

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060822\
 Data File : BP010728.D
 Acq On : 08 Jun 2022 16:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 04:12:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 07 19:29:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	39.884	0.3	97	0.00
3	Pyridine	40.000	42.866	-7.2	101	0.00
4	n-Nitrosodimethylamine	40.000	41.836	-4.6	98	0.00
5 S	2-Fluorophenol	80.000	83.397	-4.2	99	0.00
6	Aniline	40.000	43.926	-9.8	104	0.00
7 S	Phenol-d6	80.000	87.847	-9.8	104	0.00
8	2-Chlorophenol	40.000	43.165	-7.9	104	0.00
9	Benzaldehyde	40.000	39.789	0.5	108	0.00
10 C	Phenol	40.000	43.554	-8.9	102	0.00
11	bis(2-Chloroethyl)ether	40.000	42.727	-6.8	105	0.00
12	1,3-Dichlorobenzene	40.000	41.220	-3.0	101	0.00
13 C	1,4-Dichlorobenzene	40.000	41.787	-4.5	100	0.00
14	1,2-Dichlorobenzene	40.000	41.857	-4.6	103	0.00
15	Benzyl Alcohol	40.000	44.027	-10.1	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.547	-3.9	101	0.00
17	2-Methylphenol	40.000	43.732	-9.3	105	0.00
18	Hexachloroethane	40.000	40.857	-2.1	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.151	-10.4	106	0.00
20	3+4-Methylphenols	40.000	45.060	-12.7	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	40.922	-2.3	107	0.00
23 S	Nitrobenzene-d5	80.000	81.922	-2.4	106	0.00
24	Nitrobenzene	40.000	40.713	-1.8	105	0.00
25	Isophorone	40.000	42.211	-5.5	109	0.00
26 C	2-Nitrophenol	40.000	42.686	-6.7	107	0.00
27	2,4-Dimethylphenol	40.000	41.745	-4.4	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.790	-2.0	107	0.00
29 C	2,4-Dichlorophenol	40.000	42.072	-5.2	108	0.00
30	1,2,4-Trichlorobenzene	40.000	40.360	-0.9	106	0.00
31	Naphthalene	40.000	40.804	-2.0	106	0.00
32	Benzoic acid	40.000	42.822	-7.1	115	0.00
33	4-Chloroaniline	40.000	42.864	-7.2	109	0.00
34 C	Hexachlorobutadiene	40.000	40.432	-1.1	105	0.00
35	Caprolactam	40.000	46.138	-15.3	119	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.431	-8.6	112	0.00
37	2-Methylnaphthalene	40.000	41.759	-4.4	108	0.00
38	1-Methylnaphthalene	40.000	41.635	-4.1	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.632	0.9	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.469	3.8	110	0.00
42 S	2,4,6-Tribromophenol	80.000	86.414	-8.0	118	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.643	-4.1	113	0.00
44	2,4,5-Trichlorophenol	40.000	41.969	-4.9	112	0.00
45 S	2-Fluorobiphenyl	80.000	79.667	0.4	111	0.00
46	1,1'-Biphenyl	40.000	39.523	1.2	110	0.00
47	2-Chloronaphthalene	40.000	39.906	0.2	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060822\
 Data File : BP010728.D
 Acq On : 08 Jun 2022 16:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 04:12:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 07 19:29:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.099	-7.7	117	0.00
49	Acenaphthylene	40.000	40.735	-1.8	113	0.00
50	Dimethylphthalate	40.000	40.949	-2.4	115	0.00
51	2,6-Dinitrotoluene	40.000	42.687	-6.7	118	0.00
52 C	Acenaphthene	40.000	40.689	-1.7	114	0.00
53	3-Nitroaniline	40.000	42.165	-5.4	112	0.00
54 P	2,4-Dinitrophenol	40.000	42.322	-5.8	125	0.00
55	Dibenzofuran	40.000	40.689	-1.7	114	0.00
56 P	4-Nitrophenol	40.000	45.314	-13.3	121	0.00
57	2,4-Dinitrotoluene	40.000	43.372	-8.4	117	0.00
58	Fluorene	40.000	41.421	-3.6	116	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.102	-7.8	118	0.00
60	Diethylphthalate	40.000	41.642	-4.1	117	0.00
61	4-Chlorophenyl-phenylether	40.000	41.021	-2.6	114	0.00
62	4-Nitroaniline	40.000	46.774	-16.9	124	0.00
63	Azobenzene	40.000	41.207	-3.0	114	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	117	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.834	-9.6	122	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.799	-2.0	116	0.00
67	4-Bromophenyl-phenylether	40.000	40.996	-2.5	117	0.00
68	Hexachlorobenzene	40.000	40.720	-1.8	120	0.00
69	Atrazine	40.000	41.428	-3.6	120	0.00
70 C	Pentachlorophenol	40.000	45.338	-13.3	126	0.00
71	Phenanthrene	40.000	40.831	-2.1	117	0.00
72	Anthracene	40.000	41.369	-3.4	118	0.00
73	Carbazole	40.000	42.497	-6.2	120	0.00
74	Di-n-butylphthalate	40.000	41.294	-3.2	120	0.00
75 C	Fluoranthene	40.000	42.247	-5.6	120	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	123	0.00
77	Benzydine	40.000	51.624	-29.1#	128	0.00
78	Pyrene	40.000	39.114	2.2	120	0.00
79 S	Terphenyl-d14	80.000	79.498	0.6	122	0.00
80	Butylbenzylphthalate	40.000	40.394	-1.0	123	0.00
81	Benzo(a)anthracene	40.000	40.752	-1.9	122	0.00
82	3,3'-Dichlorobenzidine	40.000	44.059	-10.1	131	0.00
83	Chrysene	40.000	40.925	-2.3	126	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.314	-0.8	123	0.00
85 c	Di-n-octyl phthalate	40.000	40.692	-1.7	124	0.00
86 I	Perylene-d12	20.000	20.000	0.0	127	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.435	-6.1	131	0.00
88	Benzo(b)fluoranthene	40.000	41.143	-2.9	130	0.00
89	Benzo(k)fluoranthene	40.000	40.229	-0.6	124	0.00
90 C	Benzo(a)pyrene	40.000	40.922	-2.3	126	0.00
91	Dibenzo(a,h)anthracene	40.000	42.911	-7.3	133	0.00
92	Benzo(g,h,i)perylene	40.000	42.450	-6.1	131	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060822\
Data File : BP010728.D
Acq On : 08 Jun 2022 16:37
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 09 04:12:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 07 19:29:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0