




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060921\
Data File : BP005809.D
Acq On : 09 Jun 2021 14:31
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 09 16:37:43 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 11:54:55 2021
Response via : Initial Calibration






TIC: BP005809.D\data.ms

(77) Benzidine**19.504min (+ 0.000) 1507519.89 ng****response 658738**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	13.86
183.00	11.90	11.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060921\
 Data File : BP005809.D
 Acq On : 09 Jun 2021 14:31
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 09 16:37:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration



TIC: BP005809.D\data.ms

(70) Pentachlorophenol (C)

16.975min (+ 0.000) 139648.05 ng

response 136185

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.67
264.00	62.90	61.68
0.00	0.00	0.00