

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061021\
 Data File : BP005851.D
 Acq On : 11 Jun 2021 00:24
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 11 07:02:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	36.192	9.5	87	0.00
3	Pyridine	40.000	39.775	0.6	91	0.00
4	n-Nitrosodimethylamine	40.000	36.291	9.3	88	0.00
5 S	2-Fluorophenol	80.000	77.437	3.2	93	0.00
6	Aniline	40.000	38.237	4.4	92	0.00
7 S	Phenol-d6	80.000	78.754	1.6	94	0.00
8	2-Chlorophenol	40.000	38.809	3.0	95	0.00
9	Benzaldehyde	40.000	44.495	-11.2	97	0.00
10 C	Phenol	40.000	38.806	3.0	94	0.00
11	bis(2-Chloroethyl)ether	40.000	38.519	3.7	94	0.00
12	1,3-Dichlorobenzene	40.000	37.369	6.6	91	0.00
13 C	1,4-Dichlorobenzene	40.000	37.219	7.0	91	0.00
14	1,2-Dichlorobenzene	40.000	37.442	6.4	91	0.00
15	Benzyl Alcohol	40.000	39.055	2.4	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.432	13.9	83	0.00
17	2-Methylphenol	40.000	38.849	2.9	93	0.00
18	Hexachloroethane	40.000	37.639	5.9	91	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.612	3.5	93	0.00
20	3+4-Methylphenols	40.000	39.209	2.0	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	89	0.00
22	Acetophenone	40.000	40.638	-1.6	94	0.00
23 S	Nitrobenzene-d5	80.000	79.306	0.9	92	0.00
24	Nitrobenzene	40.000	38.521	3.7	92	0.00
25	Isophorone	40.000	38.727	3.2	92	0.00
26 C	2-Nitrophenol	40.000	40.189	-0.5	95	0.00
27	2,4-Dimethylphenol	40.000	38.379	4.1	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.871	2.8	94	0.00
29 C	2,4-Dichlorophenol	40.000	39.474	1.3	95	0.00
30	1,2,4-Trichlorobenzene	40.000	38.060	4.8	93	0.00
31	Naphthalene	40.000	38.322	4.2	93	0.00
32	Benzoic acid	40.000	39.294	1.8	90	0.01
33	4-Chloroaniline	40.000	39.515	1.2	95	0.00
34 C	Hexachlorobutadiene	40.000	38.777	3.1	95	0.00
35	Caprolactam	40.000	40.944	-2.4	92	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.511	-1.3	97	0.00
37	2-Methylnaphthalene	40.000	38.914	2.7	95	0.00
38	1-Methylnaphthalene	40.000	38.728	3.2	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	89	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.113	2.2	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.357	11.6	85	0.00
42 S	2,4,6-Tribromophenol	80.000	82.486	-3.1	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.554	3.6	91	0.00
44	2,4,5-Trichlorophenol	40.000	38.832	2.9	91	0.00
45 S	2-Fluorobiphenyl	80.000	78.655	1.7	93	0.00
46	1,1'-Biphenyl	40.000	39.920	0.2	93	0.00
47	2-Chloronaphthalene	40.000	38.510	3.7	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061021\
 Data File : BP005851.D
 Acq On : 11 Jun 2021 00:24
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 11 07:02:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.161	2.1	92	0.00
49	Acenaphthylene	40.000	38.314	4.2	91	0.00
50	Dimethylphthalate	40.000	38.928	2.7	95	0.00
51	2,6-Dinitrotoluene	40.000	38.595	3.5	92	0.00
52 C	Acenaphthene	40.000	37.984	5.0	92	0.00
53	3-Nitroaniline	40.000	39.030	2.4	92	0.00
54 P	2,4-Dinitrophenol	40.000	36.103	9.7	84	0.00
55	Dibenzofuran	40.000	38.024	4.9	92	0.00
56 P	4-Nitrophenol	40.000	40.656	-1.6	94	0.00
57	2,4-Dinitrotoluene	40.000	40.267	-0.7	94	0.00
58	Fluorene	40.000	38.121	4.7	92	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.385	6.5	88	0.00
60	Diethylphthalate	40.000	39.346	1.6	95	0.00
61	4-Chlorophenyl-phenylether	40.000	37.923	5.2	92	0.00
62	4-Nitroaniline	40.000	39.957	0.1	94	0.00
63	Azobenzene	40.000	38.047	4.9	90	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.095	4.8	87	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.286	1.8	92	0.00
67	4-Bromophenyl-phenylether	40.000	38.860	2.9	91	0.00
68	Hexachlorobenzene	40.000	38.658	3.4	91	0.00
69	Atrazine	40.000	44.602	-11.5	89	0.00
70 C	Pentachlorophenol	40.000	42.568	-6.4	91	0.00
71	Phenanthrene	40.000	38.656	3.4	91	0.00
72	Anthracene	40.000	38.636	3.4	90	0.00
73	Carbazole	40.000	39.220	2.0	91	0.00
74	Di-n-butylphthalate	40.000	40.942	-2.4	95	0.00
75 C	Fluoranthene	40.000	38.274	4.3	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
77	Benzydine	40.000	44.772	-11.9	87	0.00
78	Pyrene	40.000	37.523	6.2	91	0.00
79 S	Terphenyl-d14	80.000	79.172	1.0	95	0.00
80	Butylbenzylphthalate	40.000	40.851	-2.1	97	0.00
81	Benzo(a)anthracene	40.000	38.174	4.6	93	0.00
82	3,3'-Dichlorobenzidine	40.000	37.527	6.2	93	0.00
83	Chrysene	40.000	38.792	3.0	95	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.957	-4.9	101	0.00
85 c	Di-n-octyl phthalate	40.000	42.296	-5.7	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.223	4.4	98	0.00
88	Benzo(b)fluoranthene	40.000	38.458	3.9	98	0.00
89	Benzo(k)fluoranthene	40.000	38.710	3.2	99	0.00
90 C	Benzo(a)pyrene	40.000	38.597	3.5	99	0.00
91	Dibenzo(a,h)anthracene	40.000	38.141	4.6	98	0.00
92	Benzo(g,h,i)perylene	40.000	37.740	5.6	97	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061021\
Data File : BP005851.D
Acq On : 11 Jun 2021 00:24
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 11 07:02:06 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 11:54:55 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0