



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061021\
 Data File : BP005850.D
 Acq On : 10 Jun 2021 23:16
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 11 13:53:36 2021






AutoFind: Scans 2493, 2494, 2495; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.7	34813	PASS
68	69	0.00	2	1.7	738	PASS
69	198	0.00	100	35.1	44153	PASS
70	69	0.00	2	0.5	242	PASS
127	198	10	80	45.5	57293	PASS
197	198	0.00	2	0.2	252	PASS
198	198	100	100	100.0	125851	PASS
199	198	5	9	7.1	8894	PASS
275	198	10	60	31.3	39372	PASS
365	198	1	100	5.7	7166	PASS
441	198	0.01	100	26.3	33088	PASS
442	442	100	100	100.0	218316	PASS
443	442	15	24	19.1	41797	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061021\
 Data File : BP005850.D
 Acq On : 10 Jun 2021 23:16
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Jun 11 07:25:51 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration



TIC: BP005850.D\data.ms

(70) Pentachlorophenol (C)

16.969min (-0.006) 643983.09 ng




response 198320

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	64.49
264.00	62.90	63.32
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061021\
Data File : BP005850.D
Acq On : 10 Jun 2021 23:16
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Jun 11 07:25:51 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 11:54:55 2021
Response via : Initial Calibration



TIC: BP005850.D\data.ms

(77) Benzidine

19.504min (-19.504) 0.00 ng

response 0

Ion	Exp%	Act%
184.00	100.00	0.00
185.00	13.70	0.00#
183.00	11.90	0.00#
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/10/2021	BNA_P	BP005850.D
Compound Name	Response	Retention Time
DDT	408820	20.692
DDD	40419	20.263
DDE	1566	19.775
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
41985	450805	9.31

Instrument :
BNA_P
ClientSampleId :
DFTPP