

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061025\
 Data File : BP024892.D
 Acq On : 10 Jun 2025 13:43
 Operator : RC/JU
 Sample : Q2265-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 TP11-MHL-WC

Quant Time: Jun 10 14:10:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 06 16:20:27 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.608	152	209510	20.000	ng	0.00
21) Naphthalene-d8	10.372	136	824081	20.000	ng	0.00
39) Acenaphthene-d10	14.248	164	530605	20.000	ng	0.00
64) Phenanthrene-d10	17.060	188	1084172	20.000	ng	0.00
76) Chrysene-d12	21.483	240	1201133	20.000	ng	0.00
86) Perylene-d12	24.730	264	1427181	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.237	112	1197486	95.406	ng	0.00
7) Phenol-d6	6.814	99	1508721	90.849	ng	0.00
23) Nitrobenzene-d5	8.761	82	950157	56.027	ng	0.00
42) 2,4,6-Tribromophenol	15.778	330	742068	101.156	ng	0.00
45) 2-Fluorobiphenyl	12.849	172	2210917	56.132	ng	-0.01
79) Terphenyl-d14	19.783	244	3779952	56.402	ng	-0.01
Target Compounds						
60) Diethylphthalate	15.095	149	118256	3.031	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061025\
Data File : BP024892.D
Acq On : 10 Jun 2025 13:43
Operator : RC/JU
Sample : Q2265-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
TP11-MHL-WC

Quant Time: Jun 10 14:10:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060625.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 06 16:20:27 2025
Response via : Initial Calibration

