

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061121\
 Data File : BP005865.D
 Acq On : 11 Jun 2021 11:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 11 12:47:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	67477	20.00	ng	-0.01
21) Naphthalene-d8	10.752	136	252382	20.00	ng	-0.01
39) Acenaphthene-d10	14.575	164	154430	20.00	ng	0.00
64) Phenanthrene-d10	17.322	188	312773	20.00	ng	0.00
76) Chrysene-d12	21.392	240	362598	20.00	ng	0.00
86) Perylene-d12	23.816	264	422192	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.511	112	325627	77.44	ng	0.00
7) Phenol-d6	7.111	99	410867	75.38	ng	-0.01
23) Nitrobenzene-d5	9.111	82	374337	72.72	ng	-0.01
42) 2,4,6-Tribromophenol	16.063	330	193304	72.94	ng	0.00
45) 2-Fluorobiphenyl	13.204	172	843313	71.71	ng	-0.01
79) Terphenyl-d14	19.869	244	1389320	71.75	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.381	88	69114	36.38	ng	99
3) Pyridine	3.793	79	177124	39.67	ng	98
4) n-Nitrosodimethylamine	3.699	42	74543	35.92	ng	93
6) Aniline	7.269	93	231046	37.97	ng	98
8) 2-Chlorophenol	7.511	128	186641	38.71	ng	98
9) Benzaldehyde	7.087	77	80272	34.55	ng	98
10) Phenol	7.140	94	207817	38.17	ng	99
11) bis(2-Chloroethyl)ether	7.369	93	157058	38.07	ng	99
12) 1,3-Dichlorobenzene	7.834	146	205772	38.17	ng	99
13) 1,4-Dichlorobenzene	7.981	146	208169	37.86	ng	97
14) 1,2-Dichlorobenzene	8.299	146	197970	37.67	ng	99
15) Benzyl Alcohol	8.187	79	154075	37.53	ng	96
16) 2,2'-oxybis(1-Chloropr...	8.481	45	127110	33.33	ng	97
17) 2-Methylphenol	8.393	107	141676	38.19	ng	99
18) Hexachloroethane	9.028	117	74568	37.49	ng	95
19) n-Nitroso-di-n-propyla...	8.758	70	123613	36.87	ng	97
20) 3+4-Methylphenols	8.722	107	188787	37.67	ng	97
22) Acetophenone	8.775	105	243120	36.60	ng	99
24) Nitrobenzene	9.152	77	193506	36.20	ng	99
25) Isophorone	9.681	82	340873	35.86	ng	99
26) 2-Nitrophenol	9.869	139	96381	37.79	ng	95
27) 2,4-Dimethylphenol	9.928	122	137100	36.08	ng	99
28) bis(2-Chloroethoxy)met...	10.163	93	180319	36.34	ng	99
29) 2,4-Dichlorophenol	10.405	162	170072	37.24	ng	99
30) 1,2,4-Trichlorobenzene	10.616	180	183595	35.78	ng	99
31) Naphthalene	10.805	128	530906	36.36	ng	99
32) Benzoic acid	10.081	122	95783	35.53	ng	98
33) 4-Chloroaniline	10.916	127	215860	36.59	ng	98
34) Hexachlorobutadiene	11.087	225	125761	36.24	ng	97
35) Caprolactam	11.704	113	46432	35.30	ng	94
36) 4-Chloro-3-methylphenol	12.034	107	168232	36.28	ng	97
37) 2-Methylnaphthalene	12.410	142	373445	35.92	ng	100
38) 1-Methylnaphthalene	12.628	142	350754	35.82	ng	98
40) 1,2,4,5-Tetrachloroben...	12.775	216	194015	35.69	ng	99
41) Hexachlorocyclopentadiene	12.757	237	89380	35.98	ng	97
43) 2,4,6-Trichlorophenol	13.016	196	131103	35.16	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061121\
 Data File : BP005865.D
 Acq On : 11 Jun 2021 11:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 11 12:47:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.087	196	144370	35.95	ng	96
46) 1,1'-Biphenyl	13.416	154	442012	35.50	ng	98
47) 2-Chloronaphthalene	13.457	162	364007	35.80	ng	97
48) 2-Nitroaniline	13.657	65	95756	35.82	ng	96
49) Acenaphthylene	14.298	152	556378	35.67	ng	100
50) Dimethylphthalate	14.034	163	449688	35.42	ng	99
51) 2,6-Dinitrotoluene	14.151	165	102867	35.64	ng	93
52) Acenaphthene	14.640	154	333071	35.16	ng	100
53) 3-Nitroaniline	14.481	138	101839	36.41	ng	97
54) 2,4-Dinitrophenol	14.693	184	54351	34.59	ng	95
55) Dibenzofuran	14.975	168	547426	35.14	ng	97
56) 4-Nitrophenol	14.787	139	82451	36.37	ng	97
57) 2,4-Dinitrotoluene	14.940	165	141213	36.58	ng	98
58) Fluorene	15.622	166	429167	35.36	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.198	232	124259m	35.11	ng	
60) Diethylphthalate	15.393	149	453527	35.60	ng	100
61) 4-Chlorophenyl-phenyle...	15.610	204	219291	34.84	ng	96
62) 4-Nitroaniline	15.645	138	104590	36.33	ng	96
63) Azobenzene	15.904	77	398511	34.58	ng	97
65) 4,6-Dinitro-2-methylph...	15.704	198	85197	37.07	ng	100
66) n-Nitrosodiphenylamine	15.828	169	380644	36.70	ng	99
67) 4-Bromophenyl-phenylether	16.510	248	143060	35.64	ng	97
68) Hexachlorobenzene	16.628	284	173179	35.99	ng	99
69) Atrazine	16.781	200	120653	37.39	ng	99
70) Pentachlorophenol	16.969	266	100859	37.70	ng	99
71) Phenanthrene	17.363	178	678321	36.11	ng	99
72) Anthracene	17.451	178	674254	36.47	ng	99
73) Carbazole	17.722	167	649783	36.61	ng	100
74) Di-n-butylphthalate	18.269	149	758440	36.88	ng	100
75) Fluoranthene	19.322	202	820738	35.98	ng	98
77) Benzidine	19.498	184	308063	40.83	ng	99
78) Pyrene	19.675	202	855907	33.81	ng	99
80) Butylbenzylphthalate	20.539	149	369287	35.46	ng	97
81) Benzo(a)anthracene	21.375	228	930530	35.55	ng	100
82) 3,3'-Dichlorobenzidine	21.304	252	323824	35.78	ng	96
83) Chrysene	21.428	228	905937	35.73	ng	100
84) Bis(2-ethylhexyl)phtha...	21.298	149	566257	37.07	ng	99
85) Di-n-octyl phthalate	22.222	149	1032208	37.78	ng	99
87) Indeno(1,2,3-cd)pyrene	26.357	276	1341368	37.15	ng	99
88) Benzo(b)fluoranthene	23.074	252	1095253	37.88	ng	100
89) Benzo(k)fluoranthene	23.122	252	1011706	36.08	ng	100
90) Benzo(a)pyrene	23.704	252	1012188	37.23	ng	99
91) Dibenzo(a,h)anthracene	26.380	278	1164306	37.47	ng	99
92) Benzo(g,h,i)perylene	27.151	276	1134559	36.96	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061121\
 Data File : BP005865.D
 Acq On : 11 Jun 2021 11:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 11 12:47:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 11:54:55 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.946 min Scan# 826
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 67477
Ion Ratio Lower Upper
152 100
150 167.0 125.6 188.4
115 65.1 48.5 72.7



#2
1,4-Dioxane
Concen: 36.38 ng
RT: 3.381 min Scan# 50
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 88 Resp: 69114
Ion Ratio Lower Upper
88 100
58 59.4 46.7 70.1
43 25.1 20.3 30.5









#3
Pyridine
Concen: 39.67 ng
RT: 3.793 min Scan# 120
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 177124
Ion Ratio Lower Upper
79 100
52 50.7 41.7 62.5
51 28.3 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 35.92 ng
RT: 3.699 min Scan# 104
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 42 Resp: 74543
Ion Ratio Lower Upper
42 100
74 140.1 105.4 158.0
44 6.3 4.2 6.4









#5
2-Fluorophenol
Concen: 77.44 ng
RT: 5.511 min Scan# 412
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 325627
Ion Ratio Lower Upper
112 100
64 51.3 41.4 62.2
63 28.0 23.8 35.8



#6
Aniline
Concen: 37.97 ng
RT: 7.269 min Scan# 711
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 93 Resp: 231046
Ion Ratio Lower Upper
93 100
66 38.8 29.7 44.5
65 18.8 15.3 22.9









#7
Phenol-d6
Concen: 75.38 ng
RT: 7.111 min Scan# 684
Delta R.T. -0.011 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 99 Resp: 410867
Ion Ratio Lower Upper
99 100
42 16.3 13.4 20.0
71 39.4 30.6 46.0



#8
2-Chlorophenol
Concen: 38.71 ng
RT: 7.511 min Scan# 752
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 128 Resp: 186641
Ion Ratio Lower Upper
128 100
130 32.6 12.4 52.4
64 38.4 20.3 60.3









#9
Benzaldehyde
Concen: 34.55 ng
RT: 7.087 min Scan# 680
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 80272
Ion Ratio Lower Upper
77 100
105 99.5 76.7 116.7
106 97.0 76.2 116.2



#10
Phenol
Concen: 38.17 ng
RT: 7.140 min Scan# 689
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 94 Resp: 207817
Ion Ratio Lower Upper
94 100
65 28.4 8.8 48.8
66 41.5 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 38.07 ng
RT: 7.369 min Scan# 728
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 157058
Ion Ratio Lower Upper
93 100
63 65.9 46.2 86.2
95 32.7 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 38.17 ng
RT: 7.834 min Scan# 807
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 146 Resp: 205772
Ion Ratio Lower Upper
146 100
148 65.3 51.5 77.3
75 32.1 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 37.86 ng
RT: 7.981 min Scan# 832
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 208169
Ion Ratio Lower Upper
146 100
148 66.8 51.3 76.9
111 42.8 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 37.67 ng
RT: 8.299 min Scan# 886
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:146 Resp: 197970
Ion Ratio Lower Upper
146 100
148 63.6 50.3 75.5
111 43.6 33.7 50.5









#15
Benzyl Alcohol
Concen: 37.53 ng
RT: 8.187 min Scan# 867
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79	Resp: 154075
Ion Ratio	Lower Upper
79 100	
108 73.2	62.5 93.7
77 67.6	52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 33.33 ng
RT: 8.481 min Scan# 917
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 45	Resp: 127110
Ion Ratio	Lower Upper
45 100	
77 24.0	2.0 42.0
79 17.5	0.0 36.7









#17
2-Methylphenol
Concen: 38.19 ng
RT: 8.393 min Scan# 902
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:107	Resp: 141676
Ion Ratio	Lower Upper
107 100	
108 110.7	87.7 131.5
77 49.8	40.2 60.2
79 49.2	39.0 58.4



#18
Hexachloroethane
Concen: 37.49 ng
RT: 9.028 min Scan# 1010
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:117	Resp: 74568
Ion Ratio	Lower Upper
117 100	
119 96.4	76.2 114.2
201 112.6	83.4 125.0









#19

n-Nitroso-di-n-propylamine
 Concen: 36.87 ng
 RT: 8.758 min Scan# 964
 Delta R.T. -0.012 min
 Lab File: BP005865.D
 Acq: 11 Jun 2021 11:38

Instrument :
 BNA_P
 ClientSampleId :
 SSTCCCC040




Tgt Ion: 70	Resp: 123613
Ion Ratio	Lower Upper
70 100	
42 47.9	40.8 61.2
101 10.4	8.8 13.2
130 23.6	18.9 28.3




#20

3+4-Methylphenols
 Concen: 37.67 ng
 RT: 8.722 min Scan# 958
 Delta R.T. -0.012 min
 Lab File: BP005865.D
 Acq: 11 Jun 2021 11:38

Tgt Ion: 107	Resp: 188787
Ion Ratio	Lower Upper
107 100	
108 87.4	71.0 111.0
77 35.5	16.6 56.6
79 33.2	12.7 52.7









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.752 min Scan# 1303
Delta R.T. -0.011 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	252382
Ion Ratio	Lower	Upper	
136	100		
137	11.2	8.6	12.8
54	5.6	4.6	6.8
68	4.4	3.6	5.4



#22
Acetophenone
Concen: 36.60 ng
RT: 8.775 min Scan# 967
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:	105	Resp:	243120
Ion Ratio	Lower	Upper	
105	100		
71	2.4	1.9	2.9
51	22.1	18.4	27.6
120	22.0	17.9	26.9









#23
Nitrobenzene-d5
Concen: 72.72 ng
RT: 9.111 min Scan# 1024
Delta R.T. -0.011 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 374337
Ion Ratio Lower Upper
82 100
128 44.2 35.7 53.5
54 42.2 35.0 52.4



#24
Nitrobenzene
Concen: 36.20 ng
RT: 9.152 min Scan# 1031
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 77 Resp: 193506
Ion Ratio Lower Upper
77 100
123 47.4 38.2 57.2
65 14.1 11.4 17.2









#25
Isophorone
Concen: 35.86 ng
RT: 9.681 min Scan# 1121
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	82	Resp:	340873
Ion	Ratio	Lower	Upper
82	100		
95	8.7	6.5	9.7
138	18.7	15.0	22.4



#26
2-Nitrophenol
Concen: 37.79 ng
RT: 9.869 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:	139	Resp:	96381
Ion	Ratio	Lower	Upper
139	100		
109	35.9	28.8	43.2
65	45.8	41.0	61.6









#27
2,4-Dimethylphenol
Concen: 36.08 ng
RT: 9.928 min Scan# 1163
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 137100
Ion Ratio Lower Upper
122 100
107 118.5 94.5 141.7
121 58.7 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 36.34 ng
RT: 10.163 min Scan# 1203
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 93 Resp: 180319
Ion Ratio Lower Upper
93 100
95 32.5 25.9 38.9
123 11.9 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 37.24 ng
RT: 10.405 min Scan# 1244
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 170072
Ion Ratio Lower Upper
162 100
164 63.4 44.9 84.9
98 34.0 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 35.78 ng
RT: 10.616 min Scan# 1280
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:180 Resp: 183595
Ion Ratio Lower Upper
180 100
182 96.3 77.9 116.9
145 31.9 24.7 37.1









#31
Naphthalene
Concen: 36.36 ng
RT: 10.805 min Scan# 1312
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 530906
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.6
127 13.1 10.7 16.1



#32
Benzoic acid
Concen: 35.53 ng
RT: 10.081 min Scan# 1189
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:122 Resp: 95783
Ion Ratio Lower Upper
122 100
105 125.8 103.2 143.2
77 97.1 76.2 116.2









#33
4-Chloroaniline
Concen: 36.59 ng
RT: 10.916 min Scan# 1331
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	215860
Ion	Ratio	Lower	Upper
127	100		
129	31.8	26.1	39.1
65	29.0	22.5	33.7
92	19.7	15.0	22.6



#34
Hexachlorobutadiene
Concen: 36.24 ng
RT: 11.087 min Scan# 1360
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:	225	Resp:	125761
Ion	Ratio	Lower	Upper
225	100		
223	60.8	51.0	76.6
227	63.1	51.4	77.2









#35
Caprolactam
Concen: 35.30 ng
RT: 11.704 min Scan# 1465
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 46432
Ion Ratio Lower Upper
113 100
55 107.8 98.9 138.9
56 93.0 71.7 111.7



#36
4-Chloro-3-methylphenol
Concen: 36.28 ng
RT: 12.034 min Scan# 1521
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:107 Resp: 168232
Ion Ratio Lower Upper
107 100
144 26.3 21.1 31.7
142 79.2 66.0 99.0









#37
2-Methylnaphthalene
Concen: 35.92 ng
RT: 12.410 min Scan# 1585
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 373445
Ion Ratio Lower Upper
142 100
141 89.4 71.8 107.6
115 39.3 31.6 47.4



#38
1-Methylnaphthalene
Concen: 35.82 ng
RT: 12.628 min Scan# 1622
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:142 Resp: 350754
Ion Ratio Lower Upper
142 100
141 93.2 73.3 109.9
116 4.5 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.575 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 154430
Ion Ratio Lower Upper
164 100
162 99.8 80.2 120.2
160 45.6 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 35.69 ng
RT: 12.775 min Scan# 1647
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:216 Resp: 194015
Ion Ratio Lower Upper
216 100
214 78.2 61.9 92.9
179 21.6 16.8 25.2
108 18.1 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 35.98 ng
RT: 12.757 min Scan# 1644
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:237	Resp:	89380
Ion Ratio	Lower	Upper
237	100	
235	61.8	44.5 84.5
272	13.4	0.0 32.5



#42
2,4,6-Tribromophenol
Concen: 72.94 ng
RT: 16.063 min Scan# 2206
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:330	Resp:	193304
Ion Ratio	Lower	Upper
330	100	
332	97.8	77.8 116.8
141	32.0	25.4 38.2









#43
2,4,6-Trichlorophenol
Concen: 35.16 ng
RT: 13.016 min Scan# 1688
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 131103
Ion Ratio Lower Upper
196 100
198 97.1 79.5 119.3
200 32.1 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 35.95 ng
RT: 13.087 min Scan# 1700
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:196 Resp: 144370
Ion Ratio Lower Upper
196 100
198 95.2 74.7 112.1
97 51.5 37.4 56.0
132 28.3 20.0 30.0









#45
2-Fluorobiphenyl
Concen: 71.71 ng
RT: 13.204 min Scan# 1720
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 843313
Ion Ratio Lower Upper
172 100
171 35.6 27.9 41.9
170 23.5 18.6 27.8



#46
1,1'-Biphenyl
Concen: 35.50 ng
RT: 13.416 min Scan# 1756
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:154 Resp: 442012
Ion Ratio Lower Upper
154 100
153 39.6 20.4 60.4
76 9.9 0.0 30.8









#47
2-Chloronaphthalene
Concen: 35.80 ng
RT: 13.457 min Scan# 1763
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 364007
Ion Ratio Lower Upper
162 100
127 42.1 31.4 47.0
164 32.7 26.6 39.8



#48
2-Nitroaniline
Concen: 35.82 ng
RT: 13.657 min Scan# 1797
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion: 65 Resp: 95756
Ion Ratio Lower Upper
65 100
92 75.8 57.7 86.5
138 119.2 92.1 138.1









#49
Acenaphthylene
Concen: 35.67 ng
RT: 14.298 min Scan# 1906
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 556378
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2
153 13.2 10.6 16.0



#50
Dimethylphthalate
Concen: 35.42 ng
RT: 14.034 min Scan# 1861
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:163 Resp: 449688
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 9.8 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 35.64 ng
RT: 14.151 min Scan# 1881
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 102867
Ion Ratio Lower Upper
165 100
63 56.9 41.3 61.9
89 60.6 44.8 67.2



#52
Acenaphthene
Concen: 35.16 ng
RT: 14.640 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:154 Resp: 333071
Ion Ratio Lower Upper
154 100
153 112.4 89.4 134.2
152 53.5 43.0 64.4









#53
3-Nitroaniline
Concen: 36.41 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 101839
Ion Ratio Lower Upper
138 100
108 10.5 8.4 12.6
92 106.1 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 34.59 ng
RT: 14.693 min Scan# 1973
Delta R.T. -0.005 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:184 Resp: 54351
Ion Ratio Lower Upper
184 100
63 60.5 47.0 70.6
154 71.5 52.3 78.5









#55
Dibenzenofuran
Concen: 35.14 ng
RT: 14.975 min Scan# 2021
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 547426
Ion Ratio Lower Upper
168 100
139 41.2 31.0 46.6
169 12.9 10.8 16.2



#56
4-Nitrophenol
Concen: 36.37 ng
RT: 14.787 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:139 Resp: 82451
Ion Ratio Lower Upper
139 100
109 91.9 72.3 112.3
65 88.0 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 36.58 ng
RT: 14.940 min Scan# 2015
Delta R.T. -0.005 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	165	Resp:	141213
Ion Ratio	Lower	Upper	
165	100		
63	42.5	32.3	48.5
89	71.3	55.9	83.9
182	2.8	2.2	3.4



#58
Fluorene
Concen: 35.36 ng
RT: 15.622 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:	166	Resp:	429167
Ion Ratio	Lower	Upper	
166	100		
165	97.8	78.0	117.0
167	13.3	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 35.11 ng m
RT: 15.198 min Scan# 2059
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	232	Resp:	124259
Ion Ratio	Lower	Upper	
232	100		
131	41.9	32.2	48.4
130	2.6	2.0	3.0
166	25.7	19.4	29.2



#60
Diethylphthalate
Concen: 35.60 ng
RT: 15.393 min Scan# 2092
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:	149	Resp:	453527
Ion Ratio	Lower	Upper	
149	100		
177	21.9	17.5	26.3
150	12.7	9.9	14.9









#61
4-Chlorophenyl-phenylether
Concen: 34.84 ng
RT: 15.610 min Scan# 2129
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 219291
Ion Ratio Lower Upper
204 100
206 33.0 26.6 40.0
141 65.6 48.6 73.0



#62
4-Nitroaniline
Concen: 36.33 ng
RT: 15.645 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:138 Resp: 104590
Ion Ratio Lower Upper
138 100
92 48.9 31.2 71.2
108 103.2 78.6 118.6









#63
Azobenzene
Concen: 34.58 ng
RT: 15.904 min Scan# 2179
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
77	100		
182	27.6	6.0	46.0
105	20.3	0.0	38.8
51	29.8	8.8	48.8



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion	Ratio	Lower	Upper
188	100		
94	7.3	5.6	8.4
80	7.4	6.3	9.5









#65
4,6-Dinitro-2-methylphenol
Concen: 37.07 ng
RT: 15.704 min Scan# 2145
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 85197
Ion Ratio Lower Upper
198 100
51 34.1 14.0 54.0
105 39.5 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 36.70 ng
RT: 15.828 min Scan# 2166
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:169 Resp: 380644
Ion Ratio Lower Upper
169 100
168 66.8 54.4 81.6
167 35.8 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 35.64 ng
RT: 16.510 min Scan# 2282
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 143060
Ion Ratio Lower Upper
248 100
250 96.2 79.8 119.8
141 71.4 57.8 86.6



#68
Hexachlorobenzene
Concen: 35.99 ng
RT: 16.628 min Scan# 2302
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:284 Resp: 173179
Ion Ratio Lower Upper
284 100
142 32.8 26.9 40.3
249 30.2 24.5 36.7









#69
Atrazine
Concen: 37.39 ng
RT: 16.781 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 120653
Ion Ratio Lower Upper
200 100
173 25.0 4.1 44.1
215 47.7 27.7 67.7



#70
Pentachlorophenol
Concen: 37.70 ng
RT: 16.969 min Scan# 2360
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:266 Resp: 100859
Ion Ratio Lower Upper
266 100
268 63.7 51.6 77.4
264 61.8 50.3 75.5









#71
Phenanthrene
Concen: 36.11 ng
RT: 17.363 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 678321
Ion Ratio Lower Upper
178 100
176 18.7 15.1 22.7
179 14.9 12.4 18.6



#72
Anthracene
Concen: 36.47 ng
RT: 17.451 min Scan# 2442
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:178 Resp: 674254
Ion Ratio Lower Upper
178 100
176 18.2 14.9 22.3
179 15.4 12.2 18.2









#73
 Carbazole
 Concen: 36.61 ng
 RT: 17.722 min Scan# 2488
 Delta R.T. -0.006 min
 Lab File: BP005865.D
 Acq: 11 Jun 2021 11:38




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:167	Resp: 649783
Ion Ratio	Lower Upper
167	100
166	21.4 17.1 25.7
139	13.9 10.7 16.1



#74
 Di-n-butylphthalate
 Concen: 36.88 ng
 RT: 18.269 min Scan# 2581
 Delta R.T. -0.006 min
 Lab File: BP005865.D
 Acq: 11 Jun 2021 11:38

Tgt Ion:149	Resp: 758440
Ion Ratio	Lower Upper
149	100
150	9.2 7.3 10.9
104	5.6 4.5 6.7









#75
Fluoranthene
Concen: 35.98 ng
RT: 19.322 min Scan# 2760
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 820738
Ion Ratio Lower Upper
202 100
101 8.1 0.0 27.6
203 17.0 0.0 37.8



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.392 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:240 Resp: 362598
Ion Ratio Lower Upper
240 100
120 8.2 6.5 9.7
236 25.4 20.6 31.0









#77
Benzidine
Concen: 40.83 ng
RT: 19.498 min Scan# 2790
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 308063
Ion Ratio Lower Upper
184 100
185 14.4 11.0 16.4
183 12.1 9.5 14.3



#78
Pyrene
Concen: 33.81 ng
RT: 19.675 min Scan# 2820
Delta R.T. 0.000 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:202 Resp: 855907
Ion Ratio Lower Upper
202 100
200 20.9 16.3 24.5
203 17.2 14.2 21.2









#79
 Terphenyl-d14
 Concen: 71.75 ng
 RT: 19.869 min Scan# 2853
 Delta R.T. -0.000 min
 Lab File: BP005865.D
 Acq: 11 Jun 2021 11:38




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:244 Resp: 1389320
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.0 9.0
 122 8.2 7.1 10.7



#80
 Butylbenzylphthalate
 Concen: 35.46 ng
 RT: 20.539 min Scan# 2967
 Delta R.T. 0.000 min
 Lab File: BP005865.D
 Acq: 11 Jun 2021 11:38

Tgt Ion:149 Resp: 369287
 Ion Ratio Lower Upper
 149 100
 91 65.3 54.6 81.8
 206 22.1 16.7 25.1









#81
Benzo(a)anthracene
Concen: 35.55 ng
RT: 21.375 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 930530
Ion Ratio Lower Upper
228 100
226 27.2 21.6 32.4
229 19.6 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 35.78 ng
RT: 21.304 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:252 Resp: 323824
Ion Ratio Lower Upper
252 100
254 66.4 50.6 76.0
126 10.4 7.0 10.6









#83
Chrysene
Concen: 35.73 ng
RT: 21.428 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 905937
Ion Ratio Lower Upper
228 100
226 29.3 23.8 35.6
229 19.3 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 37.07 ng
RT: 21.298 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:149 Resp: 566257
Ion Ratio Lower Upper
149 100
167 27.9 21.9 32.9
279 5.5 4.3 6.5









#85
Di-n-octyl phthalate
Concen: 37.78 ng
RT: 22.222 min Scan# 3253
Delta R.T. 0.000 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 1032208
Ion Ratio Lower Upper
149 100
167 1.6 1.4 2.0
43 6.6 5.6 8.4



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.816 min Scan# 3524
Delta R.T. 0.000 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:264 Resp: 422192
Ion Ratio Lower Upper
264 100
260 23.9 18.6 28.0
265 22.2 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 37.15 ng
RT: 26.357 min Scan# 3956
Delta R.T. -0.012 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 1341368
Ion Ratio Lower Upper
276 100
138 20.1 15.3 22.9
277 25.5 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 37.88 ng
RT: 23.074 min Scan# 3398
Delta R.T. 0.000 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:252 Resp: 1095253
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 7.6 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 36.08 ng
RT: 23.122 min Scan# 3406
Delta R.T. 0.000 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 1011706
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 7.6 6.0 9.0



#90
Benzo(a)pyrene
Concen: 37.23 ng
RT: 23.704 min Scan# 3505
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:252 Resp: 1012188
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 8.8 6.5 9.7









#91
Dibenzo(a,h)anthracene
Concen: 37.47 ng
RT: 26.380 min Scan# 3960
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 1164306
Ion Ratio Lower Upper
278 100
139 11.7 9.3 13.9
279 22.9 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 36.96 ng
RT: 27.151 min Scan# 4091
Delta R.T. -0.006 min
Lab File: BP005865.D
Acq: 11 Jun 2021 11:38

Tgt Ion:276 Resp: 1134559
Ion Ratio Lower Upper
276 100
277 23.2 18.9 28.3
138 16.9 12.6 18.8

