




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061224\
Data File : BP020596.D
Acq On : 12 Jun 2024 10:01
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 12 12:28:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration



TIC: BP020596.D\data.ms

(70) Pentachlorophenol (C)




16.816min (-0.012) 111404.76 ng

response 428519

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	63.86
264.00	63.30	63.03
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061224\
Data File : BP020596.D
Acq On : 12 Jun 2024 10:01
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 12 12:28:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration



TIC: BP020596.D\data.ms

(77) Benzidine**19.516min (-0.017) 31494.05 ng****response 2753872**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.01
183.00	11.70	12.12
0.00	0.00	0.00