



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061324\
Data File : BP020609.D
Acq On : 13 Jun 2024 12:08
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu May 30 03:41:03 2024






AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.1	154552	PASS
68	69	0.00	2	1.9	3049	PASS
69	198	0.00	100	30.9	164005	PASS
70	69	0.00	2	0.5	831	PASS
127	198	10	80	39.5	209846	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	531383	PASS
199	198	5	9	6.8	36093	PASS
275	198	10	60	29.9	158738	PASS
365	198	1	100	3.7	19791	PASS
441	198	0.01	100	16.4	86947	PASS
442	442	100	100	100.0	556923	PASS
443	442	15	24	19.1	106404	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061324\
Data File : BP020609.D
Acq On : 13 Jun 2024 12:08
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Jun 13 13:13:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration



TIC: BP020609.D\data.ms

(70) Pentachlorophenol (C)

16.816min (-0.012) 1657217.48 ng




response 715010

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	65.47
264.00	63.30	64.61
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061324\
Data File : BP020609.D
Acq On : 13 Jun 2024 12:08
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Jun 13 13:13:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration



TIC: BP020609.D\data.ms

(77) Benzidine

19.528min (-0.005) 537596.94 ng

response 3099504

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	13.96
183.00	11.70	12.08
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/13/2024	BNA_P	<u>BP020609.D</u>
Compound Name	Response	Retention Time
DDT	1536038	20.81
DDD	25818	20.398
DDE	2162	19.822
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
27980	1564018	1.79

Instrument :
BNA_P
ClientSampleId :
DFTPP