

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061421\
 Data File : BP005915.D
 Acq On : 14 Jun 2021 20:21
 Operator : CG/JU
 Sample : M2625-25RE
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B12(154-158)RE

Quant Time: Jun 15 04:17:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 14 13:15:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	64872	20.000	ng	0.00
21) Naphthalene-d8	10.740	136	229563	20.000	ng	0.00
39) Acenaphthene-d10	14.563	164	135715	20.000	ng	0.00
64) Phenanthrene-d10	17.304	188	287780	20.000	ng	0.00
76) Chrysene-d12	21.375	240	362174	20.000	ng	0.00
86) Perylene-d12	23.786	264	424763	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.505	112	523169	129.407	ng	0.00
7) Phenol-d6	7.111	99	534256	101.956	ng	0.00
23) Nitrobenzene-d5	9.099	82	421188	89.952	ng	0.00
42) 2,4,6-Tribromophenol	16.051	330	328912	141.226	ng	0.00
45) 2-Fluorobiphenyl	13.193	172	918062	88.829	ng	0.00
79) Terphenyl-d14	19.857	244	1632634	84.409	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061421\
Data File : BP005915.D
Acq On : 14 Jun 2021 20:21
Operator : CG/JU
Sample : M2625-25RE
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
18-B12(154-158)RE

Quant Time: Jun 15 04:17:44 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 14 13:15:03 2021
Response via : Initial Calibration

