

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061521\
 Data File : BP005922.D
 Acq On : 15 Jun 2021 10:12
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 11:31:56 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 14 13:15:03 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	76	0.00
2	1,4-Dioxane	40.000	41.624	-4.1	83	0.00
3	Pyridine	40.000	47.114	-17.8	89	0.00
4	n-Nitrosodimethylamine	40.000	39.908	0.2	80	0.00
5 S	2-Fluorophenol	80.000	90.230	-12.8	89	0.00
6	Aniline	40.000	45.272	-13.2	90	0.00
7 S	Phenol-d6	80.000	90.235	-12.8	89	0.00
8	2-Chlorophenol	40.000	45.068	-12.7	91	0.00
9	Benzaldehyde	40.000	46.278	-15.7	84	0.00
10 C	Phenol	40.000	45.256	-13.1	91	0.00
11	bis(2-Chloroethyl)ether	40.000	45.534	-13.8	92	0.00
12	1,3-Dichlorobenzene	40.000	44.808	-12.0	90	0.00
13 C	1,4-Dichlorobenzene	40.000	44.074	-10.2	89	0.00
14	1,2-Dichlorobenzene	40.000	44.572	-11.4	90	0.00
15	Benzyl Alcohol	40.000	44.310	-10.8	87	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.358	-3.4	82	0.00
17	2-Methylphenol	40.000	46.033	-15.1	91	0.00
18	Hexachloroethane	40.000	43.279	-8.2	86	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.217	-10.5	88	0.00
20	3+4-Methylphenols	40.000	46.446	-16.1	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	77	0.00
22	Acetophenone	40.000	43.166	-7.9	87	0.00
23 S	Nitrobenzene-d5	80.000	82.822	-3.5	84	0.00
24	Nitrobenzene	40.000	41.293	-3.2	86	0.00
25	Isophorone	40.000	42.123	-5.3	87	0.00
26 C	2-Nitrophenol	40.000	44.161	-10.4	91	0.00
27	2,4-Dimethylphenol	40.000	42.226	-5.6	88	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.575	-8.9	92	0.00
29 C	2,4-Dichlorophenol	40.000	42.913	-7.3	90	0.00
30	1,2,4-Trichlorobenzene	40.000	41.589	-4.0	88	0.00
31	Naphthalene	40.000	42.463	-6.2	90	0.00
32	Benzoic acid	40.000	42.169	-5.4	85	0.00
33	4-Chloroaniline	40.000	43.423	-8.6	91	0.00
34 C	Hexachlorobutadiene	40.000	41.698	-4.2	89	0.00
35	Caprolactam	40.000	43.897	-9.7	85	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.290	-8.2	90	0.00
37	2-Methylnaphthalene	40.000	42.110	-5.3	89	0.00
38	1-Methylnaphthalene	40.000	42.646	-6.6	89	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	77	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.907	-2.3	84	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.309	6.7	79	0.00
42 S	2,4,6-Tribromophenol	80.000	86.707	-8.4	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.418	-1.0	83	0.00
44	2,4,5-Trichlorophenol	40.000	41.168	-2.9	84	0.00
45 S	2-Fluorobiphenyl	80.000	82.617	-3.3	86	0.00
46	1,1'-Biphenyl	40.000	41.163	-2.9	84	0.00
47	2-Chloronaphthalene	40.000	41.347	-3.4	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061521\
 Data File : BP005922.D
 Acq On : 15 Jun 2021 10:12
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 11:31:56 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 14 13:15:03 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.510	-6.3	87	0.00
49	Acenaphthylene	40.000	41.752	-4.4	87	0.00
50	Dimethylphthalate	40.000	41.757	-4.4	89	0.00
51	2,6-Dinitrotoluene	40.000	41.865	-4.7	87	0.00
52 C	Acenaphthene	40.000	41.592	-4.0	88	0.00
53	3-Nitroaniline	40.000	43.012	-7.5	89	0.00
54 P	2,4-Dinitrophenol	40.000	45.833	-14.6	96	0.00
55	Dibenzofuran	40.000	40.464	-1.2	86	0.00
56 P	4-Nitrophenol	40.000	41.988	-5.0	84	0.00
57	2,4-Dinitrotoluene	40.000	43.075	-7.7	88	0.00
58	Fluorene	40.000	41.244	-3.1	87	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.062	-0.2	83	0.00
60	Diethylphthalate	40.000	41.851	-4.6	88	0.00
61	4-Chlorophenyl-phenylether	40.000	40.876	-2.2	87	0.00
62	4-Nitroaniline	40.000	43.924	-9.8	91	0.00
63	Azobenzene	40.000	40.220	-0.5	84	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	75	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.006	5.0	76	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.677	-4.2	86	0.00
67	4-Bromophenyl-phenylether	40.000	42.071	-5.2	87	0.00
68	Hexachlorobenzene	40.000	41.397	-3.5	86	0.00
69	Atrazine	40.000	35.995	10.0	63	0.00
70 C	Pentachlorophenol	40.000	41.935	-4.8	79	0.00
71	Phenanthrene	40.000	41.405	-3.5	85	0.00
72	Anthracene	40.000	42.136	-5.3	86	0.00
73	Carbazole	40.000	42.560	-6.4	87	0.00
74	Di-n-butylphthalate	40.000	44.122	-10.3	90	0.00
75 C	Fluoranthene	40.000	42.769	-6.9	89	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	84	0.00
77	Benzydine	40.000	36.986	7.5	68	0.00
78	Pyrene	40.000	39.698	0.8	92	0.00
79 S	Terphenyl-d14	80.000	82.445	-3.1	94	0.00
80	Butylbenzylphthalate	40.000	42.809	-7.0	97	0.00
81	Benzo(a)anthracene	40.000	40.932	-2.3	95	0.00
82	3,3'-Dichlorobenzidine	40.000	42.241	-5.6	99	0.00
83	Chrysene	40.000	41.540	-3.8	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.605	-9.0	100	0.00
85 c	Di-n-octyl phthalate	40.000	43.272	-8.2	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	81	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.134	-7.8	96	0.00
88	Benzo(b)fluoranthene	40.000	44.247	-10.6	98	0.00
89	Benzo(k)fluoranthene	40.000	43.826	-9.6	97	0.00
90 C	Benzo(a)pyrene	40.000	43.927	-9.8	97	0.00
91	Dibenzo(a,h)anthracene	40.000	42.985	-7.5	96	0.00
92	Benzo(g,h,i)perylene	40.000	42.481	-6.2	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061521\
Data File : BP005922.D
Acq On : 15 Jun 2021 10:12
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 15 11:31:56 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 14 13:15:03 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0