

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061521\
 Data File : BP005936.D
 Acq On : 15 Jun 2021 18:19
 Operator : CG/JU
 Sample : M2658-02MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 T-FLOORMS

Manual Integrations
 APPROVED

mohammad
 6/16/2021 3:54:27 PM

Quant Time: Jun 15 18:55:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 14 13:15:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	66642	20.000	ng	0.00
21) Naphthalene-d8	10.734	136	246983	20.000	ng	-0.01
39) Acenaphthene-d10	14.557	164	148375	20.000	ng	0.00
64) Phenanthrene-d10	17.304	188	304637	20.000	ng	0.00
76) Chrysene-d12	21.375	240	373763	20.000	ng	0.00
86) Perylene-d12	23.786	264	446308	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.505	112	529785	127.563	ng	0.00
7) Phenol-d6	7.111	99	577608	107.302	ng	0.00
23) Nitrobenzene-d5	9.099	82	427258	84.812	ng	0.00
42) 2,4,6-Tribromophenol	16.051	330	361749	142.073	ng	0.00
45) 2-Fluorobiphenyl	13.187	172	989823	87.600	ng	0.00
79) Terphenyl-d14	19.851	244	1787228	89.537	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	67795	36.132	ng	# 66
3) Pyridine	3.793	79	129719	29.415	ng	98
4) n-Nitrosodimethylamine	3.693	42	79851	38.955	ng	91
6) Aniline	7.264	93	140716	23.415	ng	99
8) 2-Chlorophenol	7.499	128	215279	45.209	ng	97
9) Benzaldehyde	7.069	77	143086	62.352	ng	97
10) Phenol	7.134	94	219758	40.866	ng	98
11) bis(2-Chloroethyl)ether	7.358	93	190975	46.866	ng	97
12) 1,3-Dichlorobenzene	7.822	146	226170	42.477	ng	99
13) 1,4-Dichlorobenzene	7.969	146	232752	42.864	ng	99
14) 1,2-Dichlorobenzene	8.287	146	225773	43.498	ng	98
15) Benzyl Alcohol	8.181	79	188212	46.420	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.469	45	157990	41.942	ng	97
17) 2-Methylphenol	8.387	107	163578	44.645	ng	99
18) Hexachloroethane	9.011	117	84111	42.823	ng	96
19) n-Nitroso-di-n-propyla...	8.746	70	144931	43.765	ng	93
20) 3+4-Methylphenols	8.716	107	216769	43.799	ng	99
22) Acetophenone	8.763	105	309980	47.683	ng	# 98
24) Nitrobenzene	9.140	77	221392	42.322	ng	99
25) Isophorone	9.669	82	381569	41.015	ng	100
26) 2-Nitrophenol	9.852	139	112546	45.091	ng	96
27) 2,4-Dimethylphenol	9.916	122	186329	50.109	ng	98
28) bis(2-Chloroethoxy)met...	10.146	93	240166	49.464	ng	99
29) 2,4-Dichlorophenol	10.393	162	197597	44.218	ng	100
30) 1,2,4-Trichlorobenzene	10.599	180	215717	42.962	ng	97
31) Naphthalene	10.787	128	634890	44.428	ng	99
32) Benzoic acid	10.075	122	122213	44.411	ng	99
33) 4-Chloroaniline	10.910	127	50416	8.733	ng	99
34) Hexachlorobutadiene	11.069	225	142359	41.917	ng	99
35) Caprolactam	11.704	113	51990	40.391	ng	97
36) 4-Chloro-3-methylphenol	12.034	107	208946	46.050	ng	97
37) 2-Methylnaphthalene	12.393	142	451781	44.404	ng	99
38) 1-Methylnaphthalene	12.610	142	418980	43.718	ng	99
40) 1,2,4,5-Tetrachloroben...	12.763	216	244631	46.840	ng	99
41) Hexachlorocyclopentadiene	12.740	237	288722	107.961	ng	98
43) 2,4,6-Trichlorophenol	13.004	196	160332	44.757	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061521\
 Data File : BP005936.D
 Acq On : 15 Jun 2021 18:19
 Operator : CG/JU
 Sample : M2658-02MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 T-FLOORMS

Manual Integrations
 APPROVED

mohammad
 6/16/2021 3:54:27 PM

Quant Time: Jun 15 18:55:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 14 13:15:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.081	196	176796	45.827	ng	95
46) 1,1'-Biphenyl	13.399	154	561601	46.944	ng	100
47) 2-Chloronaphthalene	13.440	162	446338	45.690	ng	98
48) 2-Nitroaniline	13.646	65	121260	47.210	ng	95
49) Acenaphthylene	14.281	152	696952	46.503	ng	99
50) Dimethylphthalate	14.022	163	622455	51.023	ng	100
51) 2,6-Dinitrotoluene	14.140	165	124798	45.008	ng	98
52) Acenaphthene	14.622	154	413629	45.446	ng	99
53) 3-Nitroaniline	14.469	138	79341	29.523	ng	100
54) 2,4-Dinitrophenol	14.681	184	145006	87.449	ng	97
55) Dibenzofuran	14.957	168	662465	44.260	ng	98
56) 4-Nitrophenol	14.793	139	195535	89.764	ng	94
57) 2,4-Dinitrotoluene	14.928	165	174866	47.151	ng	99
58) Fluorene	15.604	166	535000	45.875	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.187	232	153593m	45.173	ng	
60) Diethylphthalate	15.381	149	588352	48.069	ng	100
61) 4-Chlorophenyl-phenyle...	15.598	204	275781	45.604	ng	98
62) 4-Nitroaniline	15.634	138	127106	45.951	ng	98
63) Azobenzene	15.892	77	491883	44.420	ng	95
65) 4,6-Dinitro-2-methylph...	15.693	198	94092	42.030	ng	99
66) n-Nitrosodiphenylamine	15.816	169	474445	46.964	ng	99
67) 4-Bromophenyl-phenylether	16.492	248	183399	46.904	ng	98
68) Hexachlorobenzene	16.610	284	220830	47.118	ng	97
69) Atrazine	16.769	200	181477	57.740	ng	98
70) Pentachlorophenol	16.957	266	264480	101.489	ng	99
71) Phenanthrene	17.345	178	860231	47.021	ng	100
72) Anthracene	17.439	178	878036	48.765	ng	100
73) Carbazole	17.710	167	804635	46.546	ng	99
74) Di-n-butylphthalate	18.257	149	1048770	52.363	ng	100
75) Fluoranthene	19.310	202	1088649	48.999	ng	99
77) Benzidine	19.486	184	95476	12.276	ng	99
78) Pyrene	19.657	202	1132173	43.390	ng	99
80) Butylbenzylphthalate	20.522	149	517124	48.172	ng	98
81) Benzo(a)anthracene	21.363	228	1221592	45.277	ng	100
82) 3,3'-Dichlorobenzidine	21.292	252	338208	36.258	ng	97
83) Chrysene	21.416	228	1208360	46.240	ng	99
84) Bis(2-ethylhexyl)phtha...	21.280	149	888078	56.408	ng	99
85) Di-n-octyl phthalate	22.204	149	1432900	50.876	ng	99
87) Indeno(1,2,3-cd)pyrene	26.333	276	1828610	47.912	ng	99
88) Benzo(b)fluoranthene	23.051	252	1427830	46.710	ng	100
89) Benzo(k)fluoranthene	23.098	252	1396763	47.122	ng	99
90) Benzo(a)pyrene	23.686	252	1407088	48.961	ng	98
91) Dibenzo(a,h)anthracene	26.345	278	1554964	47.336	ng	99
92) Benzo(g,h,i)perylene	27.110	276	1544416	47.593	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061521\
Data File : BP005936.D
Acq On : 15 Jun 2021 18:19
Operator : CG/JU
Sample : M2658-02MS
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
BNA_P
ClientSampled :
T-FLOORMS

Manual Integrations
APPROVED

mohammad
6/16/2021 3:54:27 PM

Quant Time: Jun 15 18:55:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 14 13:15:03 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.934 min Scan# 824
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	152	Resp:	66642
Ion	Ratio	Lower	Upper
152	100		
150	153.5	125.6	188.4
115	60.1	48.5	72.7

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#2
1,4-Dioxane
Concen: 36.132 ng
RT: 3.381 min Scan# 50
Delta R.T. 0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	88	Resp:	67795
Ion	Ratio	Lower	Upper
88	100		
58	59.3	46.7	70.1
43	80.0	20.3	30.5









#3
Pyridine
Concen: 29.415 ng
RT: 3.793 min Scan# 120
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS

Tgt Ion: 79 Resp: 129719
Ion Ratio Lower Upper
79 100
52 52.9 41.7 62.5
51 28.7 24.1 36.1




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#4
n-Nitrosodimethylamine
Concen: 38.955 ng
RT: 3.693 min Scan# 103
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 42 Resp: 79851
Ion Ratio Lower Upper
42 100
74 143.0 105.4 158.0
44 6.2 4.2 6.4









#5
2-Fluorophenol
Concen: 127.563 ng
RT: 5.505 min Scan# 411
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 112 Resp: 529785
Ion Ratio Lower Upper
112 100
64 50.8 41.4 62.2
63 27.6 23.8 35.8

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#6
Aniline
Concen: 23.415 ng
RT: 7.264 min Scan# 710
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 93 Resp: 140716
Ion Ratio Lower Upper
93 100
66 37.5 29.7 44.5
65 18.9 15.3 22.9









#7
Phenol-d6
Concen: 107.302 ng
RT: 7.111 min Scan# 684
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS

Tgt Ion: 99 Resp: 577608
Ion Ratio Lower Upper
99 100
42 15.2 13.4 20.0
71 38.1 30.6 46.0




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#8
2-Chlorophenol
Concen: 45.209 ng
RT: 7.499 min Scan# 750
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion: 128 Resp: 215279
Ion Ratio Lower Upper
128 100
130 31.2 12.4 52.4
64 38.5 20.3 60.3





#9
Benzaldehyde
Concen: 62.352 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS



Tgt Ion: 77 Resp: 143086
Ion Ratio Lower Upper
77 100
105 101.4 76.7 116.7
106 95.4 76.2 116.2




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#10
Phenol
Concen: 40.866 ng
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion: 94 Resp: 219758
Ion Ratio Lower Upper
94 100
65 27.5 8.8 48.8
66 41.5 20.7 60.7





#11
bis(2-Chloroethyl)ether
Concen: 46.866 ng
RT: 7.358 min Scan# 726
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion: 93 Resp: 190975
Ion Ratio Lower Upper
93 100
63 63.9 46.2 86.2
95 34.0 12.2 52.2

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#12
1,3-Dichlorobenzene
Concen: 42.477 ng
RT: 7.822 min Scan# 805
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion: 146 Resp: 226170
Ion Ratio Lower Upper
146 100
148 65.3 51.5 77.3
75 31.3 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 42.864 ng
RT: 7.969 min Scan# 830
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:146 Resp: 232752
Ion Ratio Lower Upper
146 100
148 65.1 51.3 76.9
111 41.1 33.0 49.6

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#14
1,2-Dichlorobenzene
Concen: 43.498 ng
RT: 8.287 min Scan# 884
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:146 Resp: 225773
Ion Ratio Lower Upper
146 100
148 64.0 50.3 75.5
111 43.2 33.7 50.5





#15
Benzyl Alcohol
Concen: 46.420 ng
RT: 8.181 min Scan# 866
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion: 79 Resp: 188212
Ion Ratio Lower Upper
79 100
108 77.0 62.5 93.7
77 63.9 52.9 79.3

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.942 ng
RT: 8.469 min Scan# 915
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Tgt Ion: 45 Resp: 157990
Ion Ratio Lower Upper
45 100
77 23.8 2.0 42.0
79 18.0 0.0 36.7





#17
2-Methylphenol
Concen: 44.645 ng
RT: 8.387 min Scan# 901
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion:107 Resp: 163578
Ion Ratio Lower Upper
107 100
108 110.2 87.7 131.5
77 49.2 40.2 60.2
79 48.7 39.0 58.4

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#18
Hexachloroethane
Concen: 42.823 ng
RT: 9.011 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Tgt Ion:117 Resp: 84111
Ion Ratio Lower Upper
117 100
119 98.4 76.2 114.2
201 108.3 83.4 125.0





#19
n-Nitroso-di-n-propylamine
Concen: 43.765 ng
RT: 8.746 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion: 70 Resp: 144931
Ion Ratio Lower Upper
70 100
42 43.8 40.8 61.2
101 10.9 8.8 13.2
130 24.7 18.9 28.3

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#20
3+4-Methylphenols
Concen: 43.799 ng
RT: 8.716 min Scan# 957
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion: 107 Resp: 216769
Ion Ratio Lower Upper
107 100
108 89.6 71.0 111.0
77 35.2 16.6 56.6
79 32.8 12.7 52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.734 min Scan# 1300
Delta R.T. -0.011 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	136	Resp:	246983
Ion Ratio	Lower	Upper	
136	100		
137	10.7	8.6	12.8
54	6.0	4.6	6.8
68	4.5	3.6	5.4

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#22
Acetophenone
Concen: 47.683 ng
RT: 8.763 min Scan# 965
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	105	Resp:	309980
Ion Ratio	Lower	Upper	
105	100		
71	1.8	1.9	2.9
51	21.0	18.4	27.6
120	22.6	17.9	26.9









#23
Nitrobenzene-d5
Concen: 84.812 ng
RT: 9.099 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 82 Resp: 427258
Ion Ratio Lower Upper
82 100
128 45.1 35.7 53.5
54 40.8 35.0 52.4

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#24
Nitrobenzene
Concen: 42.322 ng
RT: 9.140 min Scan# 1029
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 77 Resp: 221392
Ion Ratio Lower Upper
77 100
123 48.0 38.2 57.2
65 13.8 11.4 17.2









#25
Isophorone
Concen: 41.015 ng
RT: 9.669 min Scan# 1119
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	82	Resp:	381569
Ion	Ratio	Lower	Upper
82	100		
95	8.3	6.5	9.7
138	18.8	15.0	22.4

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#26
2-Nitrophenol
Concen: 45.091 ng
RT: 9.852 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	139	Resp:	112546
Ion	Ratio	Lower	Upper
139	100		
109	36.5	28.8	43.2
65	46.4	41.0	61.6









#27
2,4-Dimethylphenol
Concen: 50.109 ng
RT: 9.916 min Scan# 1161
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:122 Resp: 186329
Ion Ratio Lower Upper
122 100
107 120.7 94.5 141.7
121 59.0 47.4 71.2

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#28
bis(2-Chloroethoxy)methane
Concen: 49.464 ng
RT: 10.146 min Scan# 1200
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion: 93 Resp: 240166
Ion Ratio Lower Upper
93 100
95 32.6 25.9 38.9
123 12.3 10.2 15.2





#29
2,4-Dichlorophenol
Concen: 44.218 ng
RT: 10.393 min Scan# 1242
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion:162 Resp: 197597
Ion Ratio Lower Upper
162 100
164 64.8 44.9 84.9
98 34.9 14.3 54.3

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#30
1,2,4-Trichlorobenzene
Concen: 42.962 ng
RT: 10.599 min Scan# 1277
Delta R.T. -0.011 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Tgt Ion:180 Resp: 215717
Ion Ratio Lower Upper
180 100
182 94.1 77.9 116.9
145 31.0 24.7 37.1





#31
Naphthalene
Concen: 44.428 ng
RT: 10.787 min Scan# 1309
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS



Tgt Ion:128 Resp: 634890
Ion Ratio Lower Upper
128 100
129 11.2 9.0 13.6
127 12.9 10.7 16.1




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#32
Benzoic acid
Concen: 44.411 ng
RT: 10.075 min Scan# 1188
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:122 Resp: 122213
Ion Ratio Lower Upper
122 100
105 124.2 103.2 143.2
77 96.2 76.2 116.2









#33
4-Chloroaniline
Concen: 8.733 ng
RT: 10.910 min Scan# 1330
Delta R.T. 0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	127	Resp:	50416
Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.1	39.1
65	27.7	22.5	33.7
92	19.6	15.0	22.6

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#34
Hexachlorobutadiene
Concen: 41.917 ng
RT: 11.069 min Scan# 1357
Delta R.T. -0.011 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	225	Resp:	142359
Ion	Ratio	Lower	Upper
225	100		
223	64.1	51.0	76.6
227	63.2	51.4	77.2









#35
Caprolactam
Concen: 40.391 ng
RT: 11.704 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS

Tgt Ion:113 Resp: 51990
Ion Ratio Lower Upper
113 100
55 113.0 98.9 138.9
56 91.3 71.7 111.7




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#36
4-Chloro-3-methylphenol
Concen: 46.050 ng
RT: 12.034 min Scan# 1521
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:107 Resp: 208946
Ion Ratio Lower Upper
107 100
144 25.4 21.1 31.7
142 79.9 66.0 99.0









#37
2-Methylnaphthalene
Concen: 44.404 ng
RT: 12.393 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:142 Resp: 451781
Ion Ratio Lower Upper
142 100
141 89.1 71.8 107.6
115 38.5 31.6 47.4

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#38
1-Methylnaphthalene
Concen: 43.718 ng
RT: 12.610 min Scan# 1619
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:142 Resp: 418980
Ion Ratio Lower Upper
142 100
141 92.8 73.3 109.9
116 4.0 3.5 5.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.557 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS



Tgt Ion:164 Resp: 148375
Ion Ratio Lower Upper
164 100
162 100.4 80.2 120.2
160 44.9 34.5 51.7




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 46.840 ng
RT: 12.763 min Scan# 1645
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:216 Resp: 244631
Ion Ratio Lower Upper
216 100
214 77.4 61.9 92.9
179 21.6 16.8 25.2
108 19.6 14.9 22.3






#41
Hexachlorocyclopentadiene
Concen: 107.961 ng
RT: 12.740 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19





Tgt Ion: 237 Resp: 288722
Ion Ratio Lower Upper
237 100
235 63.2 44.5 84.5
272 13.2 0.0 32.5




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#42
2,4,6-Tribromophenol
Concen: 142.073 ng
RT: 16.051 min Scan# 2204
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Tgt Ion: 330 Resp: 361749
Ion Ratio Lower Upper
330 100
332 97.9 77.8 116.8
141 29.3 25.4 38.2





#43
2,4,6-Trichlorophenol
Concen: 44.757 ng
RT: 13.004 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampled :
T-FLOORMS




Tgt Ion:196 Resp: 160332
Ion Ratio Lower Upper
196 100
198 96.6 79.5 119.3
200 31.1 25.0 37.6

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#44
2,4,5-Trichlorophenol
Concen: 45.827 ng
RT: 13.081 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion:196 Resp: 176796
Ion Ratio Lower Upper
196 100
198 99.1 74.7 112.1
97 48.2 37.4 56.0
132 26.7 20.0 30.0









#45
2-Fluorobiphenyl
Concen: 87.600 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	172	Resp:	989823
Ion Ratio	Lower	Upper	
172	100		
171	35.9	27.9	41.9
170	23.7	18.6	27.8

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#46
1,1'-Biphenyl
Concen: 46.944 ng
RT: 13.399 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:	154	Resp:	561601
Ion Ratio	Lower	Upper	
154	100		
153	40.3	20.4	60.4
76	10.6	0.0	30.8





#47
2-Chloronaphthalene
Concen: 45.690 ng
RT: 13.440 min Scan# 1760
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion:162 Resp: 446338
Ion Ratio Lower Upper
162 100
127 40.4 31.4 47.0
164 32.5 26.6 39.8

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#48
2-Nitroaniline
Concen: 47.210 ng
RT: 13.646 min Scan# 1795
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion: 65 Resp: 121260
Ion Ratio Lower Upper
65 100
92 76.8 57.7 86.5
138 120.1 92.1 138.1









#49
Acenaphthylene
Concen: 46.503 ng
RT: 14.281 min Scan# 1903
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:152 Resp: 696952
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.2
153 13.1 10.6 16.0

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#50
Dimethylphthalate
Concen: 51.023 ng
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:163 Resp: 622455
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.2
164 10.2 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 45.008 ng
RT: 14.140 min Scan# 1879
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion:165 Resp: 124798
Ion Ratio Lower Upper
165 100
63 53.3 41.3 61.9
89 57.1 44.8 67.2

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#52
Acenaphthene
Concen: 45.446 ng
RT: 14.622 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion:154 Resp: 413629
Ion Ratio Lower Upper
154 100
153 112.4 89.4 134.2
152 53.0 43.0 64.4









#53
3-Nitroaniline
Concen: 29.523 ng
RT: 14.469 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS

Tgt Ion:138 Resp: 79341
Ion Ratio Lower Upper
138 100
108 11.5 8.4 12.6
92 109.8 87.8 131.8




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#54
2,4-Dinitrophenol
Concen: 87.449 ng
RT: 14.681 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:184 Resp: 145006
Ion Ratio Lower Upper
184 100
63 59.0 47.0 70.6
154 69.6 52.3 78.5





#55
Dibenzofuran
Concen: 44.260 ng
RT: 14.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS



Tgt Ion:168 Resp: 662465
Ion Ratio Lower Upper
168 100
139 40.3 31.0 46.6
169 13.3 10.8 16.2




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#56
4-Nitrophenol
Concen: 89.764 ng
RT: 14.793 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:139 Resp: 195535
Ion Ratio Lower Upper
139 100
109 84.6 72.3 112.3
65 88.7 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 47.151 ng
RT: 14.928 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:165 Resp: 174866
Ion Ratio Lower Upper
165 100
63 40.6 32.3 48.5
89 70.5 55.9 83.9
182 2.7 2.2 3.4

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#58
Fluorene
Concen: 45.875 ng
RT: 15.604 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:166 Resp: 535000
Ion Ratio Lower Upper
166 100
165 98.4 78.0 117.0
167 13.4 10.5 15.7





#59
2,3,4,6-Tetrachlorophenol
Concen: 45.173 ng m
RT: 15.187 min Scan# 2057
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampled :
T-FLOORMS




Tgt Ion:232 Resp: 153593
Ion Ratio Lower Upper
232 100
131 40.2 32.2 48.4
130 2.6 2.0 3.0
166 25.3 19.4 29.2

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#60
Diethylphthalate
Concen: 48.069 ng
RT: 15.381 min Scan# 2090
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion:149 Resp: 588352
Ion Ratio Lower Upper
149 100
177 21.8 17.5 26.3
150 12.5 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 45.604 ng
RT: 15.598 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	204	Resp:	275781
Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.6	40.0
141	62.7	48.6	73.0

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#62
4-Nitroaniline
Concen: 45.951 ng
RT: 15.634 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	138	Resp:	127106
Ion	Ratio	Lower	Upper
138	100		
92	51.0	31.2	71.2
108	95.7	78.6	118.6









#63
Azobenzene
Concen: 44.420 ng
RT: 15.892 min Scan# 2177
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 77	Resp: 491883
Ion Ratio	Lower Upper
77 100	
182 29.3	6.0 46.0
105 21.4	0.0 38.8
51 26.9	8.8 48.8

Instrument :
BNA_P
ClientSampled :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.304 min Scan# 2417
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:188	Resp: 304637
Ion Ratio	Lower Upper
188 100	
94 7.6	5.6 8.4
80 7.9	6.3 9.5









#65
4,6-Dinitro-2-methylphenol
Concen: 42.030 ng
RT: 15.693 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	198	Resp:	94092
Ion Ratio	Lower	Upper	
198	100		
51	33.2	14.0	54.0
105	39.7	19.5	59.5

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#66
n-Nitrosodiphenylamine
Concen: 46.964 ng
RT: 15.816 min Scan# 2164
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	169	Resp:	474445
Ion Ratio	Lower	Upper	
169	100		
168	67.2	54.4	81.6
167	37.0	29.0	43.6









#67
4-Bromophenyl-phenylether
Concen: 46.904 ng
RT: 16.492 min Scan# 2279
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 248 Resp: 183399
Ion Ratio Lower Upper
248 100
250 98.6 79.8 119.8
141 74.6 57.8 86.6

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#68
Hexachlorobenzene
Concen: 47.118 ng
RT: 16.610 min Scan# 2299
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 284 Resp: 220830
Ion Ratio Lower Upper
284 100
142 31.1 26.9 40.3
249 31.1 24.5 36.7









#69
Atrazine
Concen: 57.740 ng
RT: 16.769 min Scan# 2326
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:200 Resp: 181477
Ion Ratio Lower Upper
200 100
173 25.8 4.1 44.1
215 47.3 27.7 67.7

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM




#70
Pentachlorophenol
Concen: 101.489 ng
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:266 Resp: 264480
Ion Ratio Lower Upper
266 100
268 63.7 51.6 77.4
264 63.5 50.3 75.5









#71
Phenanthrene
Concen: 47.021 ng
RT: 17.345 min Scan# 2424
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19





Tgt Ion:178 Resp: 860231
Ion Ratio Lower Upper
178 100
176 19.0 15.1 22.7
179 15.5 12.4 18.6



#72
Anthracene
Concen: 48.765 ng
RT: 17.439 min Scan# 2440
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Tgt Ion:178 Resp: 878036
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.3 12.2 18.2



Instrument :
BNA_P
ClientSampleId :
T-FLOORMS


Manual Integrations
APPROVED

mohammad
6/16/2021 3:54:27 PM



#73
Carbazole
Concen: 46.546 ng
RT: 17.710 min Scan# 2486
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion:167 Resp: 804635
Ion Ratio Lower Upper
167 100
166 21.6 17.1 25.7
139 13.7 10.7 16.1

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#74
Di-n-butylphthalate
Concen: 52.363 ng
RT: 18.257 min Scan# 2579
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Tgt Ion:149 Resp: 1048770
Ion Ratio Lower Upper
149 100
150 9.1 7.3 10.9
104 5.2 4.5 6.7





#75
Fluoranthene
Concen: 48.999 ng
RT: 19.310 min Scan# 2758
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Tgt Ion:202 Resp: 1088649
Ion Ratio Lower Upper
202 100
101 8.3 0.0 27.6
203 17.4 0.0 37.8

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.375 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion:240 Resp: 373763
Ion Ratio Lower Upper
240 100
120 9.2 6.5 9.7
236 25.6 20.6 31.0









#77
Benzidine
Concen: 12.276 ng
RT: 19.486 min Scan# 2788
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	184	Resp:	95476
Ion Ratio	Lower	Upper	
184	100		
185	14.0	11.0	16.4
183	12.4	9.5	14.3

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#78
Pyrene
Concen: 43.390 ng
RT: 19.657 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:	202	Resp:	1132173
Ion Ratio	Lower	Upper	
202	100		
200	20.9	16.3	24.5
203	17.8	14.2	21.2





#79
Terphenyl-d14
Concen: 89.537 ng
RT: 19.851 min Scan# 2850
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19





Tgt Ion:244 Resp: 1787228

Ion	Ratio	Lower	Upper
244	100		
212	7.1	6.0	9.0
122	8.7	7.1	10.7

Instrument :
BNA_P
ClientSampleId :
T-FLOORMS

Manual Integrations
APPROVED




mohammad
6/16/2021 3:54:27 PM




#80
Butylbenzylphthalate
Concen: 48.172 ng
RT: 20.522 min Scan# 2964
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion:149 Resp: 517124


Ion	Ratio	Lower	Upper
149	100		
91	69.1	54.6	81.8
206	22.8	16.7	25.1





#81
Benzo(a)anthracene
Concen: 45.277 ng
RT: 21.363 min Scan# 3107
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampled :
T-FLOORMS



Tgt Ion:228 Resp: 1221592
Ion Ratio Lower Upper
228 100
226 26.7 21.6 32.4
229 19.8 15.8 23.6




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#82
3,3'-Dichlorobenzidine
Concen: 36.258 ng
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:252 Resp: 338208
Ion Ratio Lower Upper
252 100
254 65.8 50.6 76.0
126 9.6 7.0 10.6





#83
Chrysene
Concen: 46.240 ng
RT: 21.416 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS



Tgt Ion:228 Resp: 1208360
Ion Ratio Lower Upper
228 100
226 29.0 23.8 35.6
229 19.7 15.4 23.2




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#84
Bis(2-ethylhexyl)phthalate
Concen: 56.408 ng
RT: 21.280 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:149 Resp: 888078
Ion Ratio Lower Upper
149 100
167 28.1 21.9 32.9
279 5.3 4.3 6.5





#85
Di-n-octyl phthalate
Concen: 50.876 ng
RT: 22.204 min Scan# 3250
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampled :
T-FLOORMS




Tgt Ion:149 Resp: 1432900
Ion Ratio Lower Upper
149 100
167 1.6 1.4 2.0
43 6.6 5.6 8.4

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:27 PM




#86
Perylene-d12
Concen: 20.000 ng
RT: 23.786 min Scan# 3519
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion:264 Resp: 446308
Ion Ratio Lower Upper
264 100
260 23.1 18.6 28.0
265 22.2 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 47.912 ng
RT: 26.333 min Scan# 3952
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19

Tgt Ion: 276 Resp: 1828610
Ion Ratio Lower Upper
276 100
138 20.0 15.3 22.9
277 25.6 20.4 30.6

Instrument :
BNA_P
ClientSampled :
T-FLOORMS




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#88
Benzo(b)fluoranthene
Concen: 46.710 ng
RT: 23.051 min Scan# 3394
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion: 252 Resp: 1427830
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 8.1 6.2 9.4





#89
Benzo(k)fluoranthene
Concen: 47.122 ng
RT: 23.098 min Scan# 3402
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampleId :
T-FLOORMS



Tgt Ion:252 Resp: 1396763
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 8.4 6.0 9.0




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:27 PM



#90
Benzo(a)pyrene
Concen: 48.961 ng
RT: 23.686 min Scan# 3502
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19


Tgt Ion:252 Resp: 1407088
Ion Ratio Lower Upper
252 100
253 21.3 17.9 26.9
125 8.6 6.5 9.7





#91
Dibenzo(a,h)anthracene
Concen: 47.336 ng
RT: 26.345 min Scan# 3954
Delta R.T. -0.006 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19




Instrument :
BNA_P
ClientSampled :
T-FLOORMS




Tgt Ion:278 Resp: 1554964
Ion Ratio Lower Upper
278 100
139 12.7 9.3 13.9
279 23.5 18.8 28.2

Manual Integrations
APPROVED

mohammad
6/16/2021 3:54:27 PM



#92
Benzo(g,h,i)perylene
Concen: 47.593 ng
RT: 27.110 min Scan# 4084
Delta R.T. 0.000 min
Lab File: BP005936.D
Acq: 15 Jun 2021 18:19



Tgt Ion:276 Resp: 1544416
Ion Ratio Lower Upper
276 100
277 23.7 18.9 28.3
138 17.8 12.6 18.8

