

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061521\
 Data File : BP005937.D
 Acq On : 15 Jun 2021 18:54
 Operator : CG/JU
 Sample : M2658-02MSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 T-FLOORMSD

Manual Integrations
 APPROVED

mohammad
 6/16/2021 3:54:30 PM

Quant Time: Jun 15 19:25:48 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 14 13:15:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	71524	20.000	ng	0.00
21) Naphthalene-d8	10.740	136	271097	20.000	ng	0.00
39) Acenaphthene-d10	14.563	164	165180	20.000	ng	0.00
64) Phenanthrene-d10	17.304	188	334809	20.000	ng	0.00
76) Chrysene-d12	21.380	240	387003	20.000	ng	0.00
86) Perylene-d12	23.786	264	450369	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.505	112	580380	130.207	ng	0.00
7) Phenol-d6	7.111	99	640970	110.945	ng	0.00
23) Nitrobenzene-d5	9.099	82	471905	85.343	ng	0.00
42) 2,4,6-Tribromophenol	16.051	330	396053	139.720	ng	0.00
45) 2-Fluorobiphenyl	13.187	172	1091817	86.796	ng	0.00
79) Terphenyl-d14	19.851	244	1905293	92.186	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	72799	36.151	ng	# 69
3) Pyridine	3.799	79	143831	30.389	ng	99
4) n-Nitrosodimethylamine	3.693	42	84741	38.518	ng	89
6) Aniline	7.263	93	165191	25.611	ng	99
8) 2-Chlorophenol	7.499	128	233335	45.657	ng	100
9) Benzaldehyde	7.075	77	150244	61.003	ng	94
10) Phenol	7.134	94	242919	42.090	ng	99
11) bis(2-Chloroethyl)ether	7.358	93	206746	47.273	ng	99
12) 1,3-Dichlorobenzene	7.822	146	243385	42.590	ng	98
13) 1,4-Dichlorobenzene	7.969	146	248571	42.652	ng	99
14) 1,2-Dichlorobenzene	8.287	146	241915	43.427	ng	98
15) Benzyl Alcohol	8.181	79	207998	47.798	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.469	45	171972	42.538	ng	96
17) 2-Methylphenol	8.387	107	179450	45.634	ng	98
18) Hexachloroethane	9.016	117	90936	43.138	ng	97
19) n-Nitroso-di-n-propyla...	8.746	70	160185	45.069	ng	95
20) 3+4-Methylphenols	8.716	107	242288	45.613	ng	99
22) Acetophenone	8.763	105	340774	47.758	ng	98
24) Nitrobenzene	9.140	77	242408	42.217	ng	99
25) Isophorone	9.669	82	424356	41.557	ng	99
26) 2-Nitrophenol	9.852	139	123814	45.193	ng	94
27) 2,4-Dimethylphenol	9.916	122	206730	50.651	ng	95
28) bis(2-Chloroethoxy)met...	10.152	93	265936	49.900	ng	99
29) 2,4-Dichlorophenol	10.393	162	221105	45.078	ng	100
30) 1,2,4-Trichlorobenzene	10.599	180	237171	43.033	ng	98
31) Naphthalene	10.787	128	694007	44.245	ng	100
32) Benzoic acid	10.081	122	135500	44.795	ng	98
33) 4-Chloroaniline	10.904	127	63449	10.013	ng	97
34) Hexachlorobutadiene	11.069	225	156690	42.032	ng	98
35) Caprolactam	11.704	113	59447	42.076	ng	96
36) 4-Chloro-3-methylphenol	12.034	107	231943	46.571	ng	99
37) 2-Methylnaphthalene	12.393	142	503337	45.070	ng	98
38) 1-Methylnaphthalene	12.610	142	460326	43.760	ng	99
40) 1,2,4,5-Tetrachloroben...	12.763	216	266296	45.801	ng	99
41) Hexachlorocyclopentadiene	12.740	237	321610	108.021	ng	97
43) 2,4,6-Trichlorophenol	13.004	196	178151	44.672	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061521\
 Data File : BP005937.D
 Acq On : 15 Jun 2021 18:54
 Operator : CG/JU
 Sample : M2658-02MSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 T-FLOORMSD


Manual Integrations
 APPROVED

mohammad
 6/16/2021 3:54:30 PM

Quant Time: Jun 15 19:25:48 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 14 13:15:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.081	196	198218	46.152	ng	98
46) 1,1'-Biphenyl	13.398	154	616271	46.273	ng	100
47) 2-Chloronaphthalene	13.440	162	494756	45.493	ng	99
48) 2-Nitroaniline	13.651	65	138138	48.309	ng	95
49) Acenaphthylene	14.287	152	773425	46.355	ng	99
50) Dimethylphthalate	14.022	163	684688	50.414	ng	99
51) 2,6-Dinitrotoluene	14.140	165	139719	45.262	ng	98
52) Acenaphthene	14.622	154	454312	44.838	ng	99
53) 3-Nitroaniline	14.475	138	92171	30.808	ng	100
54) 2,4-Dinitrophenol	14.681	184	158631	86.017	ng	94
55) Dibenzofuran	14.957	168	736015	44.171	ng	99
56) 4-Nitrophenol	14.793	139	214766	88.562	ng	94
57) 2,4-Dinitrotoluene	14.928	165	193853	46.953	ng	99
58) Fluorene	15.604	166	586859	45.202	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.187	232	168469m	44.507	ng	
60) Diethylphthalate	15.381	149	646101	47.417	ng	99
61) 4-Chlorophenyl-phenyle...	15.598	204	302590	44.947	ng	98
62) 4-Nitroaniline	15.634	138	139224	45.211	ng	98
63) Azobenzene	15.892	77	540750	43.865	ng	96
65) 4,6-Dinitro-2-methylph...	15.692	198	103855	42.210	ng	99
66) n-Nitrosodiphenylamine	15.816	169	521861	47.003	ng	98
67) 4-Bromophenyl-phenylether	16.492	248	202042	47.016	ng	95
68) Hexachlorobenzene	16.610	284	244362	47.440	ng	98
69) Atrazine	16.769	200	201676	58.384	ng	98
70) Pentachlorophenol	16.957	266	287458	100.366	ng	99
71) Phenanthrene	17.351	178	941340	46.817	ng	99
72) Anthracene	17.439	178	964214	48.725	ng	100
73) Carbazole	17.710	167	870931	45.841	ng	100
74) Di-n-butylphthalate	18.257	149	1137585	51.679	ng	99
75) Fluoranthene	19.310	202	1179935	48.322	ng	98
77) Benzidine	19.486	184	93542	11.616	ng	99
78) Pyrene	19.657	202	1218604	45.105	ng	99
80) Butylbenzylphthalate	20.522	149	550497	49.526	ng	99
81) Benzo(a)anthracene	21.363	228	1275395	45.654	ng	100
82) 3,3'-Dichlorobenzidine	21.292	252	353668	36.618	ng	98
83) Chrysene	21.416	228	1268954	46.898	ng	99
84) Bis(2-ethylhexyl)phtha...	21.280	149	927878	56.920	ng	99
85) Di-n-octyl phthalate	22.204	149	1478550	50.701	ng	99
87) Indeno(1,2,3-cd)pyrene	26.327	276	1839369	47.759	ng	99
88) Benzo(b)fluoranthene	23.051	252	1475242	47.825	ng	99
89) Benzo(k)fluoranthene	23.098	252	1413064	47.242	ng	99
90) Benzo(a)pyrene	23.686	252	1425322	49.149	ng	98
91) Dibenzo(a,h)anthracene	26.345	278	1573783	47.477	ng	99
92) Benzo(g,h,i)perylene	27.104	276	1554298	47.466	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.934 min Scan# 824
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:	152	Resp:	71524
Ion Ratio	Lower	Upper	
152	100		
150	159.2	125.6	188.4
115	61.2	48.5	72.7

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#2
1,4-Dioxane
Concen: 36.151 ng
RT: 3.381 min Scan# 50
Delta R.T. 0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:	88	Resp:	72799
Ion Ratio	Lower	Upper	
88	100		
58	58.1	46.7	70.1
43	77.0	20.3	30.5





#3
Pyridine
Concen: 30.389 ng
RT: 3.799 min Scan# 121
Delta R.T. 0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion: 79 Resp: 143831
Ion Ratio Lower Upper
79 100
52 52.3 41.7 62.5
51 29.1 24.1 36.1

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#4
n-Nitrosodimethylamine
Concen: 38.518 ng
RT: 3.693 min Scan# 103
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion: 42 Resp: 84741
Ion Ratio Lower Upper
42 100
74 144.6 105.4 158.0
44 5.6 4.2 6.4









#5
2-Fluorophenol
Concen: 130.207 ng
RT: 5.505 min Scan# 411
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion: 112 Resp: 580380
Ion Ratio Lower Upper
112 100
64 51.3 41.4 62.2
63 27.5 23.8 35.8

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#6
Aniline
Concen: 25.611 ng
RT: 7.263 min Scan# 710
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion: 93 Resp: 165191
Ion Ratio Lower Upper
93 100
66 37.6 29.7 44.5
65 18.7 15.3 22.9









#7
Phenol-d6
Concen: 110.945 ng
RT: 7.111 min Scan# 684
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion: 99 Resp: 640970
Ion Ratio Lower Upper
99 100
42 15.4 13.4 20.0
71 38.3 30.6 46.0

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#8
2-Chlorophenol
Concen: 45.657 ng
RT: 7.499 min Scan# 750
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion: 128 Resp: 233335
Ion Ratio Lower Upper
128 100
130 32.3 12.4 52.4
64 40.4 20.3 60.3





#9
Benzaldehyde
Concen: 61.003 ng
RT: 7.075 min Scan# 678
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD



Tgt Ion: 77 Resp: 150244
Ion Ratio Lower Upper
77 100
105 104.3 76.7 116.7
106 99.8 76.2 116.2




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#10
Phenol
Concen: 42.090 ng
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion: 94 Resp: 242919
Ion Ratio Lower Upper
94 100
65 27.1 8.8 48.8
66 40.6 20.7 60.7





#11
bis(2-Chloroethyl)ether
Concen: 47.273 ng
RT: 7.358 min Scan# 726
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion: 93 Resp: 206746
Ion Ratio Lower Upper
93 100
63 66.5 46.2 86.2
95 33.3 12.2 52.2

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#12
1,3-Dichlorobenzene
Concen: 42.590 ng
RT: 7.822 min Scan# 805
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion: 146 Resp: 243385
Ion Ratio Lower Upper
146 100
148 63.0 51.5 77.3
75 30.8 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 42.652 ng
RT: 7.969 min Scan# 830
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:146 Resp: 248571
Ion Ratio Lower Upper
146 100
148 64.2 51.3 76.9
111 40.3 33.0 49.6

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#14
1,2-Dichlorobenzene
Concen: 43.427 ng
RT: 8.287 min Scan# 884
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:146 Resp: 241915
Ion Ratio Lower Upper
146 100
148 64.3 50.3 75.5
111 43.5 33.7 50.5





#15
Benzyl Alcohol
Concen: 47.798 ng
RT: 8.181 min Scan# 866
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion: 79 Resp: 207998
Ion Ratio Lower Upper
79 100
108 76.2 62.5 93.7
77 65.5 52.9 79.3

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.538 ng
RT: 8.469 min Scan# 915
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion: 45 Resp: 171972
Ion Ratio Lower Upper
45 100
77 23.8 2.0 42.0
79 18.4 0.0 36.7





#17
2-Methylphenol
Concen: 45.634 ng
RT: 8.387 min Scan# 901
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:107 Resp: 179450
Ion Ratio Lower Upper
107 100
108 113.2 87.7 131.5
77 49.0 40.2 60.2
79 48.4 39.0 58.4

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#18
Hexachloroethane
Concen: 43.138 ng
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:117 Resp: 90936
Ion Ratio Lower Upper
117 100
119 94.4 76.2 114.2
201 108.4 83.4 125.0





#19
n-Nitroso-di-n-propylamine
Concen: 45.069 ng
RT: 8.746 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion: 70 Resp: 160185
Ion Ratio Lower Upper
70 100
42 46.1 40.8 61.2
101 10.5 8.8 13.2
130 23.4 18.9 28.3

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#20
3+4-Methylphenols
Concen: 45.613 ng
RT: 8.716 min Scan# 957
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion: 107 Resp: 242288
Ion Ratio Lower Upper
107 100
108 89.9 71.0 111.0
77 35.5 16.6 56.6
79 31.9 12.7 52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.740 min Scan# 1301
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:	136	Resp:	271097
Ion Ratio	Lower	Upper	
136	100		
137	10.6	8.6	12.8
54	4.7	4.6	6.8
68	4.9	3.6	5.4

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#22
Acetophenone
Concen: 47.758 ng
RT: 8.763 min Scan# 965
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:	105	Resp:	340774
Ion Ratio	Lower	Upper	
105	100		
71	2.1	1.9	2.9
51	21.1	18.4	27.6
120	22.6	17.9	26.9









#23
Nitrobenzene-d5
Concen: 85.343 ng
RT: 9.099 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion: 82 Resp: 471905
Ion Ratio Lower Upper
82 100
128 44.7 35.7 53.5
54 40.3 35.0 52.4

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#24
Nitrobenzene
Concen: 42.217 ng
RT: 9.140 min Scan# 1029
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion: 77 Resp: 242408
Ion Ratio Lower Upper
77 100
123 47.3 38.2 57.2
65 13.0 11.4 17.2





#25
Isophorone
Concen: 41.557 ng
RT: 9.669 min Scan# 1119
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion: 82 Resp: 424356
Ion Ratio Lower Upper
82 100
95 8.2 6.5 9.7
138 19.1 15.0 22.4

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#26
2-Nitrophenol
Concen: 45.193 ng
RT: 9.852 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion: 139 Resp: 123814
Ion Ratio Lower Upper
139 100
109 34.1 28.8 43.2
65 46.0 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 50.651 ng
RT: 9.916 min Scan# 1161
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:122 Resp: 206730
Ion Ratio Lower Upper
122 100
107 125.0 94.5 141.7
121 62.2 47.4 71.2

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#28
bis(2-Chloroethoxy)methane
Concen: 49.900 ng
RT: 10.152 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion: 93 Resp: 265936
Ion Ratio Lower Upper
93 100
95 32.3 25.9 38.9
123 12.1 10.2 15.2





#29
2,4-Dichlorophenol
Concen: 45.078 ng
RT: 10.393 min Scan# 1242
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:162 Resp: 221105
Ion Ratio Lower Upper
162 100
164 65.0 44.9 84.9
98 34.4 14.3 54.3

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#30
1,2,4-Trichlorobenzene
Concen: 43.033 ng
RT: 10.599 min Scan# 1277
Delta R.T. -0.012 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:180 Resp: 237171
Ion Ratio Lower Upper
180 100
182 95.0 77.9 116.9
145 30.7 24.7 37.1





#31
Naphthalene
Concen: 44.245 ng
RT: 10.787 min Scan# 1309
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD



Tgt Ion:128 Resp: 694007
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.6
127 13.4 10.7 16.1




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#32
Benzoic acid
Concen: 44.795 ng
RT: 10.081 min Scan# 1189
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:122 Resp: 135500
Ion Ratio Lower Upper
122 100
105 124.6 103.2 143.2
77 92.5 76.2 116.2





#33
4-Chloroaniline
Concen: 10.013 ng
RT: 10.904 min Scan# 1329
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD



Tgt Ion:	127	Resp:	63449
Ion Ratio	Lower	Upper	
127	100		
129	32.8	26.1	39.1
65	30.6	22.5	33.7
92	21.2	15.0	22.6




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#34
Hexachlorobutadiene
Concen: 42.032 ng
RT: 11.069 min Scan# 1357
Delta R.T. -0.012 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:	225	Resp:	156690
Ion Ratio	Lower	Upper	
225	100		
223	61.5	51.0	76.6
227	64.8	51.4	77.2





#35
Caprolactam
Concen: 42.076 ng
RT: 11.704 min Scan# 1465
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:113 Resp: 59447
Ion Ratio Lower Upper
113 100
55 112.9 98.9 138.9
56 94.0 71.7 111.7

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#36
4-Chloro-3-methylphenol
Concen: 46.571 ng
RT: 12.034 min Scan# 1521
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion:107 Resp: 231943
Ion Ratio Lower Upper
107 100
144 26.2 21.1 31.7
142 81.9 66.0 99.0









#37
2-Methylnaphthalene
Concen: 45.070 ng
RT: 12.393 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:142 Resp: 503337
Ion Ratio Lower Upper
142 100
141 88.8 71.8 107.6
115 37.1 31.6 47.4

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#38
1-Methylnaphthalene
Concen: 43.760 ng
RT: 12.610 min Scan# 1619
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:142 Resp: 460326
Ion Ratio Lower Upper
142 100
141 92.6 73.3 109.9
116 4.3 3.5 5.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.563 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:164 Resp: 165180
Ion Ratio Lower Upper
164 100
162 99.7 80.2 120.2
160 43.6 34.5 51.7

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 45.801 ng
RT: 12.763 min Scan# 1645
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:216 Resp: 266296
Ion Ratio Lower Upper
216 100
214 78.4 61.9 92.9
179 21.5 16.8 25.2
108 19.9 14.9 22.3





#41
Hexachlorocyclopentadiene
Concen: 108.021 ng
RT: 12.740 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampled :
T-FLOORMSD



Tgt Ion:237 Resp: 321610
Ion Ratio Lower Upper
237 100
235 62.3 44.5 84.5
272 13.3 0.0 32.5




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#42
2,4,6-Tribromophenol
Concen: 139.720 ng
RT: 16.051 min Scan# 2204
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:330 Resp: 396053
Ion Ratio Lower Upper
330 100
332 98.7 77.8 116.8
141 29.3 25.4 38.2





#43
2,4,6-Trichlorophenol
Concen: 44.672 ng
RT: 13.004 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:196 Resp: 178151
Ion Ratio Lower Upper
196 100
198 96.6 79.5 119.3
200 32.1 25.0 37.6

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#44
2,4,5-Trichlorophenol
Concen: 46.152 ng
RT: 13.081 min Scan# 1699
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:196 Resp: 198218
Ion Ratio Lower Upper
196 100
198 95.4 74.7 112.1
97 47.8 37.4 56.0
132 27.1 20.0 30.0





#45
2-Fluorobiphenyl
Concen: 86.796 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD



Tgt Ion:172 Resp: 1091817
Ion Ratio Lower Upper
172 100
171 34.9 27.9 41.9
170 22.8 18.6 27.8




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#46
1,1'-Biphenyl
Concen: 46.273 ng
RT: 13.398 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:154 Resp: 616271
Ion Ratio Lower Upper
154 100
153 40.5 20.4 60.4
76 11.0 0.0 30.8





#47
2-Chloronaphthalene
Concen: 45.493 ng
RT: 13.440 min Scan# 1760
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:162 Resp: 494756
Ion Ratio Lower Upper
162 100
127 39.4 31.4 47.0
164 32.1 26.6 39.8

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#48
2-Nitroaniline
Concen: 48.309 ng
RT: 13.651 min Scan# 1796
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion: 65 Resp: 138138
Ion Ratio Lower Upper
65 100
92 73.1 57.7 86.5
138 123.2 92.1 138.1









#49
Acenaphthylene
Concen: 46.355 ng
RT: 14.287 min Scan# 1904
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:	152	Resp:	773425
Ion	Ratio	Lower	Upper
152	100		
151	19.8	16.2	24.2
153	12.8	10.6	16.0

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#50
Dimethylphthalate
Concen: 50.414 ng
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:	163	Resp:	684688
Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.2
164	10.4	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 45.262 ng
RT: 14.140 min Scan# 1879
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:165 Resp: 139719
Ion Ratio Lower Upper
165 100
63 52.4 41.3 61.9
89 58.2 44.8 67.2

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#52
Acenaphthene
Concen: 44.838 ng
RT: 14.622 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion:154 Resp: 454312
Ion Ratio Lower Upper
154 100
153 112.5 89.4 134.2
152 52.6 43.0 64.4









#53
3-Nitroaniline
Concen: 30.808 ng
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:138 Resp: 92171
Ion Ratio Lower Upper
138 100
108 11.3 8.4 12.6
92 110.1 87.8 131.8

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#54
2,4-Dinitrophenol
Concen: 86.017 ng
RT: 14.681 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:184 Resp: 158631
Ion Ratio Lower Upper
184 100
63 61.1 47.0 70.6
154 72.3 52.3 78.5





#55
Dibenzofuran
Concen: 44.171 ng
RT: 14.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:168 Resp: 736015
Ion Ratio Lower Upper
168 100
139 39.7 31.0 46.6
169 13.8 10.8 16.2

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#56
4-Nitrophenol
Concen: 88.562 ng
RT: 14.793 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:139 Resp: 214766
Ion Ratio Lower Upper
139 100
109 85.3 72.3 112.3
65 88.2 72.5 112.5





#57
2,4-Dinitrotoluene
Concen: 46.953 ng
RT: 14.928 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:165 Resp: 193853
Ion Ratio Lower Upper
165 100
63 39.5 32.3 48.5
89 70.2 55.9 83.9
182 2.8 2.2 3.4

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#58
Fluorene
Concen: 45.202 ng
RT: 15.604 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:166 Resp: 586859
Ion Ratio Lower Upper
166 100
165 98.6 78.0 117.0
167 13.5 10.5 15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 44.507 ng m
RT: 15.187 min Scan# 2057
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54





Tgt Ion:232 Resp: 168469
Ion Ratio Lower Upper
232 100
131 40.6 32.2 48.4
130 2.7 2.0 3.0
166 25.9 19.4 29.2



#60
Diethylphthalate
Concen: 47.417 ng
RT: 15.381 min Scan# 2090
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:149 Resp: 646101
Ion Ratio Lower Upper
149 100
177 21.8 17.5 26.3
150 12.9 9.9 14.9



Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD


Manual Integrations
APPROVED

mohammad
6/16/2021 3:54:30 PM



#61
4-Chlorophenyl-phenylether
Concen: 44.947 ng
RT: 15.598 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD



Tgt Ion:204 Resp: 302590
Ion Ratio Lower Upper
204 100
206 33.3 26.6 40.0
141 62.6 48.6 73.0




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#62
4-Nitroaniline
Concen: 45.211 ng
RT: 15.634 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:138 Resp: 139224
Ion Ratio Lower Upper
138 100
92 50.4 31.2 71.2
108 96.8 78.6 118.6









#63
Azobenzene
Concen: 43.865 ng
RT: 15.892 min Scan# 2177
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion: 77	Resp: 540750
Ion Ratio	Lower Upper
77 100	
182 27.5	6.0 46.0
105 20.8	0.0 38.8
51 26.8	8.8 48.8

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.304 min Scan# 2417
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:188	Resp: 334809
Ion Ratio	Lower Upper
188 100	
94 7.9	5.6 8.4
80 8.3	6.3 9.5





#65
4,6-Dinitro-2-methylphenol
Concen: 42.210 ng
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:198 Resp: 103855
Ion Ratio Lower Upper
198 100
51 33.9 14.0 54.0
105 40.5 19.5 59.5

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#66
n-Nitrosodiphenylamine
Concen: 47.003 ng
RT: 15.816 min Scan# 2164
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion:169 Resp: 521861
Ion Ratio Lower Upper
169 100
168 65.7 54.4 81.6
167 36.2 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 47.016 ng
RT: 16.492 min Scan# 2279
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:	248	Resp:	202042
Ion Ratio	Lower	Upper	
248	100		
250	96.4	79.8	119.8
141	77.7	57.8	86.6

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#68
Hexachlorobenzene
Concen: 47.440 ng
RT: 16.610 min Scan# 2299
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:	284	Resp:	244362
Ion Ratio	Lower	Upper	
284	100		
142	32.1	26.9	40.3
249	31.1	24.5	36.7









#69
Atrazine
Concen: 58.384 ng
RT: 16.769 min Scan# 2326
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:200 Resp: 201676
Ion Ratio Lower Upper
200 100
173 25.5 4.1 44.1
215 48.5 27.7 67.7

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#70
Pentachlorophenol
Concen: 100.366 ng
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:266 Resp: 287458
Ion Ratio Lower Upper
266 100
268 66.0 51.6 77.4
264 63.6 50.3 75.5





#71
Phenanthrene
Concen: 46.817 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD



Tgt Ion:178 Resp: 941340
Ion Ratio Lower Upper
178 100
176 18.7 15.1 22.7
179 15.0 12.4 18.6




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#72
Anthracene
Concen: 48.725 ng
RT: 17.439 min Scan# 2440
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:178 Resp: 964214
Ion Ratio Lower Upper
178 100
176 18.3 14.9 22.3
179 15.4 12.2 18.2





#73
Carbazole
Concen: 45.841 ng
RT: 17.710 min Scan# 2486
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:167 Resp: 870931
Ion Ratio Lower Upper
167 100
166 21.5 17.1 25.7
139 13.6 10.7 16.1

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#74
Di-n-butylphthalate
Concen: 51.679 ng
RT: 18.257 min Scan# 2579
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:149 Resp: 1137585
Ion Ratio Lower Upper
149 100
150 8.9 7.3 10.9
104 5.2 4.5 6.7





#75
Fluoranthene
Concen: 48.322 ng
RT: 19.310 min Scan# 2758
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:202 Resp: 1179935
Ion Ratio Lower Upper
202 100
101 8.2 0.0 27.6
203 17.1 0.0 37.8

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.380 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion:240 Resp: 387003
Ion Ratio Lower Upper
240 100
120 8.2 6.5 9.7
236 25.4 20.6 31.0









#77
Benzidine
Concen: 11.616 ng
RT: 19.486 min Scan# 2788
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:	184	Resp:	93542
Ion Ratio	Lower	Upper	
184	100		
185	14.0	11.0	16.4
183	12.7	9.5	14.3

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#78
Pyrene
Concen: 45.105 ng
RT: 19.657 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:	202	Resp:	1218604
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.3	24.5
203	17.6	14.2	21.2





#79
Terphenyl-d14
Concen: 92.186 ng
RT: 19.851 min Scan# 2850
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:244 Resp: 1905293
Ion Ratio Lower Upper
244 100
212 7.3 6.0 9.0
122 8.6 7.1 10.7

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#80
Butylbenzylphthalate
Concen: 49.526 ng
RT: 20.522 min Scan# 2964
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion:149 Resp: 550497
Ion Ratio Lower Upper
149 100
91 67.5 54.6 81.8
206 20.9 16.7 25.1









#81
Benzo(a)anthracene
Concen: 45.654 ng
RT: 21.363 min Scan# 3107
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion:228 Resp: 1275395
Ion Ratio Lower Upper
228 100
226 27.3 21.6 32.4
229 19.7 15.8 23.6

Instrument :
BNA_P
ClientSampled :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#82
3,3'-Dichlorobenzidine
Concen: 36.618 ng
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion:252 Resp: 353668
Ion Ratio Lower Upper
252 100
254 65.1 50.6 76.0
126 9.6 7.0 10.6





#83
Chrysene
Concen: 46.898 ng
RT: 21.416 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:228 Resp: 1268954
Ion Ratio Lower Upper
228 100
226 29.5 23.8 35.6
229 19.6 15.4 23.2

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#84
Bis(2-ethylhexyl)phthalate
Concen: 56.920 ng
RT: 21.280 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:149 Resp: 927878
Ion Ratio Lower Upper
149 100
167 27.9 21.9 32.9
279 5.4 4.3 6.5





#85
Di-n-octyl phthalate
Concen: 50.701 ng
RT: 22.204 min Scan# 3250
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:149 Resp: 1478550
Ion Ratio Lower Upper
149 100
167 1.6 1.4 2.0
43 6.5 5.6 8.4

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#86
Perylene-d12
Concen: 20.000 ng
RT: 23.786 min Scan# 3519
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion:264 Resp: 450369
Ion Ratio Lower Upper
264 100
260 23.1 18.6 28.0
265 22.1 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 47.759 ng
RT: 26.327 min Scan# 3951
Delta R.T. -0.012 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54

Tgt Ion: 276 Resp: 1839369
Ion Ratio Lower Upper
276 100
138 20.0 15.3 22.9
277 25.7 20.4 30.6

Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Manual Integrations
APPROVED


mohammad
6/16/2021 3:54:30 PM



#88
Benzo(b)fluoranthene
Concen: 47.825 ng
RT: 23.051 min Scan# 3394
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54


Tgt Ion: 252 Resp: 1475242
Ion Ratio Lower Upper
252 100
253 22.3 17.5 26.3
125 8.1 6.2 9.4





#89
Benzo(k)fluoranthene
Concen: 47.242 ng
RT: 23.098 min Scan# 3402
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:252 Resp: 1413064
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 8.1 6.0 9.0

Manual Integrations
APPROVED



mohammad
6/16/2021 3:54:30 PM




#90
Benzo(a)pyrene
Concen: 49.149 ng
RT: 23.686 min Scan# 3502
Delta R.T. 0.000 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Tgt Ion:252 Resp: 1425322
Ion Ratio Lower Upper
252 100
253 21.5 17.9 26.9
125 9.0 6.5 9.7





#91
Dibenzo(a,h)anthracene
Concen: 47.477 ng
RT: 26.345 min Scan# 3954
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54




Instrument :
BNA_P
ClientSampleId :
T-FLOORMSD




Tgt Ion:278 Resp: 1573783
Ion Ratio Lower Upper
278 100
139 12.2 9.3 13.9
279 23.9 18.8 28.2

Manual Integrations
APPROVED

mohammad
6/16/2021 3:54:30 PM



#92
Benzo(g,h,i)perylene
Concen: 47.466 ng
RT: 27.104 min Scan# 4083
Delta R.T. -0.006 min
Lab File: BP005937.D
Acq: 15 Jun 2021 18:54



Tgt Ion:276 Resp: 1554298
Ion Ratio Lower Upper
276 100
277 23.8 18.9 28.3
138 16.8 12.6 18.8

