

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP061620\
 Data File : BP002417.D
 Acq On : 16 Jun 2020 16:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 16 17:13:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP060520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 05 15:08:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	0.00
2	1,4-Dioxane	0.498	0.526	-5.6	66	0.00
3	Pyridine	1.353	1.494	-10.4	70	0.00
4	n-Nitrosodimethylamine	0.558	0.575	-3.0	62	0.00
5 S	2-Fluorophenol	1.081	1.087	-0.6	62	0.00
6	Aniline	1.957	2.022	-3.3	62	0.00
7 S	Phenol-d6	1.439	1.486	-3.3	63	0.00
8	2-Chlorophenol	1.215	1.220	-0.4	61	0.00
9	Benzaldehyde	0.735	0.772	-5.0	61	0.00
10 C	Phenol	1.561	1.615	-3.5	63	0.00
11	bis(2-Chloroethyl)ether	1.303	1.332	-2.2	62	0.00
12	1,3-Dichlorobenzene	1.399	1.368	2.2	60	0.00
13 C	1,4-Dichlorobenzene	1.408	1.383	1.8	60	0.00
14	1,2-Dichlorobenzene	1.349	1.338	0.8	61	0.00
15	Benzyl Alcohol	1.115	1.124	-0.8	60	0.00
16	2,2'-oxybis(1-Chloropropane	1.849	1.924	-4.1	64	0.00
17	2-Methylphenol	1.140	1.160	-1.8	61	0.00
18	Hexachloroethane	0.513	0.505	1.6	61	0.00
19 P	n-Nitroso-di-n-propylamine	0.962	1.004	-4.4	63	0.00
20	3+4-Methylphenols	1.539	1.571	-2.1	61	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	62	0.00
22	Acetophenone	0.449	0.472	-5.1	64	0.00
23 S	Nitrobenzene-d5	0.335	0.349	-4.2	63	0.00
24	Nitrobenzene	0.360	0.373	-3.6	63	0.00
25	Isophorone	0.682	0.698	-2.3	61	0.00
26 C	2-Nitrophenol	0.160	0.157	1.9	58	0.00
27	2,4-Dimethylphenol	0.252	0.253	-0.4	60	0.00
28	bis(2-Chloroethoxy)methane	0.406	0.422	-3.9	63	0.00
29 C	2,4-Dichlorophenol	0.283	0.277	2.1	59	0.00
30	1,2,4-Trichlorobenzene	0.316	0.310	1.9	60	0.00
31	Naphthalene	0.929	0.936	-0.8	62	0.00
32	Benzoic acid	0.191	0.168	12.0	47#	-0.02
33	4-Chloroaniline	0.406	0.408	-0.5	60	0.00
34 C	Hexachlorobutadiene	0.184	0.179	2.7	60	0.00
35	Caprolactam	0.100	0.098	2.0	56	-0.02
36 C	4-Chloro-3-methylphenol	0.307	0.303	1.3	58	0.00
37	2-Methylnaphthalene	0.659	0.659	0.0	61	0.00
38	1-Methylnaphthalene	0.619	0.622	-0.5	61	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	60	0.00
40	1,2,4,5-Tetrachlorobenzene	0.527	0.544	-3.2	61	0.00
41 P	Hexachlorocyclopentadiene	0.280	0.282	-0.7	57	0.00
42 S	2,4,6-Tribromophenol	0.191	0.194	-1.6	58	0.00
43 C	2,4,6-Trichlorophenol	0.373	0.365	2.1	56	0.00
44	2,4,5-Trichlorophenol	0.432	0.433	-0.2	57	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP061620\
 Data File : BP002417.D
 Acq On : 16 Jun 2020 16:33
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 16 17:13:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP060520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 05 15:08:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.189	1.199	-0.8	62	0.00
46	1,1'-Biphenyl	1.311	1.337	-2.0	60	0.00
47	2-Chloronaphthalene	1.072	1.093	-2.0	60	0.00
48	2-Nitroaniline	0.342	0.361	-5.6	59	0.00
49	Acenaphthylene	1.711	1.731	-1.2	59	0.00
50	Dimethylphthalate	1.370	1.381	-0.8	58	0.00
51	2,6-Dinitrotoluene	0.298	0.299	-0.3	56	0.00
52 C	Acenaphthene	1.002	1.024	-2.2	60	0.00
53	3-Nitroaniline	0.340	0.340	0.0	56	0.00
54 P	2,4-Dinitrophenol	0.163	0.151	7.4	50#	0.00
55	Dibenzofuran	1.614	1.650	-2.2	59	0.00
56 P	4-Nitrophenol	0.270	0.264	2.2	53	0.00
57	2,4-Dinitrotoluene	0.405	0.405	0.0	55	0.00
58	Fluorene	1.270	1.231	3.1	62	0.00
59	2,3,4,6-Tetrachlorophenol	0.343	0.332	3.2	55	0.00
60	Diethylphthalate	1.373	1.364	0.7	57	0.00
61	4-Chlorophenyl-phenylether	0.615	0.644	-4.7	63	0.00
62	4-Nitroaniline	0.327	0.332	-1.5	57	0.00
63	Azobenzene	1.367	1.445	-5.7	62	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	56	0.00
65	4,6-Dinitro-2-methylphenol	0.106	0.105	0.9	53	0.00
66 c	n-Nitrosodiphenylamine	0.511	0.536	-4.9	59	0.00
67	4-Bromophenyl-phenylether	0.189	0.199	-5.3	58	0.00
68	Hexachlorobenzene	0.201	0.204	-1.5	57	0.00
69	Atrazine	0.167	0.160	4.2	52	0.00
70 C	Pentachlorophenol	0.120	0.121	-0.8	52	0.00
71	Phenanthrene	0.936	0.970	-3.6	58	0.00
72	Anthracene	0.929	0.969	-4.3	59	0.00
73	Carbazole	0.914	0.945	-3.4	57	0.00
74	Di-n-butylphthalate	1.082	1.092	-0.9	56	0.00
75 C	Fluoranthene	1.117	1.149	-2.9	57	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
77	Benzidine	0.447	0.382	14.5	46#	0.00
78	Pyrene	1.225	1.242	-1.4	58	0.00
79 S	Terphenyl-d14	0.763	0.829	-8.7	64	0.00
80	Butylbenzylphthalate	0.545	0.513	5.9	53	0.00
81	Benzo(a)anthracene	1.142	1.155	-1.1	59	0.00
82	3,3'-Dichlorobenzidine	0.424	0.393	7.3	51	0.00
83	Chrysene	1.082	1.105	-2.1	60	0.00
84	Bis(2-ethylhexyl)phthalate	0.792	0.776	2.0	56	0.00
85 c	Di-n-octyl phthalate	1.389	1.314	5.4	54	0.00
86 I	Perylene-d12	1.000	1.000	0.0	57	0.00
87	Indeno(1,2,3-cd)pyrene	1.251	1.248	0.2	55	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP061620\
Data File : BP002417.D
Acq On : 16 Jun 2020 16:33
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 16 17:13:50 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP060520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 05 15:08:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.079	1.098	-1.8	54	0.00
89	Benzo(k)fluoranthene	1.025	1.044	-1.9	58	0.00
90 C	Benzo(a)pyrene	1.011	0.998	1.3	54	0.00
91	Dibenzo(a,h)anthracene	1.003	1.021	-1.8	56	0.00
92	Benzo(q,h,i)perylene	1.039	1.030	0.9	54	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0