

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061621\
 Data File : BP005946.D
 Acq On : 16 Jun 2021 12:21
 Operator : CG/JU
 Sample : PB137148BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB137148BSD

Manual Integrations
 APPROVED

mohammad
 6/17/2021 12:16:18 PM

Quant Time: Jun 16 13:03:32 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 11:35:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	75302	20.000	ng	0.00
21) Naphthalene-d8	10.734	136	296792	20.000	ng	0.00
39) Acenaphthene-d10	14.557	164	171918	20.000	ng	0.00
64) Phenanthrene-d10	17.304	188	347777	20.000	ng	0.00
76) Chrysene-d12	21.380	240	406805	20.000	ng	0.00
86) Perylene-d12	23.792	264	457821	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.499	112	651385	138.805	ng	0.00
7) Phenol-d6	7.105	99	764882	125.750	ng	0.00
23) Nitrobenzene-d5	9.099	82	481236	79.495	ng	0.00
42) 2,4,6-Tribromophenol	16.051	330	378499	128.294	ng	0.00
45) 2-Fluorobiphenyl	13.187	172	1103565	84.292	ng	0.00
79) Terphenyl-d14	19.857	244	1686460	77.626	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.364	88	66704	31.462	ng	97
3) Pyridine	3.775	79	171292	34.376	ng	99
4) n-Nitrosodimethylamine	3.687	42	85734	37.015	ng	83
6) Aniline	7.258	93	259819	38.261	ng	99
8) 2-Chlorophenol	7.499	128	238622	44.349	ng	98
9) Benzaldehyde	7.069	77	122821	47.366	ng	99
10) Phenol	7.134	94	276998	45.587	ng	100
11) bis(2-Chloroethyl)ether	7.352	93	204601	44.436	ng	99
12) 1,3-Dichlorobenzene	7.816	146	252069	41.897	ng	99
13) 1,4-Dichlorobenzene	7.963	146	259176	42.241	ng	100
14) 1,2-Dichlorobenzene	8.281	146	246136	41.968	ng	98
15) Benzyl Alcohol	8.175	79	201839	44.056	ng	96
16) 2,2'-oxybis(1-Chloropr...	8.469	45	174539	41.007	ng	97
17) 2-Methylphenol	8.381	107	185552	44.819	ng	99
18) Hexachloroethane	9.010	117	92463	41.661	ng	94
19) n-Nitroso-di-n-propyla...	8.746	70	153594	41.047	ng	95
20) 3+4-Methylphenols	8.710	107	251670	45.002	ng	96
22) Acetophenone	8.757	105	330301	42.282	ng	# 98
24) Nitrobenzene	9.140	77	236995	37.701	ng	99
25) Isophorone	9.669	82	408279	36.521	ng	99
26) 2-Nitrophenol	9.852	139	124274	41.434	ng	93
27) 2,4-Dimethylphenol	9.916	122	207694	46.481	ng	99
28) bis(2-Chloroethoxy)met...	10.146	93	253424	43.435	ng	99
29) 2,4-Dichlorophenol	10.387	162	218939	40.772	ng	98
30) 1,2,4-Trichlorobenzene	10.599	180	234569	38.876	ng	98
31) Naphthalene	10.787	128	678773	39.527	ng	99
32) Benzoic acid	10.081	122	130770	40.237	ng	99
33) 4-Chloroaniline	10.904	127	141054	20.333	ng	98
34) Hexachlorobutadiene	11.069	225	154055	37.748	ng	98
35) Caprolactam	11.698	113	63896m	41.310	ng	
36) 4-Chloro-3-methylphenol	12.028	107	224896	41.247	ng	98
37) 2-Methylnaphthalene	12.393	142	488145	39.926	ng	98
38) 1-Methylnaphthalene	12.610	142	448811	38.971	ng	99
40) 1,2,4,5-Tetrachloroben...	12.763	216	261087	43.145	ng	98
41) Hexachlorocyclopentadiene	12.740	237	289016	94.020	ng	98
43) 2,4,6-Trichlorophenol	13.004	196	172995	41.679	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP061621\
 Data File : BP005946.D
 Acq On : 16 Jun 2021 12:21
 Operator : CG/JU
 Sample : PB137148BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB137148BSD


Manual Integrations
 APPROVED

mohammad
 6/17/2021 12:16:18 PM

Quant Time: Jun 16 13:03:32 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 11:35:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.075	196	187341	41.910	ng	98
46) 1,1'-Biphenyl	13.398	154	605153	43.657	ng	100
47) 2-Chloronaphthalene	13.440	162	467731	41.323	ng	98
48) 2-Nitroaniline	13.645	65	126646	42.555	ng	98
49) Acenaphthylene	14.281	152	733691	42.250	ng	100
50) Dimethylphthalate	14.022	163	590976	41.809	ng	99
51) 2,6-Dinitrotoluene	14.140	165	130137	40.506	ng	98
52) Acenaphthene	14.622	154	430758	40.847	ng	98
53) 3-Nitroaniline	14.469	138	88626	28.462	ng	96
54) 2,4-Dinitrophenol	14.681	184	124458	66.035	ng	95
55) Dibenzofuran	14.957	168	684882	39.491	ng	98
56) 4-Nitrophenol	14.787	139	206172	81.686	ng	96
57) 2,4-Dinitrotoluene	14.928	165	176048	40.969	ng	# 98
58) Fluorene	15.604	166	549252	40.648	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.187	232	163212m	41.428	ng	
60) Diethylphthalate	15.381	149	589729	41.583	ng	100
61) 4-Chlorophenyl-phenyle...	15.598	204	281452	40.168	ng	97
62) 4-Nitroaniline	15.634	138	129118	40.286	ng	98
63) Azobenzene	15.892	77	497468	38.772	ng	96
65) 4,6-Dinitro-2-methylph...	15.692	198	85712	33.537	ng	98
66) n-Nitrosodiphenylamine	15.816	169	479184	41.550	ng	98
67) 4-Bromophenyl-phenylether	16.492	248	183174	41.036	ng	96
68) Hexachlorobenzene	16.610	284	223157	41.708	ng	99
69) Atrazine	16.769	200	184609	51.450	ng	98
70) Pentachlorophenol	16.957	266	262992	88.400	ng	99
71) Phenanthrene	17.351	178	860542	41.203	ng	100
72) Anthracene	17.439	178	880337	42.828	ng	99
73) Carbazole	17.710	167	784951	39.775	ng	99
74) Di-n-butylphthalate	18.257	149	981379	42.921	ng	100
75) Fluoranthene	19.310	202	1017018	40.097	ng	98
77) Benzidine	19.486	184	474001	55.997	ng	100
78) Pyrene	19.663	202	1054548	37.132	ng	99
80) Butylbenzylphthalate	20.527	149	462537	39.587	ng	97
81) Benzo(a)anthracene	21.363	228	1133459	38.598	ng	99
82) 3,3'-Dichlorobenzidine	21.298	252	337650	33.258	ng	98
83) Chrysene	21.416	228	1108879	38.987	ng	99
84) Bis(2-ethylhexyl)phtha...	21.280	149	704088	41.089	ng	99
85) Di-n-octyl phthalate	22.204	149	1267706	41.355	ng	99
87) Indeno(1,2,3-cd)pyrene	26.333	276	1719312	43.915	ng	99
88) Benzo(b)fluoranthene	23.057	252	1339595	42.721	ng	99
89) Benzo(k)fluoranthene	23.104	252	1243032	40.881	ng	99
90) Benzo(a)pyrene	23.686	252	1280384	43.432	ng	98
91) Dibenzo(a,h)anthracene	26.351	278	1464644	43.465	ng	99
92) Benzo(g,h,i)perylene	27.115	276	1468309	44.110	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.928 min Scan# 823
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion:152 Resp: 75302
Ion Ratio Lower Upper
152 100
150 149.5 125.6 188.4
115 63.2 48.5 72.7

Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#2
1,4-Dioxane
Concen: 31.462 ng
RT: 3.364 min Scan# 47
Delta R.T. -0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion: 88 Resp: 66704
Ion Ratio Lower Upper
88 100
58 59.9 46.7 70.1
43 23.6 20.3 30.5






#3
Pyridine
Concen: 34.376 ng
RT: 3.775 min Scan# 117
Delta R.T. -0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM







Tgt Ion: 79 Resp: 171292
Ion Ratio Lower Upper
79 100
52 52.5 41.7 62.5
51 29.9 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 37.015 ng
RT: 3.687 min Scan# 102
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion: 42 Resp: 85734
Ion Ratio Lower Upper
42 100
74 151.9 105.4 158.0
44 5.4 4.2 6.4





#5
2-Fluorophenol
Concen: 138.805 ng
RT: 5.499 min Scan# 410
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:112 Resp: 651385
Ion Ratio Lower Upper
112 100
64 51.3 41.4 62.2
63 28.0 23.8 35.8

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#6
Aniline
Concen: 38.261 ng
RT: 7.258 min Scan# 709
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion: 93 Resp: 259819
Ion Ratio Lower Upper
93 100
66 37.6 29.7 44.5
65 18.3 15.3 22.9





#7
Phenol-d6
Concen: 125.750 ng
RT: 7.105 min Scan# 683
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion: 99 Resp: 764882
Ion Ratio Lower Upper
99 100
42 15.3 13.4 20.0
71 38.0 30.6 46.0

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#8
2-Chlorophenol
Concen: 44.349 ng
RT: 7.499 min Scan# 750
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion: 128 Resp: 238622
Ion Ratio Lower Upper
128 100
130 33.1 12.4 52.4
64 39.2 20.3 60.3





#9
Benzaldehyde
Concen: 47.366 ng
RT: 7.069 min Scan# 677
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion: 77 Resp: 122821
Ion Ratio Lower Upper
77 100
105 98.7 76.7 116.7
106 95.6 76.2 116.2

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#10
Phenol
Concen: 45.587 ng
RT: 7.134 min Scan# 688
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21



Tgt Ion: 94 Resp: 276998
Ion Ratio Lower Upper
94 100
65 28.4 8.8 48.8
66 40.9 20.7 60.7






#11
bis(2-Chloroethyl)ether
Concen: 44.436 ng
RT: 7.352 min Scan# 725
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM







Tgt Ion: 93 Resp: 204601
Ion Ratio Lower Upper
93 100
63 65.5 46.2 86.2
95 31.8 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 41.897 ng
RT: 7.816 min Scan# 804
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion: 146 Resp: 252069
Ion Ratio Lower Upper
146 100
148 63.7 51.5 77.3
75 32.2 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 42.241 ng
RT: 7.963 min Scan# 829
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion:146 Resp: 259176
Ion Ratio Lower Upper
146 100
148 64.2 51.3 76.9
111 41.1 33.0 49.6

Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#14
1,2-Dichlorobenzene
Concen: 41.968 ng
RT: 8.281 min Scan# 883
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:146 Resp: 246136
Ion Ratio Lower Upper
146 100
148 64.2 50.3 75.5
111 43.7 33.7 50.5





#15
Benzyl Alcohol
Concen: 44.056 ng
RT: 8.175 min Scan# 865
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion: 79 Resp: 201839
Ion Ratio Lower Upper
79 100
108 73.3 62.5 93.7
77 64.5 52.9 79.3

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.007 ng
RT: 8.469 min Scan# 915
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion: 45 Resp: 174539
Ion Ratio Lower Upper
45 100
77 23.8 2.0 42.0
79 17.3 0.0 36.7





#17
2-Methylphenol
Concen: 44.819 ng
RT: 8.381 min Scan# 900
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:107 Resp: 185552
Ion Ratio Lower Upper
107 100
108 111.1 87.7 131.5
77 49.2 40.2 60.2
79 48.9 39.0 58.4

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#18
Hexachloroethane
Concen: 41.661 ng
RT: 9.010 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21



Tgt Ion:117 Resp: 92463
Ion Ratio Lower Upper
117 100
119 99.2 76.2 114.2
201 113.1 83.4 125.0






#19

n-Nitroso-di-n-propylamine
Concen: 41.047 ng
RT: 8.746 min Scan# 962
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM






Tgt Ion: 70	Resp: 153594
Ion Ratio	Lower Upper
70 100	
42 45.2	40.8 61.2
101 10.1	8.8 13.2
130 23.6	18.9 28.3




#20

3+4-Methylphenols
Concen: 45.002 ng
RT: 8.710 min Scan# 956
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt	Ion:107	Resp:	251670
Ion	Ratio	Lower	Upper
107	100		
108	88.2	71.0	111.0
77	33.8	16.6	56.6
79	29.3	12.7	52.7






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.734 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM







Tgt Ion:136 Resp: 296792
Ion Ratio Lower Upper
136 100
137 11.0 8.6 12.8
54 5.8 4.6 6.8
68 4.8 3.6 5.4



#22
Acetophenone
Concen: 42.282 ng
RT: 8.757 min Scan# 964
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion:105 Resp: 330301
Ion Ratio Lower Upper
105 100
71 3.5 1.9 2.9#
51 21.3 18.4 27.6
120 22.0 17.9 26.9









#23
Nitrobenzene-d5
Concen: 79.495 ng
RT: 9.099 min Scan# 1022
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion: 82 Resp: 481236
Ion Ratio Lower Upper
82 100
128 45.6 35.7 53.5
54 40.2 35.0 52.4

Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#24
Nitrobenzene
Concen: 37.701 ng
RT: 9.140 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion: 77 Resp: 236995
Ion Ratio Lower Upper
77 100
123 47.5 38.2 57.2
65 13.5 11.4 17.2






#25
Isophorone
Concen: 36.521 ng
RT: 9.669 min Scan# 1119
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM







Tgt Ion: 82 Resp: 408279
Ion Ratio Lower Upper
82 100
95 8.5 6.5 9.7
138 19.1 15.0 22.4



#26
2-Nitrophenol
Concen: 41.434 ng
RT: 9.852 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion: 139 Resp: 124274
Ion Ratio Lower Upper
139 100
109 33.7 28.8 43.2
65 45.0 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 46.481 ng
RT: 9.916 min Scan# 1161
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion:122 Resp: 207694
Ion Ratio Lower Upper
122 100
107 116.2 94.5 141.7
121 58.8 47.4 71.2

Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#28
bis(2-Chloroethoxy)methane
Concen: 43.435 ng
RT: 10.146 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion: 93 Resp: 253424
Ion Ratio Lower Upper
93 100
95 32.9 25.9 38.9
123 12.3 10.2 15.2





#29
2,4-Dichlorophenol
Concen: 40.772 ng
RT: 10.387 min Scan# 1241
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:162 Resp: 218939
Ion Ratio Lower Upper
162 100
164 62.8 44.9 84.9
98 35.1 14.3 54.3

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#30
1,2,4-Trichlorobenzene
Concen: 38.876 ng
RT: 10.599 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21



Tgt Ion:180 Resp: 234569
Ion Ratio Lower Upper
180 100
182 94.8 77.9 116.9
145 31.5 24.7 37.1









#31
Naphthalene
Concen: 39.527 ng
RT: 10.787 min Scan# 1309
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion:128 Resp: 678773
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 13.3 10.7 16.1

Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#32
Benzoic acid
Concen: 40.237 ng
RT: 10.081 min Scan# 1189
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:122 Resp: 130770
Ion Ratio Lower Upper
122 100
105 124.5 103.2 143.2
77 96.8 76.2 116.2





#33
4-Chloroaniline
Concen: 20.333 ng
RT: 10.904 min Scan# 1329
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD



Tgt Ion:	127	Resp:	141054
Ion	Ratio	Lower	Upper
127	100		
129	33.6	26.1	39.1
65	28.6	22.5	33.7
92	19.6	15.0	22.6




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#34
Hexachlorobutadiene
Concen: 37.748 ng
RT: 11.069 min Scan# 1357
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:	225	Resp:	154055
Ion	Ratio	Lower	Upper
225	100		
223	63.3	51.0	76.6
227	66.7	51.4	77.2





#35
Caprolactam
Concen: 41.310 ng m
RT: 11.698 min Scan# 1464
Delta R.T. -0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:113 Resp: 63896
Ion Ratio Lower Upper
113 100
55 110.1 98.9 138.9
56 96.2 71.7 111.7

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#36
4-Chloro-3-methylphenol
Concen: 41.247 ng
RT: 12.028 min Scan# 1520
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21



Tgt Ion:107 Resp: 224896
Ion Ratio Lower Upper
107 100
144 25.8 21.1 31.7
142 80.6 66.0 99.0









#37
2-Methylnaphthalene
Concen: 39.926 ng
RT: 12.393 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21

Tgt Ion:142 Resp: 488145
Ion Ratio Lower Upper
142 100
141 88.1 71.8 107.6
115 37.7 31.6 47.4

Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#38
1-Methylnaphthalene
Concen: 38.971 ng
RT: 12.610 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:142 Resp: 448811
Ion Ratio Lower Upper
142 100
141 92.6 73.3 109.9
116 3.9 3.5 5.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:164 Resp: 171918
Ion Ratio Lower Upper
164 100
162 101.0 80.2 120.2
160 44.8 34.5 51.7

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 43.145 ng
RT: 12.763 min Scan# 1645
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:216 Resp: 261087
Ion Ratio Lower Upper
216 100
214 78.4 61.9 92.9
179 21.9 16.8 25.2
108 20.0 14.9 22.3





#41
Hexachlorocyclopentadiene
Concen: 94.020 ng
RT: 12.740 min Scan# 1641
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:237 Resp: 289016
Ion Ratio Lower Upper
237 100
235 62.6 44.5 84.5
272 13.2 0.0 32.5

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#42
2,4,6-Tribromophenol
Concen: 128.294 ng
RT: 16.051 min Scan# 2204
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:330 Resp: 378499
Ion Ratio Lower Upper
330 100
332 97.2 77.8 116.8
141 29.2 25.4 38.2





#43
2,4,6-Trichlorophenol
Concen: 41.679 ng
RT: 13.004 min Scan# 1686
Delta R.T. 0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:196 Resp: 172995
Ion Ratio Lower Upper
196 100
198 95.9 79.5 119.3
200 32.0 25.0 37.6

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#44
2,4,5-Trichlorophenol
Concen: 41.910 ng
RT: 13.075 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:196 Resp: 187341
Ion Ratio Lower Upper
196 100
198 93.8 74.7 112.1
97 49.2 37.4 56.0
132 26.9 20.0 30.0





#45
2-Fluorobiphenyl
Concen: 84.292 ng
RT: 13.187 min Scan# 1717
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:172 Resp: 1103565
Ion Ratio Lower Upper
172 100
171 35.7 27.9 41.9
170 23.2 18.6 27.8

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#46
1,1'-Biphenyl
Concen: 43.657 ng
RT: 13.398 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:154 Resp: 605153
Ion Ratio Lower Upper
154 100
153 40.6 20.4 60.4
76 10.9 0.0 30.8





#47
2-Chloronaphthalene
Concen: 41.323 ng
RT: 13.440 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:162 Resp: 467731
Ion Ratio Lower Upper
162 100
127 41.1 31.4 47.0
164 32.6 26.6 39.8

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#48
2-Nitroaniline
Concen: 42.555 ng
RT: 13.645 min Scan# 1795
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion: 65 Resp: 126646
Ion Ratio Lower Upper
65 100
92 70.8 57.7 86.5
138 113.1 92.1 138.1





#49
Acenaphthylene
Concen: 42.250 ng
RT: 14.281 min Scan# 1903
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD



Tgt Ion:152 Resp: 733691
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2
153 13.3 10.6 16.0




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#50
Dimethylphthalate
Concen: 41.809 ng
RT: 14.022 min Scan# 1859
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:163 Resp: 590976
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.2
164 10.3 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 40.506 ng
RT: 14.140 min Scan# 1879
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:165 Resp: 130137
Ion Ratio Lower Upper
165 100
63 51.1 41.3 61.9
89 58.1 44.8 67.2

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#52
Acenaphthene
Concen: 40.847 ng
RT: 14.622 min Scan# 1961
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:154 Resp: 430758
Ion Ratio Lower Upper
154 100
153 114.4 89.4 134.2
152 53.0 43.0 64.4





#53
3-Nitroaniline
Concen: 28.462 ng
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:138 Resp: 88626
Ion Ratio Lower Upper
138 100
108 11.3 8.4 12.6
92 114.3 87.8 131.8

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#54
2,4-Dinitrophenol
Concen: 66.035 ng
RT: 14.681 min Scan# 1971
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21



Tgt Ion:184 Resp: 124458
Ion Ratio Lower Upper
184 100
63 59.3 47.0 70.6
154 71.7 52.3 78.5






#55
Dibenzofuran
Concen: 39.491 ng
RT: 14.957 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM







Tgt Ion:168 Resp: 684882
Ion Ratio Lower Upper
168 100
139 40.3 31.0 46.6
169 13.3 10.8 16.2



#56
4-Nitrophenol
Concen: 81.686 ng
RT: 14.787 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:139 Resp: 206172
Ion Ratio Lower Upper
139 100
109 86.2 72.3 112.3
65 91.0 72.5 112.5





#57
2,4-Dinitrotoluene
Concen: 40.969 ng
RT: 14.928 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:165 Resp: 176048
Ion Ratio Lower Upper
165 100
63 41.6 32.3 48.5
89 70.9 55.9 83.9
182 3.4 2.2 3.4#

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#58
Fluorene
Concen: 40.648 ng
RT: 15.604 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:166 Resp: 549252
Ion Ratio Lower Upper
166 100
165 98.5 78.0 117.0
167 13.5 10.5 15.7





#59
2,3,4,6-Tetrachlorophenol
Concen: 41.428 ng m
RT: 15.187 min Scan# 2057
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:232 Resp: 163212
Ion Ratio Lower Upper
232 100
131 38.4 32.2 48.4
130 2.5 2.0 3.0
166 24.6 19.4 29.2

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#60
Diethylphthalate
Concen: 41.583 ng
RT: 15.381 min Scan# 2090
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:149 Resp: 589729
Ion Ratio Lower Upper
149 100
177 22.0 17.5 26.3
150 12.2 9.9 14.9





#61
4-Chlorophenyl-phenylether
Concen: 40.168 ng
RT: 15.598 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:204 Resp: 281452
Ion Ratio Lower Upper
204 100
206 33.0 26.6 40.0
141 64.0 48.6 73.0

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#62
4-Nitroaniline
Concen: 40.286 ng
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21



Tgt Ion:138 Resp: 129118
Ion Ratio Lower Upper
138 100
92 50.1 31.2 71.2
108 96.7 78.6 118.6






#63
Azobenzene
Concen: 38.772 ng
RT: 15.892 min Scan# 2177
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM







Tgt Ion: 77 Resp: 497468
Ion Ratio Lower Upper
77 100
182 28.5 6.0 46.0
105 20.7 0.0 38.8
51 27.2 8.8 48.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.304 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion: 188 Resp: 347777
Ion Ratio Lower Upper
188 100
94 7.6 5.6 8.4
80 8.0 6.3 9.5





#65
4,6-Dinitro-2-methylphenol
Concen: 33.537 ng
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:198 Resp: 85712
Ion Ratio Lower Upper
198 100
51 33.1 14.0 54.0
105 41.2 19.5 59.5

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#66
n-Nitrosodiphenylamine
Concen: 41.550 ng
RT: 15.816 min Scan# 2164
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:169 Resp: 479184
Ion Ratio Lower Upper
169 100
168 66.1 54.4 81.6
167 35.8 29.0 43.6





#67
4-Bromophenyl-phenylether
Concen: 41.036 ng
RT: 16.492 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:248 Resp: 183174
Ion Ratio Lower Upper
248 100
250 98.1 79.8 119.8
141 77.9 57.8 86.6

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#68
Hexachlorobenzene
Concen: 41.708 ng
RT: 16.610 min Scan# 2299
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:284 Resp: 223157
Ion Ratio Lower Upper
284 100
142 33.1 26.9 40.3
249 31.1 24.5 36.7









#69
Atrazine
Concen: 51.450 ng
RT: 16.769 min Scan# 2326
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21





Tgt Ion:200 Resp: 184609
Ion Ratio Lower Upper
200 100
173 25.4 4.1 44.1
215 47.0 27.7 67.7



#70
Pentachlorophenol
Concen: 88.400 ng
RT: 16.957 min Scan# 2358
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:266 Resp: 262992
Ion Ratio Lower Upper
266 100
268 64.0 51.6 77.4
264 64.1 50.3 75.5



Instrument :
BNA_P
ClientSampleId :
PB137148BSD


Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM



#71
Phenanthrene
Concen: 41.203 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD



Tgt Ion:178 Resp: 860542
Ion Ratio Lower Upper
178 100
176 19.0 15.1 22.7
179 15.3 12.4 18.6




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#72
Anthracene
Concen: 42.828 ng
RT: 17.439 min Scan# 2440
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:178 Resp: 880337
Ion Ratio Lower Upper
178 100
176 18.1 14.9 22.3
179 15.0 12.2 18.2





#73
 Carbazole
 Concen: 39.775 ng
 RT: 17.710 min Scan# 2486
 Delta R.T. 0.000 min
 Lab File: BP005946.D
 Acq: 16 Jun 2021 12:21




Instrument :
 BNA_P
 ClientSampleId :
 PB137148BSD



Tgt Ion:167 Resp: 784951
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 13.7 10.7 16.1




Manual Integrations
 APPROVED


mohammad
 6/17/2021 12:16:18 PM



#74
 Di-n-butylphthalate
 Concen: 42.921 ng
 RT: 18.257 min Scan# 2579
 Delta R.T. 0.000 min
 Lab File: BP005946.D
 Acq: 16 Jun 2021 12:21

Tgt Ion:149 Resp: 981379
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 5.2 4.5 6.7






#75
Fluoranthene
Concen: 40.097 ng
RT: 19.310 min Scan# 2758
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM







Tgt Ion:202 Resp: 1017018
Ion Ratio Lower Upper
202 100
101 8.5 0.0 27.6
203 17.2 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.380 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:240 Resp: 406805
Ion Ratio Lower Upper
240 100
120 8.5 6.5 9.7
236 25.4 20.6 31.0





#77
Benzidine
Concen: 55.997 ng
RT: 19.486 min Scan# 2788
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:184 Resp: 474001
Ion Ratio Lower Upper
184 100
185 13.9 11.0 16.4
183 12.0 9.5 14.3

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#78
Pyrene
Concen: 37.132 ng
RT: 19.663 min Scan# 2818
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:202 Resp: 1054548
Ion Ratio Lower Upper
202 100
200 21.0 16.3 24.5
203 17.2 14.2 21.2





#79
Terphenyl-d14
Concen: 77.626 ng
RT: 19.857 min Scan# 2851
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:244 Resp: 1686460
Ion Ratio Lower Upper
244 100
212 7.1 6.0 9.0
122 8.2 7.1 10.7

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#80
Butylbenzylphthalate
Concen: 39.587 ng
RT: 20.527 min Scan# 2965
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:149 Resp: 462537
Ion Ratio Lower Upper
149 100
91 65.5 54.6 81.8
206 22.5 16.7 25.1





#81
Benzo(a)anthracene
Concen: 38.598 ng
RT: 21.363 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD



Tgt Ion:228 Resp: 1133459
Ion Ratio Lower Upper
228 100
226 27.2 21.6 32.4
229 20.1 15.8 23.6




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#82
3,3'-Dichlorobenzidine
Concen: 33.258 ng
RT: 21.298 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:252 Resp: 337650
Ion Ratio Lower Upper
252 100
254 65.0 50.6 76.0
126 9.1 7.0 10.6





#83
Chrysene
Concen: 38.987 ng
RT: 21.416 min Scan# 3116
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:228 Resp: 1108879
Ion Ratio Lower Upper
228 100
226 29.3 23.8 35.6
229 19.6 15.4 23.2

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#84
Bis(2-ethylhexyl)phthalate
Concen: 41.089 ng
RT: 21.280 min Scan# 3093
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:149 Resp: 704088
Ion Ratio Lower Upper
149 100
167 27.3 21.9 32.9
279 4.7 4.3 6.5





#85
Di-n-octyl phthalate
Concen: 41.355 ng
RT: 22.204 min Scan# 3250
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:149 Resp: 1267706
Ion Ratio Lower Upper
149 100
167 1.6 1.4 2.0
43 6.6 5.6 8.4

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#86
Perylene-d12
Concen: 20.000 ng
RT: 23.792 min Scan# 3520
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:264 Resp: 457821
Ion Ratio Lower Upper
264 100
260 23.8 18.6 28.0
265 22.0 17.6 26.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 43.915 ng
RT: 26.333 min Scan# 3952
Delta R.T. -0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD



Tgt Ion:276 Resp: 1719312
Ion Ratio Lower Upper
276 100
138 19.8 15.3 22.9
277 25.7 20.4 30.6




Manual Integrations
APPROVED


mohammad
6/17/2021 12:16:18 PM



#88
Benzo(b)fluoranthene
Concen: 42.721 ng
RT: 23.057 min Scan# 3395
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21


Tgt Ion:252 Resp: 1339595
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 7.9 6.2 9.4





#89
Benzo(k)fluoranthene
Concen: 40.881 ng
RT: 23.104 min Scan# 3403
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:252 Resp: 1243032
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 8.1 6.0 9.0

Manual Integrations
APPROVED



mohammad
6/17/2021 12:16:18 PM




#90
Benzo(a)pyrene
Concen: 43.432 ng
RT: 23.686 min Scan# 3502
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Tgt Ion:252 Resp: 1280384
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 8.8 6.5 9.7





#91
Dibenzo(a,h)anthracene
Concen: 43.465 ng
RT: 26.351 min Scan# 3955
Delta R.T. 0.000 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21




Instrument :
BNA_P
ClientSampleId :
PB137148BSD




Tgt Ion:278 Resp: 1464644
Ion Ratio Lower Upper
278 100
139 12.4 9.3 13.9
279 23.4 18.8 28.2

Manual Integrations
APPROVED

mohammad
6/17/2021 12:16:18 PM



#92
Benzo(g,h,i)perylene
Concen: 44.110 ng
RT: 27.115 min Scan# 4085
Delta R.T. -0.006 min
Lab File: BP005946.D
Acq: 16 Jun 2021 12:21



Tgt Ion:276 Resp: 1468309
Ion Ratio Lower Upper
276 100
277 23.6 18.9 28.3
138 17.1 12.6 18.8

