

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/16/2025	BNA_P	<u>BP024954.D</u>
Compound Name	Response	Retention Time
DDT	2334210	20.671
DDD	110076	20.218
DDE	7204	19.683
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
117280	2451490	4.78