

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062022\
 Data File : BP010760.D
 Acq On : 20 Jun 2022 13:28
 Operator : CG/JU
 Sample : SST002005
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST0020605

Quant Time: Jun 20 14:51:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 14:42:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	138471	20.000	ng/ul	0.00
20) Naphthalene-d8	10.616	136	521699	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.457	164	318481	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.216	188	664507	20.000	ng/ul	0.00
79) Chrysene-d12	21.310	240	753443	20.000	ng/ul	0.00
88) Perylene-d12	23.698	264	807163	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.276	96	26434	8.024	ng/uL	0.00
4) Pyridine-d5	3.693	84	158189	18.299	ng/ul	0.00
7) Phenol-d5	6.999	99	194532	18.518	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	120905	17.168	ng/ul	0.00
11) 2-Chlorophenol-d4	7.352	132	169556	20.067	ng/ul	0.00
15) 4-Methylphenol-d8	8.534	113	158806	18.589	ng/ul	0.00
21) Nitrobenzene-d5	8.975	128	78579	19.107	ng/ul	0.00
24) 2-Nitrophenol-d4	9.699	143	84261	18.225	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.246	165	155938	19.605	ng/ul	0.00
31) 4-Chloroaniline-d4	10.758	131	213505	19.572	ng/ul	0.00
46) Dimethylphthalate-d6	13.869	166	452766	20.003	ng/ul	0.00
49) Acenaphthylene-d8	14.151	160	543749	20.301	ng/ul	0.00
54) 4-Nitrophenol-d4	14.687	143	58178	16.035	ng/ul	0.00
60) Fluorene-d10	15.451	176	401648	20.526	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.587	200	70309	17.701	ng/ul	0.00
73) Anthracene-d10	17.216	188	664507	22.082	ng/ul	-0.10
81) Pyrene-d10	19.551	212	735683	18.799	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.545	264	819255	20.562	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	27416	8.037	ng/ul#	75
5) Pyridine	3.717	79	168046	18.820	ng/ul#	71
6) Benzaldehyde	6.964	77	108589	19.942	ng/ul	91
8) Phenol	7.022	94	208882	18.543	ng/ul	89
10) Bis(2-Chloroethyl)ether	7.246	93	168626	19.110	ng/ul#	84
12) 2-Chlorophenol	7.387	128	179026	20.261	ng/ul#	88
13) 2-Methylphenol	8.269	108	152734	18.242	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.346	45	199711	13.800	ng/ul#	92
16) Acetophenone	8.758	105	533	0.038	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.628	70	118088	16.440	ng/ul#	86
18) 4-Methylphenol	8.599	108	166599	18.702	ng/ul	96
19) Hexachloroethane	8.893	117	72785	17.708	ng/ul	95
22) Nitrobenzene	9.022	77	179971	17.009	ng/ul	92
23) Isophorone	9.552	82	319568	17.089	ng/ul	97
25) 2-Nitrophenol	9.734	139	93692	19.517	ng/ul#	80
26) 2,4-Dimethylphenol	9.799	107	181369	17.934	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.122	93	342	0.031	ng/ul	94
29) 2,4-Dichlorophenol	10.275	162	161551	19.859	ng/ul	96
30) Naphthalene	10.663	128	552144	20.506	ng/ul	99
32) 4-Chloroaniline	10.787	127	215291	19.737	ng/ul	100
33) Hexachlorobutadiene	10.952	225	119405	18.638	ng/ul	95
34) Caprolactam	11.752	113	138	0.054	ng/ul	92
35) 4-Chloro-3-methylphenol	11.922	107	153193	16.971	ng/ul#	89

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062022\
 Data File : BP010760.D
 Acq On : 20 Jun 2022 13:28
 Operator : CG/JU
 Sample : SST002005
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST0020605

Quant Time: Jun 20 14:51:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 14:42:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.275	142	365878	20.129	ng/ul	100
37) 1-Methylnaphthalene	12.499	142	364016	19.946	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.652	216	210136	20.044	ng/ul	97
40) Hexachlorocyclopentadiene	12.628	237	75455	15.340	ng/ul	95
41) 2,4,6-Trichlorophenol	12.899	196	123213	19.394	ng/ul	98
42) 2,4,5-Trichlorophenol	12.975	196	127356	18.949	ng/ul	98
43) 1,1'-Biphenyl	13.293	154	486049	20.561	ng/ul	100
44) 2-Chloronaphthalene	13.328	162	377941	20.570	ng/ul	98
45) 2-Nitroaniline	13.546	65	84826	13.644	ng/ul#	72
47) Dimethylphthalate	13.916	163	445563	19.798	ng/ul	98
48) 2,6-Dinitrotoluene	14.040	165	84851	17.733	ng/ul#	86
50) Acenaphthylene	14.175	152	595655	20.669	ng/ul	98
51) 3-Nitroaniline	14.375	138	77378	17.640	ng/ul	88
52) Acenaphthene	14.522	153	390522	20.957	ng/ul	99
53) 2,4-Dinitrophenol	14.675	184	799	0.330	ng/ul	93
55) 4-Nitrophenol	14.704	109	43721	11.959	ng/ul#	77
56) Dibenzofuran	14.857	168	565136	20.666	ng/ul	91
57) 2,4-Dinitrotoluene	14.834	165	129300	19.625	ng/ul#	76
58) 2,3,4,6-Tetrachlorophenol	15.093	232	110814	19.271	ng/ul	95
59) Diethylphthalate	15.281	149	430192	18.390	ng/ul	99
61) Fluorene	15.510	166	455173	20.513	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.504	204	240151	20.143	ng/ul	93
63) 4-Nitroaniline	15.540	138	70016	16.910	ng/ul#	76
66) 4,6-Dinitro-2-methylph...	15.604	198	72136	18.819	ng/ul#	90
67) N-Nitrosodiphenylamine	15.722	169	387513	20.721	ng/ul	99
68) 4-Bromophenyl-phenylether	16.398	248	145788	20.009	ng/ul	90
69) Hexachlorobenzene	16.522	284	174964	26.563	ng/ul	95
70) Atrazine	16.681	200	139622	18.353	ng/ul	99
71) Pentachlorophenol	16.875	266	81206	17.152	ng/ul	94
72) Phenanthrene	17.257	178	712804	20.547	ng/ul	98
74) Anthracene	17.345	178	731229	20.844	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.257	216	219941	19.346	ng/uL	97
76) Pentachlorobenzene	14.781	250	208134	20.984	ng/uL	98
77) Carbazole	17.622	167	622283	20.399	ng/ul	99
78) Di-n-butylphthalate	18.175	149	694122	16.612	ng/ul	99
80) Fluoranthene	19.228	202	863949	18.697	ng/ul	97
82) Pyrene	19.581	202	900405	19.112	ng/ul	95
83) Butylbenzylphthalate	20.457	149	323180	16.468	ng/ul	90
84) 3,3'-Dichlorobenzidine	21.228	252	307207	22.237	ng/ul	98
85) Benzo(a)anthracene	21.292	228	914302	19.659	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.216	149	481470	16.677	ng/ul	99
87) Chrysene	21.345	228	909547	19.970	ng/ul	99
89) Di-n-octyl phthalate	22.133	149	848900	16.441	ng/ul	100
90) Benzo(b)fluoranthene	22.969	252	1039451	20.841	ng/ul	99
91) Benzo(k)fluoranthene	23.022	252	953630	20.041	ng/ul	99
93) Benzo(a)pyrene	23.592	252	1003737	20.422	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.192	276	1203105	20.998	ng/ul	96
95) Dibenzo(a,h)anthracene	26.204	278	1030938	20.898	ng/ul	98
96) Benzo(g,h,i)perylene	26.951	276	1031374	20.956	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_P
ClientSampleId :
SSTD020605



1,4-Dioxane

Concen: 8.037 ng/uL

RT: 3.311 min Scan#

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion: 88 Resp: 27416

Ion Ratio Lower Upper

88 100

43 25.2 35.8 53.6#

58 65.2 68.1 102.1#



Pyridine

Concen: 18.820 ng/uL

RT: 3.717 min Scan# 107

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28




Tgt Ion: 79 Resp: 168046


Ion Ratio Lower Upper

79 100

52 71.8 83.8 125.8#

51 34.6 41.0 61.4#





Benzaldehyde

Concen: 19.942 ng/ul

RT: 6.964 min Scan# 659

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion: 77 Resp: 108589

Ion Ratio Lower Upper

77 100

105 104.0 76.0 114.0

106 98.9 73.2 109.8



Phenol

Concen: 18.543 ng/ul

RT: 7.022 min Scan# 669

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28




Tgt Ion: 94 Resp: 208882


Ion Ratio Lower Upper

94 100

65 28.0 27.4 41.2

66 39.5 37.4 56.2









Bis(2-Chloroethyl)ether
 Concen: 19.110 ng/ul
 RT: 7.246 min Scan#
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:	93	Resp:	168626
Ion	Ratio	Lower	Upper
93	100		
63	69.9	72.4	108.6#
95	32.9	27.3	40.9



2-Chlorophenol
 Concen: 20.261 ng/ul
 RT: 7.387 min Scan# 731
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	128	Resp:	179026
Ion	Ratio	Lower	Upper
128	100		
64	42.6	44.8	67.2#
130	32.4	25.8	38.6





2-Methylphenol

Concen: 18.242 ng/uI

RT: 8.269 min Scan# 881

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :





SSTD020605

Tgt Ion:108 Resp: 152734

Ion Ratio Lower Upper

108 100

107 89.7 71.4 107.0



2,2'-oxybis(1-Chloropropane)

Concen: 13.800 ng/uI

RT: 8.346 min Scan# 894

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28




Tgt Ion: 45 Resp: 199711


Ion Ratio Lower Upper

45 100

77 16.6 11.2 16.8

79 12.1 6.9 10.3#





Acetophenone

Concen: 0.038 ng/u1

RT: 8.758 min Scan# 945

Delta R.T. 0.112 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:105 Resp: 533

Ion Ratio Lower Upper

105 100

77 85.4 67.5 101.3

51 41.1 29.6 44.4



N-Nitroso-di-n-propylamine

Concen: 16.440 ng/u1

RT: 8.628 min Scan# 942

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Tgt Ion: 70 Resp: 118088




Ion Ratio Lower Upper


70 100

42 54.2 53.8 80.8

101 10.7 7.8 11.6

130 24.3 15.4 23.2#





4-Methylphenol

Concen: 18.702 ng/ul

RT: 8.599 min Scan# 937

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :





SSTD020605

Tgt Ion:108 Resp: 166599

Ion Ratio Lower Upper

108 100

107 114.3 88.3 132.5



Hexachloroethane

Concen: 17.708 ng/ul

RT: 8.893 min Scan# 987

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28




Tgt Ion:117 Resp: 72785


Ion Ratio Lower Upper

117 100

201 113.6 85.7 128.5

199 73.5 57.4 86.0





Nitrobenzene

Concen: 17.009 ng/ul

RT: 9.022 min Scan# 1009

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion: 77 Resp: 179971

Ion Ratio Lower Upper

77 100

123 45.1 31.5 47.3

65 12.6 11.4 17.2



Isophorone

Concen: 17.089 ng/ul

RT: 9.552 min Scan# 1099

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28




Tgt Ion: 82 Resp: 319568

Ion Ratio Lower Upper


82 100

95 7.3 6.6 10.0

138 18.3 13.8 20.6



Abundance Scan 1130 (9.734 min): BP010760.D\data.ms (-1130)



2-Nitrophenol

Concen: 19.517 ng/ul

RT: 9.734 min Scan# 1130

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :

SSTD020605




Tgt Ion:139 Resp: 93692

Ion Ratio Lower Upper


139 100

65 48.7 54.2 81.4#

109 29.7 30.6 45.8#



Abundance Scan 1141 (9.799 min): BP010760.D\data.ms (-1141)



2,4-Dimethylphenol

Concen: 17.934 ng/ul

RT: 9.799 min Scan# 1141

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28




Tgt Ion:107 Resp: 181369

Ion Ratio Lower Upper

107 100

121 51.4 39.1 58.7


122 85.0 66.1 99.1






Abundance Scan 1180 (10.028 min): BP010760.D\data.ms (-1123) (-)

Bis(2-Chloroethoxy)methane
Concen: 0.031 ng/ul
RT: 10.122 min Scan# 1180
Delta R.T. 0.094 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28

Instrument :
BNA_P
ClientSampleId :
SSTD020605







Tgt Ion:	93	Resp:	342
Ion	Ratio	Lower	Upper
93	100		
95	29.3	26.6	39.8
123	10.4	7.8	11.8




Abundance Scan 1222 (10.275 min): BP010760.D\data.ms (-1123) (-)

2,4-Dichlorophenol
Concen: 19.859 ng/ul
RT: 10.275 min Scan# 1222
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28

Tgt Ion:	162	Resp:	161551
Ion	Ratio	Lower	Upper
162	100		
164	66.1	52.5	78.7
98	34.0	31.4	47.0









Scan 1288 (10.663 min): BP010760.D\data.ms (-1280) (-)
 Naphthalene
 Concen: 20.506 ng/ul
 RT: 10.663 min Scan# 1288
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument : BNA_P
 ClientSampleId : SST020605


Tgt Ion:	128	Resp:	552144
Ion	Ratio	Lower	Upper
128	100		
129	10.6	8.7	13.1
127	12.9	10.2	15.4



Scan 1309 (10.787 min): BP010760.D\data.ms (-1302) (-)
 4-Chloroaniline
 Concen: 19.737 ng/ul
 RT: 10.787 min Scan# 1309
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	127	Resp:	215291
Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.2	39.2









Scan 1337 (10.952 min): BP010760.D\data.ms (-1328) (-)
 Hexachlorobutadiene
 Concen: 18.638 ng/ul
 RT: 10.952 min Scan# 1337
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020605


Tgt Ion:	225	Resp:	119405
Ion	Ratio	Lower	Upper
225	100		
223	64.4	53.7	80.5
227	67.5	49.4	74.2



Scan 1443 (11.752 min): BP010760.D\data.ms (-1438) (-)
 Caprolactam
 Concen: 0.054 ng/ul
 RT: 11.752 min Scan# 1473
 Delta R.T. 0.176 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	113	Resp:	138
Ion	Ratio	Lower	Upper
113	100		
55	205.7	153.8	230.8
56	129.0	110.2	165.4









Scan 1502 (11.922 min): BP010760.D\data.ms (-1435) (-)
 4-Chloro-3-methylphenol
 Concen: 16.971 ng/ul
 RT: 11.922 min Scan# 1502
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument : BNA_P
 ClientSampleId : SST020605


Tgt Ion:	107	Resp:	153193
Ion	Ratio	Lower	Upper
107	100		
144	28.0	18.6	27.8#
142	83.9	59.6	89.4



Scan 1562 (12.275 min): BP010760.D\data.ms (-1536) (-)
 2-Methylnaphthalene
 Concen: 20.129 ng/ul
 RT: 12.275 min Scan# 1562
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	142	Resp:	365878
Ion	Ratio	Lower	Upper
142	100		
141	87.4	70.1	105.1









1-Methylnaphthalene
 Concen: 19.946 ng/u1
 RT: 12.499 min Scan# 1600
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:	142	Resp:	364016
Ion Ratio	Lower	Upper	
142	100		
141	92.8	73.3	109.9
116	3.4	4.0	6.0#



1,2,4,5-Tetrachlorobenzene
 Concen: 20.044 ng/u1
 RT: 12.652 min Scan# 1626
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	216	Resp:	210136
Ion Ratio	Lower	Upper	
216	100		
214	77.1	63.0	94.6
179	19.8	18.1	27.1
108	14.3	13.6	20.4









Hexachlorocyclopentadiene
 Concen: 15.340 ng/ul
 RT: 12.628 min Scan# 1610
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020605


Tgt Ion:	237	Resp:	75455
Ion	Ratio	Lower	Upper
237	100		
235	62.1	52.8	79.2
272	12.4	11.5	17.3



2,4,6-Trichlorophenol
 Concen: 19.394 ng/ul
 RT: 12.899 min Scan# 1668
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	196	Resp:	123213
Ion	Ratio	Lower	Upper
196	100		
198	96.3	78.2	117.4
200	30.0	25.0	37.6









2,4,5-Trichlorophenol
Concen: 18.949 ng/ul
RT: 12.975 min Scan# 1681
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28




Instrument :
BNA_P
ClientSampleId :
SSTD020605

Tgt Ion:	196	Resp:	127356
Ion	Ratio	Lower	Upper
196	100		
198	98.7	77.0	115.4
200	30.2	24.8	37.2



1,1'-Biphenyl
Concen: 20.561 ng/ul
RT: 13.293 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28

Tgt Ion:	154	Resp:	486049
Ion	Ratio	Lower	Upper
154	100		
153	40.0	31.9	47.9
76	11.3	9.2	13.8



Abundance Scan 1741 (13.328 min): BP010760.D\data.ms (-1741) (-)

2-Chloronaphthalene

Concen: 20.570 ng/ul

RT: 13.328 min Scan# 1741

Delta R.T. 0.000 min

Lab File: BP010760.D


Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :

SSTD020605






Tgt Ion:162 Resp: 377941

Ion Ratio Lower Upper

162 100

164 33.2 25.5 38.3

127 38.5 31.9 47.9



Abundance Scan 1778 (13.546 min): BP010760.D\data.ms (-1778) (-)

2-Nitroaniline

Concen: 13.644 ng/ul

RT: 13.546 min Scan# 1778

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28




Tgt Ion: 65 Resp: 84826


Ion Ratio Lower Upper

65 100

92 73.0 49.0 73.6

138 116.1 66.4 99.6#









Scan 1841 (13.916 min): BP010760.D\data.ms (-1841)
 Dimethylphthalate
 Concen: 19.798 ng/ul
 RT: 13.916 min Scan# 1841
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020605


Tgt Ion:	163	Resp:	445563
Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.4	5.2
164	10.3	7.6	11.4



Scan 1862 (14.040 min): BP010760.D\data.ms (-1862)
 2,6-Dinitrotoluene
 Concen: 17.733 ng/ul
 RT: 14.040 min Scan# 1862
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	165	Resp:	84851
Ion	Ratio	Lower	Upper
165	100		
89	52.3	52.9	79.3#
121	21.0	19.4	29.2









Acenaphthylene
 Concen: 20.669 ng/ul
 RT: 14.175 min Scan# 1885
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020605


Tgt Ion:	152	Resp:	595655
Ion	Ratio	Lower	Upper
152	100		
151	19.3	16.5	24.7
153	12.9	10.1	15.1



3-Nitroaniline
 Concen: 17.640 ng/ul
 RT: 14.375 min Scan# 1919
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	138	Resp:	77378
Ion	Ratio	Lower	Upper
138	100		
108	11.5	11.0	16.4
92	116.4	104.6	156.8





Acenaphthene

Concen: 20.957 ng/ul

RT: 14.522 min Scan# 1952

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :

SSTD020605





Tgt Ion:153 Resp: 390522

Ion Ratio Lower Upper

153 100

152 46.0 36.2 54.2

154 88.2 70.8 106.2



2,4-Dinitrophenol

Concen: 0.330 ng/ul

RT: 14.675 min Scan# 1970

Delta R.T. 0.076 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28




Tgt Ion:184 Resp: 799


Ion Ratio Lower Upper

184 100

63 88.6 73.8 110.6

154 61.6 57.2 85.8









4-Nitrophenol
 Concen: 11.959 ng/u1
 RT: 14.704 min Scan# 1975
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:109 Resp: 43721
 Ion Ratio Lower Upper
 109 100
 139 142.8 75.6 113.4#
 65 138.7 107.6 161.4



Dibenzofuran
 Concen: 20.666 ng/u1
 RT: 14.857 min Scan# 2001
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:168 Resp: 565136
 Ion Ratio Lower Upper
 168 100
 139 35.2 32.8 49.2









2,4-Dinitrotoluene
Concen: 19.625 ng/ul
RT: 14.834 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:	165	Resp:	129300
Ion	Ratio	Lower	Upper
165	100		
63	40.4	47.0	70.6#
182	3.0	2.2	3.2



2,3,4,6-Tetrachlorophenol
Concen: 19.271 ng/ul
RT: 15.093 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28

Tgt Ion:	232	Resp:	110814
Ion	Ratio	Lower	Upper
232	100		
131	39.9	35.8	53.8
130	2.5	1.8	2.8
166	29.2	23.7	35.5









Diethylphthalate
 Concen: 18.390 ng/ul
 RT: 15.281 min Scan# 2073
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:	149	Resp:	430192
Ion	Ratio	Lower	Upper
149	100		
177	22.4	17.6	26.4
150	12.6	10.3	15.5



Fluorene
 Concen: 20.513 ng/ul
 RT: 15.510 min Scan# 2112
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	166	Resp:	455173
Ion	Ratio	Lower	Upper
166	100		
165	97.4	77.4	116.2
167	13.3	10.8	16.2









4-Chlorophenyl-phenylether
 Concen: 20.143 ng/ul
 RT: 15.504 min Scan# 2111
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:	204	Resp:	240151
Ion	Ratio	Lower	Upper
204	100		
206	33.0	26.0	39.0
141	54.0	49.3	73.9



4-Nitroaniline
 Concen: 16.910 ng/ul
 RT: 15.540 min Scan# 2117
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	138	Resp:	70016
Ion	Ratio	Lower	Upper
138	100		
92	50.8	46.2	69.2
108	73.9	85.8	128.8#









Scan 2128 (15.604 min): BP010760.D\data.ms (-2128) (-)
 4,6-Dinitro-2-methylphenol
 Concen: 18.819 ng/ul
 RT: 15.604 min Scan# 2128
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605

Tgt Ion:	198	Resp:	72136
Ion Ratio	Lower	Upper	
198	100		
182	5.8	4.3	6.5
77	23.1	23.7	35.5#



Scan 2148 (15.722 min): BP010760.D\data.ms (-2148) (-)
 N-Nitrosodiphenylamine
 Concen: 20.721 ng/ul
 RT: 15.722 min Scan# 2148
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	169	Resp:	387513
Ion Ratio	Lower	Upper	
169	100		
168	67.2	52.6	78.8
167	34.4	27.8	41.6



Abundance Scan 2263 (16.398 min): BP010760.D\data.ms (-2259) (-)

4-Bromophenyl-phenylether

Concen: 20.009 ng/ul

RT: 16.398 min Scan# 2263

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Instrument :

BNA_P

ClientSampleId :

SSTD020605

Ref

Abundance Scan 2263 (16.398 min): BP010760.D\data.ms

Raw

Sub

m/z-->

Abundance Scan 2263 (16.398 min): BP010760.D\data.ms (-2229) (-)

Raw

Sub

m/z-->

Tgt Ion:248 Resp: 145788

Ion Ratio Lower Upper

248 100

250 95.8 80.8 121.2

141 66.7 63.9 95.9

Abundance

16.398

100000

80000

60000

40000

20000

0

Time-->

Abundance Scan 2284 (16.522 min): BP010760.D\data.ms (-2279) (-)

Hexachlorobenzene

Concen: 26.563 ng/ul

RT: 16.522 min Scan# 2284

Delta R.T. 0.000 min

Lab File: BP010760.D

Acq: 20 Jun 2022 13:28

Ref

Abundance Scan 2284 (16.522 min): BP010760.D\data.ms

Raw

Sub

m/z-->

Abundance Scan 2284 (16.522 min): BP010760.D\data.ms (-2250) (-)

Raw

Sub

m/z-->

Tgt Ion:284 Resp: 174964

Ion Ratio Lower Upper

284 100

142 30.5 27.8 41.6

249 30.7 25.4 38.2

Abundance


16.522

100000

50000

0





Time-->



Atrazine
Concen: 18.353 ng/ul
RT: 16.681 min Scan# 2311
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:	200	Resp:	139622
Ion Ratio	Lower	Upper	
200	100		
173	25.5	20.7	31.1
215	48.7	38.6	57.8



Pentachlorophenol
Concen: 17.152 ng/ul
RT: 16.875 min Scan# 2344
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28

Tgt Ion:	266	Resp:	81206
Ion Ratio	Lower	Upper	
266	100		
264	59.1	51.4	77.2
268	62.4	53.5	80.3









Scan 2409 (17.257 min): BP010760.D\data.ms (-2375)
 Phenanthrene
 Concen: 20.547 ng/ul
 RT: 17.257 min Scan# 2409
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument : BNA_P
 ClientSampleId : SSTD020605

Tgt Ion:	178	Resp:	712804
Ion	Ratio	Lower	Upper
178	100		
179	15.3	12.7	19.1
176	19.1	14.6	22.0




Scan 2424 (17.345 min): BP010760.D\data.ms (-2375)
 Anthracene
 Concen: 20.844 ng/ul
 RT: 17.345 min Scan# 2424
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	178	Resp:	731229
Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.2	18.2
176	17.8	14.5	21.7





Abundance Scan 1729 (13.257 min): BP010760.D\data.ms (-1729) (-)




1,2,3,4-Tetrachlorobenzene
Concen: 19.346 ng/uL
RT: 13.257 min Scan# 1729
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28

Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:	216	Resp:	219941
Ion	Ratio	Lower	Upper
216	100		
214	77.6	65.0	97.4
218	48.8	39.1	58.7



Abundance Scan 1729 (13.257 min): BP010760.D\data.ms (-1695) (-)





Abundance Scan 1988 (14.781 min): BP010760.D\data.ms (-1976) (-)





Pentachlorobenzene
Concen: 20.984 ng/uL
RT: 14.781 min Scan# 1988
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28

Tgt Ion:	250	Resp:	208134
Ion	Ratio	Lower	Upper
250	100		
248	63.6	49.8	74.6
252	65.0	51.2	76.8



Abundance Scan 1988 (14.781 min): BP010760.D\data.ms (-1954) (-)









Carbazole
Concen: 20.399 ng/ul
RT: 17.622 min Scan# 2471
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28




Instrument :
BNA_P
ClientSampleId :
SSTD020605


Tgt Ion:	167	Resp:	622283
Ion	Ratio	Lower	Upper
167	100		
166	21.7	16.9	25.3
139	12.8	10.6	16.0



Di-n-butylphthalate
Concen: 16.612 ng/ul
RT: 18.175 min Scan# 2565
Delta R.T. 0.000 min
Lab File: BP010760.D
Acq: 20 Jun 2022 13:28

Tgt Ion:	149	Resp:	694122
Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.1	10.7
104	4.9	4.6	6.8









Scan 2744 (19.228 min): BP010760.D\data.ms (-2738) (-)
 Fluoranthene
 Concen: 18.697 ng/ul
 RT: 19.228 min Scan# 2744
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument : BNA_P
 ClientSampleId : SST020605


Tgt Ion:	202	Resp:	863949
Ion Ratio	Lower	Upper	
202	100		
101	10.1	7.1	10.7
100	8.2	5.7	8.5



Scan 2804 (19.581 min): BP010760.D\data.ms (-2794) (-)
 Pyrene
 Concen: 19.112 ng/ul
 RT: 19.581 min Scan# 2804
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	202	Resp:	900405
Ion Ratio	Lower	Upper	
202	100		
101	11.7	7.9	11.9
100	10.0	6.7	10.1









Scan 2953 (20.457 min): BP010760.D\data.ms (-2953)
 Butylbenzylphthalate
 Concen: 16.468 ng/u1
 RT: 20.457 min Scan# 2953
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:	149	Resp:	323180
Ion	Ratio	Lower	Upper
149	100		
91	64.5	60.0	90.0
206	23.1	17.4	26.2



Scan 3084 (21.228 min): BP010760.D\data.ms (-3084)
 3,3'-Dichlorobenzidine
 Concen: 22.237 ng/u1
 RT: 21.228 min Scan# 3084
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	252	Resp:	307207
Ion	Ratio	Lower	Upper
252	100		
254	62.8	51.0	76.4
126	9.1	8.6	12.8









Scan 3095 (21.292 min): BP010760.D\data.ms (-3085) (-)
 Benzo(a)anthracene
 Concen: 19.659 ng/u1
 RT: 21.292 min Scan# 3095
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:	228	Resp:	914302
Ion	Ratio	Lower	Upper
228	100		
229	19.3	15.0	22.4
226	25.8	20.6	31.0



Scan 3082 (21.216 min): BP010760.D\data.ms (-3086) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 16.677 ng/u1
 RT: 21.216 min Scan# 3082
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	149	Resp:	481470
Ion	Ratio	Lower	Upper
149	100		
167	27.8	22.6	33.8
279	5.2	4.2	6.4









Scan 3104 (21.345 min): BP010760.D\data.ms (-3087) (-)
 Chrysene
 Concen: 19.970 ng/ul
 RT: 21.345 min Scan# 3104
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:	228	Resp:	909547
Ion	Ratio	Lower	Upper
228	100		
226	29.2	23.6	35.4
229	19.3	15.6	23.4



Scan 3238 (22.133 min): BP010760.D\data.ms (-3238) (-)
 Di-n-octyl phthalate
 Concen: 16.441 ng/ul
 RT: 22.133 min Scan# 3238
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:149 Resp: 848900









Scan 3380 (22.969 min): BP010760.D\data.ms (-3379) (-)
 Benzo(b)fluoranthene
 Concen: 20.841 ng/u1
 RT: 22.969 min Scan# 3380
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SST020605


Tgt Ion:	252	Resp:	1039451
Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.8	26.6
125	9.1	6.2	9.2



Scan 3389 (23.022 min): BP010760.D\data.ms (-3385) (-)
 Benzo(k)fluoranthene
 Concen: 20.041 ng/u1
 RT: 23.022 min Scan# 3389
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	252	Resp:	953630
Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.0	25.4
125	8.2	6.2	9.4









Scan 3486 (23.592 min): BP010760.D\data.ms (-3486) (-)
 Benzo(a)pyrene
 Concen: 20.422 ng/u1
 RT: 23.592 min Scan# 3486
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020605


Tgt Ion:	252	Resp:	1003737
Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.0
125	9.4	6.6	10.0



Scan 3928 (26.192 min): BP010760.D\data.ms (-3928) (-)
 Indeno(1,2,3-cd)pyrene
 Concen: 20.998 ng/u1
 RT: 26.192 min Scan# 3928
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	276	Resp:	1203105
Ion	Ratio	Lower	Upper
276	100		
138	21.6	14.6	22.0
277	25.2	19.6	29.4









Scan 3930 (26.204 min): BP010760.D\data.ms (-3915) (-)
 Dibenzo(a,h)anthracene
 Concen: 20.898 ng/ul
 RT: 26.204 min Scan# 3930
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020605

Tgt Ion:	278	Resp:	1030938
Ion	Ratio	Lower	Upper
278	100		
139	13.4	9.4	14.2
279	23.9	18.9	28.3



Scan 4057 (26.951 min): BP010760.D\data.ms (-4046) (-)
 Benzo(g,h,i)perylene
 Concen: 20.956 ng/ul
 RT: 26.951 min Scan# 4057
 Delta R.T. 0.000 min
 Lab File: BP010760.D
 Acq: 20 Jun 2022 13:28

Tgt Ion:	276	Resp:	1031374
Ion	Ratio	Lower	Upper
276	100		
138	18.0	12.7	19.1
277	23.6	18.2	27.4

