

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062022\
 Data File : BP010766.D
 Acq On : 20 Jun 2022 17:58
 Operator : CG/JU
 Sample : PB145601BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS601

Quant Time: Jun 21 00:46:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 16:18:42 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2022
 Supervised By :mohammad ahmed 06/27/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	137624	20.000	ng/u1	0.00
20) Naphthalene-d8	10.616	136	516102	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	322045	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.216	188	676281	20.000	ng/u1	0.00
79) Chrysene-d12	21.310	240	758941	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	824558	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	18067	5.642	ng/uL	0.00
4) Pyridine-d5	3.693	84	246518	29.601	ng/u1	0.00
7) Phenol-d5	6.999	99	303573	30.400	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	184431	30.333	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	268182	31.553	ng/u1	0.00
15) 4-Methylphenol-d8	8.540	113	247900	30.723	ng/u1	0.00
21) Nitrobenzene-d5	8.981	128	125736	32.813	ng/u1	0.00
24) 2-Nitrophenol-d4	9.699	143	139065	33.874	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.246	165	253084	33.271	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	329578	30.760	ng/u1	0.00
46) Dimethylphthalate-d6	13.869	166	733078	33.156	ng/u1	0.00
49) Acenaphthylene-d8	14.151	160	874337	32.653	ng/u1	0.00
54) 4-Nitrophenol-d4	14.681	143	108889	33.339	ng/u1	0.00
60) Fluorene-d10	15.451	176	649909	33.068	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.586	200	134543	33.345	ng/u1	0.00
73) Anthracene-d10	17.310	188	1011355m	32.976	ng/u1	0.00
81) Pyrene-d10	19.551	212	1229170	32.579	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.545	264	1384108	33.711	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	40866	12.309	ng/uL#	82
5) Pyridine	3.711	79	270201	31.370	ng/u1#	68
6) Benzaldehyde	6.957	77	207408	45.196	ng/u1	94
8) Phenol	7.022	94	349275	32.685	ng/u1#	86
10) Bis(2-Chloroethyl)ether	7.246	93	273862	32.445	ng/u1#	83
12) 2-Chlorophenol	7.381	128	292959	33.349	ng/u1#	88
13) 2-Methylphenol	8.269	108	254183	32.094	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.346	45	317795	32.049	ng/u1#	93
16) Acetophenone	8.646	105	418062	33.437	ng/u1#	88
17) N-Nitroso-di-n-propyla...	8.628	70	200059	34.176	ng/u1#	86
18) 4-Methylphenol	8.599	108	278011	32.849	ng/u1	97
19) Hexachloroethane	8.893	117	118132	33.125	ng/u1	93
22) Nitrobenzene	9.022	77	305025	34.579	ng/u1#	90
23) Isophorone	9.551	82	548009	35.060	ng/u1	98
25) 2-Nitrophenol	9.734	139	162189	35.920	ng/u1#	80
26) 2,4-Dimethylphenol	9.799	107	300346	33.450	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.028	93	343996	34.063	ng/u1	98
29) 2,4-Dichlorophenol	10.269	162	275044	35.261	ng/u1	95
30) Naphthalene	10.663	128	909494	33.727	ng/u1	99
32) 4-Chloroaniline	10.781	127	349230	32.342	ng/u1	100
33) Hexachlorobutadiene	10.951	225	199304	34.114	ng/u1	95
34) Caprolactam	11.575	113	82055m	36.678	ng/u1	
35) 4-Chloro-3-methylphenol	11.922	107	278219	37.502	ng/u1	93
36) 2-Methylnaphthalene	12.275	142	614323	34.629	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062022\
 Data File : BP010766.D
 Acq On : 20 Jun 2022 17:58
 Operator : CG/JU
 Sample : PB145601BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS601


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2022
 Supervised By :mohammad ahmed 06/27/2022

Quant Time: Jun 21 00:46:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 16:18:42 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.498	142	623086	34.771	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.651	216	358818	33.890	ng/ul	96
40) Hexachlorocyclopentadiene	12.628	237	148139	30.274	ng/ul	95
41) 2,4,6-Trichlorophenol	12.898	196	215410	35.425	ng/ul	99
42) 2,4,5-Trichlorophenol	12.975	196	225582	35.480	ng/ul	97
43) 1,1'-Biphenyl	13.292	154	824874	33.864	ng/ul	99
44) 2-Chloronaphthalene	13.328	162	638808	33.815	ng/ul	98
45) 2-Nitroaniline	13.545	65	157054	38.118	ng/ul#	75
47) Dimethylphthalate	13.922	163	779077	35.285	ng/ul	99
48) 2,6-Dinitrotoluene	14.039	165	161295	38.588	ng/ul#	83
50) Acenaphthylene	14.181	152	1019001	34.685	ng/ul	99
51) 3-Nitroaniline	14.375	138	150649	37.274	ng/ul#	87
52) Acenaphthene	14.522	153	666966	34.298	ng/ul	99
53) 2,4-Dinitrophenol	14.592	184	83115	34.820	ng/ul#	70
55) 4-Nitrophenol	14.698	109	87120	36.212	ng/ul#	76
56) Dibenzofuran	14.857	168	964506	34.667	ng/ul	91
57) 2,4-Dinitrotoluene	14.834	165	236491	39.315	ng/ul#	74
58) 2,3,4,6-Tetrachlorophenol	15.092	232	201702	37.946	ng/ul	94
59) Diethylphthalate	15.286	149	762703	36.434	ng/ul	98
61) Fluorene	15.510	166	785301	35.008	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.504	204	413330	35.160	ng/ul	92
63) 4-Nitroaniline	15.539	138	161322m	43.071	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.604	198	141985	35.463	ng/ul#	88
67) N-Nitrosodiphenylamine	15.722	169	678223	34.740	ng/ul	99
68) 4-Bromophenyl-phenylether	16.398	248	258520	34.718	ng/ul	92
69) Hexachlorobenzene	16.522	284	307202	34.097	ng/ul	92
70) Atrazine	16.681	200	248860	34.294	ng/ul	98
71) Pentachlorophenol	16.875	266	158654	34.272	ng/ul	97
72) Phenanthrene	17.257	178	1249026	34.724	ng/ul	99
74) Anthracene	17.351	178	1267408	34.971	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.257	216	373844	31.280	ng/uL	98
76) Pentachlorobenzene	14.781	250	363744	33.574	ng/uL	100
77) Carbazole	17.622	167	1118305	35.130	ng/ul	99
78) Di-n-butylphthalate	18.175	149	1269376	37.198	ng/ul	99
80) Fluoranthene	19.227	202	1541078	34.612	ng/ul	97
82) Pyrene	19.580	202	1599263	34.571	ng/ul	95
83) Butylbenzylphthalate	20.457	149	602181	37.196	ng/ul	88
84) 3,3'-Dichlorobenzidine	21.227	252	598233	38.424	ng/ul	98
85) Benzo(a)anthracene	21.292	228	1620620	35.235	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.216	149	895899	37.051	ng/ul	100
87) Chrysene	21.345	228	1632029	35.819	ng/ul	99
89) Di-n-octyl phthalate	22.133	149	1544388	34.828	ng/ul	100
90) Benzo(b)fluoranthene	22.974	252	1772497	34.536	ng/ul	99
91) Benzo(k)fluoranthene	23.021	252	1774236	36.919	ng/ul	99
93) Benzo(a)pyrene	23.598	252	1786499	35.520	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.192	276	2152744	35.062	ng/ul	96
95) Dibenzo(a,h)anthracene	26.204	278	1848264	35.220	ng/ul	98
96) Benzo(g,h,i)perylene	26.951	276	1837899	34.637	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.309 ng/uL
RT: 3.305 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion: 88 Resp: 40860
Ion Ratio Lower Upper
88 100
43 30.7 35.8 53.6
58 70.9 68.1 102.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#5
Pyridine
Concen: 31.370 ng/uL
RT: 3.711 min Scan# 106
Delta R.T. -0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion: 79 Resp: 270201
Ion Ratio Lower Upper
79 100
52 68.1 83.8 125.8
51 34.0 41.0 61.4





#6
Benzaldehyde
Concen: 45.196 ng/ul
RT: 6.957 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58



Instrument :
BNA_P
ClientSampleId :
SLCS601



Tgt Ion: 77 Resp: 207408
Ion Ratio Lower Upper
77 100
105 100.3 76.0 114.0
106 96.7 73.2 109.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#8
Phenol
Concen: 32.685 ng/ul
RT: 7.022 min Scan# 669
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58


Tgt Ion: 94 Resp: 349275
Ion Ratio Lower Upper
94 100
65 27.4 27.4 41.2#
66 36.9 37.4 56.2#





#10
Bis(2-Chloroethyl)ether
Concen: 32.445 ng/ul
RT: 7.246 min Scan# 707
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601



Tgt Ion: 93 Resp: 273862
Ion Ratio Lower Upper
93 100
63 69.1 72.4 108.6
95 31.9 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#12
2-Chlorophenol
Concen: 33.349 ng/ul
RT: 7.381 min Scan# 730
Delta R.T. -0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58

Tgt Ion:128 Resp: 292959
Ion Ratio Lower Upper
128 100
64 43.0 44.8 67.2
130 30.6 25.8 38.6









#13
2-Methylphenol
Concen: 32.094 ng/ul
RT: 8.269 min Scan# 881
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58

Instrument :
BNA_P
ClientSampleId :
SLCS601

Tgt Ion:108 Resp: 25418
Ion Ratio Lower Upper
108 100
107 91.2 71.4 107.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.049 ng/ul
RT: 8.346 min Scan# 894
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58


Tgt Ion: 45 Resp: 317795
Ion Ratio Lower Upper
45 100
77 15.9 11.2 16.8
79 12.3 6.9 10.3#





#16
Acetophenone
Concen: 33.437 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601



Tgt Ion: 105 Resp: 418062
Ion Ratio Lower Upper
105 100
77 77.7 67.5 101.3
51 24.0 29.6 44.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#17
N-Nitroso-di-n-propylamine
Concen: 34.176 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58


Tgt Ion: 70 Resp: 200059
Ion Ratio Lower Upper
70 100
42 54.1 53.8 80.8
101 10.7 7.8 11.6
130 24.1 15.4 23.2





#18
4-Methylphenol
Concen: 32.849 ng/ul
RT: 8.599 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:108 Resp: 27801
Ion Ratio Lower Upper
108 100
107 113.3 88.3 132.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#19
Hexachloroethane
Concen: 33.125 ng/ul
RT: 8.893 min Scan# 987
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:117 Resp: 118132
Ion Ratio Lower Upper
117 100
201 118.5 85.7 128.5
199 71.5 57.4 86.0





#22
Nitrobenzene
Concen: 34.579 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion: 77 Resp: 305025
Ion Ratio Lower Upper
77 100
123 47.4 31.5 47.3
65 13.2 11.4 17.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#23
Isophorone
Concen: 35.060 ng/ul
RT: 9.551 min Scan# 1099
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58



Tgt Ion: 82 Resp: 548009
Ion Ratio Lower Upper
82 100
95 7.4 6.6 10.0
138 18.2 13.8 20.6









#25
2-Nitrophenol
Concen: 35.920 ng/ul
RT: 9.734 min Scan# 1130
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58

Instrument :
BNA_P
ClientSampleId :
SLCS601

Tgt Ion:139 Resp: 162189
Ion Ratio Lower Upper
139 100
65 48.6 54.2 81.4
109 30.1 30.6 45.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#26
2,4-Dimethylphenol
Concen: 33.450 ng/ul
RT: 9.799 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58


Tgt Ion:107 Resp: 300346
Ion Ratio Lower Upper
107 100
121 50.6 39.1 58.7
122 86.3 66.1 99.1





#27
 Bis(2-Chloroethoxy)methane
 Concen: 34.063 ng/ul
 RT: 10.028 min Scan# 1180
 Delta R.T. -0.000 min
 Lab File: BP010766.D
 Acq: 20 Jun 2022 17:58




Instrument :
 BNA_P
 ClientSampleId :
 SLCS601



Tgt Ion: 93 Resp: 343990
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.6 39.8
 123 11.6 7.8 11.8




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
 Supervised By :mohammad ahmed 06/27/2022



#29
 2,4-Dichlorophenol
 Concen: 35.261 ng/ul
 RT: 10.269 min Scan# 1221
 Delta R.T. -0.006 min
 Lab File: BP010766.D
 Acq: 20 Jun 2022 17:58


Tgt Ion:162 Resp: 275044
 Ion Ratio Lower Upper
 162 100
 164 63.9 52.5 78.7
 98 33.1 31.4 47.0





#30
Naphthalene
Concen: 33.727 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:128 Resp: 909494
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.1 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#32
4-Chloroaniline
Concen: 32.342 ng/ul
RT: 10.781 min Scan# 1308
Delta R.T. -0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:127 Resp: 349230
Ion Ratio Lower Upper
127 100
129 32.6 26.2 39.2





#33
Hexachlorobutadiene
Concen: 34.114 ng/ul
RT: 10.951 min Scan# 1337
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601



Tgt Ion:225 Resp: 199304
Ion Ratio Lower Upper
225 100
223 62.3 53.7 80.5
227 65.5 49.4 74.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#34
Caprolactam
Concen: 36.678 ng/ul m
RT: 11.575 min Scan# 1443
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58


Tgt Ion:113 Resp: 82055
Ion Ratio Lower Upper
113 100
55 144.5 153.8 230.8#
56 108.4 110.2 165.4#





#35
4-Chloro-3-methylphenol
Concen: 37.502 ng/ul
RT: 11.922 min Scan# 1502
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:107 Resp: 278219
Ion Ratio Lower Upper
107 100
144 25.9 18.6 27.8
142 80.3 59.6 89.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#36
2-Methylnaphthalene
Concen: 34.629 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:142 Resp: 614323
Ion Ratio Lower Upper
142 100
141 88.2 70.1 105.1





#37
1-Methylnaphthalene
Concen: 34.771 ng/ul
RT: 12.498 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:142 Resp: 623080
Ion Ratio Lower Upper
142 100
141 92.6 73.3 109.9
116 3.6 4.0 6.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.890 ng/ul
RT: 12.651 min Scan# 1626
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:216 Resp: 358818
Ion Ratio Lower Upper
216 100
214 75.9 63.0 94.6
179 19.8 18.1 27.1
108 14.3 13.6 20.4





#40
Hexachlorocyclopentadiene
Concen: 30.274 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument : BNA_P
ClientSampleId : SLCS601




Tgt Ion:237 Resp: 148139
Ion Ratio Lower Upper
237 100
235 61.6 52.8 79.2
272 13.1 11.5 17.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#41
2,4,6-Trichlorophenol
Concen: 35.425 ng/ul
RT: 12.898 min Scan# 1668
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:196 Resp: 215410
Ion Ratio Lower Upper
196 100
198 97.9 78.2 117.4
200 32.3 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 35.480 ng/ul
RT: 12.975 min Scan# 1681
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:196 Resp: 22558
Ion Ratio Lower Upper
196 100
198 100.4 77.0 115.4
200 31.3 24.8 37.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#43
1,1'-Biphenyl
Concen: 33.864 ng/ul
RT: 13.292 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58



Tgt Ion:154 Resp: 824874
Ion Ratio Lower Upper
154 100
153 40.7 31.9 47.9
76 11.5 9.2 13.8





#44

2-Chloronaphthalene

Concen: 33.815 ng/ul

RT: 13.328 min Scan# 1741

Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58

Instrument :

BNA_P

ClientSampleId :





SLCS601

Tgt Ion:	162	Resp:	638800
Ion Ratio	Lower	Upper	
162	100		
164	32.0	25.5	38.3
127	38.3	31.9	47.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#45

2-Nitroaniline

Concen: 38.118 ng/ul




RT: 13.545 min Scan# 1778


Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58


Tgt Ion:	65	Resp:	157054
Ion Ratio	Lower	Upper	
65	100		
92	71.3	49.0	73.6
138	113.5	66.4	99.6





#47
Dimethylphthalate
Concen: 35.285 ng/ul
RT: 13.922 min Scan# 1842
Delta R.T. 0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:163 Resp: 77907
Ion Ratio Lower Upper
163 100
194 4.6 3.4 5.2
164 9.6 7.6 11.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#48
2,6-Dinitrotoluene
Concen: 38.588 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58



Tgt Ion:165 Resp: 161295
Ion Ratio Lower Upper
165 100
89 50.6 52.9 79.3#
121 18.7 19.4 29.2#





#50

Acenaphthylene

Concen: 34.685 ng/ul

RT: 14.181 min Scan# 1886

Delta R.T. 0.006 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58

Instrument :

BNA_P

ClientSampleId :

SLCS601

Tgt Ion:152 Resp: 101900

Ion Ratio Lower Upper

152 100

151 19.8 16.5 24.7





153 13.0 10.1 15.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2022

Supervised By :mohammad ahmed 06/27/2022



#51

3-Nitroaniline

Concen: 37.274 ng/ul

RT: 14.375 min Scan# 1919

Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58




Tgt Ion:138 Resp: 150649


Ion Ratio Lower Upper

138 100

108 10.7 11.0 16.4


92 114.9 104.6 156.8





#52
Acenaphthene
Concen: 34.298 ng/ul
RT: 14.522 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:153 Resp: 666960
Ion Ratio Lower Upper
153 100
152 46.2 36.2 54.2
154 89.3 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#53
2,4-Dinitrophenol
Concen: 34.820 ng/ul
RT: 14.592 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:184 Resp: 83115
Ion Ratio Lower Upper
184 100
63 54.5 73.8 110.6#
154 56.6 57.2 85.8#





#55
4-Nitrophenol
Concen: 36.212 ng/ul
RT: 14.698 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601



Tgt Ion:109 Resp: 87120
Ion Ratio Lower Upper
109 100
139 147.3 75.6 113.4
65 132.0 107.6 161.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#56
Dibenzofuran
Concen: 34.667 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58

Tgt Ion:168 Resp: 964506
Ion Ratio Lower Upper
168 100
139 35.6 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 39.315 ng/ul
RT: 14.834 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58

Instrument :
BNA_P
ClientSampleId :
SLCS601

Tgt Ion:165 Resp: 23649
Ion Ratio Lower Upper
165 100
63 38.4 47.0 70.6
182 3.3 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.946 ng/ul
RT: 15.092 min Scan# 2041
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58

Tgt Ion:232 Resp: 201702
Ion Ratio Lower Upper
232 100
131 38.2 35.8 53.8
130 2.1 1.8 2.8
166 29.2 23.7 35.5









#59
Diethylphthalate
Concen: 36.434 ng/ul
RT: 15.286 min Scan# 2073
Delta R.T. 0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58

Instrument :
BNA_P
ClientSampleId :
SLCS601

Tgt Ion:149 Resp: 76270
Ion Ratio Lower Upper
149 100
177 23.5 17.6 26.4
150 12.6 10.3 15.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#61
Fluorene
Concen: 35.008 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58


Tgt Ion:166 Resp: 785301
Ion Ratio Lower Upper
166 100
165 98.1 77.4 116.2
167 13.7 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 35.160 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601



Tgt Ion:204 Resp: 413330
Ion Ratio Lower Upper
204 100
206 33.8 26.0 39.0
141 53.2 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#63
4-Nitroaniline
Concen: 43.071 ng/ul m
RT: 15.539 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58


Tgt Ion:138 Resp: 161322
Ion Ratio Lower Upper
138 100
92 50.4 46.2 69.2
108 73.4 85.8 128.8#





#66
4,6-Dinitro-2-methylphenol
Concen: 35.463 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601



Tgt Ion:198 Resp: 141985
Ion Ratio Lower Upper
198 100
182 4.1 4.3 6.5
77 22.2 23.7 35.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#67
N-Nitrosodiphenylamine
Concen: 34.740 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58


Tgt Ion:169 Resp: 678223
Ion Ratio Lower Upper
169 100
168 65.3 52.6 78.8
167 35.1 27.8 41.6





#68
4-Bromophenyl-phenylether
Concen: 34.718 ng/ul
RT: 16.398 min Scan# 2263
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:248 Resp: 258520
Ion Ratio Lower Upper
248 100
250 98.5 80.8 121.2
141 66.1 63.9 95.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#69
Hexachlorobenzene
Concen: 34.097 ng/ul
RT: 16.522 min Scan# 2284
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58



Tgt Ion:284 Resp: 307202
Ion Ratio Lower Upper
284 100
142 27.8 27.8 41.6
249 29.8 25.4 38.2





#70

Atrazine

Concen: 34.294 ng/ul

RT: 16.681 min Scan# 2311

Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58

Instrument :

BNA_P

ClientSampleId :

SLCS601

Tgt Ion:200 Resp: 248860

Ion Ratio Lower Upper

200 100

173 26.4 20.7 31.1





215 50.3 38.6 57.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2022

Supervised By :mohammad ahmed 06/27/2022



#71

Pentachlorophenol

Concen: 34.272 ng/ul

RT: 16.875 min Scan# 2344

Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58




Tgt Ion:266 Resp: 158654


Ion Ratio Lower Upper

266 100

264 61.7 51.4 77.2


268 64.7 53.5 80.3





#72
Phenanthrene
Concen: 34.724 ng/ul
RT: 17.257 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:178 Resp: 1249020
Ion Ratio Lower Upper
178 100
179 15.3 12.7 19.1
176 18.4 14.6 22.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#74
Anthracene
Concen: 34.971 ng/ul
RT: 17.351 min Scan# 2425
Delta R.T. 0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:178 Resp: 1267408
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 18.3 14.5 21.7





#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.280 ng/uL
RT: 13.257 min Scan# 1729
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:216 Resp: 373844
Ion Ratio Lower Upper
216 100
214 79.1 65.0 97.4
218 47.7 39.1 58.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#76
Pentachlorobenzene
Concen: 33.574 ng/uL
RT: 14.781 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:250 Resp: 363744
Ion Ratio Lower Upper
250 100
248 62.1 49.8 74.6
252 63.6 51.2 76.8





#77
 Carbazole
 Concen: 35.130 ng/ul
 RT: 17.622 min Scan# 2471
 Delta R.T. -0.000 min
 Lab File: BP010766.D
 Acq: 20 Jun 2022 17:58




Instrument : BNA_P
 ClientSampleId : SLCS601




Tgt Ion:167 Resp: 111830
 Ion Ratio Lower Upper
 167 100
 166 21.2 16.9 25.3
 139 12.4 10.6 16.0

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
 Supervised By :mohammad ahmed 06/27/2022




#78
 Di-n-butylphthalate
 Concen: 37.198 ng/ul
 RT: 18.175 min Scan# 2565
 Delta R.T. -0.000 min
 Lab File: BP010766.D
 Acq: 20 Jun 2022 17:58




Tgt Ion:149 Resp: 1269376
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.1 10.7
 104 4.7 4.6 6.8





#80
Fluoranthene
Concen: 34.612 ng/u1
RT: 19.227 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:202 Resp: 1541078
Ion Ratio Lower Upper
202 100
101 10.3 7.1 10.7
100 8.0 5.7 8.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#82
Pyrene
Concen: 34.571 ng/u1
RT: 19.580 min Scan# 2804
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:202 Resp: 1599263
Ion Ratio Lower Upper
202 100
101 11.8 7.9 11.9
100 10.0 6.7 10.1





#83
Butylbenzylphthalate
Concen: 37.196 ng/ul
RT: 20.457 min Scan# 2953
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601




Tgt Ion:149 Resp: 60218
Ion Ratio Lower Upper
149 100
91 63.2 60.0 90.0
206 24.4 17.4 26.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#84
3,3'-Dichlorobenzidine
Concen: 38.424 ng/ul
RT: 21.227 min Scan# 3084
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Tgt Ion:252 Resp: 598233
Ion Ratio Lower Upper
252 100
254 64.8 51.0 76.4
126 8.9 8.6 12.8





#85
Benzo(a)anthracene
Concen: 35.235 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument : BNA_P
ClientSampleId : SLCS601




Tgt Ion:228 Resp: 1620620
Ion Ratio Lower Upper
228 100
229 19.5 15.0 22.4
226 26.9 20.6 31.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 37.051 ng/ul
RT: 21.216 min Scan# 3082
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58



Tgt Ion:149 Resp: 895899
Ion Ratio Lower Upper
149 100
167 28.3 22.6 33.8
279 5.3 4.2 6.4





#87

Chrysene

Concen: 35.819 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58

Instrument :

BNA_P

ClientSampleId :

SLCS601

Tgt Ion:228 Resp: 1632029

Ion Ratio Lower Upper

228 100





226 29.1 23.6 35.4

229 19.5 15.6 23.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#89

Di-n-octyl phthalate

Concen: 34.828 ng/ul




RT: 22.133 min Scan# 3238


Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58


Tgt Ion:149 Resp: 1544388





#90
Benzo(b)fluoranthene
Concen: 34.536 ng/u1
RT: 22.974 min Scan# 3381
Delta R.T. 0.006 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58




Instrument :
BNA_P
ClientSampleId :
SLCS601



Tgt Ion:252 Resp: 1772497
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.6
125 8.5 6.2 9.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#91
Benzo(k)fluoranthene
Concen: 36.919 ng/u1
RT: 23.021 min Scan# 3389
Delta R.T. -0.000 min
Lab File: BP010766.D
Acq: 20 Jun 2022 17:58

Tgt Ion:252 Resp: 1774236
Ion Ratio Lower Upper
252 100
253 21.6 17.0 25.4
125 8.3 6.2 9.4





#93

Benzo(a)pyrene

Concen: 35.520 ng/ul

RT: 23.598 min Scan# 3486

Delta R.T. 0.006 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58

Instrument :

BNA_P

ClientSampleId :

SLCS601

Tgt Ion:252 Resp: 1786499

Ion Ratio Lower Upper

252 100





253 21.4 17.4 26.0

125 9.0 6.6 10.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2022
Supervised By :mohammad ahmed 06/27/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 35.062 ng/ul

RT: 26.192 min Scan# 3928

Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58




Tgt Ion:276 Resp: 2152744


Ion Ratio Lower Upper

276 100

138 21.0 14.6 22.0

277 25.7 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 35.220 ng/ul

RT: 26.204 min Scan# 3930

Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58

Instrument :

BNA_P

ClientSampleId :

SLCS601

Tgt Ion:278 Resp: 1848264

Ion Ratio Lower Upper

278 100

139 13.5 9.4 14.2





279 23.8 18.9 28.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/21/2022

Supervised By :mohammad ahmed 06/27/2022



#96

Benzo(g,h,i)perylene

Concen: 34.637 ng/ul

RT: 26.951 min Scan# 4057

Delta R.T. -0.000 min

Lab File: BP010766.D

Acq: 20 Jun 2022 17:58

Tgt Ion:276 Resp: 1837899

Ion Ratio Lower Upper

276 100

138 17.3 12.7 19.1

277 23.4 18.2 27.4

