

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062025\
 Data File : BP025015.D
 Acq On : 20 Jun 2025 12:51
 Operator : RC/JU
 Sample : PB168536BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB168536BL

Quant Time: Jun 20 13:15:49 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 06 16:20:27 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.602	152	271573	20.000	ng	-0.01
21) Naphthalene-d8	10.378	136	988460	20.000	ng	0.00
39) Acenaphthene-d10	14.248	164	608741	20.000	ng	0.00
64) Phenanthrene-d10	17.054	188	1213121	20.000	ng	0.00
76) Chrysene-d12	21.501	240	1414680	20.000	ng	0.02
86) Perylene-d12	24.777	264	1630780	20.000	ng	0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.237	112	2457875	151.071	ng	0.00
7) Phenol-d6	6.813	99	2947532	136.926	ng	0.00
23) Nitrobenzene-d5	8.755	82	1753246	86.190	ng	0.00
42) 2,4,6-Tribromophenol	15.784	330	1346188	159.952	ng	0.00
45) 2-Fluorobiphenyl	12.854	172	3938645	87.161	ng	0.00
79) Terphenyl-d14	19.783	244	7096996	89.911	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062025\
Data File : BP025015.D
Acq On : 20 Jun 2025 12:51
Operator : RC/JU
Sample : PB168536BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB168536BL

Quant Time: Jun 20 13:15:49 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060625.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 06 16:20:27 2025
Response via : Initial Calibration

