

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062121\
 Data File : BP006006.D
 Acq On : 21 Jun 2021 21:59
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 22 00:55:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 11:35:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	109	0.00
2	1,4-Dioxane	40.000	40.771	-1.9	117	0.00
3	Pyridine	40.000	46.739	-16.8	127	0.00
4	n-Nitrosodimethylamine	40.000	38.757	3.1	111	0.00
5 S	2-Fluorophenol	80.000	88.200	-10.3	125	0.00
6	Aniline	40.000	44.819	-12.0	128	0.00
7 S	Phenol-d6	80.000	88.784	-11.0	125	0.00
8	2-Chlorophenol	40.000	44.546	-11.4	129	0.00
9	Benzaldehyde	40.000	46.787	-17.0	121	0.00
10 C	Phenol	40.000	44.733	-11.8	129	0.00
11	bis(2-Chloroethyl)ether	40.000	44.874	-12.2	130	0.00
12	1,3-Dichlorobenzene	40.000	43.935	-9.8	127	0.00
13 C	1,4-Dichlorobenzene	40.000	43.767	-9.4	127	0.00
14	1,2-Dichlorobenzene	40.000	43.933	-9.8	127	0.00
15	Benzyl Alcohol	40.000	43.721	-9.3	123	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.831	-4.6	119	-0.01
17	2-Methylphenol	40.000	45.966	-14.9	131	0.00
18	Hexachloroethane	40.000	43.092	-7.7	123	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.386	-11.0	126	0.00
20	3+4-Methylphenols	40.000	46.039	-15.1	130	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	111	0.00
22	Acetophenone	40.000	42.310	-5.8	123	0.00
23 S	Nitrobenzene-d5	80.000	82.910	-3.6	121	0.00
24	Nitrobenzene	40.000	41.717	-4.3	125	0.00
25	Isophorone	40.000	42.178	-5.4	125	0.00
26 C	2-Nitrophenol	40.000	44.176	-10.4	131	0.00
27	2,4-Dimethylphenol	40.000	41.790	-4.5	125	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.386	-8.5	132	0.00
29 C	2,4-Dichlorophenol	40.000	41.987	-5.0	126	0.00
30	1,2,4-Trichlorobenzene	40.000	41.444	-3.6	127	0.00
31	Naphthalene	40.000	42.263	-5.7	129	0.00
32	Benzoic acid	40.000	40.721	-1.8	118	0.01
33	4-Chloroaniline	40.000	42.761	-6.9	129	0.00
34 C	Hexachlorobutadiene	40.000	41.181	-3.0	126	0.00
35	Caprolactam	40.000	43.069	-7.7	121	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.795	-9.5	132	0.00
37	2-Methylnaphthalene	40.000	42.305	-5.8	129	0.00
38	1-Methylnaphthalene	40.000	42.340	-5.9	128	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	112	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.873	-2.2	121	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.678	10.8	109	0.00
42 S	2,4,6-Tribromophenol	80.000	85.445	-6.8	126	-0.01
43 C	2,4,6-Trichlorophenol	40.000	40.689	-1.7	121	0.00
44	2,4,5-Trichlorophenol	40.000	40.701	-1.8	121	0.00
45 S	2-Fluorobiphenyl	80.000	81.410	-1.8	122	0.00
46	1,1'-Biphenyl	40.000	41.051	-2.6	121	0.00
47	2-Chloronaphthalene	40.000	41.379	-3.4	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062121\
 Data File : BP006006.D
 Acq On : 21 Jun 2021 21:59
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 22 00:55:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 11:35:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.708	-4.3	124	0.00
49	Acenaphthylene	40.000	41.395	-3.5	125	0.00
50	Dimethylphthalate	40.000	41.066	-2.7	126	0.00
51	2,6-Dinitrotoluene	40.000	41.535	-3.8	125	0.00
52 C	Acenaphthene	40.000	41.174	-2.9	126	0.00
53	3-Nitroaniline	40.000	42.380	-6.0	126	0.00
54 P	2,4-Dinitrophenol	40.000	42.877	-7.2	129	0.00
55	Dibenzofuran	40.000	40.457	-1.1	124	0.00
56 P	4-Nitrophenol	40.000	38.781	3.0	113	0.00
57	2,4-Dinitrotoluene	40.000	42.411	-6.0	125	0.00
58	Fluorene	40.000	40.843	-2.1	125	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.938	2.7	116	0.00
60	Diethylphthalate	40.000	40.776	-1.9	124	0.00
61	4-Chlorophenyl-phenylether	40.000	40.505	-1.3	124	0.00
62	4-Nitroaniline	40.000	43.415	-8.5	130	0.00
63	Azobenzene	40.000	40.353	-0.9	121	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.390	-1.0	113	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.652	-6.6	123	0.00
67	4-Bromophenyl-phenylether	40.000	42.769	-6.9	123	0.00
68	Hexachlorobenzene	40.000	42.428	-6.1	123	0.00
69	Atrazine	40.000	36.141	9.6	89	0.00
70 C	Pentachlorophenol	40.000	42.343	-5.9	111	0.00
71	Phenanthrene	40.000	41.633	-4.1	120	-0.01
72	Anthracene	40.000	42.296	-5.7	121	0.00
73	Carbazole	40.000	42.014	-5.0	120	0.00
74	Di-n-butylphthalate	40.000	43.553	-8.9	124	-0.01
75 C	Fluoranthene	40.000	41.400	-3.5	120	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	110	-0.01
77	Benzydine	40.000	34.380	14.0	83	0.00
78	Pyrene	40.000	40.760	-1.9	122	-0.01
79 S	Terphenyl-d14	80.000	84.225	-5.3	125	-0.01
80	Butylbenzylphthalate	40.000	42.489	-6.2	125	-0.01
81	Benzo(a)anthracene	40.000	40.695	-1.7	123	0.00
82	3,3'-Dichlorobenzidine	40.000	42.149	-5.4	129	-0.01
83	Chrysene	40.000	41.006	-2.5	124	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.703	-6.8	127	-0.01
85 c	Di-n-octyl phthalate	40.000	42.450	-6.1	126	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	104	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	42.036	-5.1	121	-0.02
88	Benzo(b)fluoranthene	40.000	43.362	-8.4	123	-0.02
89	Benzo(k)fluoranthene	40.000	43.838	-9.6	125	-0.01
90 C	Benzo(a)pyrene	40.000	43.298	-8.2	124	-0.02
91	Dibenzo(a,h)anthracene	40.000	42.136	-5.3	121	-0.02
92	Benzo(g,h,i)perylene	40.000	41.454	-3.6	119	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062121\
Data File : BP006006.D
Acq On : 21 Jun 2021 21:59
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 22 00:55:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 16 11:35:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0