

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP062420\
 Data File : BP002530.D
 Acq On : 24 Jun 2020 19:17
 Operator : CG/JU
 Sample : L3089-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 RB47184-RT

Manual Integrations
 APPROVED

mohammad
 6/25/2020 1:07:37 PM

Quant Time: Jun 25 01:34:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP060520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 24 13:54:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	236023	20.00	ng	0.00
21) Naphthalene-d8	10.35	136	941172	20.00	ng	0.00
39) Acenaphthene-d10	14.23	164	560558	20.00	ng	0.00
64) Phenanthrene-d10	16.99	188	1181015	20.00	ng	0.00
76) Chrysene-d12	21.09	240	845578	20.00	ng	0.00
86) Perylene-d12	23.29	264	816565	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.20	112	1468092	115.10	ng	0.00
7) Phenol-d6	6.78	99	1992918	117.35	ng	0.00
23) Nitrobenzene-d5	8.73	82	1404879	89.14	ng	0.00
42) 2,4,6-Tribromophenol	15.73	330	711616	132.59	ng	0.00
45) 2-Fluorobiphenyl	12.85	172	2797790	83.92	ng	0.00
79) Terphenyl-d14	19.57	244	3476046	115.68	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
49) Acenaphthylene	13.95	152	105658	2.203	ng	98
50) Dimethylphthalate	13.69	163	342633	8.921	ng	99
71) Phenanthrene	17.03	178	319496	5.783	ng	99
75) Fluoranthene	19.01	202	876824	13.296	ng	96
78) Pyrene	19.36	202	773438	14.933	ng	99
81) Benzo(a)anthracene	21.07	228	366092	7.585	ng	95
83) Chrysene	21.13	228	388996	8.507	ng	97
87) Indeno(1,2,3-cd)pyrene	25.49	276	269380	5.274	ng	98
88) Benzo(b)fluoranthene	22.63	252	559310m	12.701	ng	
89) Benzo(k)fluoranthene	22.67	252	157861m	3.773	ng	
90) Benzo(a)pyrene	23.19	252	338619	8.206	ng	98
92) Benzo(a,h,i)perylene	26.16	276	255551	6.023	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP062420\
Data File : BP002530.D
Acq On : 24 Jun 2020 19:17
Operator : CG/JU
Sample : L3089-01
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
RB47184-RT

Manual Integrations
APPROVED

mohammad
6/25/2020 1:07:37 PM

Quant Time: Jun 25 01:34:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP060520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 24 13:54:06 2020
Response via : Initial Calibration

