

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062421\
 Data File : BP006067.D
 Acq On : 24 Jun 2021 23:08
 Operator : CG/JU
 Sample : M2810-01 2X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 NB-08-062221

Quant Time: Jun 25 01:13:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 22 17:52:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.910	152	82960	20.000	ng	-0.01
21) Naphthalene-d8	10.716	136	317731	20.000	ng	0.00
39) Acenaphthene-d10	14.539	164	192660	20.000	ng	-0.01
64) Phenanthrene-d10	17.286	188	393152	20.000	ng	0.00
76) Chrysene-d12	21.363	240	435617	20.000	ng	0.00
86) Perylene-d12	23.768	264	501656	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	147994	28.625	ng	0.00
7) Phenol-d6	7.093	99	176793	26.383	ng	0.00
23) Nitrobenzene-d5	9.081	82	106183	16.384	ng	0.00
42) 2,4,6-Tribromophenol	16.033	330	86049	26.027	ng	0.00
45) 2-Fluorobiphenyl	13.169	172	241348	16.450	ng	0.00
79) Terphenyl-d14	19.833	244	396890	17.060	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062421\
Data File : BP006067.D
Acq On : 24 Jun 2021 23:08
Operator : CG/JU
Sample : M2810-01 2X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
NB-08-062221

Quant Time: Jun 25 01:13:06 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 22 17:52:38 2021
Response via : Initial Calibration

