

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062422\
 Data File : BP010790.D
 Acq On : 24 Jun 2022 13:19
 Operator : CG/JU
 Sample : N3395-09
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0156

Quant Time: Jun 24 22:45:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	308530	20.000	ng/u1	0.00
20) Naphthalene-d8	10.516	136	1267218	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.381	164	685616	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.139	188	1271667	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.251	240	1138238	20.000	ng/u1	# 0.00
88) Perylene-d12	23.621	264	1246793	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	53436	6.552	ng/uL	0.00
4) Pyridine-d5	3.611	84	142003	6.541	ng/u1	0.00
7) Phenol-d5	6.887	99	234331	9.595	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	633587	39.667	ng/u1	0.00
11) 2-Chlorophenol-d4	7.252	132	633938	31.285	ng/u1	0.00
15) 4-Methylphenol-d8	8.428	113	410928	21.284	ng/u1	0.00
21) Nitrobenzene-d5	8.881	128	348984	43.260	ng/u1	0.00
24) 2-Nitrophenol-d4	9.604	143	287112	43.019	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.140	165	578227	35.064	ng/u1	0.00
31) 4-Chloroaniline-d4	10.657	131	713245	26.496	ng/u1	0.00
46) Dimethylphthalate-d6	13.798	166	1966666	41.875	ng/u1	0.00
49) Acenaphthylene-d8	14.069	160	2423629	40.434	ng/u1	0.00
54) 4-Nitrophenol-d4	14.575	143	50842	6.931	ng/u1	0.00
60) Fluorene-d10	15.381	176	1688179	41.887	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.492	200	129952	30.844	ng/u1	0.00
73) Anthracene-d10	17.239	188	2523600	44.604	ng/u1	0.00
81) Pyrene-d10	19.492	212	2744019	44.250	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.474	264	2826085	45.870	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062422\
Data File : BP010790.D
Acq On : 24 Jun 2022 13:19
Operator : CG/JU
Sample : N3395-09
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0156

Quant Time: Jun 24 22:45:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 24 22:42:26 2022
Response via : Initial Calibration

