

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062422\
 Data File : BP010800.D
 Acq On : 24 Jun 2022 18:58
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020679

Quant Time: Jun 24 22:47:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	432670	20.000	ng/ul	0.00
20) Naphthalene-d8	10.516	136	1795156	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.375	164	996703	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.139	188	1831419	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.251	240	1433919	20.000	ng/ul	# 0.00
88) Perylene-d12	23.621	264	1456337	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	90565	7.918	ng/uL	0.00
4) Pyridine-d5	3.605	84	589214	19.353	ng/ul	0.00
7) Phenol-d5	6.887	99	677467	19.780	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	447245	19.967	ng/ul	0.00
11) 2-Chlorophenol-d4	7.252	132	570522	20.077	ng/ul	0.00
15) 4-Methylphenol-d8	8.434	113	552538	20.407	ng/ul	0.00
21) Nitrobenzene-d5	8.881	128	254517	22.272	ng/ul	0.00
24) 2-Nitrophenol-d4	9.604	143	218897	23.152	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.140	165	473315	20.261	ng/ul	0.00
31) 4-Chloroaniline-d4	10.657	131	769402	20.177	ng/ul	0.00
46) Dimethylphthalate-d6	13.798	166	1345055	19.701	ng/ul	0.00
49) Acenaphthylene-d8	14.069	160	1695780	19.461	ng/ul	0.00
54) 4-Nitrophenol-d4	14.575	143	213445	20.015	ng/ul	0.00
60) Fluorene-d10	15.375	176	1145664	19.554	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.492	200	111462	18.370	ng/ul	0.00
73) Anthracene-d10	17.239	188	1591339	19.530	ng/ul	0.00
81) Pyrene-d10	19.486	212	1600246	20.484	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.468	264	1382440	19.210	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	95363	7.818	ng/uL#	88
5) Pyridine	3.622	79	602783	19.544	ng/ul#	77
6) Benzaldehyde	6.863	77	423226	23.419	ng/ul	96
8) Phenol	6.916	94	730635	19.945	ng/ul#	86
10) Bis(2-Chloroethyl)ether	7.152	93	589691	20.043	ng/ul	91
12) 2-Chlorophenol	7.287	128	589124	20.084	ng/ul	91
13) 2-Methylphenol	8.163	108	544223	19.956	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.263	45	817149	20.701	ng/ul#	98
16) Acetophenone	8.546	105	860899	20.704	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.534	70	430715	20.485	ng/ul	89
18) 4-Methylphenol	8.499	108	583415	20.542	ng/ul	99
19) Hexachloroethane	8.799	117	235658	19.838	ng/ul#	83
22) Nitrobenzene	8.922	77	613387	21.508	ng/ul	92
23) Isophorone	9.452	82	1182797	19.777	ng/ul	97
25) 2-Nitrophenol	9.634	139	262420	22.920	ng/ul#	81
26) 2,4-Dimethylphenol	9.699	107	620507	19.837	ng/ul	95
27) Bis(2-Chloroethoxy)met...	9.940	93	751844	19.928	ng/ul	97
29) 2,4-Dichlorophenol	10.163	162	489439	20.163	ng/ul	99
30) Naphthalene	10.569	128	1847047	19.741	ng/ul	100
32) 4-Chloroaniline	10.681	127	761198	20.148	ng/ul	99
33) Hexachlorobutadiene	10.857	225	301140	19.418	ng/ul	93
34) Caprolactam	11.446	113	154532	20.265	ng/ul	86
35) 4-Chloro-3-methylphenol	11.816	107	530642	20.169	ng/ul	93
36) 2-Methylnaphthalene	12.187	142	1203390	19.863	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062422\
 Data File : BP010800.D
 Acq On : 24 Jun 2022 18:58
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020679

Quant Time: Jun 24 22:47:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.410	142	1205746	20.032	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.557	216	560498	19.569	ng/ul	98
40) Hexachlorocyclopentadiene	12.540	237	329271	18.142	ng/ul	96
41) 2,4,6-Trichlorophenol	12.798	196	334822	20.276	ng/ul	97
42) 2,4,5-Trichlorophenol	12.869	196	365953	20.652	ng/ul	99
43) 1,1'-Biphenyl	13.210	154	1538535	19.608	ng/ul#	98
44) 2-Chloronaphthalene	13.245	162	1154978	19.540	ng/ul	99
45) 2-Nitroaniline	13.457	65	293167	23.377	ng/ul#	80
47) Dimethylphthalate	13.845	163	1344629	19.666	ng/ul	98
48) 2,6-Dinitrotoluene	13.957	165	235919	23.028	ng/ul	96
50) Acenaphthylene	14.098	152	1880690	19.707	ng/ul	99
51) 3-Nitroaniline	14.287	138	261403	21.284	ng/ul	89
52) Acenaphthene	14.440	153	1201387	19.510	ng/ul	99
53) 2,4-Dinitrophenol	14.487	184	69786	17.744	ng/ul#	84
55) 4-Nitrophenol	14.592	109	169340	19.970	ng/ul#	75
56) Dibenzofuran	14.781	168	1649755	19.559	ng/ul	95
57) 2,4-Dinitrotoluene	14.745	165	319975	23.375	ng/ul#	81
58) 2,3,4,6-Tetrachlorophenol	15.004	232	260020	20.491	ng/ul	92
59) Diethylphthalate	15.222	149	1324267	19.481	ng/ul	98
61) Fluorene	15.434	166	1315021	19.553	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.434	204	612329	19.449	ng/ul	97
63) 4-Nitroaniline	15.451	138	253325	22.308	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.504	198	122651	18.706	ng/ul	99
67) N-Nitrosodiphenylamine	15.651	169	1103032	19.885	ng/ul	98
68) 4-Bromophenyl-phenylether	16.334	248	358795	19.548	ng/ul	97
69) Hexachlorobenzene	16.434	284	412262	19.392	ng/ul	96
70) Atrazine	16.610	200	349877	19.024	ng/ul	98
71) Pentachlorophenol	16.786	266	200836	18.347	ng/ul	97
72) Phenanthrene	17.186	178	1924892	19.737	ng/ul	99
74) Anthracene	17.275	178	1918600	19.559	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.169	216	577876	19.121	ng/ul	96
76) Pentachlorobenzene	14.692	250	518563	20.007	ng/ul	99
77) Carbazole	17.551	167	1705212	19.312	ng/ul	100
78) Di-n-butylphthalate	18.128	149	1984927	18.865	ng/ul	99
80) Fluoranthene	19.163	202	1981468	20.623	ng/ul#	88
82) Pyrene	19.516	202	2020990	20.829	ng/ul#	84
83) Butylbenzylphthalate	20.416	149	686334	20.096	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.174	252	537368	18.149	ng/ul#	96
85) Benzo(a)anthracene	21.239	228	1730458	19.296	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.186	149	1061282	19.623	ng/ul#	96
87) Chrysene	21.286	228	1684869	19.606	ng/ul	100
89) Di-n-octyl phthalate	22.110	149	1592820	18.426	ng/ul	100
90) Benzo(b)fluoranthene	22.898	252	1724489	19.041	ng/ul#	96
91) Benzo(k)fluoranthene	22.951	252	1701928	19.770	ng/ul#	95
93) Benzo(a)pyrene	23.515	252	1726889	19.601	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.086	276	2073863	19.508	ng/ul#	90
95) Dibenzo(a,h)anthracene	26.109	278	1789119	19.620	ng/ul#	93
96) Benzo(g,h,i)perylene	26.833	276	1766057	19.530	ng/ul#	88





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.818 ng/uL
 RT: 3.240 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: BP010800.D
 Acq: 24 Jun 2022 18:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020679


Tgt Ion: 88 Resp: 95363
 Ion Ratio Lower Upper
 88 100
 43 28.8 35.8 53.6#
 58 80.6 68.1 102.1



#5
 Pyridine
 Concen: 19.544 ng/uL
 RT: 3.622 min Scan# 91
 Delta R.T. -0.006 min
 Lab File: BP010800.D
 Acq: 24 Jun 2022 18:58

Tgt Ion: 79 Resp: 602783
 Ion Ratio Lower Upper
 79 100
 52 80.4 83.8 125.8#
 51 36.1 41.0 61.4#









#6
Benzaldehyde
Concen: 23.419 ng/ul
RT: 6.863 min Scan# 643
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion: 77 Resp: 423226
Ion Ratio Lower Upper
77 100
105 99.4 76.0 114.0
106 95.4 73.2 109.8



#8
Phenol
Concen: 19.945 ng/ul
RT: 6.916 min Scan# 651
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion: 94 Resp: 730635
Ion Ratio Lower Upper
94 100
65 28.4 27.4 41.2
66 35.9 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 20.043 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:	93	Resp:	589691
Ion	Ratio	Lower	Upper
93	100		
63	79.7	72.4	108.6
95	33.2	27.3	40.9



#12
2-Chlorophenol
Concen: 20.084 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:	128	Resp:	589124
Ion	Ratio	Lower	Upper
128	100		
64	45.8	44.8	67.2
130	32.6	25.8	38.6









#13
2-Methylphenol
Concen: 19.956 ng/ul
RT: 8.163 min Scan# 80
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:108 Resp: 544223
Ion Ratio Lower Upper
108 100
107 90.7 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.701 ng/ul
RT: 8.263 min Scan# 880
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion: 45 Resp: 817149
Ion Ratio Lower Upper
45 100
77 13.8 11.2 16.8
79 10.3 6.9 10.3#









#16
Acetophenone
Concen: 20.704 ng/ul
RT: 8.546 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:105 Resp: 860899
Ion Ratio Lower Upper
105 100
77 75.8 67.5 101.3
51 26.2 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 20.485 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion: 70 Resp: 430715
Ion Ratio Lower Upper
70 100
42 56.3 53.8 80.8
101 10.3 7.8 11.6
130 22.1 15.4 23.2









#18
4-Methylphenol
Concen: 20.542 ng/ul
RT: 8.499 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:108 Resp: 583415
Ion Ratio Lower Upper
108 100
107 109.9 88.3 132.5



#19
Hexachloroethane
Concen: 19.838 ng/ul
RT: 8.799 min Scan# 971
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:117 Resp: 235658
Ion Ratio Lower Upper
117 100
201 90.8 85.7 128.5
199 56.8 57.4 86.0#









#22
Nitrobenzene
Concen: 21.508 ng/ul
RT: 8.922 min Scan# 993
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion: 77 Resp: 613387
Ion Ratio Lower Upper
77 100
123 45.7 31.5 47.3
65 13.1 11.4 17.2



#23
Isophorone
Concen: 19.777 ng/ul
RT: 9.452 min Scan# 1082
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion: 82 Resp: 1182797
Ion Ratio Lower Upper
82 100
95 6.8 6.6 10.0
138 18.3 13.8 20.6









#25
2-Nitrophenol
Concen: 22.920 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:139 Resp: 262420
Ion Ratio Lower Upper
139 100
65 51.4 54.2 81.4#
109 28.8 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 19.837 ng/ul
RT: 9.699 min Scan# 1124
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:107 Resp: 620507
Ion Ratio Lower Upper
107 100
121 52.5 39.1 58.7
122 87.6 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 19.928 ng/ul

RT: 9.940 min Scan# 1166

Delta R.T. -0.006 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion: 93 Resp: 751844

Ion Ratio Lower Upper

93 100

95 32.0 26.6 39.8

123 11.4 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 20.163 ng/ul

RT: 10.163 min Scan# 1203

Delta R.T. -0.000 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58




Tgt Ion: 162 Resp: 489439


Ion Ratio Lower Upper

162 100

164 66.1 52.5 78.7

98 38.6 31.4 47.0









#30
Naphthalene
Concen: 19.741 ng/ul
RT: 10.569 min Scan# 1272
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:128 Resp: 1847047
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 20.148 ng/ul
RT: 10.681 min Scan# 1291
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:127 Resp: 761198
Ion Ratio Lower Upper
127 100
129 32.3 26.2 39.2





#33

Hexachlorobutadiene

Concen: 19.418 ng/ul

RT: 10.857 min Scan# 1321

Delta R.T. -0.000 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:225 Resp: 301140

Ion Ratio Lower Upper

225 100

223 63.2 53.7 80.5

227 68.3 49.4 74.2



#34

Caprolactam

Concen: 20.265 ng/ul

RT: 11.446 min Scan# 1421

Delta R.T. -0.000 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58




Tgt Ion:113 Resp: 154532


Ion Ratio Lower Upper

113 100

55 167.4 153.8 230.8

56 124.6 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 20.169 ng/ul
RT: 11.816 min Scan# 1484
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:107 Resp: 530642
Ion Ratio Lower Upper
107 100
144 26.3 18.6 27.8
142 80.7 59.6 89.4



#36
2-Methylnaphthalene
Concen: 19.863 ng/ul
RT: 12.187 min Scan# 1547
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:142 Resp: 1203390
Ion Ratio Lower Upper
142 100
141 88.9 70.1 105.1









#37
1-Methylnaphthalene
Concen: 20.032 ng/ul
RT: 12.410 min Scan# 1585
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:142 Resp: 1205746
Ion Ratio Lower Upper
142 100
141 92.0 73.3 109.9
116 3.5 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.569 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:216 Resp: 560498
Ion Ratio Lower Upper
216 100
214 77.6 63.0 94.6
179 21.5 18.1 27.1
108 19.4 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 18.142 ng/ul
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:237 Resp: 329271
Ion Ratio Lower Upper
237 100
235 63.1 52.8 79.2
272 12.6 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 20.276 ng/ul
RT: 12.798 min Scan# 1651
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:196 Resp: 334822
Ion Ratio Lower Upper
196 100
198 93.9 78.2 117.4
200 30.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 20.652 ng/ul
RT: 12.869 min Scan# 1663
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:196 Resp: 365953
Ion Ratio Lower Upper
196 100
198 94.9 77.0 115.4
200 30.9 24.8 37.2



#43
1,1'-Biphenyl
Concen: 19.608 ng/ul
RT: 13.210 min Scan# 1721
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:154 Resp: 1538535
Ion Ratio Lower Upper
154 100
153 40.7 31.9 47.9
76 13.9 9.2 13.8#





#44

2-Chloronaphthalene

Concen: 19.540 ng/ul

RT: 13.245 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:162 Resp: 1154978

Ion Ratio Lower Upper

162 100

164 32.5 25.5 38.3

127 41.0 31.9 47.9



#45

2-Nitroaniline

Concen: 23.377 ng/ul

RT: 13.457 min Scan# 1763

Delta R.T. -0.000 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58




Tgt Ion: 65 Resp: 293167


Ion Ratio Lower Upper

65 100

92 70.1 49.0 73.6

138 106.2 66.4 99.6#









#47
Dimethylphthalate
Concen: 19.666 ng/ul
RT: 13.845 min Scan# 1829
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:163 Resp: 1344629
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.2
164 10.2 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 23.028 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:165 Resp: 235919
Ion Ratio Lower Upper
165 100
89 61.8 52.9 79.3
121 24.2 19.4 29.2









#50
Acenaphthylene
Concen: 19.707 ng/ul
RT: 14.098 min Scan# 1872
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:152 Resp: 1880690
Ion Ratio Lower Upper
152 100
151 19.9 16.5 24.7
153 12.9 10.1 15.1



#51
3-Nitroaniline
Concen: 21.284 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:138 Resp: 261403
Ion Ratio Lower Upper
138 100
108 15.3 11.0 16.4
92 116.8 104.6 156.8









#52
Acenaphthene
Concen: 19.510 ng/ul
RT: 14.440 min Scan# 1931
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:153 Resp: 1201387
Ion Ratio Lower Upper
153 100
152 46.6 36.2 54.2
154 88.3 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 17.744 ng/ul
RT: 14.487 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:184 Resp: 69786
Ion Ratio Lower Upper
184 100
63 72.5 73.8 110.6#
154 62.7 57.2 85.8









#55
4-Nitrophenol
Concen: 19.970 ng/ul
RT: 14.592 min Scan# 1956
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:109 Resp: 169340
Ion Ratio Lower Upper
109 100
139 147.3 75.6 113.4#
65 140.1 107.6 161.4



#56
Dibenzofuran
Concen: 19.559 ng/ul
RT: 14.781 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:168 Resp: 1649755
Ion Ratio Lower Upper
168 100
139 37.9 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 23.375 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:165 Resp: 319975
Ion Ratio Lower Upper
165 100
63 44.1 47.0 70.6#
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.491 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:232 Resp: 260020
Ion Ratio Lower Upper
232 100
131 51.6 35.8 53.8
130 2.6 1.8 2.8
166 32.4 23.7 35.5









#59
Diethylphthalate
Concen: 19.481 ng/ul
RT: 15.222 min Scan# 2063
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:149 Resp: 1324267
Ion Ratio Lower Upper
149 100
177 21.1 17.6 26.4
150 12.2 10.3 15.5



#61
Fluorene
Concen: 19.553 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:166 Resp: 1315021
Ion Ratio Lower Upper
166 100
165 97.3 77.4 116.2
167 13.4 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 19.449 ng/ul
RT: 15.434 min Scan# 2100
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:204 Resp: 612329
Ion Ratio Lower Upper
204 100
206 32.4 26.0 39.0
141 64.9 49.3 73.9



#63
4-Nitroaniline
Concen: 22.308 ng/ul
RT: 15.451 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:138 Resp: 253325
Ion Ratio Lower Upper
138 100
92 51.6 46.2 69.2
108 76.7 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 18.706 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:198 Resp: 122651
Ion Ratio Lower Upper
198 100
182 5.4 4.3 6.5
77 30.5 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 19.885 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:169 Resp: 1103032
Ion Ratio Lower Upper
169 100
168 67.1 52.6 78.8
167 36.6 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 19.548 ng/ul
RT: 16.334 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:248 Resp: 358795
Ion Ratio Lower Upper
248 100
250 98.3 80.8 121.2
141 82.3 63.9 95.9



#69
Hexachlorobenzene
Concen: 19.392 ng/ul
RT: 16.434 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:284 Resp: 412262
Ion Ratio Lower Upper
284 100
142 39.5 27.8 41.6
249 31.8 25.4 38.2









#70
Atrazine
Concen: 19.024 ng/ul
RT: 16.610 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:200 Resp: 349877
Ion Ratio Lower Upper
200 100
173 27.0 20.7 31.1
215 47.0 38.6 57.8



#71
Pentachlorophenol
Concen: 18.347 ng/ul
RT: 16.786 min Scan# 2329
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:266 Resp: 200836
Ion Ratio Lower Upper
266 100
264 62.8 51.4 77.2
268 63.1 53.5 80.3









#72
Phenanthrene
Concen: 19.737 ng/ul
RT: 17.186 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:178 Resp: 1924892
Ion Ratio Lower Upper
178 100
179 15.2 12.7 19.1
176 18.6 14.6 22.0



#74
Anthracene
Concen: 19.559 ng/ul
RT: 17.275 min Scan# 2412
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:178 Resp: 1918600
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 18.1 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.121 ng/uL

RT: 13.169 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:216 Resp: 577876

Ion Ratio Lower Upper

216 100

214 77.4 65.0 97.4

218 46.3 39.1 58.7



#76

Pentachlorobenzene

Concen: 20.007 ng/uL

RT: 14.692 min Scan# 1973

Delta R.T. -0.006 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58




Tgt Ion:250 Resp: 518563


Ion Ratio Lower Upper

250 100

248 63.1 49.8 74.6

252 64.0 51.2 76.8









#77
 Carbazole
 Concen: 19.312 ng/ul
 RT: 17.551 min Scan# 2459
 Delta R.T. -0.000 min
 Lab File: BP010800.D
 Acq: 24 Jun 2022 18:58

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020679

Tgt Ion:167 Resp: 1705212




Ion	Ratio	Lower	Upper
167	100		
166	21.3	16.9	25.3
139	13.4	10.6	16.0




#78
 Di-n-butylphthalate
 Concen: 18.865 ng/ul
 RT: 18.128 min Scan# 2557
 Delta R.T. -0.006 min
 Lab File: BP010800.D
 Acq: 24 Jun 2022 18:58

Tgt Ion:149 Resp: 1984927

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.1	10.7
104	5.0	4.6	6.8









#80
Fluoranthene
Concen: 20.623 ng/u1
RT: 19.163 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:202 Resp: 1981468
Ion Ratio Lower Upper
202 100
101 13.9 7.1 10.7#
100 10.4 5.7 8.5#



#82
Pyrene
Concen: 20.829 ng/u1
RT: 19.516 min Scan# 2793
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:202 Resp: 2020990
Ion Ratio Lower Upper
202 100
101 16.3 7.9 11.9#
100 13.9 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 20.096 ng/ul
RT: 20.416 min Scan# 2946
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:149 Resp: 686334
Ion Ratio Lower Upper
149 100
91 74.6 60.0 90.0
206 18.5 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 18.149 ng/ul
RT: 21.174 min Scan# 3075
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:252 Resp: 537368
Ion Ratio Lower Upper
252 100
254 65.9 51.0 76.4
126 14.3 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 19.296 ng/ul
RT: 21.239 min Scan# 3086
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:228 Resp: 1730458
Ion Ratio Lower Upper
228 100
229 19.4 15.0 22.4
226 27.0 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.623 ng/ul
RT: 21.186 min Scan# 3077
Delta R.T. -0.012 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:149 Resp: 1061282
Ion Ratio Lower Upper
149 100
167 26.0 22.6 33.8
279 3.4 4.2 6.4#









#87
Chrysene
Concen: 19.606 ng/ul
RT: 21.286 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:228 Resp: 1684869
Ion Ratio Lower Upper
228 100
226 29.5 23.6 35.4
229 19.5 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 18.426 ng/ul
RT: 22.110 min Scan# 3234
Delta R.T. -0.012 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:149 Resp: 1592820









#90
Benzo(b)fluoranthene
Concen: 19.041 ng/ul
RT: 22.898 min Scan# 3369
Delta R.T. -0.006 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58




Instrument :
BNA_P
ClientSampleId :
SSTD020679


Tgt Ion:252 Resp: 1724489
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 12.8 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 19.770 ng/ul
RT: 22.951 min Scan# 3377
Delta R.T. -0.000 min
Lab File: BP010800.D
Acq: 24 Jun 2022 18:58

Tgt Ion:252 Resp: 1701928
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.4
125 12.3 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 19.601 ng/ul

RT: 23.515 min Scan# 34

Delta R.T. -0.006 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:252 Resp: 1726889

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0

125 14.1 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.508 ng/ul

RT: 26.086 min Scan# 3910

Delta R.T. -0.006 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58




Tgt Ion:276 Resp: 2073863


Ion Ratio Lower Upper

276 100

138 28.8 14.6 22.0#

277 24.6 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 19.620 ng/ul

RT: 26.109 min Scan# 3916

Delta R.T. -0.012 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58

Instrument :

BNA_P

ClientSampleId :

SSTD020679





Tgt Ion:278 Resp: 1789119

Ion Ratio Lower Upper

278 100

139 19.7 9.4 14.2#

279 23.7 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 19.530 ng/ul

RT: 26.833 min Scan# 4037

Delta R.T. -0.012 min

Lab File: BP010800.D

Acq: 24 Jun 2022 18:58

Tgt Ion:276 Resp: 1766057

Ion Ratio Lower Upper

276 100

138 27.0 12.7 19.1#

277 23.9 18.2 27.4

