

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062422\
 Data File : BP010811.D
 Acq On : 25 Jun 2022 01:11
 Operator : CG/JU
 Sample : N3374-03MSD
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EW4S6MSD

Quant Time: Jun 25 03:04:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut
 Upadhyay
 06/27/2022

Supervised By :mohammad
 ahmed
 06/29/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	351841	20.000	ng/u1	0.00
20) Naphthalene-d8	10.528	136	17506799m	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.381	164	706889	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.139	188	1323828	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.257	240	1369087	20.000	ng/u1	# 0.00
88) Perylene-d12	23.627	264	1529567	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	51172	5.502	ng/uL	0.00
4) Pyridine-d5	3.652	84	281133m	11.355	ng/u1	0.05
7) Phenol-d5	6.905	99	254618	9.142	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.069	67	564500	30.991	ng/u1	0.00
11) 2-Chlorophenol-d4	7.258	132	706929	30.593	ng/u1	0.00
15) 4-Methylphenol-d8	8.440	113	470359	21.363	ng/u1	0.00
21) Nitrobenzene-d5	8.881	128	357761	3.210	ng/u1	0.00
24) 2-Nitrophenol-d4	9.604	143	417657	4.530	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.140	165	657458	2.886	ng/u1	0.00
31) 4-Chloroaniline-d4	10.716	131	718247	1.931	ng/u1	0.06
46) Dimethylphthalate-d6	13.816	166	1594381	32.927	ng/u1	0.02
49) Acenaphthylene-d8	14.075	160	2056348	33.274	ng/u1	0.00
54) 4-Nitrophenol-d4	14.598	143	98293	12.996	ng/u1	0.02
60) Fluorene-d10	15.381	176	1414174	34.032	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.498	200	268061	61.118	ng/u1	0.00
73) Anthracene-d10	17.239	188	2111997	35.858	ng/u1	0.00
81) Pyrene-d10	19.492	212	2391102	32.057	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.474	264	2645376	34.999	ng/u1	0.00
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.246	88	93858	9.463	ng/uL#	74
5) Pyridine	3.664	79	335711m	13.386	ng/u1	
6) Benzaldehyde	6.869	77	530062	36.069	ng/u1	97
8) Phenol	6.934	94	9473595	318.026	ng/u1#	87
10) Bis(2-Chloroethyl)ether	7.157	93	734410	30.697	ng/u1	87
12) 2-Chlorophenol	7.287	128	702584	29.455	ng/u1	91
13) 2-Methylphenol	8.169	108	537403	24.232	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.263	45	913155	28.447	ng/u1#	96
16) Acetophenone	8.557	105	219371	6.488	ng/u1#	88
18) 4-Methylphenol	8.504	108	986962	42.735	ng/u1	100
19) Hexachloroethane	8.799	117	268441	27.789	ng/u1	86
22) Nitrobenzene	8.928	77	758526	2.727	ng/u1#	90
25) 2-Nitrophenol	9.634	139	452503	4.053	ng/u1#	78
26) 2,4-Dimethylphenol	9.710	107	4056130m	13.296	ng/u1	
27) Bis(2-Chloroethoxy)met...	9.957	93	909928	2.473	ng/u1#	94
29) 2,4-Dichlorophenol	10.169	162	638387	2.697	ng/u1	99
30) Naphthalene	10.575	128	2024964	2.219	ng/u1	99
32) 4-Chloroaniline	10.740	127	695595	1.888	ng/u1	98
33) Hexachlorobutadiene	10.863	225	312173	2.064	ng/u1	95
34) Caprolactam	11.510	113	373549	5.023	ng/u1#	19
35) 4-Chloro-3-methylphenol	11.969	107	38885702	151.557	ng/u1#	22
36) 2-Methylnaphthalene	12.198	142	1302494	2.204	ng/u1	98
37) 1-Methylnaphthalene	12.422	142	1335062	2.274	ng/u1#	99
39) 1,2,4,5-Tetrachloroben...	12.563	216	605265	29.796	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062422\
 Data File : BP010811.D
 Acq On : 25 Jun 2022 01:11
 Operator : CG/JU
 Sample : N3374-03MSD
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EW4S6MSD

Quant Time: Jun 25 03:04:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut
 Upadhyay
 06/27/2022

Supervised By :mohammad
 ahmed
 06/29/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
40) Hexachlorocyclopentadiene	12.545	237	310593	24.130	ng/ul	94
41) 2,4,6-Trichlorophenol	12.810	196	508079	43.383	ng/ul	97
42) 2,4,5-Trichlorophenol	12.881	196	539647	42.939	ng/ul	100
43) 1,1'-Biphenyl	13.216	154	1664920	29.918	ng/ul	98
44) 2-Chloronaphthalene	13.251	162	720930	17.198	ng/ul#	1
45) 2-Nitroaniline	13.481	65	409766	46.071	ng/ul#	79
47) Dimethylphthalate	13.857	163	1574921	32.477	ng/ul	98
48) 2,6-Dinitrotoluene	13.969	165	342283	47.109	ng/ul	91
50) Acenaphthylene	14.104	152	2104094	31.087	ng/ul	99
51) 3-Nitroaniline	14.298	138	335051	38.464	ng/ul	84
52) Acenaphthene	14.445	153	1483049	33.959	ng/ul	99
53) 2,4-Dinitrophenol	14.492	184	159363	57.132	ng/ul#	87
55) 4-Nitrophenol	14.610	109	81960	13.628	ng/ul#	82
56) Dibenzofuran	14.787	168	1930407	32.270	ng/ul	95
57) 2,4-Dinitrotoluene	14.751	165	468295	48.235	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.010	232	380613	42.291	ng/ul#	91
59) Diethylphthalate	15.222	149	1666292	34.563	ng/ul	99
61) Fluorene	15.434	166	1564774	32.805	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.434	204	731358	32.753	ng/ul	96
63) 4-Nitroaniline	15.463	138	358125	44.467	ng/ul#	74
66) 4,6-Dinitro-2-methylph...	15.510	198	260707	55.008	ng/ul#	88
67) N-Nitrosodiphenylamine	15.651	169	1359691	33.911	ng/ul	98
68) 4-Bromophenyl-phenylether	16.333	248	437681	32.989	ng/ul	96
69) Hexachlorobenzene	16.433	284	484664	31.539	ng/ul	95
70) Atrazine	16.639	200	271482	20.422	ng/ul	97
71) Pentachlorophenol	16.786	266	196928	24.888	ng/ul	96
72) Phenanthrene	17.186	178	2374016	33.675	ng/ul	99
74) Anthracene	17.275	178	2446899	34.510	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.169	216	639721	29.284	ng/uL	96
76) Pentachlorobenzene	14.698	250	582296	31.080	ng/uL	98
77) Carbazole	17.557	167	2567876	40.234	ng/ul	99
78) Di-n-butylphthalate	18.128	149	2826131	37.159	ng/ul	98
80) Fluoranthene	19.163	202	2731519	29.776	ng/ul#	88
82) Pyrene	19.522	202	2823109	30.473	ng/ul#	87
83) Butylbenzylphthalate	20.416	149	1396556	42.827	ng/ul	97
85) Benzo(a)anthracene	21.239	228	2855564	33.350	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.186	149	2025005	39.215	ng/ul#	97
87) Chrysene	21.292	228	2720359	33.154	ng/ul	99
89) Di-n-octyl phthalate	22.110	149	3607017	39.729	ng/ul	100
90) Benzo(b)fluoranthene	22.904	252	3148227	33.098	ng/ul#	96
91) Benzo(k)fluoranthene	22.951	252	3060545	33.851	ng/ul#	95
93) Benzo(a)pyrene	23.521	252	2803932	30.303	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.092	276	3820533	34.218	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.121	278	3175500	33.157	ng/ul#	94
96) Benzo(g,h,i)perylene	26.845	276	3108257	32.727	ng/ul#	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062422\
 Data File : BP010811.D
 Acq On : 25 Jun 2022 01:11
 Operator : CG/JU
 Sample : N3374-03MSD
 Misc :
 ALS Vial : 30 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 EW4S6MSD


Quant Time: Jun 25 03:04:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut
 Upadhyay
 06/27/2022


Supervised By :mohammad
 ahmed
 06/29/2022





#2
1,4-Dioxane
Concen: 9.463 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion: 88 Resp: 93858
Ion Ratio Lower Upper
88 100
43 91.9 35.8 53.6
58 83.1 68.1 102.1

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#5
Pyridine
Concen: 13.386 ng/uL m
RT: 3.664 min Scan# 98
Delta R.T. 0.035 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion: 79 Resp: 335711
Ion Ratio Lower Upper
79 100
52 92.6 83.8 125.8
51 35.0 41.0 61.4





#6
Benzaldehyde
Concen: 36.069 ng/ul
RT: 6.869 min Scan# 642
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion: 77 Resp: 530062
Ion Ratio Lower Upper
77 100
105 98.5 76.0 114.0
106 93.3 73.2 109.8

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#8
Phenol
Concen: 318.026 ng/ul
RT: 6.934 min Scan# 654
Delta R.T. 0.018 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Tgt Ion: 94 Resp: 9473595
Ion Ratio Lower Upper
94 100
65 27.8 27.4 41.2
66 36.9 37.4 56.2#






#10
Bis(2-Chloroethyl)ether
Concen: 30.697 ng/ul
RT: 7.157 min Scan# 691
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :

EW4S6MSD






Tgt Ion: 93 Resp: 734410
Ion Ratio Lower Upper
93 100
63 74.9 72.4 108.6
95 31.9 27.3 40.9

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#12
2-Chlorophenol
Concen: 29.455 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Tgt Ion:128 Resp: 702584
Ion Ratio Lower Upper
128 100
64 45.8 44.8 67.2
130 31.6 25.8 38.6





#13
2-Methylphenol
Concen: 24.232 ng/ul
RT: 8.169 min Scan# 80
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11





Instrument :
BNA_P
ClientSampleId :
EW4S6MSD

Tgt Ion:108 Resp: 53740
Ion Ratio Lower Upper
108 100
107 90.4 71.4 107.0

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 28.447 ng/ul
RT: 8.263 min Scan# 880
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion: 45 Resp: 913155
Ion Ratio Lower Upper
45 100
77 14.8 11.2 16.8
79 11.6 6.9 10.3#





#16
Acetophenone
Concen: 6.488 ng/ul
RT: 8.557 min Scan# 910
Delta R.T. 0.012 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:105 Resp: 21937
Ion Ratio Lower Upper
105 100
77 76.3 67.5 101.3
51 26.0 29.6 44.4

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#18
4-Methylphenol
Concen: 42.735 ng/ul
RT: 8.504 min Scan# 921
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Tgt Ion:108 Resp: 986962
Ion Ratio Lower Upper
108 100
107 110.0 88.3 132.5






#19
Hexachloroethane
Concen: 27.789 ng/ul
RT: 8.799 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :

EW4S6MSD






Tgt Ion:117 Resp: 26844
Ion Ratio Lower Upper
117 100
201 94.0 85.7 128.5
199 58.3 57.4 86.0

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#22
Nitrobenzene
Concen: 2.727 ng/ul
RT: 8.928 min Scan# 993
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion: 77 Resp: 758526
Ion Ratio Lower Upper
77 100
123 47.8 31.5 47.3#
65 13.7 11.4 17.2





#25
2-Nitrophenol
Concen: 4.053 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:139 Resp: 452500
Ion Ratio Lower Upper
139 100
65 48.4 54.2 81.4
109 27.3 30.6 45.8


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay
06/27/2022



Supervised By :mohammad
ahmed
06/29/2022




#26
2,4-Dimethylphenol
Concen: 13.296 ng/ul m
RT: 9.710 min Scan# 1126
Delta R.T. 0.012 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11



Tgt Ion:107 Resp: 4056130
Ion Ratio Lower Upper
107 100
121 52.2 39.1 58.7
122 87.4 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 2.473 ng/ul

RT: 9.957 min Scan# 1165

Delta R.T. 0.012 min

Lab File: BP010811.D

Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :

EW4S6MSD

Tgt Ion: 93 Resp: 909928

Ion Ratio Lower Upper

93 100

95 35.2 26.6 39.8

123 14.9 7.8 11.8

Manual Integrations

APPROVED

Reviewed By :Jagrut





Upadhyay

06/27/2022

Supervised By :mohammad

ahmed

06/29/2022



#29

2,4-Dichlorophenol

Concen: 2.697 ng/ul

RT: 10.169 min Scan# 1204

Delta R.T. 0.006 min

Lab File: BP010811.D

Acq: 25 Jun 2022 01:11




Tgt Ion:162 Resp: 638387


Ion Ratio Lower Upper

162 100

164 65.7 52.5 78.7


98 40.7 31.4 47.0





#30
Naphthalene
Concen: 2.219 ng/ul
RT: 10.575 min Scan# 1301
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:128 Resp: 2024964
Ion Ratio Lower Upper
128 100
129 11.3 8.7 13.1
127 13.3 10.2 15.4

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#32
4-Chloroaniline
Concen: 1.888 ng/ul
RT: 10.740 min Scan# 1301
Delta R.T. 0.059 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:127 Resp: 695595
Ion Ratio Lower Upper
127 100
129 34.0 26.2 39.2





#33
Hexachlorobutadiene
Concen: 2.064 ng/ul
RT: 10.863 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:225 Resp: 31217
Ion Ratio Lower Upper
225 100
223 61.5 53.7 80.5
227 63.9 49.4 74.2

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#34
Caprolactam
Concen: 5.023 ng/ul
RT: 11.510 min Scan# 1432
Delta R.T. 0.065 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:113 Resp: 373549
Ion Ratio Lower Upper
113 100
55 375.0 153.8 230.8#
56 110.5 110.2 165.4





#35
4-Chloro-3-methylphenol
Concen: 151.557 ng/ul
RT: 11.969 min Scan# 11
Delta R.T. 0.153 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:107 Resp:3888570
Ion Ratio Lower Upper
107 100
144 0.0 18.6 27.8
142 0.0 59.6 89.4

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#36
2-Methylnaphthalene
Concen: 2.204 ng/ul
RT: 12.198 min Scan# 1549
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:142 Resp: 1302494
Ion Ratio Lower Upper
142 100
141 89.3 70.1 105.1





#37
1-Methylnaphthalene
Concen: 2.274 ng/ul
RT: 12.422 min Scan# 1585
Delta R.T. 0.012 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:142 Resp: 1335062
Ion Ratio Lower Upper
142 100
141 92.2 73.3 109.9
116 3.9 4.0 6.0

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.796 ng/ul
RT: 12.563 min Scan# 1611
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:216 Resp: 605265
Ion Ratio Lower Upper
216 100
214 77.0 63.0 94.6
179 21.8 18.1 27.1
108 18.6 13.6 20.4





#40
Hexachlorocyclopentadiene
Concen: 24.130 ng/ul
RT: 12.545 min Scan# 1607
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:237 Resp: 31059
Ion Ratio Lower Upper
237 100
235 61.0 52.8 79.2
272 12.8 11.5 17.3


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay
06/27/2022



Supervised By :mohammad
ahmed
06/29/2022




#41
2,4,6-Trichlorophenol
Concen: 43.383 ng/ul
RT: 12.810 min Scan# 1653
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11




Tgt Ion:196 Resp: 508079
Ion Ratio Lower Upper
196 100
198 95.2 78.2 117.4
200 29.4 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 42.939 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. 0.012 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:196 Resp: 53964
Ion Ratio Lower Upper
196 100
198 95.9 77.0 115.4
200 30.8 24.8 37.2

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#43
1,1'-Biphenyl
Concen: 29.918 ng/ul
RT: 13.216 min Scan# 1722
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:154 Resp: 1664920
Ion Ratio Lower Upper
154 100
153 40.2 31.9 47.9
76 13.6 9.2 13.8





#44
2-Chloronaphthalene
Concen: 17.198 ng/ul
RT: 13.251 min Scan# 1727
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:162 Resp: 720930
Ion Ratio Lower Upper
162 100
164 359.3 25.5 38.3
127 68.6 31.9 47.9

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#45
2-Nitroaniline
Concen: 46.071 ng/ul
RT: 13.481 min Scan# 1767
Delta R.T. 0.023 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion: 65 Resp: 409766
Ion Ratio Lower Upper
65 100
92 70.4 49.0 73.6
138 108.0 66.4 99.6





#47
Dimethylphthalate
Concen: 32.477 ng/ul
RT: 13.857 min Scan# 1831
Delta R.T. 0.012 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:163 Resp: 157492
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 10.7 7.6 11.4

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#48
2,6-Dinitrotoluene
Concen: 47.109 ng/ul
RT: 13.969 min Scan# 1850
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:165 Resp: 342283
Ion Ratio Lower Upper
165 100
89 56.6 52.9 79.3
121 25.4 19.4 29.2





#50
Acenaphthylene
Concen: 31.087 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:152 Resp: 2104094
Ion Ratio Lower Upper
152 100
151 20.1 16.5 24.7
153 13.4 10.1 15.1

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#51
3-Nitroaniline
Concen: 38.464 ng/ul
RT: 14.298 min Scan# 1906
Delta R.T. 0.012 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Tgt Ion:138 Resp: 335051
Ion Ratio Lower Upper
138 100
108 14.6 11.0 16.4
92 110.5 104.6 156.8





#52
Acenaphthene
Concen: 33.959 ng/ul
RT: 14.445 min Scan# 1930
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :





EW4S6MSD

Tgt Ion:153 Resp: 1483049
Ion Ratio Lower Upper
153 100
152 46.8 36.2 54.2
154 88.8 70.8 106.2

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#53
2,4-Dinitrophenol
Concen: 57.132 ng/ul
RT: 14.492 min Scan# 1939
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:184 Resp: 159363
Ion Ratio Lower Upper
184 100
63 73.6 73.8 110.6#
154 67.8 57.2 85.8





#55
4-Nitrophenol
Concen: 13.628 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. 0.018 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:109 Resp: 81960
Ion Ratio Lower Upper
109 100
139 130.5 75.6 113.4
65 128.6 107.6 161.4

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#56
Dibenzofuran
Concen: 32.270 ng/ul
RT: 14.787 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:168 Resp: 1930407
Ion Ratio Lower Upper
168 100
139 37.6 32.8 49.2





#57
2,4-Dinitrotoluene
Concen: 48.235 ng/ul
RT: 14.751 min Scan# 1982
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:165 Resp: 468295
Ion Ratio Lower Upper
165 100
63 42.5 47.0 70.6
182 3.1 2.2 3.2

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 42.291 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Tgt Ion:232 Resp: 380613
Ion Ratio Lower Upper
232 100
131 52.2 35.8 53.8
130 3.7 1.8 2.8
166 33.4 23.7 35.5






#59
Diethylphthalate
Concen: 34.563 ng/ul
RT: 15.222 min Scan# 2063
Delta R.T. -0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :

EW4S6MSD



Tgt Ion:149 Resp: 1666292
Ion Ratio Lower Upper
149 100
177 21.3 17.6 26.4
150 12.8 10.3 15.5

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay

06/27/2022


Supervised By :mohammad

ahmed



06/29/2022




#61
Fluorene
Concen: 32.805 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11



Tgt Ion:166 Resp: 1564774
Ion Ratio Lower Upper
166 100
165 98.1 77.4 116.2
167 13.8 10.8 16.2






#62
4-Chlorophenyl-phenylether
Concen: 32.753 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :

EW4S6MSD



Tgt Ion:204 Resp: 731358
Ion Ratio Lower Upper
204 100
206 32.8 26.0 39.0
141 65.8 49.3 73.9

Manual Integrations

APPROVED

Reviewed By :Jagrut




Upadhyay

06/27/2022

Supervised By :mohammad




ahmed


06/29/2022



#63
4-Nitroaniline
Concen: 44.467 ng/ul
RT: 15.463 min Scan# 2104
Delta R.T. 0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:138 Resp: 358125
Ion Ratio Lower Upper
138 100
92 49.9 46.2 69.2
108 72.3 85.8 128.8#





#66
4,6-Dinitro-2-methylphenol
Concen: 55.008 ng/ul
RT: 15.510 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:198 Resp: 26070
Ion Ratio Lower Upper
198 100
182 5.6 4.3 6.5
77 37.2 23.7 35.5

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#67
N-Nitrosodiphenylamine
Concen: 33.911 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Tgt Ion:169 Resp: 1359691
Ion Ratio Lower Upper
169 100
168 66.7 52.6 78.8
167 36.0 27.8 41.6





#68
4-Bromophenyl-phenylether
Concen: 32.989 ng/ul
RT: 16.333 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :





EW4S6MSD

Tgt Ion:248 Resp: 43768
Ion Ratio Lower Upper
248 100
250 95.9 80.8 121.2
141 82.6 63.9 95.9

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#69
Hexachlorobenzene
Concen: 31.539 ng/ul
RT: 16.433 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:284 Resp: 484664
Ion Ratio Lower Upper
284 100
142 39.4 27.8 41.6
249 30.8 25.4 38.2





#70
Atrazine
Concen: 20.422 ng/ul
RT: 16.639 min Scan# 211
Delta R.T. 0.023 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:200 Resp: 27148
Ion Ratio Lower Upper
200 100
173 28.3 20.7 31.1
215 47.2 38.6 57.8

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#71
Pentachlorophenol
Concen: 24.888 ng/ul
RT: 16.786 min Scan# 2329
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:266 Resp: 196928
Ion Ratio Lower Upper
266 100
264 62.1 51.4 77.2
268 62.5 53.5 80.3





#72
Phenanthrene
Concen: 33.675 ng/u1
RT: 17.186 min Scan# 2397
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:178 Resp: 2374010
Ion Ratio Lower Upper
178 100
179 15.6 12.7 19.1
176 19.0 14.6 22.0

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#74
Anthracene
Concen: 34.510 ng/u1
RT: 17.275 min Scan# 2412
Delta R.T. -0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:178 Resp: 2446899
Ion Ratio Lower Upper
178 100
179 15.8 12.2 18.2
176 18.8 14.5 21.7





#75
1,2,3,4-Tetrachlorobenzene
Concen: 29.284 ng/uL
RT: 13.169 min Scan# 1714
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:216 Resp: 63972
Ion Ratio Lower Upper
216 100
214 77.6 65.0 97.4
218 46.8 39.1 58.7

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#76
Pentachlorobenzene
Concen: 31.080 ng/uL
RT: 14.698 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:250 Resp: 582296
Ion Ratio Lower Upper
250 100
248 64.4 49.8 74.6
252 64.6 51.2 76.8





#77
 Carbazole
 Concen: 40.234 ng/ul
 RT: 17.557 min Scan# 2459
 Delta R.T. 0.006 min
 Lab File: BP010811.D
 Acq: 25 Jun 2022 01:11

Instrument :
 BNA_P
 ClientSampleId :
 EW4S6MSD






Tgt Ion:167 Resp: 2567870
 Ion Ratio Lower Upper
 167 100
 166 21.5 16.9 25.3
 139 13.9 10.6 16.0

Manual Integrations
 APPROVED




Reviewed By :Jagrut
 Upadhyay
 06/27/2022


Supervised By :mohammad
 ahmed
 06/29/2022



#78
 Di-n-butylphthalate
 Concen: 37.159 ng/ul
 RT: 18.128 min Scan# 2557
 Delta R.T. -0.006 min
 Lab File: BP010811.D
 Acq: 25 Jun 2022 01:11


Tgt Ion:149 Resp: 2826131
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 4.9 4.6 6.8





#80
Fluoranthene
Concen: 29.776 ng/ul
RT: 19.163 min Scan# 2733
Delta R.T. -0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:202 Resp: 2731519
Ion Ratio Lower Upper
202 100
101 13.7 7.1 10.7
100 10.8 5.7 8.5


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay
06/27/2022



Supervised By :mohammad
ahmed
06/29/2022




#82
Pyrene
Concen: 30.473 ng/ul
RT: 19.522 min Scan# 2794
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11



Tgt Ion:202 Resp: 2823109
Ion Ratio Lower Upper
202 100
101 15.5 7.9 11.9#
100 12.4 6.7 10.1#






#83
Butylbenzylphthalate
Concen: 42.827 ng/ul
RT: 20.416 min Scan# 2946
Delta R.T. -0.006 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :

EW4S6MSD



Tgt Ion:149 Resp: 1396556
Ion Ratio Lower Upper
149 100
91 73.6 60.0 90.0
206 19.1 17.4 26.2

Manual Integrations

APPROVED

Reviewed By :Jagrut




Upadhyay

06/27/2022

Supervised By :mohammad




ahmed


06/29/2022



#85
Benzo(a)anthracene
Concen: 33.350 ng/ul
RT: 21.239 min Scan# 3086
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11


Tgt Ion:228 Resp: 2855564
Ion Ratio Lower Upper
228 100
229 20.3 15.0 22.4
226 27.0 20.6 31.0





#86
Bis(2-ethylhexyl)phthalate
Concen: 39.215 ng/ul
RT: 21.186 min Scan# 3077
Delta R.T. -0.012 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD






Tgt Ion:149 Resp: 202500
Ion Ratio Lower Upper
149 100
167 27.1 22.6 33.8
279 3.7 4.2 6.4

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#87
Chrysene
Concen: 33.154 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Tgt Ion:228 Resp: 2720359
Ion Ratio Lower Upper
228 100
226 30.0 23.6 35.4
229 20.1 15.6 23.4





#89
Di-n-octyl phthalate
Concen: 39.729 ng/ul
RT: 22.110 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11





Instrument :
BNA_P
ClientSampleId :
EW4S6MSD

Tgt Ion:149 Resp: 360701

Manual Integrations
APPROVED




Reviewed By :Jagrut
Upadhyay
06/27/2022


Supervised By :mohammad
ahmed
06/29/2022



#90
Benzo(b)fluoranthene
Concen: 33.098 ng/ul
RT: 22.904 min Scan# 3369
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Tgt Ion:252 Resp: 3148227
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.6
125 12.3 6.2 9.2#






#91
Benzo(k)fluoranthene
Concen: 33.851 ng/u1
RT: 22.951 min Scan# 31
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :

EW4S6MSD



Tgt Ion:252 Resp: 306054
Ion Ratio Lower Upper
252 100
253 21.9 17.0 25.4
125 13.0 6.2 9.4

Manual Integrations

APPROVED

Reviewed By :Jagrut




Upadhyay

06/27/2022


Supervised By :mohammad

ahmed



06/29/2022




#93
Benzo(a)pyrene
Concen: 30.303 ng/u1
RT: 23.521 min Scan# 3474
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11



Tgt Ion:252 Resp: 2803932
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 14.0 6.6 10.0





#94

Indeno(1,2,3-cd)pyrene

Concen: 34.218 ng/ul

RT: 26.092 min Scan# 3911

Delta R.T. -0.000 min

Lab File: BP010811.D

Acq: 25 Jun 2022 01:11

Instrument :

BNA_P

ClientSampleId :

EW4S6MSD

Tgt Ion:276 Resp: 382053

Ion Ratio Lower Upper

276 100

138 27.4 14.6 22.0

277 24.3 19.6 29.4

Manual Integrations

APPROVED

Reviewed By :Jagrut





Upadhyay

06/27/2022

Supervised By :mohammad

ahmed

06/29/2022



#95

Dibenzo(a,h)anthracene

Concen: 33.157 ng/ul

RT: 26.121 min Scan# 3916

Delta R.T. -0.000 min

Lab File: BP010811.D

Acq: 25 Jun 2022 01:11




Tgt Ion:278 Resp: 3175500


Ion Ratio Lower Upper

278 100

139 19.0 9.4 14.2


279 23.8 18.9 28.3





#96
Benzo(g,h,i)perylene
Concen: 32.727 ng/ul
RT: 26.845 min Scan# 4039
Delta R.T. -0.000 min
Lab File: BP010811.D
Acq: 25 Jun 2022 01:11

Instrument :
BNA_P
ClientSampleId :
EW4S6MSD



Tgt Ion: 276 Resp: 310825
Ion Ratio Lower Upper
276 100
138 26.0 12.7 19.1
277 24.2 18.2 27.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay
06/27/2022

Supervised By :mohammad
ahmed
06/29/2022

