


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062422\
 Data File : BP010781.D
 Acq On : 24 Jun 2022 08:15
 Operator : CG/JU
 Sample : DFTPP661
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP661

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jun 27 23:12:49 2022



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.2	238105	PASS
68	69	0.00	2	1.7	4172	PASS
69	198	0.00	100	42.0	242727	PASS
70	69	0.00	2	0.5	1207	PASS
127	198	10	80	52.9	305242	PASS
197	198	0.00	2	0.2	1141	PASS
198	198	100	100	100.0	577251	PASS
199	198	5	9	6.7	38645	PASS
275	198	10	60	27.2	156843	PASS
365	198	1	100	3.7	21496	PASS
441	198	0.01	100	14.9	86147	PASS
442	198	50	100	96.0	554325	PASS
443	442	15	24	19.4	107358	PASS

m/z	Abundance
35.90	272.0
36.90	89.0
38.00	206.0
39.00	396.0
40.00	1772.0
41.10	642.0
42.00	289.0
43.00	548.0
44.00	9931.0
44.90	321.0
46.00	63.0
46.90	210.0
47.90	20.0
48.90	720.0
50.00	278.0
50.90	444.0
51.80	226.0
53.00	134.0
54.00	180.0
55.10	367.0
56.00	273.0
56.90	306.0
57.90	150.0
59.00	58.0
60.00	73.0
61.00	89.0
62.20	33.0
63.10	106.0
64.10	93.0
64.90	216.0
65.70	157.0
66.90	228.0
68.00	135.0
68.90	196.0
70.10	166.0
71.00	72.0
71.90	28.0
72.90	595.0
74.10	67.0
74.90	227.0
76.00	153.0
77.10	401.0
78.00	394.0
78.80	143.0
79.90	76.0
81.10	136.0
81.90	169.0
82.90	153.0
84.00	484.0
84.90	139.0
85.90	380.0
87.70	66.0
88.80	164.0
90.00	26.0
91.10	316.0
92.00	161.0
93.10	60.0
94.00	182.0
94.80	75.0
96.00	808.0
97.10	123.0
98.00	96.0
98.90	20.0
99.60	11.0
101.30	9.0
101.80	8.0
102.80	170.0
103.90	178.0
104.90	250.0
105.70	46.0
107.00	130.0
108.00	58.0
109.10	18.0
109.50	18.0
110.00	14.0
111.00	63.0
112.00	46.0
114.00	15.0
114.90	300.0
116.10	60.0
116.90	165.0
117.90	134.0
118.90	264.0
119.90	7.0
120.90	77.0
121.90	24.0
123.10	58.0
123.60	20.0
124.20	16.0
125.00	20.0
126.00	14.0
126.50	22.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

127.10	17.0
128.10	87.0
129.00	77.0
130.00	57.0
130.70	8.0
131.80	99.0
132.90	510.0
134.00	59.0
135.00	69.0
136.10	90.0
136.70	72.0
138.00	16.0
138.50	14.0
139.30	58.0
140.00	41.0
141.00	75.0
141.70	11.0
142.50	16.0
143.10	2.0
143.80	17.0
144.80	14.0
145.50	13.0
146.10	6.0
146.90	270.0
147.90	37.0
148.80	190.0
150.10	9.0
150.90	27.0
151.70	9.0
152.80	12.0
153.60	17.0
154.20	17.0
155.00	19.0
156.30	20.0
157.20	14.0
157.90	16.0
159.70	10.0
160.50	8.0
161.00	26.0
162.00	24.0
162.70	59.0
164.00	1.0
164.70	69.0
165.20	64.0
165.80	11.0
166.90	94.0
167.70	15.0
169.20	17.0
170.50	7.0
171.30	3.0
172.20	13.0
172.80	17.0
173.70	12.0
175.10	11.0
175.90	18.0
177.00	210.0
178.10	77.0
179.10	165.0
180.10	6.0
181.00	12.0
181.60	6.0
182.20	13.0
183.10	59.0
185.00	14.0
185.70	1.0
186.30	7.0
187.20	2.0
187.80	10.0
189.00	54.0
191.00	503.0
192.10	125.0
193.00	434.0
193.80	42.0
194.30	19.0
195.10	2.0
195.80	10.0
196.60	1.0
197.40	6.0
197.90	8.0
199.00	16.0
200.80	10.0
202.10	23.0
203.10	7.0
203.70	7.0
204.10	10.0
204.70	13.0
205.20	9.0
207.00	4904.0
208.00	1033.0
209.00	637.0
210.10	174.0
210.90	12.0
211.40	10.0
212.10	12.0
212.40	14.0
213.00	23.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

214.40	11.0
214.80	15.0
216.60	9.0
217.80	14.0
218.70	29.0
221.10	71.0
221.80	11.0
222.60	20.0
223.40	16.0
224.00	19.0
225.00	0.0
225.70	11.0
227.30	7.0
228.90	8.0
230.30	10.0
231.50	5.0
232.00	6.0
233.30	9.0
233.70	7.0
235.20	8.0
235.80	2.0
237.00	19.0
238.30	6.0
239.80	10.0
241.40	6.0
242.10	7.0
243.00	7.0
243.70	7.0
244.20	17.0
244.90	21.0
246.20	9.0
247.40	11.0
248.10	17.0
248.80	21.0
249.60	6.0
250.60	12.0
251.10	21.0
252.60	24.0
254.00	8.0
255.60	8.0
256.10	8.0
257.10	8.0
258.00	0.0
258.80	18.0
260.00	12.0
261.80	8.0
263.00	12.0
263.60	6.0
264.90	83.0
265.80	1.0
266.80	85.0
268.20	18.0
268.90	3.0
269.80	9.0
271.40	6.0
272.90	7.0
274.50	12.0
276.10	5.0
277.60	8.0
279.10	11.0
281.00	528.0
282.20	151.0
282.90	138.0
284.00	33.0
285.00	16.0
285.80	11.0
287.10	6.0
288.60	7.0
290.30	9.0
290.90	9.0
291.80	7.0
292.60	3.0
293.40	11.0
293.80	10.0
295.10	21.0
295.80	15.0
296.60	19.0
298.00	6.0
299.40	9.0
301.20	7.0
302.00	0.0
302.70	8.0
304.30	8.0
305.20	2.0
306.00	7.0
307.20	9.0
307.60	7.0
308.60	9.0
309.40	13.0
310.50	15.0
311.10	12.0
312.10	12.0
312.60	10.0
313.80	7.0
315.50	15.0
316.90	10.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

318.50	22.0
320.50	11.0
321.70	7.0
323.30	7.0
324.40	13.0
325.30	10.0
326.00	22.0
326.70	12.0
327.10	15.0
328.00	7.0
329.10	7.0
329.80	8.0
330.60	4.0
331.50	12.0
332.80	6.0
334.20	21.0
335.90	11.0
337.50	5.0
338.70	13.0
340.90	92.0
341.70	0.0
342.20	16.0
343.70	8.0
345.30	6.0
346.80	6.0
348.60	11.0
350.00	7.0
351.70	9.0
353.20	7.0
355.00	109.0
357.00	23.0
357.70	15.0
359.50	8.0
361.10	6.0
362.70	6.0
364.20	9.0
364.70	10.0
365.70	7.0
366.10	5.0
366.70	8.0
367.40	8.0
368.90	7.0
370.00	23.0
370.70	14.0
372.10	7.0
373.70	9.0
374.80	7.0
376.80	7.0
377.40	2.0
378.30	10.0
380.00	8.0
381.60	15.0
383.00	8.0
384.60	8.0
386.30	6.0
387.80	6.0
389.40	7.0
391.00	6.0
392.60	8.0
393.70	21.0
394.70	0.0
395.60	7.0
397.30	7.0
398.80	8.0
399.30	12.0
400.10	15.0
400.70	10.0
401.50	11.0
402.10	23.0
402.70	14.0
403.50	9.0
405.10	17.0
406.30	16.0
406.70	10.0
408.20	8.0
409.80	6.0
411.50	7.0
413.10	11.0
414.50	6.0
415.10	18.0
415.90	10.0
416.40	4.0
417.30	10.0
419.10	22.0
419.90	0.0
420.80	8.0
422.50	6.0
424.00	8.0
424.70	7.0
425.60	6.0
427.10	7.0
428.90	23.0
429.80	12.0
430.50	9.0
431.80	8.0
432.50	3.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

433.40	8.0
434.80	12.0
435.40	7.0
436.30	12.0
438.10	7.0
439.90	11.0
440.60	4.0
441.30	9.0
443.00	7.0
444.40	8.0
446.00	9.0
447.60	6.0
449.10	7.0
450.80	5.0
451.40	8.0
452.30	7.0
453.90	14.0
455.50	8.0
456.80	6.0
457.20	8.0
458.70	8.0
460.00	6.0
461.70	7.0
463.40	7.0
464.90	7.0
466.50	8.0
467.20	11.0
467.80	11.0
468.10	6.0
469.80	16.0
471.10	9.0
472.80	7.0
474.10	12.0
474.60	7.0
475.40	19.0
476.00	7.0
476.90	8.0
477.50	8.0
479.10	8.0
480.70	7.0
482.20	10.0
483.90	7.0
484.70	0.0
485.30	15.0
487.00	8.0
488.50	6.0
489.20	16.0
490.00	9.0
490.90	4.0
491.70	7.0
493.20	6.0
494.80	9.0
496.40	6.0
497.90	7.0
499.60	8.0
501.10	6.0
501.90	7.0
502.90	10.0
504.20	10.0
504.90	10.0
505.90	5.0
506.60	17.0
507.40	9.0
509.10	7.0
510.50	10.0
512.20	7.0
513.70	8.0
515.20	7.0
516.80	8.0
518.40	5.0
518.90	4.0
520.10	13.0
521.60	7.0
523.10	7.0
524.00	6.0
524.70	6.0
526.20	7.0
527.30	0.0
527.90	8.0
529.50	8.0
531.40	21.0
532.70	9.0
534.20	8.0
535.30	13.0
537.20	7.0
538.80	7.0
540.50	8.0
541.70	13.0
543.20	17.0
545.00	11.0
545.60	20.0
546.70	8.0
548.30	8.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

m/z	Abundance
35.90	477.0
37.00	1047.0
38.00	3276.0
39.10	18672.0
40.00	2252.0
41.00	1388.0
42.00	298.0
43.10	634.0
44.00	8187.0
45.00	628.0
46.00	143.0
47.10	123.0
48.00	259.0
49.00	2034.0
50.00	62952.0
51.00	253760.0
52.00	13412.0
53.00	742.0
54.00	141.0
55.00	1671.0
56.00	7663.0
57.00	17576.0
58.00	710.0
59.00	365.0
60.10	260.0
61.00	3722.0
62.00	3574.0
63.00	9979.0
64.00	1475.0
65.00	5177.0
65.90	428.0
66.90	499.0
68.00	4568.0
69.00	250176.0
70.10	1342.0
70.90	180.0
71.80	148.0
73.00	2520.0
74.00	24816.0
75.00	41536.0
76.10	13836.0
77.00	292096.0
78.00	19752.0
79.00	18904.0
80.00	14360.0
81.00	19744.0
82.00	5185.0
83.00	5104.0
84.00	1235.0
85.00	3590.0
86.00	5802.0
87.00	2766.0
87.90	731.0
88.90	571.0
91.00	4769.0
92.00	5067.0
93.00	32160.0
93.90	2553.0
94.90	578.0
96.00	1841.0
97.00	881.0
98.00	25088.0
99.00	21264.0
100.00	1763.0
101.00	11529.0
102.00	545.0
103.00	3612.0
104.00	7298.0
105.00	7252.0
106.00	2044.0
107.00	82240.0
108.00	12530.0
109.00	2264.0
110.00	152512.0
111.00	24160.0
112.00	3378.0
113.00	1012.0
113.80	250.0
114.90	568.0
116.00	5050.0
117.00	69936.0
118.00	5808.0
119.00	992.0
120.00	1499.0
121.00	366.0
122.00	5916.0
123.00	8842.0
124.00	3933.0
125.00	3282.0
127.00	297408.0
128.00	23024.0
129.00	109912.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

130.00	9362.0
131.00	2057.0
131.90	1203.0
133.00	802.0
133.90	3310.0
135.00	9432.0
135.90	3866.0
137.00	5480.0
137.90	1128.0
139.00	536.0
140.00	1548.0
141.00	14235.0
142.00	4963.0
143.00	3167.0
144.00	742.0
145.00	831.0
146.00	2898.0
147.00	7405.0
148.00	17784.0
149.00	3965.0
150.00	1096.0
151.10	2356.0
151.70	870.0
152.10	813.0
153.00	4454.0
154.00	3393.0
155.00	7929.0
156.00	10750.0
157.10	2398.0
157.90	2884.0
158.90	2098.0
160.00	4438.0
161.00	7663.0
161.90	1924.0
162.90	712.0
164.10	921.0
164.90	5154.0
166.00	3832.0
167.00	29192.0
168.00	18640.0
169.00	2790.0
169.90	820.0
171.00	1413.0
172.00	2140.0
173.00	3079.0
174.00	5770.0
175.00	10310.0
175.90	3619.0
177.00	5980.0
178.00	1663.0
179.00	20584.0
180.00	14808.0
181.00	6421.0
182.00	1054.0
183.00	767.0
184.00	1469.0
185.00	10209.0
186.00	75400.0
187.00	21744.0
188.10	2819.0
189.00	4818.0
190.00	660.0
191.00	2869.0
192.00	6886.0
193.00	6476.0
194.00	2066.0
195.00	1214.0
196.00	15896.0
198.00	526912.0
199.00	36088.0
200.00	3028.0
201.50	2315.0
203.00	4145.0
204.00	18808.0
205.00	33264.0
206.00	139392.0
207.00	22880.0
208.00	5484.0
209.00	2202.0
210.10	2573.0
211.00	5248.0
212.90	474.0
213.80	174.0
214.90	1500.0
216.00	3036.0
217.00	35272.0
218.00	4425.0
219.00	454.0
219.90	415.0
221.00	21040.0
221.90	4717.0
223.00	7657.0
224.00	74424.0
225.00	19648.0
226.00	2321.0
227.00	31920.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

228.00	4698.0
229.00	6844.0
229.90	913.0
231.00	2795.0
231.80	425.0
233.00	488.0
233.90	1912.0
235.00	2181.0
236.00	1249.0
237.00	3615.0
238.10	496.0
238.90	1164.0
240.00	1147.0
241.00	1669.0
242.00	3904.0
243.00	4055.0
244.00	56296.0
245.00	8169.0
246.00	12261.0
247.00	2557.0
248.00	438.0
248.90	2363.0
250.00	426.0
251.00	647.0
251.90	605.0
253.10	986.0
255.00	298688.0
256.00	45408.0
257.00	4358.0
258.00	17032.0
259.00	2999.0
260.00	603.0
260.90	541.0
262.10	71.0
262.80	157.0
263.90	462.0
265.00	6677.0
266.00	985.0
266.90	352.0
267.80	82.0
268.90	52.0
269.90	294.0
271.00	588.0
273.00	8717.0
274.00	22256.0
275.00	131136.0
276.00	18448.0
277.00	10453.0
277.90	1995.0
279.00	425.0
279.90	151.0
281.00	849.0
281.90	456.0
283.10	1570.0
283.90	616.0
285.00	1962.0
285.90	277.0
287.30	10.0
287.90	132.0
289.00	445.0
289.90	593.0
291.00	193.0
292.10	496.0
293.00	2644.0
294.00	755.0
296.00	37024.0
297.00	4375.0
297.90	247.0
299.10	48.0
299.80	14.0
301.10	428.0
301.80	746.0
303.00	4490.0
304.00	1108.0
305.00	234.0
305.90	27.0
306.80	54.0
308.00	520.0
308.90	624.0
310.00	608.0
311.00	184.0
312.00	150.0
313.00	275.0
314.00	2189.0
315.00	3985.0
316.00	2094.0
316.90	492.0
317.80	7.0
319.00	67.0
320.10	125.0
321.00	1354.0
322.00	673.0
323.00	11901.0
324.00	2285.0
324.90	264.0
326.10	238.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

326.90	2344.0
327.90	1087.0
328.90	258.0
329.90	132.0
330.50	4.0
332.00	828.0
333.00	1205.0
334.00	7840.0
335.00	1860.0
336.00	289.0
337.00	72.0
337.50	16.0
337.90	9.0
338.70	106.0
339.80	88.0
341.00	1444.0
341.80	317.0
342.10	268.0
342.80	126.0
343.80	7.0
344.90	71.0
345.90	2617.0
347.00	638.0
347.90	56.0
348.90	6.0
350.10	132.0
350.80	234.0
352.00	2952.0
352.90	2476.0
354.00	3161.0
355.00	876.0
355.90	128.0
356.90	28.0
358.00	124.0
358.70	211.0
359.20	308.0
359.90	89.0
361.20	49.0
361.90	10.0
362.50	1.0
363.00	57.0
364.90	16073.0
365.90	2496.0
366.90	337.0
368.10	15.0
370.10	395.0
371.00	849.0
372.00	6653.0
373.00	1285.0
373.90	184.0
375.50	27.0
376.90	163.0
378.00	132.0
378.70	7.0
379.60	5.0
380.20	8.0
381.90	10.0
383.00	1551.0
383.90	390.0
385.00	134.0
386.00	67.0
386.70	29.0
387.80	22.0
388.50	41.0
389.10	29.0
390.00	820.0
390.90	480.0
391.90	391.0
393.00	71.0
393.70	23.0
394.30	11.0
395.00	37.0
395.60	7.0
396.00	9.0
396.70	55.0
397.70	24.0
399.40	15.0
400.90	317.0
401.90	2627.0
403.00	3529.0
404.00	1185.0
405.10	292.0
405.70	8.0
406.20	0.0
407.10	13.0
408.10	21.0
408.70	13.0
410.00	80.0
411.20	11.0
411.60	9.0
413.20	5.0
413.80	8.0
415.10	172.0
415.90	16.0
416.60	8.0
416.80	8.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

418.00	7.0
418.80	3.0
419.20	14.0
419.80	49.0
421.00	2599.0
422.00	3310.0
423.00	21120.0
424.00	6938.0
424.90	931.0
425.90	87.0
427.30	18.0
428.90	20.0
429.40	13.0
430.00	8.0
430.50	8.0
431.20	19.0
432.20	9.0
433.70	8.0
435.30	10.0
436.70	18.0
438.00	57.0
438.80	151.0
441.00	59848.0
442.00	384320.0
443.00	77360.0
444.00	7286.0
445.00	466.0
445.90	19.0
447.80	7.0
448.40	17.0
449.50	8.0
450.30	7.0
451.00	7.0
452.70	6.0
454.20	11.0
455.70	7.0
457.40	7.0
458.10	13.0
458.90	6.0
460.30	14.0
461.70	16.0
462.10	14.0
463.60	9.0
464.60	18.0
465.30	10.0
466.70	8.0
468.40	8.0
470.10	11.0
471.40	7.0
473.10	7.0
474.60	7.0
476.10	11.0
476.60	22.0
477.80	6.0
479.50	8.0
481.10	8.0
481.70	8.0
482.50	8.0
483.40	13.0
484.00	6.0
485.70	8.0
487.60	15.0
489.00	14.0
490.40	7.0
491.90	7.0
493.50	7.0
495.30	15.0
496.70	7.0
498.30	6.0
499.80	7.0
501.20	15.0
502.90	22.0
504.20	9.0
504.60	7.0
506.10	7.0
507.60	18.0
509.30	7.0
511.00	6.0
512.40	9.0
514.00	7.0
515.50	7.0
516.30	6.0
517.10	6.0
518.30	22.0
518.80	6.0
519.80	14.0
520.40	7.0
521.80	6.0
523.50	11.0
524.80	11.0
526.60	9.0
528.10	7.0
529.00	0.0
529.70	7.0
531.30	7.0
532.90	7.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

534.40	6.0
535.10	15.0
536.00	9.0
536.70	10.0
537.60	7.0
539.20	6.0
540.80	9.0
542.30	7.0
543.90	8.0
545.60	7.0
547.10	6.0
547.90	2.0
548.50	15.0
549.00	14.0
549.70	5.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

m/z	Abundance
36.00	359.0
37.10	1322.0
38.00	3202.0
39.00	19632.0
40.00	2512.0
41.10	1214.0
42.00	343.0
43.00	539.0
44.00	8357.0
45.00	634.0
46.00	123.0
46.80	356.0
47.90	199.0
49.00	2544.0
50.00	70592.0
51.00	275392.0
52.00	13823.0
53.00	529.0
54.00	213.0
55.00	2046.0
56.00	7817.0
57.00	19488.0
58.00	878.0
59.00	342.0
60.10	213.0
61.00	3910.0
62.00	4320.0
63.00	10714.0
64.00	1759.0
65.00	4830.0
66.00	550.0
67.10	743.0
68.00	4868.0
69.00	288320.0
70.00	1632.0
70.90	334.0
73.00	2857.0
74.00	29848.0
75.00	45832.0
76.00	16064.0
77.00	329344.0
78.00	23592.0
79.00	22984.0
80.00	15670.0
81.00	22776.0
82.00	5707.0
83.00	5393.0
84.00	1090.0
85.00	3739.0
86.00	6748.0
87.00	3287.0
88.00	957.0
88.80	514.0
90.00	84.0
91.00	5541.0
92.00	5587.0
93.00	37040.0
94.00	2975.0
95.00	811.0
96.00	2296.0
97.10	717.0
98.00	27736.0
99.00	24584.0
100.00	2207.0
101.00	12834.0
102.00	770.0
103.00	3958.0
104.00	8125.0
105.00	7576.0
106.00	2200.0
107.00	98120.0
108.00	15497.0
109.00	3478.0
110.00	184512.0
111.00	28000.0
112.00	3552.0
113.00	957.0
113.90	327.0
114.90	573.0
116.00	5671.0
117.00	85808.0
118.00	6313.0
119.00	1010.0
120.00	1317.0
120.90	610.0
122.00	7137.0
123.00	11074.0
124.00	5210.0
124.90	3831.0
127.00	362176.0
128.00	28408.0
129.00	137024.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

130.00	12756.0
131.00	2494.0
131.90	1157.0
132.90	921.0
134.00	3906.0
135.00	11218.0
136.00	4138.0
137.00	6841.0
137.90	1108.0
139.00	678.0
140.00	1434.0
141.00	18824.0
142.00	6259.0
143.00	3892.0
144.00	1051.0
145.00	801.0
146.00	2594.0
147.00	9197.0
148.00	21368.0
149.00	3838.0
150.00	1496.0
151.10	2216.0
151.80	1482.0
153.00	5446.0
154.00	3938.0
155.00	9310.0
156.00	15128.0
157.10	3005.0
158.00	3356.0
159.00	2158.0
160.00	5286.0
161.00	8043.0
161.90	2322.0
163.00	834.0
164.10	1020.0
164.90	7211.0
166.00	5082.0
167.00	35688.0
168.00	21352.0
168.90	2772.0
169.90	1000.0
170.90	1608.0
172.00	3300.0
173.00	4205.0
174.00	7593.0
175.00	13514.0
176.00	4245.0
177.00	6676.0
177.90	2774.0
178.90	26976.0
180.00	18016.0
181.00	7308.0
182.10	1263.0
183.00	776.0
184.00	2547.0
185.00	11214.0
186.00	97032.0
187.00	29328.0
188.10	3138.0
189.00	5747.0
190.00	1185.0
191.00	3529.0
192.00	8335.0
193.00	8846.0
194.00	2111.0
195.10	1430.0
196.00	20272.0
198.00	680960.0
199.00	46200.0
200.00	3769.0
201.30	3075.0
203.00	4730.0
204.00	24488.0
205.00	42448.0
206.00	176704.0
207.00	27544.0
208.00	6611.0
209.00	2538.0
211.00	7089.0
213.00	466.0
213.80	352.0
215.00	2478.0
216.00	3758.0
217.00	46840.0
218.00	5680.0
218.90	687.0
221.00	27672.0
223.00	10984.0
224.00	102256.0
225.00	24120.0
226.10	2954.0
227.00	40840.0
228.00	5327.0
229.00	8347.0
230.00	928.0
231.00	3506.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

232.00	670.0
232.90	1038.0
234.00	2664.0
235.00	2893.0
235.90	1750.0
237.00	3765.0
238.00	686.0
239.00	1689.0
240.00	1478.0
240.90	2163.0
242.00	5385.0
243.10	6399.0
244.00	80384.0
245.00	9940.0
246.00	15456.0
247.00	2970.0
247.90	687.0
249.00	3061.0
250.00	676.0
251.10	478.0
252.00	833.0
253.00	1529.0
255.00	417344.0
256.00	58096.0
257.00	4526.0
258.00	24736.0
259.00	3377.0
259.90	704.0
260.90	673.0
261.90	199.0
263.00	311.0
264.00	637.0
265.00	10225.0
266.00	1520.0
267.00	280.0
267.80	84.0
269.10	191.0
269.90	372.0
270.90	869.0
272.00	872.0
273.00	12622.0
274.00	32568.0
275.00	184832.0
276.00	25080.0
277.00	16384.0
277.90	2601.0
279.10	507.0
280.10	149.0
281.00	803.0
282.00	582.0
283.00	2059.0
284.00	1130.0
285.00	2794.0
285.80	341.0
286.90	9.0
288.10	266.0
289.00	704.0
290.00	617.0
291.00	478.0
291.90	882.0
293.00	3415.0
294.00	926.0
294.90	995.0
296.00	52272.0
297.00	7134.0
297.90	571.0
299.10	69.0
300.00	30.0
300.80	432.0
302.00	958.0
303.00	6664.0
303.90	1598.0
305.00	412.0
306.10	64.0
306.80	51.0
308.00	1030.0
308.90	501.0
310.00	519.0
310.90	171.0
311.80	325.0
313.10	628.0
314.00	2453.0
315.00	5867.0
316.00	2789.0
317.00	691.0
317.70	19.0
318.90	87.0
321.00	1758.0
322.10	1035.0
323.00	17088.0
324.00	2968.0
324.90	356.0
325.80	323.0
326.90	3645.0
328.00	1510.0
328.90	409.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

329.90	89.0
330.90	61.0
332.00	1446.0
332.90	1484.0
334.00	11348.0
335.00	2658.0
336.00	324.0
337.00	60.0
338.10	22.0
338.80	384.0
339.90	304.0
341.00	2037.0
341.90	628.0
342.80	70.0
343.30	67.0
343.90	12.0
344.90	94.0
345.90	4087.0
347.00	692.0
347.60	50.0
348.00	39.0
349.30	27.0
350.00	94.0
351.00	450.0
352.00	5935.0
353.00	3215.0
354.00	5674.0
355.00	1288.0
355.70	98.0
356.10	114.0
356.90	35.0
358.00	109.0
358.70	371.0
359.80	72.0
360.30	21.0
361.20	57.0
362.00	22.0
363.10	47.0
365.00	26336.0
366.00	3947.0
366.90	186.0
368.30	7.0
368.90	5.0
370.00	638.0
371.00	1433.0
372.00	9946.0
373.00	2805.0
374.10	238.0
375.00	40.0
376.10	17.0
377.00	243.0
377.70	74.0
379.10	13.0
381.00	9.0
381.70	3.0
383.00	2865.0
383.90	769.0
384.90	360.0
386.10	87.0
386.90	20.0
387.40	15.0
388.00	9.0
388.50	29.0
389.00	92.0
390.00	1376.0
391.00	863.0
392.00	849.0
393.00	67.0
394.00	10.0
395.00	48.0
395.90	1.0
396.80	165.0
398.00	10.0
398.80	13.0
399.40	8.0
399.90	10.0
401.00	884.0
402.00	4202.0
403.00	6566.0
403.90	1931.0
404.90	534.0
405.80	12.0
406.20	21.0
407.10	16.0
407.60	14.0
408.50	2.0
409.00	15.0
409.80	103.0
410.60	5.0
411.00	31.0
411.90	18.0
413.30	27.0
414.00	19.0
415.00	280.0
416.10	44.0
417.00	5.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

418.40	6.0
418.80	13.0
419.80	60.0
421.00	4932.0
422.00	4797.0
423.00	35144.0
424.00	9571.0
424.90	1113.0
426.00	150.0
426.90	47.0
428.40	41.0
429.40	36.0
430.10	14.0
431.20	6.0
432.80	7.0
433.80	24.0
436.00	11.0
438.00	41.0
438.70	130.0
439.20	100.0
441.00	101864.0
442.00	647936.0
443.00	128304.0
444.00	11950.0
445.00	858.0
446.10	21.0
446.90	8.0
448.50	8.0
450.00	17.0
451.60	5.0
453.20	10.0
454.00	0.0
454.80	8.0
455.50	16.0
456.30	16.0
458.00	7.0
458.60	8.0
459.50	6.0
461.00	8.0
461.60	11.0
462.60	30.0
464.40	14.0
465.80	9.0
467.40	7.0
468.90	8.0
469.80	0.0
470.50	8.0
472.10	9.0
473.00	17.0
473.70	6.0
474.30	12.0
475.00	13.0
477.00	22.0
478.40	6.0
479.60	8.0
480.00	10.0
481.30	4.0
481.60	5.0
483.20	7.0
484.70	7.0
486.20	8.0
487.70	7.0
488.10	9.0
488.80	12.0
490.80	11.0
492.60	7.0
494.10	8.0
494.80	0.0
495.70	8.0
497.00	14.0
498.80	9.0
499.90	9.0
500.30	7.0
501.50	6.0
501.90	7.0
503.00	12.0
503.40	7.0
504.00	17.0
504.50	10.0
505.10	5.0
505.80	16.0
506.60	5.0
506.80	5.0
508.20	19.0
509.70	7.0
511.50	7.0
513.10	6.0
514.80	8.0
516.10	7.0
517.20	15.0
519.30	8.0
519.90	15.0
520.80	7.0
522.40	7.0
523.90	6.0
525.50	8.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

527.10	7.0
528.30	17.0
528.80	6.0
530.10	6.0
530.50	5.0
531.90	8.0
533.50	8.0
535.00	16.0
536.30	28.0
538.10	8.0
538.70	15.0
540.00	12.0
541.30	7.0
542.20	1.0
542.90	8.0
544.40	7.0
545.50	17.0
546.00	9.0
546.60	8.0
547.60	8.0
548.60	15.0
549.50	14.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

m/z	Abundance
36.00	466.0
37.00	1054.0
38.00	2688.0
39.00	13139.0
40.00	2041.0
41.00	1019.0
42.10	292.0
43.00	693.0
44.00	8421.0
45.00	578.0
45.80	59.0
47.00	163.0
47.90	238.0
49.00	1855.0
50.00	48248.0
51.00	186496.0
52.00	9506.0
53.10	424.0
54.00	181.0
55.00	1519.0
56.00	5576.0
57.00	12440.0
57.90	515.0
58.90	226.0
60.00	135.0
61.00	2393.0
62.00	2703.0
63.00	7779.0
64.00	889.0
65.00	4129.0
65.90	493.0
67.00	751.0
68.00	3484.0
69.00	190272.0
70.00	1145.0
71.00	393.0
72.10	188.0
73.00	2054.0
74.00	19312.0
75.00	30992.0
76.10	10749.0
77.00	225344.0
78.00	15094.0
79.00	14332.0
80.00	11477.0
81.00	15536.0
81.90	4031.0
83.00	3903.0
83.90	932.0
85.00	3180.0
86.00	4687.0
87.00	2140.0
87.90	795.0
89.00	411.0
89.90	148.0
91.00	3326.0
92.00	3820.0
93.00	26544.0
94.00	2218.0
94.90	596.0
96.10	1208.0
97.00	714.0
98.00	21616.0
99.00	16568.0
100.00	1435.0
101.00	8425.0
102.10	530.0
103.00	3091.0
104.00	5788.0
105.00	5708.0
106.00	1808.0
107.00	68144.0
108.00	10116.0
109.10	1843.0
110.00	127568.0
111.00	19576.0
112.00	2436.0
112.90	639.0
114.00	149.0
114.90	650.0
116.00	4041.0
117.00	63176.0
118.00	4704.0
118.90	613.0
120.00	1020.0
120.80	235.0
121.90	5137.0
123.00	7798.0
123.90	3311.0
125.00	2996.0
127.00	256192.0
128.00	18968.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

129.00	97968.0
130.00	7222.0
131.00	1867.0
132.00	951.0
133.00	813.0
133.90	2863.0
135.00	8093.0
136.00	2784.0
137.00	5152.0
137.90	711.0
138.90	747.0
140.00	1650.0
141.00	13492.0
142.00	4022.0
142.90	2924.0
144.00	1012.0
144.90	675.0
146.00	2203.0
147.00	7471.0
148.00	16640.0
149.00	3416.0
150.00	902.0
151.00	1412.0
152.00	1390.0
152.90	3817.0
153.90	2980.0
154.90	7081.0
156.00	10457.0
157.10	2252.0
158.00	2323.0
159.00	1715.0
159.90	4083.0
161.00	6380.0
161.90	1764.0
163.00	763.0
164.00	635.0
165.00	4609.0
166.00	4155.0
167.00	25176.0
168.00	15261.0
169.00	2323.0
169.90	973.0
170.90	1037.0
172.00	2313.0
173.00	3288.0
174.00	5374.0
175.00	9921.0
176.10	2827.0
177.00	5287.0
177.90	1502.0
179.00	19376.0
180.00	14312.0
181.00	6224.0
182.00	1312.0
182.90	545.0
183.90	1812.0
185.00	9554.0
186.00	72696.0
187.00	20848.0
188.00	2196.0
188.90	4940.0
189.90	682.0
191.00	3008.0
192.00	7165.0
193.00	7414.0
194.00	1521.0
195.00	888.0
196.00	14302.0
197.00	3422.0
198.00	523904.0
199.00	33696.0
200.00	2327.0
201.30	2172.0
203.00	3749.0
204.00	19432.0
205.00	32248.0
206.00	138624.0
207.00	23304.0
208.00	5392.0
209.00	2372.0
210.10	2231.0
211.00	4922.0
213.00	594.0
213.80	106.0
214.10	97.0
214.90	1735.0
216.00	2700.0
217.00	37400.0
218.00	4700.0
219.00	421.0
221.00	20568.0
221.80	4956.0
223.00	8423.0
224.00	80272.0
225.00	19608.0
226.10	2117.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

227.00	31384.0
228.00	4225.0
229.00	6245.0
230.00	818.0
231.00	2843.0
231.90	514.0
233.00	692.0
234.00	2165.0
235.00	2611.0
236.00	1556.0
237.00	3254.0
237.80	375.0
238.90	1357.0
240.00	1507.0
241.00	1731.0
242.00	4213.0
243.00	4049.0
244.00	63064.0
245.00	9099.0
246.00	12583.0
246.90	2669.0
247.90	562.0
249.00	2081.0
249.90	510.0
250.80	523.0
251.90	525.0
253.00	1070.0
255.00	334784.0
256.00	46272.0
257.00	3304.0
258.00	18784.0
259.00	2470.0
259.90	658.0
261.00	523.0
261.90	123.0
262.90	246.0
263.90	404.0
265.00	7346.0
265.90	1039.0
267.00	193.0
267.90	194.0
269.10	78.0
270.00	462.0
271.00	873.0
272.00	902.0
273.00	9984.0
274.00	28216.0
275.00	154560.0
276.00	20472.0
277.00	13119.0
277.90	1912.0
279.00	474.0
280.90	874.0
282.00	485.0
283.00	1770.0
284.00	1024.0
285.00	2114.0
286.10	359.0
287.10	16.0
287.80	97.0
288.20	109.0
288.90	654.0
289.90	521.0
291.00	281.0
292.00	577.0
293.00	3110.0
294.00	615.0
296.00	46096.0
297.00	6502.0
298.00	308.0
299.00	95.0
300.00	36.0
301.00	566.0
302.00	948.0
303.00	5556.0
304.00	1372.0
305.00	199.0
306.10	18.0
306.80	25.0
308.00	406.0
308.90	380.0
309.90	739.0
310.90	144.0
312.10	284.0
312.80	395.0
314.00	2362.0
315.00	5017.0
316.00	3094.0
316.80	528.0
317.80	10.0
318.90	161.0
320.00	224.0
321.00	1754.0
322.10	884.0
323.00	14624.0
323.90	2675.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

325.00	219.0
325.90	203.0
326.90	2407.0
327.90	1405.0
329.00	203.0
329.70	52.0
330.30	12.0
332.00	1048.0
332.90	1340.0
334.00	9894.0
335.00	2433.0
335.90	339.0
337.10	43.0
338.00	8.0
338.80	111.0
341.00	1790.0
342.00	473.0
342.80	55.0
343.30	56.0
343.90	45.0
345.00	136.0
346.00	3341.0
346.90	552.0
348.10	30.0
348.90	11.0
349.90	246.0
350.80	286.0
352.00	4400.0
353.00	3294.0
354.00	5206.0
355.00	976.0
356.00	182.0
356.80	0.0
357.30	7.0
358.10	140.0
358.90	303.0
359.80	160.0
360.40	16.0
361.00	18.0
361.60	15.0
362.30	12.0
365.00	22080.0
365.90	3419.0
366.80	272.0
367.80	14.0
370.00	720.0
371.10	1283.0
372.00	9080.0
373.00	2396.0
373.90	213.0
374.00	210.0
375.00	48.0
376.10	11.0
376.70	263.0
378.00	65.0
379.10	15.0
380.00	7.0
381.50	16.0
383.00	2511.0
384.00	725.0
384.80	206.0
386.10	65.0
386.90	9.0
387.50	5.0
387.80	5.0
388.50	10.0
389.00	118.0
389.90	1141.0
390.90	901.0
392.10	558.0
393.10	159.0
393.90	7.0
395.00	37.0
395.90	23.0
397.00	147.0
398.00	12.0
398.80	9.0
399.80	0.0
401.00	580.0
402.00	4320.0
403.00	5555.0
404.00	2211.0
405.00	423.0
405.90	17.0
406.50	8.0
406.80	10.0
408.40	20.0
408.90	1.0
409.80	274.0
411.10	28.0
411.90	5.0
412.80	7.0
413.90	1.0
415.00	289.0
416.10	61.0
416.90	9.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

417.60	7.0
418.70	5.0
419.90	86.0
421.00	4594.0
422.00	5132.0
423.00	31448.0
424.00	7480.0
425.00	692.0
426.20	21.0
427.00	38.0
427.90	3.0
428.40	43.0
429.40	8.0
430.10	7.0
431.80	7.0
432.60	5.0
433.60	15.0
434.90	10.0
435.90	16.0
436.40	10.0
437.80	19.0
438.90	227.0
441.00	96728.0
442.00	630720.0
443.00	116432.0
444.00	11070.0
445.10	559.0
446.00	6.0
447.40	8.0
449.10	10.0
450.60	8.0
452.30	8.0
453.80	7.0
455.00	8.0
455.30	8.0
456.90	13.0
457.80	0.0
458.60	18.0
459.50	1.0
460.00	8.0
460.60	12.0
461.70	9.0
463.20	8.0
464.80	11.0
466.40	6.0
467.80	8.0
469.90	13.0
471.00	5.0
472.70	7.0
474.00	8.0
474.20	7.0
475.50	18.0
476.00	18.0
477.40	8.0
478.90	7.0
480.60	21.0
482.10	5.0
483.60	7.0
485.30	8.0
486.80	7.0
488.40	21.0
490.10	14.0
490.60	12.0
491.50	28.0
493.00	10.0
494.60	7.0
496.30	5.0
497.80	7.0
499.40	7.0
500.40	9.0
500.90	10.0
502.60	8.0
504.30	16.0
505.80	16.0
507.30	7.0
508.10	0.0
508.80	7.0
510.40	7.0
512.00	6.0
512.60	5.0
513.60	7.0
515.20	6.0
516.80	11.0
518.30	8.0
519.90	7.0
521.40	6.0
521.90	16.0
523.10	7.0
524.60	7.0
526.20	8.0
527.80	15.0
529.30	8.0
530.90	6.0
531.40	8.0
532.40	7.0
534.00	8.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661

534.80	15.0
535.30	10.0
536.90	10.0
538.80	8.0
540.40	7.0
541.90	9.0
543.50	8.0
545.00	10.0
546.60	8.0
548.40	24.0
549.60	17.0

Instrument :
BNA_P
ClientSampleId :
DFTPP661