

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BP010823.D
 Acq On : 25 Jun 2022 12:13
 Operator : CG/JU
 Sample : PB145568BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS658

Quant Time: Jun 26 22:51:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	443740	20.000	ng/u1	0.00
20) Naphthalene-d8	10.516	136	1778557	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.375	164	915127	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.139	188	1605693	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.251	240	1518915	20.000	ng/u1	# 0.00
88) Perylene-d12	23.621	264	1720895	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	69931	5.961	ng/uL	0.00
4) Pyridine-d5	3.617	84	923102	29.563	ng/u1	0.01
7) Phenol-d5	6.905	99	1101782	31.366	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.063	67	687913	29.945	ng/u1	0.00
11) 2-Chlorophenol-d4	7.258	132	936752	32.143	ng/u1	0.00
15) 4-Methylphenol-d8	8.446	113	871969	31.402	ng/u1	0.01
21) Nitrobenzene-d5	8.887	128	429023	37.892	ng/u1	0.00
24) 2-Nitrophenol-d4	9.605	143	420203	44.859	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.146	165	756462	32.684	ng/u1	0.00
31) 4-Chloroaniline-d4	10.663	131	1046243	27.693	ng/u1	0.00
46) Dimethylphthalate-d6	13.798	166	1941004	30.964	ng/u1	0.00
49) Acenaphthylene-d8	14.069	160	2507797	31.345	ng/u1	0.00
54) 4-Nitrophenol-d4	14.598	143	327704	33.469	ng/u1	0.02
60) Fluorene-d10	15.375	176	1642736	30.537	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.498	200	193981	36.464	ng/u1	0.00
73) Anthracene-d10	17.239	188	2276830	31.871	ng/u1	0.00
81) Pyrene-d10	19.486	212	2497390	30.180	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.468	264	2798587	32.909	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	140708	11.248	ng/uL#	82
5) Pyridine	3.634	79	909435	28.751	ng/u1#	71
6) Benzaldehyde	6.869	77	741555	40.011	ng/u1	94
8) Phenol	6.934	94	1104813	29.407	ng/u1#	84
10) Bis(2-Chloroethyl)ether	7.158	93	894818	29.656	ng/u1	88
12) 2-Chlorophenol	7.287	128	913063	30.351	ng/u1	92
13) 2-Methylphenol	8.175	108	813293	29.078	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.263	45	1136486	28.072	ng/u1#	95
16) Acetophenone	8.552	105	1275594	29.912	ng/u1#	86
17) N-Nitroso-di-n-propyla...	8.540	70	606862	28.143	ng/u1	87
18) 4-Methylphenol	8.510	108	864434	29.678	ng/u1	100
19) Hexachloroethane	8.793	117	364370	29.908	ng/u1#	82
22) Nitrobenzene	8.928	77	936706	33.152	ng/u1#	88
23) Isophorone	9.457	82	1672665	28.229	ng/u1	97
25) 2-Nitrophenol	9.634	139	452113	39.857	ng/u1#	78
26) 2,4-Dimethylphenol	9.705	107	898018	28.976	ng/u1	95
27) Bis(2-Chloroethoxy)met...	9.946	93	1094375	29.278	ng/u1	98
29) 2,4-Dichlorophenol	10.175	162	729262	30.323	ng/u1	97
30) Naphthalene	10.569	128	2745121	29.614	ng/u1	99
32) 4-Chloroaniline	10.687	127	996873	26.632	ng/u1	99
33) Hexachlorobutadiene	10.852	225	450666	29.331	ng/u1	97
34) Caprolactam	11.475	113	213425	28.249	ng/u1	81
35) 4-Chloro-3-methylphenol	11.828	107	755396	28.980	ng/u1	94
36) 2-Methylnaphthalene	12.187	142	1690385	28.161	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BP010823.D
 Acq On : 25 Jun 2022 12:13
 Operator : CG/JU
 Sample : PB145568BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS658

Quant Time: Jun 26 22:51:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) 1,2,4,5-Tetrachloroben...	12.557	216	797839	30.339	ng/ul	98
40) Hexachlorocyclopentadiene	12.534	237	478640	28.723	ng/ul	97
41) 2,4,6-Trichlorophenol	12.804	196	502868	33.168	ng/ul	97
42) 2,4,5-Trichlorophenol	12.881	196	526240	32.344	ng/ul	100
43) 1,1'-Biphenyl	13.210	154	2122552	29.463	ng/ul	98
44) 2-Chloronaphthalene	13.246	162	1640200	30.223	ng/ul	98
45) 2-Nitroaniline	13.463	65	424368	36.856	ng/ul#	79
47) Dimethylphthalate	13.845	163	1878067	29.916	ng/ul	98
48) 2,6-Dinitrotoluene	13.963	165	359504	38.220	ng/ul	90
50) Acenaphthylene	14.098	152	2515288	28.706	ng/ul	98
51) 3-Nitroaniline	14.293	138	386074	34.236	ng/ul	88
52) Acenaphthene	14.440	153	1710223	30.250	ng/ul	99
53) 2,4-Dinitrophenol	14.493	184	100207	27.750	ng/ul#	79
55) 4-Nitrophenol	14.616	109	233002	29.928	ng/ul#	74
56) Dibenzofuran	14.781	168	2275319	29.381	ng/ul	96
57) 2,4-Dinitrotoluene	14.751	165	494194	39.320	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.010	232	382273	32.810	ng/ul	94
59) Diethylphthalate	15.222	149	1829465	29.312	ng/ul	99
61) Fluorene	15.434	166	1804042	29.215	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.434	204	847621	29.322	ng/ul	98
63) 4-Nitroaniline	15.463	138	414311	39.737	ng/ul#	74
66) 4,6-Dinitro-2-methylph...	15.510	198	195695	34.042	ng/ul	98
67) N-Nitrosodiphenylamine	15.651	169	1504894	30.944	ng/ul	98
68) 4-Bromophenyl-phenylether	16.334	248	504375	31.343	ng/ul	97
69) Hexachlorobenzene	16.434	284	572952	30.740	ng/ul	96
70) Atrazine	16.616	200	488500	30.296	ng/ul	97
71) Pentachlorophenol	16.792	266	196077	20.430	ng/ul	97
72) Phenanthrene	17.186	178	2623379	30.680	ng/ul	98
74) Anthracene	17.275	178	2628167	30.560	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.169	216	818302	30.883	ng/uL	96
76) Pentachlorobenzene	14.693	250	725109	31.908	ng/uL	98
77) Carbazole	17.557	167	2406997	31.093	ng/ul	99
78) Di-n-butylphthalate	18.122	149	2872397	31.138	ng/ul	99
80) Fluoranthene	19.163	202	2851001	28.013	ng/ul#	89
82) Pyrene	19.516	202	2956315	28.763	ng/ul#	84
83) Butylbenzylphthalate	20.416	149	1296660	35.842	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.174	252	1086913	34.654	ng/ul#	98
85) Benzo(a)anthracene	21.239	228	2971078	31.276	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.180	149	1891686	33.020	ng/ul#	96
87) Chrysene	21.286	228	2803578	30.798	ng/ul	99
89) Di-n-octyl phthalate	22.104	149	3321023	32.512	ng/ul	100
90) Benzo(b)fluoranthene	22.898	252	3297819	30.816	ng/ul#	96
91) Benzo(k)fluoranthene	22.951	252	3261192	32.060	ng/ul#	96
93) Benzo(a)pyrene	23.516	252	2944587	28.285	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.086	276	4089341	32.553	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.110	278	3407937	31.628	ng/ul#	93
96) Benzo(g,h,i)perylene	26.839	276	3346032	31.313	ng/ul#	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BP010823.D
 Acq On : 25 Jun 2022 12:13
 Operator : CG/JU
 Sample : PB145568BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS658

Quant Time: Jun 26 22:51:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 11.248 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion: 88 Resp: 140708
Ion Ratio Lower Upper
88 100
43 27.4 35.8 53.6#
58 72.5 68.1 102.1



#5
Pyridine
Concen: 28.751 ng/uL
RT: 3.634 min Scan# 93
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion: 79 Resp: 909435
Ion Ratio Lower Upper
79 100
52 73.3 83.8 125.8#
51 34.1 41.0 61.4#









#6
Benzaldehyde
Concen: 40.011 ng/ul
RT: 6.869 min Scan# 643
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion: 77 Resp: 741555
Ion Ratio Lower Upper
77 100
105 101.2 76.0 114.0
106 97.1 73.2 109.8



#8
Phenol
Concen: 29.407 ng/ul
RT: 6.934 min Scan# 654
Delta R.T. 0.018 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion: 94 Resp: 1104813
Ion Ratio Lower Upper
94 100
65 26.7 27.4 41.2#
66 34.4 37.4 56.2#









#10
 Bis(2-Chloroethyl)ether
 Concen: 29.656 ng/ul
 RT: 7.158 min Scan# 692
 Delta R.T. 0.000 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SLCS658


Tgt Ion:	93	Resp:	894818
Ion	Ratio	Lower	Upper
93	100		
63	75.6	72.4	108.6
95	32.2	27.3	40.9



#12
 2-Chlorophenol
 Concen: 30.351 ng/ul
 RT: 7.287 min Scan# 714
 Delta R.T. 0.000 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13

Tgt Ion:	128	Resp:	913063
Ion	Ratio	Lower	Upper
128	100		
64	46.3	44.8	67.2
130	32.2	25.8	38.6









#13
2-Methylphenol
Concen: 29.078 ng/ul
RT: 8.175 min Scan# 80
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:108 Resp: 813293
Ion Ratio Lower Upper
108 100
107 90.0 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 28.072 ng/ul
RT: 8.263 min Scan# 880
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion: 45 Resp: 1136486
Ion Ratio Lower Upper
45 100
77 15.0 11.2 16.8
79 11.7 6.9 10.3#









#16
Acetophenone
Concen: 29.912 ng/ul
RT: 8.552 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:105 Resp: 1275594
Ion Ratio Lower Upper
105 100
77 74.1 67.5 101.3
51 24.3 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 28.143 ng/ul
RT: 8.540 min Scan# 927
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion: 70 Resp: 606862
Ion Ratio Lower Upper
70 100
42 54.5 53.8 80.8
101 10.1 7.8 11.6
130 22.9 15.4 23.2









#18
4-Methylphenol
Concen: 29.678 ng/ul
RT: 8.510 min Scan# 910
Delta R.T. 0.012 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:108 Resp: 864434
Ion Ratio Lower Upper
108 100
107 110.1 88.3 132.5



#19
Hexachloroethane
Concen: 29.908 ng/ul
RT: 8.793 min Scan# 970
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:117 Resp: 364370
Ion Ratio Lower Upper
117 100
201 89.0 85.7 128.5
199 55.8 57.4 86.0#









#22
Nitrobenzene
Concen: 33.152 ng/ul
RT: 8.928 min Scan# 993
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion: 77 Resp: 936706
Ion Ratio Lower Upper
77 100
123 48.8 31.5 47.3#
65 13.0 11.4 17.2



#23
Isophorone
Concen: 28.229 ng/ul
RT: 9.457 min Scan# 1083
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion: 82 Resp: 1672665
Ion Ratio Lower Upper
82 100
95 7.0 6.6 10.0
138 18.5 13.8 20.6









#25
2-Nitrophenol
Concen: 39.857 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:139 Resp: 452113
Ion Ratio Lower Upper
139 100
65 47.9 54.2 81.4#
109 28.0 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 28.976 ng/ul
RT: 9.705 min Scan# 1125
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:107 Resp: 898018
Ion Ratio Lower Upper
107 100
121 51.7 39.1 58.7
122 87.3 66.1 99.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 29.278 ng/ul
 RT: 9.946 min Scan# 1166
 Delta R.T. 0.000 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SLCS658


Tgt Ion: 93 Resp: 1094375
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.6 39.8
 123 11.3 7.8 11.8



#29
 2,4-Dichlorophenol
 Concen: 30.323 ng/ul
 RT: 10.175 min Scan# 1205
 Delta R.T. 0.012 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13

Tgt Ion: 162 Resp: 729262
 Ion Ratio Lower Upper
 162 100
 164 64.1 52.5 78.7
 98 36.7 31.4 47.0









#30
Naphthalene
Concen: 29.614 ng/ul
RT: 10.569 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:128 Resp: 2745121
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 26.632 ng/ul
RT: 10.687 min Scan# 1292
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:127 Resp: 996873
Ion Ratio Lower Upper
127 100
129 32.4 26.2 39.2









#33
Hexachlorobutadiene
Concen: 29.331 ng/ul
RT: 10.852 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:225 Resp: 450666
Ion Ratio Lower Upper
225 100
223 63.1 53.7 80.5
227 62.8 49.4 74.2



#34
Caprolactam
Concen: 28.249 ng/ul
RT: 11.475 min Scan# 1426
Delta R.T. 0.029 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:113 Resp: 213425
Ion Ratio Lower Upper
113 100
55 162.2 153.8 230.8
56 115.9 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 28.980 ng/ul
RT: 11.828 min Scan# 1486
Delta R.T. 0.012 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:107 Resp: 755396
Ion Ratio Lower Upper
107 100
144 25.8 18.6 27.8
142 80.3 59.6 89.4



#36
2-Methylnaphthalene
Concen: 28.161 ng/ul
RT: 12.187 min Scan# 1547
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:142 Resp: 1690385
Ion Ratio Lower Upper
142 100
141 90.0 70.1 105.1









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 30.339 ng/ul
 RT: 12.557 min Scan# 1610
 Delta R.T. 0.000 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13

Instrument :
 BNA_P
 ClientSampleId :
 SLCS658

Tgt Ion:216 Resp: 797839




Ion	Ratio	Lower	Upper
216	100		
214	77.5	63.0	94.6
179	21.7	18.1	27.1
108	18.5	13.6	20.4




#40
 Hexachlorocyclopentadiene
 Concen: 28.723 ng/ul
 RT: 12.534 min Scan# 1606
 Delta R.T. -0.006 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13

Tgt Ion:237 Resp: 478640

Ion	Ratio	Lower	Upper
237	100		
235	63.4	52.8	79.2
272	12.9	11.5	17.3









#41
2,4,6-Trichlorophenol
Concen: 33.168 ng/ul
RT: 12.804 min Scan# 1652
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:196 Resp: 502868
Ion Ratio Lower Upper
196 100
198 94.2 78.2 117.4
200 30.8 25.0 37.6



#42
2,4,5-Trichlorophenol
Concen: 32.344 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. 0.012 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:196 Resp: 526240
Ion Ratio Lower Upper
196 100
198 96.2 77.0 115.4
200 31.6 24.8 37.2









#43
1,1'-Biphenyl
Concen: 29.463 ng/ul
RT: 13.210 min Scan# 1721
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:154 Resp: 2122552
Ion Ratio Lower Upper
154 100
153 40.4 31.9 47.9
76 13.8 9.2 13.8



#44
2-Chloronaphthalene
Concen: 30.223 ng/ul
RT: 13.246 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:162 Resp: 1640200
Ion Ratio Lower Upper
162 100
164 32.6 25.5 38.3
127 41.1 31.9 47.9









#45
2-Nitroaniline
Concen: 36.856 ng/ul
RT: 13.463 min Scan# 1763
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion: 65 Resp: 424368
Ion Ratio Lower Upper
65 100
92 69.2 49.0 73.6
138 109.7 66.4 99.6#



#47
Dimethylphthalate
Concen: 29.916 ng/ul
RT: 13.845 min Scan# 1829
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:163 Resp: 1878067
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.3 7.6 11.4









#48
2,6-Dinitrotoluene
Concen: 38.220 ng/ul
RT: 13.963 min Scan# 1849
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:165 Resp: 359504
Ion Ratio Lower Upper
165 100
89 56.3 52.9 79.3
121 22.1 19.4 29.2



#50
Acenaphthylene
Concen: 28.706 ng/ul
RT: 14.098 min Scan# 1872
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:152 Resp: 2515288
Ion Ratio Lower Upper
152 100
151 19.5 16.5 24.7
153 12.9 10.1 15.1









#51
3-Nitroaniline
Concen: 34.236 ng/ul
RT: 14.293 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:138 Resp: 386074
Ion Ratio Lower Upper
138 100
108 13.0 11.0 16.4
92 116.0 104.6 156.8



#52
Acenaphthene
Concen: 30.250 ng/ul
RT: 14.440 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:153 Resp: 1710223
Ion Ratio Lower Upper
153 100
152 46.8 36.2 54.2
154 88.5 70.8 106.2









#53
2,4-Dinitrophenol
Concen: 27.750 ng/ul
RT: 14.493 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:184 Resp: 100207
Ion Ratio Lower Upper
184 100
63 68.8 73.8 110.6#
154 58.5 57.2 85.8



#55
4-Nitrophenol
Concen: 29.928 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. 0.024 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:109 Resp: 233002
Ion Ratio Lower Upper
109 100
139 152.5 75.6 113.4#
65 136.7 107.6 161.4









#56
Dibenzenofuran
Concen: 29.381 ng/ul
RT: 14.781 min Scan# 1988
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:168 Resp: 2275319
Ion Ratio Lower Upper
168 100
139 38.3 32.8 49.2



#57
2,4-Dinitrotoluene
Concen: 39.320 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:165 Resp: 494194
Ion Ratio Lower Upper
165 100
63 42.0 47.0 70.6#
182 2.8 2.2 3.2









#58
2,3,4,6-Tetrachlorophenol
Concen: 32.810 ng/ul
RT: 15.010 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:232	Resp:	382273
Ion	Ratio	Lower Upper
232	100	
131	48.6	35.8 53.8
130	2.6	1.8 2.8
166	32.4	23.7 35.5



#59
Diethylphthalate
Concen: 29.312 ng/ul
RT: 15.222 min Scan# 2063
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:149	Resp:	1829465
Ion	Ratio	Lower Upper
149	100	
177	21.6	17.6 26.4
150	12.4	10.3 15.5









#61
Fluorene
Concen: 29.215 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:166 Resp: 1804042
Ion Ratio Lower Upper
166 100
165 98.0 77.4 116.2
167 13.2 10.8 16.2



#62
4-Chlorophenyl-phenylether
Concen: 29.322 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:204 Resp: 847621
Ion Ratio Lower Upper
204 100
206 32.6 26.0 39.0
141 63.9 49.3 73.9









#63
4-Nitroaniline
Concen: 39.737 ng/ul
RT: 15.463 min Scan# 2104
Delta R.T. 0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:138 Resp: 414311
Ion Ratio Lower Upper
138 100
92 49.9 46.2 69.2
108 71.9 85.8 128.8#



#66
4,6-Dinitro-2-methylphenol
Concen: 34.042 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:198 Resp: 195695
Ion Ratio Lower Upper
198 100
182 4.6 4.3 6.5
77 30.7 23.7 35.5









#67
N-Nitrosodiphenylamine
Concen: 30.944 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:169 Resp: 1504894
Ion Ratio Lower Upper
169 100
168 66.8 52.6 78.8
167 35.9 27.8 41.6



#68
4-Bromophenyl-phenylether
Concen: 31.343 ng/ul
RT: 16.334 min Scan# 2252
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:248 Resp: 504375
Ion Ratio Lower Upper
248 100
250 97.0 80.8 121.2
141 78.7 63.9 95.9









#69
Hexachlorobenzene
Concen: 30.740 ng/ul
RT: 16.434 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:284 Resp: 572952
Ion Ratio Lower Upper
284 100
142 38.1 27.8 41.6
249 30.9 25.4 38.2



#70
Atrazine
Concen: 30.296 ng/ul
RT: 16.616 min Scan# 2300
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:200 Resp: 488500
Ion Ratio Lower Upper
200 100
173 28.0 20.7 31.1
215 47.0 38.6 57.8





#71

Pentachlorophenol

Concen: 20.430 ng/ul

RT: 16.792 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP010823.D

Acq: 25 Jun 2022 12:13

Instrument :

BNA_P

ClientSampleId :

SLCS658





Tgt Ion:266 Resp: 196077

Ion Ratio Lower Upper

266 100

264 62.6 51.4 77.2

268 63.1 53.5 80.3



#72

Phenanthrene

Concen: 30.680 ng/ul

RT: 17.186 min Scan# 2397

Delta R.T. 0.000 min

Lab File: BP010823.D

Acq: 25 Jun 2022 12:13




Tgt Ion:178 Resp: 2623379


Ion Ratio Lower Upper

178 100

179 15.0 12.7 19.1

176 18.7 14.6 22.0









#74
 Anthracene
 Concen: 30.560 ng/u1
 RT: 17.275 min Scan# 2413
 Delta R.T. -0.006 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SLCS658


Tgt Ion:178 Resp: 2628167
 Ion Ratio Lower Upper
 178 100
 179 15.3 12.2 18.2
 176 18.6 14.5 21.7



#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 30.883 ng/uL
 RT: 13.169 min Scan# 1714
 Delta R.T. 0.000 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13

Tgt Ion:216 Resp: 818302
 Ion Ratio Lower Upper
 216 100
 214 77.0 65.0 97.4
 218 47.0 39.1 58.7









#76
 Pentachlorobenzene
 Concen: 31.908 ng/uL
 RT: 14.693 min Scan# 1973
 Delta R.T. -0.006 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SLCS658


Tgt Ion:250 Resp: 725109
 Ion Ratio Lower Upper
 250 100
 248 61.0 49.8 74.6
 252 62.8 51.2 76.8



#77
 Carbazole
 Concen: 31.093 ng/uL
 RT: 17.557 min Scan# 2460
 Delta R.T. 0.006 min
 Lab File: BP010823.D
 Acq: 25 Jun 2022 12:13

Tgt Ion:167 Resp: 2406997
 Ion Ratio Lower Upper
 167 100
 166 21.7 16.9 25.3
 139 13.1 10.6 16.0









#78
Di-n-butylphthalate
Concen: 31.138 ng/ul
RT: 18.122 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:149 Resp: 2872397
Ion Ratio Lower Upper
149 100
150 9.0 7.1 10.7
104 4.9 4.6 6.8



#80
Fluoranthene
Concen: 28.013 ng/ul
RT: 19.163 min Scan# 2733
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:202 Resp: 2851001
Ion Ratio Lower Upper
202 100
101 13.5 7.1 10.7#
100 10.1 5.7 8.5#





#82

Pyrene

Concen: 28.763 ng/ul

RT: 19.516 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP010823.D

Acq: 25 Jun 2022 12:13

Instrument :

BNA_P

ClientSampleId :

SLCS658





Tgt Ion:202 Resp: 2956315

Ion Ratio Lower Upper

202 100

101 16.3 7.9 11.9#

100 13.6 6.7 10.1#



#83

Butylbenzylphthalate

Concen: 35.842 ng/ul

RT: 20.416 min Scan# 2946

Delta R.T. -0.006 min

Lab File: BP010823.D

Acq: 25 Jun 2022 12:13




Tgt Ion:149 Resp: 1296660


Ion Ratio Lower Upper

149 100

91 71.0 60.0 90.0

206 19.9 17.4 26.2









#84
3,3'-Dichlorobenzidine
Concen: 34.654 ng/ul
RT: 21.174 min Scan# 3075
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:252 Resp: 1086913
Ion Ratio Lower Upper
252 100
254 64.3 51.0 76.4
126 14.0 8.6 12.8#



#85
Benzo(a)anthracene
Concen: 31.276 ng/ul
RT: 21.239 min Scan# 3086
Delta R.T. 0.000 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:228 Resp: 2971078
Ion Ratio Lower Upper
228 100
229 19.7 15.0 22.4
226 26.7 20.6 31.0









#86
Bis(2-ethylhexyl)phthalate
Concen: 33.020 ng/ul
RT: 21.180 min Scan# 3079
Delta R.T. -0.018 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:149 Resp: 1891686
Ion Ratio Lower Upper
149 100
167 26.0 22.6 33.8
279 3.7 4.2 6.4#



#87
Chrysene
Concen: 30.798 ng/ul
RT: 21.286 min Scan# 3094
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:228 Resp: 2803578
Ion Ratio Lower Upper
228 100
226 29.8 23.6 35.4
229 19.8 15.6 23.4





#89

Di-n-octyl phthalate

Concen: 32.512 ng/ul

RT: 22.104 min Scan# 31

Delta R.T. -0.018 min

Lab File: BP010823.D

Acq: 25 Jun 2022 12:13





Instrument :

BNA_P

ClientSampleId :

SLCS658

Tgt Ion:149 Resp: 3321023



#90

Benzo(b)fluoranthene

Concen: 30.816 ng/ul

RT: 22.898 min Scan# 3368

Delta R.T. -0.006 min

Lab File: BP010823.D

Acq: 25 Jun 2022 12:13




Tgt Ion:252 Resp: 3297819


Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.6

125 13.0 6.2 9.2#





#91

Benzo(k)fluoranthene

Concen: 32.060 ng/u1

RT: 22.951 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP010823.D

Acq: 25 Jun 2022 12:13

Instrument :

BNA_P

ClientSampleId :

SLCS658





Tgt Ion:252 Resp: 3261192

Ion Ratio Lower Upper

252 100

253 21.7 17.0 25.4

125 12.1 6.2 9.4#



#93

Benzo(a)pyrene

Concen: 28.285 ng/u1

RT: 23.516 min Scan# 3473

Delta R.T. -0.006 min

Lab File: BP010823.D

Acq: 25 Jun 2022 12:13




Tgt Ion:252 Resp: 2944587


Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 14.1 6.6 10.0#









#94
Indeno(1,2,3-cd)pyrene
Concen: 32.553 ng/ul
RT: 26.086 min Scan# 3910
Delta R.T. -0.006 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13




Instrument :
BNA_P
ClientSampleId :
SLCS658


Tgt Ion:276 Resp: 4089341
Ion Ratio Lower Upper
276 100
138 26.8 14.6 22.0#
277 24.8 19.6 29.4



#95
Dibenzo(a,h)anthracene
Concen: 31.628 ng/ul
RT: 26.110 min Scan# 3914
Delta R.T. -0.012 min
Lab File: BP010823.D
Acq: 25 Jun 2022 12:13

Tgt Ion:278 Resp: 3407937
Ion Ratio Lower Upper
278 100
139 19.3 9.4 14.2#
279 24.3 18.9 28.3





#96

Benzo(g,h,i)perylene

Concen: 31.313 ng/ul

RT: 26.839 min Scan# 4038

Delta R.T. -0.006 min

Lab File: BP010823.D


Acq: 25 Jun 2022 12:13

Instrument :

BNA_P

ClientSampleId :

SLCS658



Tgt Ion:276 Resp: 3346032

Ion Ratio Lower Upper

276 100

138 26.2 12.7 19.1#

277 23.6 18.2 27.4

