

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BP010851.D
 Acq On : 26 Jun 2022 04:34
 Operator : CG/JU
 Sample : PB145767BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS767

Quant Time: Jun 26 22:56:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	387608	20.000	ng/ul	0.00
20) Naphthalene-d8	10.510	136	1552599	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.369	164	851059	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.134	188	1692250	20.000	ng/ul	#-0.01
79) Chrysene-d12	21.245	240	1659315	20.000	ng/ul	#-0.01
88) Perylene-d12	23.610	264	1750072	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	72685	7.094	ng/uL	0.00
4) Pyridine-d5	3.611	84	970823	35.593	ng/ul	0.00
7) Phenol-d5	6.893	99	1148537	37.432	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.058	67	706454	35.205	ng/ul	0.00
11) 2-Chlorophenol-d4	7.252	132	959628	37.696	ng/ul	0.00
15) 4-Methylphenol-d8	8.434	113	903948	37.268	ng/ul	0.00
21) Nitrobenzene-d5	8.875	128	437270	44.241	ng/ul	0.00
24) 2-Nitrophenol-d4	9.599	143	425504	52.036	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.134	165	791051	39.152	ng/ul	0.00
31) 4-Chloroaniline-d4	10.652	131	1129789	34.256	ng/ul	0.00
46) Dimethylphthalate-d6	13.793	166	2171398	37.247	ng/ul	0.00
49) Acenaphthylene-d8	14.063	160	2723677	36.606	ng/ul	0.00
54) 4-Nitrophenol-d4	14.587	143	443193	48.672	ng/ul	0.00
60) Fluorene-d10	15.369	176	1869204	37.363	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.492	200	257802	45.982	ng/ul	0.00
73) Anthracene-d10	17.234	188	2828152	37.563	ng/ul	-0.01
81) Pyrene-d10	19.486	212	3281665	36.302	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.463	264	3455939	39.962	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	142021	12.997	ng/uL#	85
5) Pyridine	3.628	79	909933	32.933	ng/ul#	75
6) Benzaldehyde	6.864	77	742029	45.834	ng/ul	95
8) Phenol	6.922	94	1117367	34.048	ng/ul#	84
10) Bis(2-Chloroethyl)ether	7.152	93	886364	33.629	ng/ul	88
12) 2-Chlorophenol	7.281	128	914202	34.790	ng/ul	90
13) 2-Methylphenol	8.163	108	816873	33.435	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.252	45	1142621	32.311	ng/ul#	97
16) Acetophenone	8.540	105	1253829	33.660	ng/ul#	87
17) N-Nitroso-di-n-propyla...	8.534	70	596911	31.690	ng/ul	88
18) 4-Methylphenol	8.499	108	876437	34.447	ng/ul	99
19) Hexachloroethane	8.787	117	356226	33.474	ng/ul#	82
22) Nitrobenzene	8.916	77	940086	38.114	ng/ul	91
23) Isophorone	9.446	82	1654519	31.987	ng/ul	97
25) 2-Nitrophenol	9.628	139	444349	44.873	ng/ul#	79
26) 2,4-Dimethylphenol	9.699	107	871761	32.223	ng/ul	94
27) Bis(2-Chloroethoxy)met...	9.934	93	1088011	33.344	ng/ul	98
29) 2,4-Dichlorophenol	10.163	162	743878	35.432	ng/ul	98
30) Naphthalene	10.563	128	2746278	33.938	ng/ul	99
32) 4-Chloroaniline	10.681	127	1044591	31.968	ng/ul	99
33) Hexachlorobutadiene	10.846	225	440355	32.831	ng/ul	96
34) Caprolactam	11.463	113	246291	37.343	ng/ul	87
35) 4-Chloro-3-methylphenol	11.816	107	807327	35.480	ng/ul	94
36) 2-Methylnaphthalene	12.181	142	1722606	32.874	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BP010851.D
 Acq On : 26 Jun 2022 04:34
 Operator : CG/JU
 Sample : PB145767BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS767

Quant Time: Jun 26 22:56:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.399	142	1721576	33.071	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.546	216	803059	32.836	ng/ul	99
40) Hexachlorocyclopentadiene	12.528	237	456137	29.434	ng/ul	96
41) 2,4,6-Trichlorophenol	12.799	196	523774	37.147	ng/ul	99
42) 2,4,5-Trichlorophenol	12.869	196	573000	37.870	ng/ul	99
43) 1,1'-Biphenyl	13.204	154	2174722	32.459	ng/ul	98
44) 2-Chloronaphthalene	13.240	162	1696224	33.608	ng/ul	99
45) 2-Nitroaniline	13.451	65	480786	44.899	ng/ul#	80
47) Dimethylphthalate	13.840	163	2036533	34.882	ng/ul	99
48) 2,6-Dinitrotoluene	13.957	165	397154	45.401	ng/ul	90
50) Acenaphthylene	14.093	152	2658489	32.625	ng/ul	99
51) 3-Nitroaniline	14.287	138	463327	44.180	ng/ul	87
52) Acenaphthene	14.434	153	1808645	34.399	ng/ul	99
53) 2,4-Dinitrophenol	14.487	184	141478	42.128	ng/ul#	82
55) 4-Nitrophenol	14.604	109	304915	42.113	ng/ul#	76
56) Dibenzofuran	14.775	168	2450104	34.019	ng/ul	95
57) 2,4-Dinitrotoluene	14.745	165	583403	49.912	ng/ul#	78
58) 2,3,4,6-Tetrachlorophenol	14.998	232	448168	41.361	ng/ul	90
59) Diethylphthalate	15.216	149	2064244	35.564	ng/ul	99
61) Fluorene	15.428	166	1983403	34.537	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.428	204	913434	33.978	ng/ul	99
63) 4-Nitroaniline	15.457	138	521257	53.758	ng/ul#	75
66) 4,6-Dinitro-2-methylph...	15.504	198	252880	41.740	ng/ul	97
67) N-Nitrosodiphenylamine	15.640	169	1698988	33.148	ng/ul	98
68) 4-Bromophenyl-phenylether	16.328	248	562328	33.157	ng/ul	98
69) Hexachlorobenzene	16.428	284	651823	33.182	ng/ul	97
70) Atrazine	16.610	200	592914	34.891	ng/ul	98
71) Pentachlorophenol	16.781	266	334261	33.047	ng/ul	98
72) Phenanthrene	17.175	178	3188304	35.379	ng/ul	99
74) Anthracene	17.269	178	3165524	34.925	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.157	216	833061	29.832	ng/uL	97
76) Pentachlorobenzene	14.687	250	760825	31.768	ng/uL	99
77) Carbazole	17.545	167	3051701	37.405	ng/ul	99
78) Di-n-butylphthalate	18.116	149	3521918	36.226	ng/ul	99
80) Fluoranthene	19.157	202	3678214	33.083	ng/ul#	88
82) Pyrene	19.510	202	3772560	33.599	ng/ul#	83
83) Butylbenzylphthalate	20.404	149	1618374	40.949	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.175	252	1329465	38.801	ng/ul	99
85) Benzo(a)anthracene	21.233	228	3758782	36.220	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.175	149	2377835	37.994	ng/ul#	97
87) Chrysene	21.286	228	3546301	35.661	ng/ul	99
89) Di-n-octyl phthalate	22.092	149	4069319	39.174	ng/ul	100
90) Benzo(b)fluoranthene	22.892	252	4022357	36.959	ng/ul#	96
91) Benzo(k)fluoranthene	22.939	252	3875627	37.465	ng/ul#	95
93) Benzo(a)pyrene	23.510	252	3520907	33.257	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.074	276	4770015	37.339	ng/ul#	89
95) Dibenzo(a,h)anthracene	26.098	278	3989055	36.403	ng/ul#	93
96) Benzo(g,h,i)perylene	26.827	276	3880354	35.708	ng/ul#	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BP010851.D
 Acq On : 26 Jun 2022 04:34
 Operator : CG/JU
 Sample : PB145767BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS767

Quant Time: Jun 26 22:56:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 12.997 ng/uL
RT: 3.240 min Scan# 20
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion: 88 Resp: 142021
Ion Ratio Lower Upper
88 100
43 28.3 35.8 53.6#
58 76.3 68.1 102.1



#5
Pyridine
Concen: 32.933 ng/uL
RT: 3.628 min Scan# 92
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion: 79 Resp: 909933
Ion Ratio Lower Upper
79 100
52 77.5 83.8 125.8#
51 35.3 41.0 61.4#









#6
Benzaldehyde
Concen: 45.834 ng/ul
RT: 6.864 min Scan# 643
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion: 77 Resp: 742029
Ion Ratio Lower Upper
77 100
105 99.3 76.0 114.0
106 96.5 73.2 109.8



#8
Phenol
Concen: 34.048 ng/ul
RT: 6.922 min Scan# 652
Delta R.T. 0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion: 94 Resp: 1117367
Ion Ratio Lower Upper
94 100
65 26.6 27.4 41.2#
66 35.1 37.4 56.2#









#10
 Bis(2-Chloroethyl)ether
 Concen: 33.629 ng/ul
 RT: 7.152 min Scan# 691
 Delta R.T. -0.006 min
 Lab File: BP010851.D
 Acq: 26 Jun 2022 04:34

Instrument :
 BNA_P
 ClientSampleId :
 SLCS767

Tgt Ion: 93 Resp: 886364




Ion	Ratio	Lower	Upper
93	100		
63	76.2	72.4	108.6
95	32.3	27.3	40.9




#12
 2-Chlorophenol
 Concen: 34.790 ng/ul
 RT: 7.281 min Scan# 713
 Delta R.T. -0.006 min
 Lab File: BP010851.D
 Acq: 26 Jun 2022 04:34

Tgt Ion:128 Resp: 914202

Ion	Ratio	Lower	Upper
128	100		
64	45.1	44.8	67.2
130	32.3	25.8	38.6





#13

2-Methylphenol

Concen: 33.435 ng/ul

RT: 8.163 min Scan# 8

Delta R.T. -0.006 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34

Instrument :

BNA_P

ClientSampleId :





SLCS767

Tgt Ion:108 Resp: 816873

Ion Ratio Lower Upper

108 100

107 89.8 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 32.311 ng/ul

RT: 8.252 min Scan# 878

Delta R.T. -0.012 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34




Tgt Ion: 45 Resp: 1142621


Ion Ratio Lower Upper

45 100

77 14.3 11.2 16.8

79 11.1 6.9 10.3#









#16
Acetophenone
Concen: 33.660 ng/ul
RT: 8.540 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:105 Resp: 1253829
Ion Ratio Lower Upper
105 100
77 74.6 67.5 101.3
51 25.6 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 31.690 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion: 70 Resp: 596911
Ion Ratio Lower Upper
70 100
42 55.6 53.8 80.8
101 10.4 7.8 11.6
130 22.6 15.4 23.2









#18
4-Methylphenol
Concen: 34.447 ng/ul
RT: 8.499 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:108 Resp: 876437
Ion Ratio Lower Upper
108 100
107 111.3 88.3 132.5



#19
Hexachloroethane
Concen: 33.474 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:117 Resp: 356226
Ion Ratio Lower Upper
117 100
201 88.5 85.7 128.5
199 56.0 57.4 86.0#









#22
Nitrobenzene
Concen: 38.114 ng/ul
RT: 8.916 min Scan# 991
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion: 77 Resp: 940086
Ion Ratio Lower Upper
77 100
123 46.2 31.5 47.3
65 13.3 11.4 17.2



#23
Isophorone
Concen: 31.987 ng/ul
RT: 9.446 min Scan# 1081
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion: 82 Resp: 1654519
Ion Ratio Lower Upper
82 100
95 6.9 6.6 10.0
138 18.0 13.8 20.6









#25
2-Nitrophenol
Concen: 44.873 ng/ul
RT: 9.628 min Scan# 1112
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:139 Resp: 444349
Ion Ratio Lower Upper
139 100
65 48.9 54.2 81.4#
109 28.2 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 32.223 ng/ul
RT: 9.699 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:107 Resp: 871761
Ion Ratio Lower Upper
107 100
121 52.9 39.1 58.7
122 88.1 66.1 99.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 33.344 ng/ul
 RT: 9.934 min Scan# 1166
 Delta R.T. -0.012 min
 Lab File: BP010851.D
 Acq: 26 Jun 2022 04:34

Instrument :
 BNA_P
 ClientSampleId :
 SLCS767

Tgt Ion: 93 Resp: 1088011




Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.6	39.8
123	11.1	7.8	11.8




#29
 2,4-Dichlorophenol
 Concen: 35.432 ng/ul
 RT: 10.163 min Scan# 1203
 Delta R.T. 0.000 min
 Lab File: BP010851.D
 Acq: 26 Jun 2022 04:34

Tgt Ion: 162 Resp: 743878

Ion	Ratio	Lower	Upper
162	100		
164	65.0	52.5	78.7
98	37.5	31.4	47.0









#30
Naphthalene
Concen: 33.938 ng/ul
RT: 10.563 min Scan# 1272
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:128 Resp: 2746278
Ion Ratio Lower Upper
128 100
129 10.6 8.7 13.1
127 13.0 10.2 15.4



#32
4-Chloroaniline
Concen: 31.968 ng/ul
RT: 10.681 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:127 Resp: 1044591
Ion Ratio Lower Upper
127 100
129 32.3 26.2 39.2









#33
Hexachlorobutadiene
Concen: 32.831 ng/ul
RT: 10.846 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:225 Resp: 440355
Ion Ratio Lower Upper
225 100
223 62.9 53.7 80.5
227 64.1 49.4 74.2



#34
Caprolactam
Concen: 37.343 ng/ul
RT: 11.463 min Scan# 1424
Delta R.T. 0.018 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:113 Resp: 246291
Ion Ratio Lower Upper
113 100
55 169.3 153.8 230.8
56 125.7 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 35.480 ng/ul
RT: 11.816 min Scan# 1484
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:107 Resp: 807327
Ion Ratio Lower Upper
107 100
144 26.1 18.6 27.8
142 80.0 59.6 89.4



#36
2-Methylnaphthalene
Concen: 32.874 ng/ul
RT: 12.181 min Scan# 1546
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:142 Resp: 1722606
Ion Ratio Lower Upper
142 100
141 88.6 70.1 105.1









#37
1-Methylnaphthalene
Concen: 33.071 ng/ul
RT: 12.399 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:142 Resp: 1721576
Ion Ratio Lower Upper
142 100
141 92.9 73.3 109.9
116 3.6 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.836 ng/ul
RT: 12.546 min Scan# 1608
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:216 Resp: 803059
Ion Ratio Lower Upper
216 100
214 78.4 63.0 94.6
179 22.3 18.1 27.1
108 19.2 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 29.434 ng/ul
RT: 12.528 min Scan# 1605
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:237 Resp: 456137
Ion Ratio Lower Upper
237 100
235 62.7 52.8 79.2
272 12.9 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 37.147 ng/ul
RT: 12.799 min Scan# 1651
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:196 Resp: 523774
Ion Ratio Lower Upper
196 100
198 97.2 78.2 117.4
200 31.5 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 37.870 ng/ul
RT: 12.869 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:196 Resp: 573000
Ion Ratio Lower Upper
196 100
198 95.5 77.0 115.4
200 30.9 24.8 37.2



#43
1,1'-Biphenyl
Concen: 32.459 ng/ul
RT: 13.204 min Scan# 1720
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:154 Resp: 2174722
Ion Ratio Lower Upper
154 100
153 41.0 31.9 47.9
76 13.4 9.2 13.8





#44

2-Chloronaphthalene

Concen: 33.608 ng/ul

RT: 13.240 min Scan# 1726

Delta R.T. -0.012 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34

Instrument :

BNA_P

ClientSampleId :

SLCS767





Tgt Ion:162 Resp: 1696224

Ion Ratio Lower Upper

162 100

164 32.3 25.5 38.3

127 41.1 31.9 47.9



#45

2-Nitroaniline

Concen: 44.899 ng/ul

RT: 13.451 min Scan# 1762

Delta R.T. -0.006 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34




Tgt Ion: 65 Resp: 480786


Ion Ratio Lower Upper

65 100

92 70.2 49.0 73.6

138 107.0 66.4 99.6#









#47
Dimethylphthalate
Concen: 34.882 ng/ul
RT: 13.840 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:163 Resp: 2036533
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.2
164 9.9 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 45.401 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:165 Resp: 397154
Ion Ratio Lower Upper
165 100
89 56.8 52.9 79.3
121 22.2 19.4 29.2









#50
Acenaphthylene
Concen: 32.625 ng/ul
RT: 14.093 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:152 Resp: 2658489
Ion Ratio Lower Upper
152 100
151 20.0 16.5 24.7
153 13.0 10.1 15.1



#51
3-Nitroaniline
Concen: 44.180 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:138 Resp: 463327
Ion Ratio Lower Upper
138 100
108 12.1 11.0 16.4
92 115.0 104.6 156.8









#52
Acenaphthene
Concen: 34.399 ng/ul
RT: 14.434 min Scan# 1929
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:153 Resp: 1808645
Ion Ratio Lower Upper
153 100
152 46.6 36.2 54.2
154 87.8 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 42.128 ng/ul
RT: 14.487 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:184 Resp: 141478
Ion Ratio Lower Upper
184 100
63 69.1 73.8 110.6#
154 63.9 57.2 85.8









#55
4-Nitrophenol
Concen: 42.113 ng/ul
RT: 14.604 min Scan# 1956
Delta R.T. 0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:109 Resp: 304915
Ion Ratio Lower Upper
109 100
139 146.5 75.6 113.4#
65 137.7 107.6 161.4



#56
Dibenzofuran
Concen: 34.019 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:168 Resp: 2450104
Ion Ratio Lower Upper
168 100
139 37.9 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 49.912 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:165 Resp: 583403
Ion Ratio Lower Upper
165 100
63 41.4 47.0 70.6#
182 3.0 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.361 ng/ul
RT: 14.998 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:232 Resp: 448168
Ion Ratio Lower Upper
232 100
131 52.6 35.8 53.8
130 2.7 1.8 2.8
166 33.8 23.7 35.5









#59
Diethylphthalate
Concen: 35.564 ng/ul
RT: 15.216 min Scan# 2062
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:149 Resp: 2064244
Ion Ratio Lower Upper
149 100
177 21.6 17.6 26.4
150 12.3 10.3 15.5



#61
Fluorene
Concen: 34.537 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:166 Resp: 1983403
Ion Ratio Lower Upper
166 100
165 98.7 77.4 116.2
167 13.6 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 33.978 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:204 Resp: 913434
Ion Ratio Lower Upper
204 100
206 32.5 26.0 39.0
141 63.3 49.3 73.9



#63
4-Nitroaniline
Concen: 53.758 ng/ul
RT: 15.457 min Scan# 2103
Delta R.T. 0.000 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:138 Resp: 521257
Ion Ratio Lower Upper
138 100
92 50.7 46.2 69.2
108 72.3 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 41.740 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:198 Resp: 252880
Ion Ratio Lower Upper
198 100
182 4.9 4.3 6.5
77 27.9 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 33.148 ng/ul
RT: 15.640 min Scan# 2134
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:169 Resp: 1698988
Ion Ratio Lower Upper
169 100
168 66.6 52.6 78.8
167 35.9 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 33.157 ng/ul
RT: 16.328 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:248 Resp: 562328
Ion Ratio Lower Upper
248 100
250 98.0 80.8 121.2
141 78.3 63.9 95.9



#69
Hexachlorobenzene
Concen: 33.182 ng/ul
RT: 16.428 min Scan# 2268
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:284 Resp: 651823
Ion Ratio Lower Upper
284 100
142 37.9 27.8 41.6
249 31.4 25.4 38.2









#70
Atrazine
Concen: 34.891 ng/ul
RT: 16.610 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:200 Resp: 592914
Ion Ratio Lower Upper
200 100
173 28.0 20.7 31.1
215 47.7 38.6 57.8



#71
Pentachlorophenol
Concen: 33.047 ng/ul
RT: 16.781 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:266 Resp: 334261
Ion Ratio Lower Upper
266 100
264 63.5 51.4 77.2
268 65.1 53.5 80.3









#72
Phenanthrene
Concen: 35.379 ng/ul
RT: 17.175 min Scan# 2111
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:178 Resp: 3188304
Ion Ratio Lower Upper
178 100
179 15.6 12.7 19.1
176 19.0 14.6 22.0



#74
Anthracene
Concen: 34.925 ng/ul
RT: 17.269 min Scan# 2411
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:178 Resp: 3165524
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 18.8 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 29.832 ng/uL

RT: 13.157 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34

Instrument :

BNA_P

ClientSampleId :

SLCS767





Tgt Ion:216 Resp: 833061

Ion Ratio Lower Upper

216 100

214 77.6 65.0 97.4

218 48.0 39.1 58.7



#76

Pentachlorobenzene

Concen: 31.768 ng/uL

RT: 14.687 min Scan# 1972

Delta R.T. -0.012 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34




Tgt Ion:250 Resp: 760825


Ion Ratio Lower Upper

250 100

248 63.1 49.8 74.6

252 63.7 51.2 76.8









#77
 Carbazole
 Concen: 37.405 ng/ul
 RT: 17.545 min Scan# 2459
 Delta R.T. -0.006 min
 Lab File: BP010851.D
 Acq: 26 Jun 2022 04:34

Instrument :
 BNA_P
 ClientSampleId :
 SLCS767

Tgt Ion:167 Resp: 3051701




Ion	Ratio	Lower	Upper
167	100		
166	21.5	16.9	25.3
139	13.2	10.6	16.0




#78
 Di-n-butylphthalate
 Concen: 36.226 ng/ul
 RT: 18.116 min Scan# 2555
 Delta R.T. -0.018 min
 Lab File: BP010851.D
 Acq: 26 Jun 2022 04:34

Tgt Ion:149 Resp: 3521918

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.1	10.7
104	4.9	4.6	6.8









#80
Fluoranthene
Concen: 33.083 ng/u1
RT: 19.157 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:202 Resp: 3678214
Ion Ratio Lower Upper
202 100
101 13.6 7.1 10.7#
100 10.6 5.7 8.5#



#82
Pyrene
Concen: 33.599 ng/u1
RT: 19.510 min Scan# 2792
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:202 Resp: 3772560
Ion Ratio Lower Upper
202 100
101 16.9 7.9 11.9#
100 13.9 6.7 10.1#









#83
 Butylbenzylphthalate
 Concen: 40.949 ng/ul
 RT: 20.404 min Scan# 2944
 Delta R.T. -0.018 min
 Lab File: BP010851.D
 Acq: 26 Jun 2022 04:34




Instrument :
 BNA_P
 ClientSampleId :
 SLCS767


Tgt Ion:149 Resp: 1618374
 Ion Ratio Lower Upper
 149 100
 91 74.1 60.0 90.0
 206 18.9 17.4 26.2



#84
 3,3'-Dichlorobenzidine
 Concen: 38.801 ng/ul
 RT: 21.175 min Scan# 3075
 Delta R.T. -0.006 min
 Lab File: BP010851.D
 Acq: 26 Jun 2022 04:34

Tgt Ion:252 Resp: 1329465
 Ion Ratio Lower Upper
 252 100
 254 63.9 51.0 76.4
 126 11.9 8.6 12.8





#85

Benzo(a)anthracene

Concen: 36.220 ng/ul

RT: 21.233 min Scan# 3086

Delta R.T. -0.006 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34

Instrument :

BNA_P

ClientSampleId :

SLCS767





Tgt Ion:228 Resp: 3758782

Ion Ratio Lower Upper

228 100

229 19.7 15.0 22.4

226 27.0 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 37.994 ng/ul

RT: 21.175 min Scan# 3075

Delta R.T. -0.024 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34




Tgt Ion:149 Resp: 2377835


Ion Ratio Lower Upper

149 100

167 26.8 22.6 33.8

279 3.8 4.2 6.4#





#87

Chrysene

Concen: 35.661 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.006 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34

Instrument :

BNA_P

ClientSampleId :

SLCS767





Tgt Ion:228 Resp: 3546301

Ion Ratio Lower Upper

228 100

226 29.9 23.6 35.4

229 19.8 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 39.174 ng/ul




RT: 22.092 min Scan# 3231


Delta R.T. -0.029 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34

Tgt Ion:149 Resp: 4069319









#90
Benzo(b)fluoranthene
Concen: 36.959 ng/ul
RT: 22.892 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:252 Resp: 4022357
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 12.7 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 37.465 ng/ul
RT: 22.939 min Scan# 3375
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:252 Resp: 3875627
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.4
125 12.9 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 33.257 ng/ul
RT: 23.510 min Scan# 3474
Delta R.T. -0.012 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34




Instrument :
BNA_P
ClientSampleId :
SLCS767


Tgt Ion:252 Resp: 3520907
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 14.1 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 37.339 ng/ul
RT: 26.074 min Scan# 3908
Delta R.T. -0.018 min
Lab File: BP010851.D
Acq: 26 Jun 2022 04:34

Tgt Ion:276 Resp: 4770015
Ion Ratio Lower Upper
276 100
138 29.0 14.6 22.0#
277 24.7 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 36.403 ng/ul

RT: 26.098 min Scan# 3916

Delta R.T. -0.024 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34

Instrument :

BNA_P

ClientSampleId :

SLCS767





Tgt Ion:278 Resp: 3989055

Ion Ratio Lower Upper

278 100

139 19.2 9.4 14.2#

279 24.0 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 35.708 ng/ul

RT: 26.827 min Scan# 4036

Delta R.T. -0.018 min

Lab File: BP010851.D

Acq: 26 Jun 2022 04:34

Tgt Ion:276 Resp: 3880354

Ion Ratio Lower Upper

276 100

138 26.0 12.7 19.1#

277 23.9 18.2 27.4

