

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BPCCC836.D
 Acq On : 25 Jun 2022 20:07
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jun 26 22:59:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	519811	20.000	ng/ul	0.00
20) Naphthalene-d8	10.516	136	2154638	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.375	164	1144992	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.139	188	2063625	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.245	240	2058518	20.000	ng/ul	#-0.01
88) Perylene-d12	23.615	264	2250302	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	103423	7.526	ng/uL	0.00
4) Pyridine-d5	3.616	84	671311	18.353	ng/ul	0.01
7) Phenol-d5	6.899	99	782440	19.015	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.058	67	503482	18.709	ng/ul	0.00
11) 2-Chlorophenol-d4	7.252	132	668673	19.587	ng/ul	0.00
15) 4-Methylphenol-d8	8.440	113	629361	19.348	ng/ul	0.00
21) Nitrobenzene-d5	8.881	128	302166	22.030	ng/ul	0.00
24) 2-Nitrophenol-d4	9.599	143	291270	25.667	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.140	165	547845	19.539	ng/ul	0.00
31) 4-Chloroaniline-d4	10.657	131	834503	18.233	ng/ul	0.00
46) Dimethylphthalate-d6	13.792	166	1416700	18.063	ng/ul	0.00
49) Acenaphthylene-d8	14.063	160	1808379	18.065	ng/ul	0.00
54) 4-Nitrophenol-d4	14.592	143	241849	19.742	ng/ul	0.01
60) Fluorene-d10	15.369	176	1207225	17.936	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.492	200	131900	19.292	ng/ul	0.00
73) Anthracene-d10	17.233	188	1704137	18.561	ng/ul	-0.01
81) Pyrene-d10	19.486	212	1892006	16.871	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.457	264	2079238	18.698	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	107928	7.365	ng/uL#	83
5) Pyridine	3.634	79	695729	18.776	ng/ul#	71
6) Benzaldehyde	6.863	77	419936	19.342	ng/ul	95
8) Phenol	6.922	94	838228	19.046	ng/ul#	86
10) Bis(2-Chloroethyl)ether	7.152	93	669733	18.948	ng/ul	89
12) 2-Chlorophenol	7.287	128	686713	19.487	ng/ul#	89
13) 2-Methylphenol	8.169	108	619928	18.921	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.257	45	871462	18.376	ng/ul#	96
16) Acetophenone	8.546	105	966580	19.349	ng/ul#	86
17) N-Nitroso-di-n-propyla...	8.534	70	473759	18.755	ng/ul	87
18) 4-Methylphenol	8.499	108	663421	19.443	ng/ul	98
19) Hexachloroethane	8.793	117	265891	18.631	ng/ul	85
22) Nitrobenzene	8.922	77	691984	20.216	ng/ul	90
23) Isophorone	9.451	82	1265503	17.630	ng/ul	97
25) 2-Nitrophenol	9.628	139	332207	24.174	ng/ul#	82
26) 2,4-Dimethylphenol	9.699	107	697736	18.584	ng/ul	95
27) Bis(2-Chloroethoxy)met...	9.940	93	843599	18.630	ng/ul	98
29) 2,4-Dichlorophenol	10.163	162	561028	19.256	ng/ul	99
30) Naphthalene	10.563	128	2064746	18.386	ng/ul	100
32) 4-Chloroaniline	10.681	127	831685	18.340	ng/ul	100
33) Hexachlorobutadiene	10.851	225	342089	18.378	ng/ul	96
34) Caprolactam	11.463	113	156300	17.077	ng/ul	86
35) 4-Chloro-3-methylphenol	11.822	107	582370	18.442	ng/ul	91
36) 2-Methylnaphthalene	12.181	142	1315510	18.091	ng/ul	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BPCCC836.D
 Acq On : 25 Jun 2022 20:07
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jun 26 22:59:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.404	142	1317527	18.237	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.551	216	623712	18.956	ng/ul	98
40) Hexachlorocyclopentadiene	12.534	237	386741	18.549	ng/ul	96
41) 2,4,6-Trichlorophenol	12.798	196	384268	20.257	ng/ul	97
42) 2,4,5-Trichlorophenol	12.869	196	404263	19.859	ng/ul	99
43) 1,1'-Biphenyl	13.204	154	1669524	18.522	ng/ul#	98
44) 2-Chloronaphthalene	13.240	162	1254094	18.469	ng/ul	97
45) 2-Nitroaniline	13.457	65	318621	22.117	ng/ul#	78
47) Dimethylphthalate	13.839	163	1422850	18.115	ng/ul	98
48) 2,6-Dinitrotoluene	13.957	165	261461	22.216	ng/ul	91
50) Acenaphthylene	14.092	152	1994466	18.193	ng/ul	99
51) 3-Nitroaniline	14.287	138	282225	20.003	ng/ul	87
52) Acenaphthene	14.439	153	1289725	18.232	ng/ul	99
53) 2,4-Dinitrophenol	14.487	184	94732	20.967	ng/ul#	83
55) 4-Nitrophenol	14.604	109	186808	19.177	ng/ul#	75
56) Dibenzofuran	14.775	168	1768250	18.249	ng/ul	95
57) 2,4-Dinitrotoluene	14.745	165	356276	22.656	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.004	232	285769	19.603	ng/ul#	93
59) Diethylphthalate	15.216	149	1393276	17.842	ng/ul	98
61) Fluorene	15.428	166	1387484	17.958	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.428	204	654005	18.082	ng/ul	98
63) 4-Nitroaniline	15.451	138	266556	20.433	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.510	198	138862	18.796	ng/ul	94
67) N-Nitrosodiphenylamine	15.645	169	1165270	18.643	ng/ul	99
68) 4-Bromophenyl-phenylether	16.328	248	375554	18.159	ng/ul	98
69) Hexachlorobenzene	16.428	284	435939	18.199	ng/ul	95
70) Atrazine	16.610	200	363716	17.551	ng/ul	98
71) Pentachlorophenol	16.786	266	268253	21.748	ng/ul	97
72) Phenanthrene	17.180	178	2043933	18.599	ng/ul	99
74) Anthracene	17.269	178	2052878	18.573	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.163	216	629508	18.486	ng/ul	100
76) Pentachlorobenzene	14.687	250	557110	19.075	ng/ul	100
77) Carbazole	17.551	167	1888179	18.978	ng/ul	99
78) Di-n-butylphthalate	18.116	149	2150878	18.142	ng/ul	99
80) Fluoranthene	19.157	202	2288165	16.589	ng/ul#	88
82) Pyrene	19.516	202	2367193	16.994	ng/ul#	88
83) Butylbenzylphthalate	20.410	149	970336	19.791	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.169	252	784321	18.452	ng/ul#	98
85) Benzo(a)anthracene	21.233	228	2365446	18.373	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.174	149	1427894	18.391	ng/ul#	97
87) Chrysene	21.280	228	2293950	18.594	ng/ul	100
89) Di-n-octyl phthalate	22.092	149	2499129	18.710	ng/ul	100
90) Benzo(b)fluoranthene	22.892	252	2602197	18.595	ng/ul#	97
91) Benzo(k)fluoranthene	22.939	252	2419099	18.186	ng/ul#	95
93) Benzo(a)pyrene	23.510	252	2544493	18.691	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	26.068	276	3024427	18.412	ng/ul#	90
95) Dibenzo(a,h)anthracene	26.092	278	2602400	18.470	ng/ul#	93
96) Benzo(g,h,i)perylene	26.827	276	2559754	18.319	ng/ul#	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#2
 1,4-Dioxane
 Concen: 7.365 ng/uL
 RT: 3.240 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: BPCCC836.D
 Acq: 25 Jun 2022 20:07

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion: 88 Resp: 107928
 Ion Ratio Lower Upper
 88 100
 43 28.2 35.8 53.6#
 58 74.0 68.1 102.1



#5
 Pyridine
 Concen: 18.776 ng/uL
 RT: 3.634 min Scan# 93
 Delta R.T. 0.006 min
 Lab File: BPCCC836.D
 Acq: 25 Jun 2022 20:07

Tgt Ion: 79 Resp: 695729
 Ion Ratio Lower Upper
 79 100
 52 72.7 83.8 125.8#
 51 33.9 41.0 61.4#









#6
Benzaldehyde
Concen: 19.342 ng/ul
RT: 6.863 min Scan# 643
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 419936
Ion Ratio Lower Upper
77 100
105 99.1 76.0 114.0
106 96.8 73.2 109.8



#8
Phenol
Concen: 19.046 ng/ul
RT: 6.922 min Scan# 652
Delta R.T. 0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07


Tgt Ion: 94 Resp: 838228
Ion Ratio Lower Upper
94 100
65 27.8 27.4 41.2
66 36.1 37.4 56.2#








#10
Bis(2-Chloroethyl)ether
Concen: 18.948 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Instrument :
BNA_P
ClientSampleId :







Tgt Ion: 93 Resp: 669733
Ion Ratio Lower Upper
93 100
63 77.3 72.4 108.6
95 32.2 27.3 40.9



#12
2-Chlorophenol
Concen: 19.487 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:128 Resp: 686713
Ion Ratio Lower Upper
128 100
64 44.4 44.8 67.2#
130 32.8 25.8 38.6









#13
2-Methylphenol
Concen: 18.921 ng/ul
RT: 8.169 min Scan# 801
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 619928
Ion Ratio Lower Upper
108 100
107 90.4 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.376 ng/ul
RT: 8.257 min Scan# 879
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion: 45 Resp: 871462
Ion Ratio Lower Upper
45 100
77 14.6 11.2 16.8
79 11.4 6.9 10.3#









#16
Acetophenone
Concen: 19.349 ng/ul
RT: 8.546 min Scan# 91
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 966580
Ion Ratio Lower Upper
105 100
77 74.5 67.5 101.3
51 25.2 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 18.755 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion: 70 Resp: 473759
Ion Ratio Lower Upper
70 100
42 54.3 53.8 80.8
101 10.3 7.8 11.6
130 22.6 15.4 23.2









#18
4-Methylphenol
Concen: 19.443 ng/ul
RT: 8.499 min Scan# 91
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 663421
Ion Ratio Lower Upper
108 100
107 112.8 88.3 132.5



#19
Hexachloroethane
Concen: 18.631 ng/ul
RT: 8.793 min Scan# 970
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:117 Resp: 265891
Ion Ratio Lower Upper
117 100
201 92.3 85.7 128.5
199 58.2 57.4 86.0









#22
Nitrobenzene
Concen: 20.216 ng/ul
RT: 8.922 min Scan# 993
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 691984
Ion Ratio Lower Upper
77 100
123 47.1 31.5 47.3
65 12.7 11.4 17.2



#23
Isophorone
Concen: 17.630 ng/ul
RT: 9.451 min Scan# 1082
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion: 82 Resp: 1265503
Ion Ratio Lower Upper
82 100
95 6.8 6.6 10.0
138 18.1 13.8 20.6









#25
2-Nitrophenol
Concen: 24.174 ng/ul
RT: 9.628 min Scan# 1112
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 332207
Ion Ratio Lower Upper
139 100
65 51.7 54.2 81.4#
109 28.8 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 18.584 ng/ul
RT: 9.699 min Scan# 1124
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:107 Resp: 697736
Ion Ratio Lower Upper
107 100
121 51.7 39.1 58.7
122 88.2 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 18.630 ng/ul

RT: 9.940 min Scan# 1166

Delta R.T. -0.006 min

Lab File: BPCCC836.D

Acq: 25 Jun 2022 20:07

Instrument :

BNA_P

ClientSampleId :





Tgt Ion: 93 Resp: 843599

Ion Ratio Lower Upper

93 100

95 32.3 26.6 39.8

123 11.3 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 19.256 ng/ul

RT: 10.163 min Scan# 1203

Delta R.T. -0.000 min

Lab File: BPCCC836.D

Acq: 25 Jun 2022 20:07




Tgt Ion:162 Resp: 561028


Ion Ratio Lower Upper

162 100

164 65.4 52.5 78.7


98 38.3 31.4 47.0








#30
Naphthalene
Concen: 18.386 ng/ul
RT: 10.563 min Scan# 1271
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:128 Resp: 2064746
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 12.8 10.2 15.4



#32
4-Chloroaniline
Concen: 18.340 ng/ul
RT: 10.681 min Scan# 1291
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07


Tgt Ion:127 Resp: 831685
Ion Ratio Lower Upper
127 100
129 32.7 26.2 39.2








#33
Hexachlorobutadiene
Concen: 18.378 ng/ul
RT: 10.851 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07


Instrument :
BNA_P
ClientSampleId :





Tgt Ion:225 Resp: 342089
Ion Ratio Lower Upper
225 100
223 63.3 53.7 80.5
227 64.3 49.4 74.2




#34
Caprolactam
Concen: 17.077 ng/ul
RT: 11.463 min Scan# 1424
Delta R.T. 0.018 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07



Tgt Ion:113 Resp: 156300
Ion Ratio Lower Upper
113 100
55 167.7 153.8 230.8
56 125.5 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 18.442 ng/ul
RT: 11.822 min Scan# 1485
Delta R.T. 0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 582370
Ion Ratio Lower Upper
107 100
144 26.9 18.6 27.8
142 82.6 59.6 89.4



#36
2-Methylnaphthalene
Concen: 18.091 ng/ul
RT: 12.181 min Scan# 1546
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:142 Resp: 1315510
Ion Ratio Lower Upper
142 100
141 89.4 70.1 105.1









#37
1-Methylnaphthalene
Concen: 18.237 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 1317527
Ion Ratio Lower Upper
142 100
141 93.2 73.3 109.9
116 3.6 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.956 ng/ul
RT: 12.551 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:216 Resp: 623712
Ion Ratio Lower Upper
216 100
214 78.0 63.0 94.6
179 21.8 18.1 27.1
108 19.5 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 18.549 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 386741
Ion Ratio Lower Upper
237 100
235 62.7 52.8 79.2
272 13.0 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 20.257 ng/ul
RT: 12.798 min Scan# 1651
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:196 Resp: 384268
Ion Ratio Lower Upper
196 100
198 95.1 78.2 117.4
200 29.7 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 19.859 ng/ul

RT: 12.869 min Scan# 1

Delta R.T. -0.000 min


Lab File: BPCCC836.D

Acq: 25 Jun 2022 20:07

Instrument :

BNA_P

ClientSampleId :






Tgt Ion:196 Resp: 404263

Ion Ratio Lower Upper

196 100

198 95.1 77.0 115.4

200 30.8 24.8 37.2



#43

1,1'-Biphenyl

Concen: 18.522 ng/ul

RT: 13.204 min Scan# 1720

Delta R.T. -0.012 min

Lab File: BPCCC836.D

Acq: 25 Jun 2022 20:07




Tgt Ion:154 Resp: 1669524


Ion Ratio Lower Upper

154 100

153 40.6 31.9 47.9

76 14.1 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 18.469 ng/ul
RT: 13.240 min Scan# 1726
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:162 Resp: 1254094
Ion Ratio Lower Upper
162 100
164 33.0 25.5 38.3
127 41.8 31.9 47.9



#45
2-Nitroaniline
Concen: 22.117 ng/ul
RT: 13.457 min Scan# 1763
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion: 65 Resp: 318621
Ion Ratio Lower Upper
65 100
92 70.9 49.0 73.6
138 109.8 66.4 99.6#









#47
Dimethylphthalate
Concen: 18.115 ng/ul
RT: 13.839 min Scan# 1828
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 1422850
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.4 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 22.216 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:165 Resp: 261461
Ion Ratio Lower Upper
165 100
89 57.6 52.9 79.3
121 22.3 19.4 29.2









#50
Acenaphthylene
Concen: 18.193 ng/ul
RT: 14.092 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 1994466
Ion Ratio Lower Upper
152 100
151 19.8 16.5 24.7
153 13.1 10.1 15.1



#51
3-Nitroaniline
Concen: 20.003 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:138 Resp: 282225
Ion Ratio Lower Upper
138 100
108 12.5 11.0 16.4
92 114.2 104.6 156.8









#52
Acenaphthene
Concen: 18.232 ng/ul
RT: 14.439 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 1289725
Ion Ratio Lower Upper
153 100
152 47.3 36.2 54.2
154 88.4 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 20.967 ng/ul
RT: 14.487 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:184 Resp: 94732
Ion Ratio Lower Upper
184 100
63 71.9 73.8 110.6#
154 62.3 57.2 85.8









#55
4-Nitrophenol
Concen: 19.177 ng/ul
RT: 14.604 min Scan# 1956
Delta R.T. 0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 186808
Ion Ratio Lower Upper
109 100
139 145.4 75.6 113.4#
65 141.4 107.6 161.4



#56
Dibenzofuran
Concen: 18.249 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:168 Resp: 1768250
Ion Ratio Lower Upper
168 100
139 38.1 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 22.656 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 356276
Ion Ratio Lower Upper
165 100
63 43.4 47.0 70.6#
182 2.8 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.603 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07


Tgt Ion:232 Resp: 285769
Ion Ratio Lower Upper
232 100
131 50.0 35.8 53.8
130 2.9 1.8 2.8#
166 32.9 23.7 35.5








#59
Diethylphthalate
Concen: 17.842 ng/ul
RT: 15.216 min Scan# 2062
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:149 Resp: 1393276
Ion Ratio Lower Upper
149 100
177 21.2 17.6 26.4
150 12.0 10.3 15.5



#61
Fluorene
Concen: 17.958 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:166 Resp: 1387484
Ion Ratio Lower Upper
166 100
165 98.0 77.4 116.2
167 13.4 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 18.082 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 654005
Ion Ratio Lower Upper
204 100
206 32.7 26.0 39.0
141 64.3 49.3 73.9



#63
4-Nitroaniline
Concen: 20.433 ng/ul
RT: 15.451 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:138 Resp: 266556
Ion Ratio Lower Upper
138 100
92 52.9 46.2 69.2
108 74.1 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 18.796 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 138862
Ion Ratio Lower Upper
198 100
182 4.8 4.3 6.5
77 25.7 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 18.643 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:169 Resp: 1165270
Ion Ratio Lower Upper
169 100
168 66.4 52.6 78.8
167 35.5 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 18.159 ng/ul
RT: 16.328 min Scan# 21
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 375554
Ion Ratio Lower Upper
248 100
250 98.0 80.8 121.2
141 81.1 63.9 95.9



#69
Hexachlorobenzene
Concen: 18.199 ng/ul
RT: 16.428 min Scan# 2268
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:284 Resp: 435939
Ion Ratio Lower Upper
284 100
142 39.5 27.8 41.6
249 31.0 25.4 38.2





#70

Atrazine

Concen: 17.551 ng/ul

RT: 16.610 min Scan# 21

Delta R.T. -0.006 min

Lab File: BPCCC836.D

Acq: 25 Jun 2022 20:07

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 363716

Ion Ratio Lower Upper

200 100

173 28.7 20.7 31.1

215 47.8 38.6 57.8



#71

Pentachlorophenol

Concen: 21.748 ng/ul

RT: 16.786 min Scan# 2329

Delta R.T. -0.000 min

Lab File: BPCCC836.D

Acq: 25 Jun 2022 20:07




Tgt Ion:266 Resp: 268253


Ion Ratio Lower Upper

266 100

264 62.5 51.4 77.2

268 63.6 53.5 80.3









#72
Phenanthrene
Concen: 18.599 ng/ul
RT: 17.180 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:178 Resp: 2043933
Ion Ratio Lower Upper
178 100
179 15.5 12.7 19.1
176 19.1 14.6 22.0



#74
Anthracene
Concen: 18.573 ng/ul
RT: 17.269 min Scan# 2411
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:178 Resp: 2052878
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 18.6 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.486 ng/uL
RT: 13.163 min Scan# 1714
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 629508
Ion Ratio Lower Upper
216 100
214 80.6 65.0 97.4
218 49.0 39.1 58.7



#76
Pentachlorobenzene
Concen: 19.075 ng/uL
RT: 14.687 min Scan# 1972
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07


Tgt Ion:250 Resp: 557110
Ion Ratio Lower Upper
250 100
248 62.6 49.8 74.6
252 64.1 51.2 76.8








#77
 Carbazole
 Concen: 18.978 ng/ul
 RT: 17.551 min Scan# 2459
 Delta R.T. -0.000 min
 Lab File: BPCCC836.D
 Acq: 25 Jun 2022 20:07

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion:167 Resp: 1888179
 Ion Ratio Lower Upper
 167 100
 166 21.3 16.9 25.3
 139 12.7 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 18.142 ng/ul
 RT: 18.116 min Scan# 2555
 Delta R.T. -0.018 min
 Lab File: BPCCC836.D
 Acq: 25 Jun 2022 20:07

Tgt Ion:149 Resp: 2150878
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.1 10.7
 104 5.0 4.6 6.8





#80

Fluoranthene

Concen: 16.589 ng/u1

RT: 19.157 min Scan# 21

Delta R.T. -0.012 min

Lab File: BPCCC836.D

Acq: 25 Jun 2022 20:07

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:202 Resp: 2288165

Ion Ratio Lower Upper

202 100

101 13.8 7.1 10.7#

100 10.5 5.7 8.5#



#82

Pyrene

Concen: 16.994 ng/u1

RT: 19.516 min Scan# 2793

Delta R.T. -0.006 min

Lab File: BPCCC836.D

Acq: 25 Jun 2022 20:07




Tgt Ion:202 Resp: 2367193


Ion Ratio Lower Upper

202 100

101 14.7 7.9 11.9#

100 12.0 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 19.791 ng/ul
RT: 20.410 min Scan# 2945
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 970336
Ion Ratio Lower Upper
149 100
91 70.9 60.0 90.0
206 20.4 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 18.452 ng/ul
RT: 21.169 min Scan# 3074
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:252 Resp: 784321
Ion Ratio Lower Upper
252 100
254 64.0 51.0 76.4
126 14.5 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 18.373 ng/ul
RT: 21.233 min Scan# 3086
Delta R.T. -0.006 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 2365446
Ion Ratio Lower Upper
228 100
229 19.9 15.0 22.4
226 26.6 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.391 ng/ul
RT: 21.174 min Scan# 3075
Delta R.T. -0.024 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:149 Resp: 1427894
Ion Ratio Lower Upper
149 100
167 26.9 22.6 33.8
279 3.7 4.2 6.4#









#87
Chrysene
Concen: 18.594 ng/ul
RT: 21.280 min Scan# 3093
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 2293950
Ion Ratio Lower Upper
228 100
226 29.5 23.6 35.4
229 19.7 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 18.710 ng/ul
RT: 22.092 min Scan# 3231
Delta R.T. -0.030 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:149 Resp: 2499129









#90
Benzo(b)fluoranthene
Concen: 18.595 ng/ul
RT: 22.892 min Scan# 3367
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 2602197
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 12.3 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 18.186 ng/ul
RT: 22.939 min Scan# 3375
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:252 Resp: 2419099
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.4
125 12.4 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 18.691 ng/ul
RT: 23.510 min Scan# 3472
Delta R.T. -0.012 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 2544493
Ion Ratio Lower Upper
252 100
253 21.5 17.4 26.0
125 13.5 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.412 ng/ul
RT: 26.068 min Scan# 3907
Delta R.T. -0.024 min
Lab File: BPCCC836.D
Acq: 25 Jun 2022 20:07

Tgt Ion:276 Resp: 3024427
Ion Ratio Lower Upper
276 100
138 28.0 14.6 22.0#
277 25.3 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 18.470 ng/ul
 RT: 26.092 min Scan# 3916
 Delta R.T. -0.030 min
 Lab File: BPCCC836.D
 Acq: 25 Jun 2022 20:07

Instrument :
 BNA_P
 ClientSampleId :

Tgt Ion:278 Resp: 2602400
 Ion Ratio Lower Upper
 278 100
 139 19.7 9.4 14.2#
 279 23.8 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 18.319 ng/ul
 RT: 26.827 min Scan# 4036
 Delta R.T. -0.018 min
 Lab File: BPCCC836.D
 Acq: 25 Jun 2022 20:07

Tgt Ion:276 Resp: 2559754
 Ion Ratio Lower Upper
 276 100
 138 25.4 12.7 19.1#
 277 23.9 18.2 27.4

