

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BPCCC852.D
 Acq On : 26 Jun 2022 05:41
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Jun 26 22:59:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	959883	20.000	ng/u1	0.00
20) Naphthalene-d8	10.510	136	4066433	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.375	164	2562976	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.134	188	5281039	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.251	240	5256184	20.000	ng/u1	# 0.00
88) Perylene-d12	23.621	264	6056404	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	187499	7.389	ng/uL	0.00
4) Pyridine-d5	3.611	84	1370648	20.292	ng/u1	0.00
7) Phenol-d5	6.893	99	1587389	20.891	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.058	67	983260	19.786	ng/u1	0.00
11) 2-Chlorophenol-d4	7.252	132	1301055	20.638	ng/u1	0.00
15) 4-Methylphenol-d8	8.440	113	1314060	21.877	ng/u1	0.00
21) Nitrobenzene-d5	8.875	128	615493	23.776	ng/u1	0.00
24) 2-Nitrophenol-d4	9.599	143	617998	28.856	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.134	165	1152648	21.782	ng/u1	0.00
31) 4-Chloroaniline-d4	10.657	131	1739078	20.133	ng/u1	0.00
46) Dimethylphthalate-d6	13.793	166	3584069	20.414	ng/u1	0.00
49) Acenaphthylene-d8	14.063	160	4111138	18.348	ng/u1	0.00
54) 4-Nitrophenol-d4	14.593	143	717140	26.152	ng/u1	0.01
60) Fluorene-d10	15.369	176	2963510	19.670	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.498	200	472375	26.998	ng/u1	0.00
73) Anthracene-d10	17.234	188	4368536	18.593	ng/u1	-0.01
81) Pyrene-d10	19.486	212	5107343	17.835	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.469	264	5727606	19.138	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	192003	7.095	ng/uL#	86
5) Pyridine	3.628	79	1390773	20.326	ng/u1#	74
6) Benzaldehyde	6.864	77	981645	24.485	ng/u1	95
8) Phenol	6.922	94	1677990	20.647	ng/u1#	85
10) Bis(2-Chloroethyl)ether	7.152	93	1309851	20.068	ng/u1	89
12) 2-Chlorophenol	7.281	128	1341734	20.618	ng/u1	90
13) 2-Methylphenol	8.169	108	1259317	20.814	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.258	45	1716427	19.600	ng/u1#	96
16) Acetophenone	8.546	105	1901319	20.611	ng/u1#	86
17) N-Nitroso-di-n-propyla...	8.540	70	980996	21.031	ng/u1#	88
18) 4-Methylphenol	8.499	108	1371387	21.765	ng/u1	97
19) Hexachloroethane	8.787	117	515594	19.564	ng/u1#	82
22) Nitrobenzene	8.922	77	1372788	21.250	ng/u1	90
23) Isophorone	9.452	82	2746385	20.272	ng/u1	97
25) 2-Nitrophenol	9.628	139	694126	26.764	ng/u1#	79
26) 2,4-Dimethylphenol	9.699	107	1433964	20.237	ng/u1	94
27) Bis(2-Chloroethoxy)met...	9.940	93	1711357	20.025	ng/u1	98
29) 2,4-Dichlorophenol	10.163	162	1180526	21.469	ng/u1	98
30) Naphthalene	10.563	128	3970019	18.732	ng/u1	100
32) 4-Chloroaniline	10.681	127	1724406	20.149	ng/u1	98
33) Hexachlorobutadiene	10.846	225	655592	18.662	ng/u1	96
34) Caprolactam	11.475	113	314997	18.235	ng/u1	83
35) 4-Chloro-3-methylphenol	11.816	107	1445224	24.250	ng/u1	94
36) 2-Methylnaphthalene	12.181	142	2676964	19.506	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BPCCC852.D
 Acq On : 26 Jun 2022 05:41
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Jun 26 22:59:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.399	142	2709486	19.872	ng/ul#	100
39) 1,2,4,5-Tetrachloroben...	12.551	216	1256903	17.066	ng/ul	99
40) Hexachlorocyclopentadiene	12.528	237	767827	16.452	ng/ul	95
41) 2,4,6-Trichlorophenol	12.799	196	914510	21.537	ng/ul	98
42) 2,4,5-Trichlorophenol	12.869	196	1022936	22.449	ng/ul	98
43) 1,1'-Biphenyl	13.204	154	3580512	17.746	ng/ul#	97
44) 2-Chloronaphthalene	13.240	162	2656684	17.479	ng/ul	98
45) 2-Nitroaniline	13.457	65	874389	27.115	ng/ul#	75
47) Dimethylphthalate	13.840	163	3564159	20.271	ng/ul	98
48) 2,6-Dinitrotoluene	13.957	165	740036	28.092	ng/ul	90
50) Acenaphthylene	14.093	152	4480413	18.258	ng/ul	99
51) 3-Nitroaniline	14.287	138	829514	26.265	ng/ul	87
52) Acenaphthene	14.434	153	2945443	18.602	ng/ul	98
53) 2,4-Dinitrophenol	14.493	184	381647	37.736	ng/ul#	78
55) 4-Nitrophenol	14.610	109	536632	24.611	ng/ul#	76
56) Dibenzofuran	14.775	168	4125122	19.019	ng/ul	96
57) 2,4-Dinitrotoluene	14.745	165	1031400	29.301	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.004	232	779637	23.893	ng/ul	94
59) Diethylphthalate	15.216	149	3581327	20.488	ng/ul	99
61) Fluorene	15.428	166	3347690	19.357	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.428	204	1558098	19.245	ng/ul	97
63) 4-Nitroaniline	15.463	138	860858	29.481	ng/ul#	73
66) 4,6-Dinitro-2-methylph...	15.510	198	482735	25.532	ng/ul#	94
67) N-Nitrosodiphenylamine	15.645	169	2910702	18.197	ng/ul	97
68) 4-Bromophenyl-phenylether	16.328	248	971562	18.357	ng/ul	98
69) Hexachlorobenzene	16.428	284	1125100	18.353	ng/ul	96
70) Atrazine	16.610	200	702106	13.239	ng/ul	98
71) Pentachlorophenol	16.781	266	866369	27.447	ng/ul	96
72) Phenanthrene	17.181	178	5243520	18.645	ng/ul	99
74) Anthracene	17.269	178	5223108	18.466	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.163	216	1338827	15.363	ng/uL	98
76) Pentachlorobenzene	14.687	250	1308296	17.505	ng/uL	99
77) Carbazole	17.545	167	4954319	19.459	ng/ul	99
78) Di-n-butylphthalate	18.116	149	5968134	19.671	ng/ul	99
80) Fluoranthene	19.157	202	6152612	17.470	ng/ul#	88
82) Pyrene	19.516	202	6191515	17.408	ng/ul#	86
83) Butylbenzylphthalate	20.410	149	2835514	22.649	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.175	252	2080867	19.172	ng/ul	98
85) Benzo(a)anthracene	21.233	228	6270854	19.076	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.175	149	4048268	20.420	ng/ul#	97
87) Chrysene	21.286	228	5972927	18.961	ng/ul	98
89) Di-n-octyl phthalate	22.098	149	7194876	20.014	ng/ul	100
90) Benzo(b)fluoranthene	22.898	252	7249276	19.248	ng/ul#	96
91) Benzo(k)fluoranthene	22.945	252	6448472	18.013	ng/ul#	94
93) Benzo(a)pyrene	23.516	252	6913784	18.871	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.080	276	8450593	19.115	ng/ul#	90
95) Dibenzo(a,h)anthracene	26.104	278	7203771	18.996	ng/ul#	93
96) Benzo(g,h,i)perylene	26.839	276	7138274	18.982	ng/ul#	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Jun 26 22:59:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 24 22:42:26 2022
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.095 ng/uL
RT: 3.240 min Scan# 20
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Instrument :
BNA_P
ClientSampleId :





Tgt Ion: 88 Resp: 192003
Ion Ratio Lower Upper
88 100
43 27.3 35.8 53.6#
58 78.2 68.1 102.1




#5
Pyridine
Concen: 20.326 ng/uL
RT: 3.628 min Scan# 92
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41



Tgt Ion: 79 Resp: 1390773
Ion Ratio Lower Upper
79 100
52 76.8 83.8 125.8#
51 34.7 41.0 61.4#









#6
Benzaldehyde
Concen: 24.485 ng/ul
RT: 6.864 min Scan# 643
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 981645
Ion Ratio Lower Upper
77 100
105 100.7 76.0 114.0
106 95.1 73.2 109.8



#8
Phenol
Concen: 20.647 ng/ul
RT: 6.922 min Scan# 652
Delta R.T. 0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Tgt Ion: 94 Resp: 1677990
Ion Ratio Lower Upper
94 100
65 27.4 27.4 41.2#
66 35.2 37.4 56.2#








#10
Bis(2-Chloroethyl)ether
Concen: 20.068 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Instrument :
BNA_P
ClientSampleId :



Tgt Ion: 93 Resp: 1309851




Ion	Ratio	Lower	Upper
93	100		
63	76.4	72.4	108.6
95	33.0	27.3	40.9




#12
2-Chlorophenol
Concen: 20.618 ng/ul
RT: 7.281 min Scan# 713
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:128 Resp: 1341734

Ion	Ratio	Lower	Upper
128	100		
64	45.3	44.8	67.2
130	32.6	25.8	38.6









#13
2-Methylphenol
Concen: 20.814 ng/ul
RT: 8.169 min Scan# 80
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 1259317
Ion Ratio Lower Upper
108 100
107 90.0 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.600 ng/ul
RT: 8.258 min Scan# 879
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion: 45 Resp: 1716427
Ion Ratio Lower Upper
45 100
77 14.9 11.2 16.8
79 11.3 6.9 10.3#









#16
Acetophenone
Concen: 20.611 ng/ul
RT: 8.546 min Scan# 91
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 1901319
Ion Ratio Lower Upper
105 100
77 74.6 67.5 101.3
51 24.5 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 21.031 ng/ul
RT: 8.540 min Scan# 927
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion: 70 Resp: 980996
Ion Ratio Lower Upper
70 100
42 55.3 53.8 80.8
101 10.3 7.8 11.6
130 23.2 15.4 23.2#









#18
4-Methylphenol
Concen: 21.765 ng/ul
RT: 8.499 min Scan# 91
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 1371387
Ion Ratio Lower Upper
108 100
107 113.5 88.3 132.5



#19
Hexachloroethane
Concen: 19.564 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Tgt Ion:117 Resp: 515594
Ion Ratio Lower Upper
117 100
201 88.4 85.7 128.5
199 55.7 57.4 86.0#








#22
Nitrobenzene
Concen: 21.250 ng/ul
RT: 8.922 min Scan# 993
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Instrument :
BNA_P
ClientSampleId :



Tgt Ion: 77 Resp: 1372788




Ion	Ratio	Lower	Upper
77	100		
123	47.3	31.5	47.3
65	13.3	11.4	17.2




#23
Isophorone
Concen: 20.272 ng/ul
RT: 9.452 min Scan# 1082
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion: 82 Resp: 2746385

Ion	Ratio	Lower	Upper
82	100		
95	7.0	6.6	10.0
138	18.7	13.8	20.6









#25
2-Nitrophenol
Concen: 26.764 ng/ul
RT: 9.628 min Scan# 1112
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 694126
Ion Ratio Lower Upper
139 100
65 49.2 54.2 81.4#
109 28.6 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 20.237 ng/ul
RT: 9.699 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:107 Resp: 1433964
Ion Ratio Lower Upper
107 100
121 52.3 39.1 58.7
122 88.2 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 20.025 ng/ul

RT: 9.940 min Scan# 1166

Delta R.T. -0.006 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Instrument :

BNA_P

ClientSampleId :





Tgt Ion: 93 Resp: 1711357

Ion Ratio Lower Upper

93 100

95 32.3 26.6 39.8

123 11.7 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 21.469 ng/ul

RT: 10.163 min Scan# 1203

Delta R.T. 0.000 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41




Tgt Ion:162 Resp: 1180526


Ion Ratio Lower Upper

162 100

164 64.5 52.5 78.7

98 36.6 31.4 47.0









#30
Naphthalene
Concen: 18.732 ng/ul
RT: 10.563 min Scan# 1271
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 3970019
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 20.149 ng/ul
RT: 10.681 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Tgt Ion:127 Resp: 1724406
Ion Ratio Lower Upper
127 100
129 31.7 26.2 39.2








#33
Hexachlorobutadiene
Concen: 18.662 ng/ul
RT: 10.846 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:225 Resp: 655592
Ion Ratio Lower Upper
225 100
223 63.1 53.7 80.5
227 63.6 49.4 74.2



#34
Caprolactam
Concen: 18.235 ng/ul
RT: 11.475 min Scan# 1426
Delta R.T. 0.029 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:113 Resp: 314997
Ion Ratio Lower Upper
113 100
55 162.1 153.8 230.8
56 122.2 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 24.250 ng/ul
RT: 11.816 min Scan# 1484
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 1445224
Ion Ratio Lower Upper
107 100
144 25.9 18.6 27.8
142 80.0 59.6 89.4



#36
2-Methylnaphthalene
Concen: 19.506 ng/ul
RT: 12.181 min Scan# 1546
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:142 Resp: 2676964
Ion Ratio Lower Upper
142 100
141 89.5 70.1 105.1









#37
1-Methylnaphthalene
Concen: 19.872 ng/ul
RT: 12.399 min Scan# 11
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 2709486
Ion Ratio Lower Upper
142 100
141 91.9 73.3 109.9
116 3.8 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.066 ng/ul
RT: 12.551 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:216 Resp: 1256903
Ion Ratio Lower Upper
216 100
214 78.6 63.0 94.6
179 22.2 18.1 27.1
108 18.7 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 16.452 ng/ul
RT: 12.528 min Scan# 1605
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 767827
Ion Ratio Lower Upper
237 100
235 62.3 52.8 79.2
272 12.6 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 21.537 ng/ul
RT: 12.799 min Scan# 1651
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Tgt Ion:196 Resp: 914510
Ion Ratio Lower Upper
196 100
198 96.0 78.2 117.4
200 31.0 25.0 37.6








#42
2,4,5-Trichlorophenol
Concen: 22.449 ng/ul
RT: 12.869 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:196 Resp: 1022936
Ion Ratio Lower Upper
196 100
198 97.8 77.0 115.4
200 31.7 24.8 37.2



#43
1,1'-Biphenyl
Concen: 17.746 ng/ul
RT: 13.204 min Scan# 1720
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:154 Resp: 3580512
Ion Ratio Lower Upper
154 100
153 40.8 31.9 47.9
76 13.8 9.2 13.8#





#44

2-Chloronaphthalene

Concen: 17.479 ng/ul

RT: 13.240 min Scan# 1726

Delta R.T. -0.012 min


Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Instrument :

BNA_P

ClientSampleId :






Tgt Ion:162 Resp: 2656684

Ion Ratio Lower Upper

162 100

164 32.9 25.5 38.3

127 41.3 31.9 47.9



#45

2-Nitroaniline

Concen: 27.115 ng/ul

RT: 13.457 min Scan# 1763

Delta R.T. 0.000 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41




Tgt Ion: 65 Resp: 874389


Ion Ratio Lower Upper

65 100

92 71.2 49.0 73.6

138 113.3 66.4 99.6#









#47
Dimethylphthalate
Concen: 20.271 ng/ul
RT: 13.840 min Scan# 1829
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 3564159
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.2 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 28.092 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Tgt Ion:165 Resp: 740036
Ion Ratio Lower Upper
165 100
89 56.9 52.9 79.3
121 21.8 19.4 29.2








#50
Acenaphthylene
Concen: 18.258 ng/ul
RT: 14.093 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:152 Resp: 4480413
Ion Ratio Lower Upper
152 100
151 20.1 16.5 24.7
153 13.4 10.1 15.1



#51
3-Nitroaniline
Concen: 26.265 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:138 Resp: 829514
Ion Ratio Lower Upper
138 100
108 11.3 11.0 16.4
92 114.1 104.6 156.8









#52
Acenaphthene
Concen: 18.602 ng/ul
RT: 14.434 min Scan# 1929
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 2945443
Ion Ratio Lower Upper
153 100
152 47.2 36.2 54.2
154 89.2 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 37.736 ng/ul
RT: 14.493 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:184 Resp: 381647
Ion Ratio Lower Upper
184 100
63 63.2 73.8 110.6#
154 61.7 57.2 85.8









#55
4-Nitrophenol
Concen: 24.611 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. 0.018 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 536632
Ion Ratio Lower Upper
109 100
139 148.8 75.6 113.4#
65 132.7 107.6 161.4



#56
Dibenzofuran
Concen: 19.019 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:168 Resp: 4125122
Ion Ratio Lower Upper
168 100
139 38.6 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 29.301 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 1031400
Ion Ratio Lower Upper
165 100
63 42.2 47.0 70.6#
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.893 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Tgt Ion:232 Resp: 779637
Ion Ratio Lower Upper
232 100
131 48.9 35.8 53.8
130 2.4 1.8 2.8
166 32.5 23.7 35.5








#59
Diethylphthalate
Concen: 20.488 ng/ul
RT: 15.216 min Scan# 2062
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Instrument :
BNA_P
ClientSampleId :





Tgt Ion:149 Resp: 3581327
Ion Ratio Lower Upper
149 100
177 21.2 17.6 26.4
150 12.5 10.3 15.5




#61
Fluorene
Concen: 19.357 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41



Tgt Ion:166 Resp: 3347690
Ion Ratio Lower Upper
166 100
165 96.2 77.4 116.2
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 19.245 ng/ul
RT: 15.428 min Scan# 20
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 1558098
Ion Ratio Lower Upper
204 100
206 32.7 26.0 39.0
141 64.9 49.3 73.9



#63
4-Nitroaniline
Concen: 29.481 ng/ul
RT: 15.463 min Scan# 2104
Delta R.T. 0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:138 Resp: 860858
Ion Ratio Lower Upper
138 100
92 49.0 46.2 69.2
108 71.2 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 25.532 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 482735
Ion Ratio Lower Upper
198 100
182 4.1 4.3 6.5#
77 26.5 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 18.197 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:169 Resp: 2910702
Ion Ratio Lower Upper
169 100
168 68.1 52.6 78.8
167 36.3 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 18.357 ng/ul
RT: 16.328 min Scan# 21
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 971562
Ion Ratio Lower Upper
248 100
250 99.5 80.8 121.2
141 77.0 63.9 95.9



#69
Hexachlorobenzene
Concen: 18.353 ng/ul
RT: 16.428 min Scan# 2268
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:284 Resp: 1125100
Ion Ratio Lower Upper
284 100
142 39.1 27.8 41.6
249 31.6 25.4 38.2









#70
Atrazine
Concen: 13.239 ng/ul
RT: 16.610 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:200 Resp: 702106
Ion Ratio Lower Upper
200 100
173 28.0 20.7 31.1
215 48.5 38.6 57.8



#71
Pentachlorophenol
Concen: 27.447 ng/ul
RT: 16.781 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:266 Resp: 866369
Ion Ratio Lower Upper
266 100
264 62.9 51.4 77.2
268 62.3 53.5 80.3









#72
Phenanthrene
Concen: 18.645 ng/ul
RT: 17.181 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:178 Resp: 5243520
Ion Ratio Lower Upper
178 100
179 15.6 12.7 19.1
176 19.2 14.6 22.0



#74
Anthracene
Concen: 18.466 ng/ul
RT: 17.269 min Scan# 2411
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:178 Resp: 5223108
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 18.5 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 15.363 ng/uL
RT: 13.163 min Scan# 1713
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 1338827
Ion Ratio Lower Upper
216 100
214 78.4 65.0 97.4
218 48.0 39.1 58.7



#76
Pentachlorobenzene
Concen: 17.505 ng/uL
RT: 14.687 min Scan# 1972
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41


Tgt Ion:250 Resp: 1308296
Ion Ratio Lower Upper
250 100
248 62.5 49.8 74.6
252 63.2 51.2 76.8








#77
 Carbazole
 Concen: 19.459 ng/ul
 RT: 17.545 min Scan# 2458
 Delta R.T. -0.006 min
 Lab File: BPCCC852.D
 Acq: 26 Jun 2022 05:41


Instrument :
 BNA_P
 ClientSampleId :





Tgt Ion:167 Resp: 4954319
 Ion Ratio Lower Upper
 167 100
 166 21.9 16.9 25.3
 139 13.7 10.6 16.0




#78
 Di-n-butylphthalate
 Concen: 19.671 ng/ul
 RT: 18.116 min Scan# 2555
 Delta R.T. -0.018 min
 Lab File: BPCCC852.D
 Acq: 26 Jun 2022 05:41




Tgt Ion:149 Resp: 5968134
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 5.1 4.6 6.8








#80
Fluoranthene
Concen: 17.470 ng/ul
RT: 19.157 min Scan# 21
Delta R.T. -0.012 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:202 Resp: 6152612
Ion Ratio Lower Upper
202 100
101 13.8 7.1 10.7#
100 10.5 5.7 8.5#



#82
Pyrene
Concen: 17.408 ng/ul
RT: 19.516 min Scan# 2793
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:202 Resp: 6191515
Ion Ratio Lower Upper
202 100
101 15.5 7.9 11.9#
100 12.8 6.7 10.1#





#83

Butylbenzylphthalate

Concen: 22.649 ng/ul

RT: 20.410 min Scan# 2945

Delta R.T. -0.012 min


Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Instrument :

BNA_P

ClientSampleId :






Tgt Ion:149 Resp: 2835514

Ion Ratio Lower Upper

149 100

91 69.3 60.0 90.0

206 20.6 17.4 26.2



#84

3,3'-Dichlorobenzidine

Concen: 19.172 ng/ul

RT: 21.175 min Scan# 3075

Delta R.T. -0.006 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41




Tgt Ion:252 Resp: 2080867


Ion Ratio Lower Upper

252 100

254 64.7 51.0 76.4

126 12.4 8.6 12.8





#85

Benzo(a)anthracene

Concen: 19.076 ng/ul

RT: 21.233 min Scan# 3086

Delta R.T. -0.006 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 6270854

Ion Ratio Lower Upper

228 100

229 20.5 15.0 22.4

226 27.8 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.420 ng/ul

RT: 21.175 min Scan# 3075

Delta R.T. -0.023 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41




Tgt Ion:149 Resp: 4048268


Ion Ratio Lower Upper

149 100

167 26.8 22.6 33.8

279 3.8 4.2 6.4#





#87

Chrysene

Concen: 18.961 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.006 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 5972927

Ion Ratio Lower Upper

228 100

226 30.4 23.6 35.4

229 20.2 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 20.014 ng/ul




RT: 22.098 min Scan# 3232


Delta R.T. -0.023 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Tgt Ion:149 Resp: 7194876









#90
Benzo(b)fluoranthene
Concen: 19.248 ng/ul
RT: 22.898 min Scan# 31
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 7249276
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 12.4 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 18.013 ng/ul
RT: 22.945 min Scan# 3376
Delta R.T. -0.006 min
Lab File: BPCCC852.D
Acq: 26 Jun 2022 05:41

Tgt Ion:252 Resp: 6448472
Ion Ratio Lower Upper
252 100
253 22.5 17.0 25.4
125 12.7 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 18.871 ng/ul

RT: 23.516 min Scan# 34

Delta R.T. -0.006 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 6913784

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 14.0 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.115 ng/ul

RT: 26.080 min Scan# 3909

Delta R.T. -0.012 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41




Tgt Ion:276 Resp: 8450593


Ion Ratio Lower Upper

276 100

138 28.2 14.6 22.0#

277 25.0 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 18.996 ng/ul

RT: 26.104 min Scan# 3916

Delta R.T. -0.018 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 7203771

Ion Ratio Lower Upper

278 100

139 19.4 9.4 14.2#

279 24.2 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 18.982 ng/ul

RT: 26.839 min Scan# 4038

Delta R.T. -0.006 min

Lab File: BPCCC852.D

Acq: 26 Jun 2022 05:41

Tgt Ion:276 Resp: 7138274

Ion Ratio Lower Upper

276 100

138 26.0 12.7 19.1#

277 23.9 18.2 27.4

