

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BP010834.D
 Acq On : 25 Jun 2022 18:25
 Operator : CG/JU
 Sample : PB145662BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS662

Quant Time: Jun 26 22:53:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	402140	20.000	ng/ul	0.00
20) Naphthalene-d8	10.516	136	1567135	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.375	164	800210	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.139	188	1466382	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.251	240	1466567	20.000	ng/ul	# 0.00
88) Perylene-d12	23.615	264	1611386	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	69542	6.542	ng/uL	0.00
4) Pyridine-d5	3.611	84	897252	31.707	ng/ul	0.00
7) Phenol-d5	6.899	99	1036356	32.555	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	667247	32.050	ng/ul	0.00
11) 2-Chlorophenol-d4	7.252	132	892116	33.778	ng/ul	0.00
15) 4-Methylphenol-d8	8.440	113	821237	32.634	ng/ul	0.00
21) Nitrobenzene-d5	8.881	128	399682	40.063	ng/ul	0.00
24) 2-Nitrophenol-d4	9.599	143	387847	46.991	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.140	165	709033	34.767	ng/ul	0.00
31) 4-Chloroaniline-d4	10.657	131	1016252	30.528	ng/ul	0.00
46) Dimethylphthalate-d6	13.792	166	1861160	33.954	ng/ul	0.00
49) Acenaphthylene-d8	14.063	160	2317365	33.125	ng/ul	0.00
54) 4-Nitrophenol-d4	14.592	143	333384	38.939	ng/ul	0.01
60) Fluorene-d10	15.375	176	1567241	33.317	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.498	200	190415	39.194	ng/ul	0.00
73) Anthracene-d10	17.239	188	2260254	34.644	ng/ul	0.00
81) Pyrene-d10	19.486	212	2578985	32.278	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.468	264	2875798	36.116	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	139614	12.315	ng/uL#	84
5) Pyridine	3.634	79	870684	30.374	ng/ul#	72
6) Benzaldehyde	6.863	77	722761	43.031	ng/ul	96
8) Phenol	6.928	94	1046396	30.734	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.152	93	850468	31.101	ng/ul	89
12) 2-Chlorophenol	7.287	128	870822	31.942	ng/ul#	90
13) 2-Methylphenol	8.169	108	769830	30.371	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.252	45	1104920	30.116	ng/ul#	96
16) Acetophenone	8.546	105	1196464	30.959	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.534	70	569038	29.119	ng/ul	88
18) 4-Methylphenol	8.505	108	813121	30.804	ng/ul	99
19) Hexachloroethane	8.793	117	349449	31.650	ng/ul	83
22) Nitrobenzene	8.922	77	882772	35.458	ng/ul#	89
23) Isophorone	9.452	82	1555244	29.788	ng/ul	97
25) 2-Nitrophenol	9.634	139	416248	41.645	ng/ul#	80
26) 2,4-Dimethylphenol	9.704	107	820353	30.041	ng/ul	94
27) Bis(2-Chloroethoxy)met...	9.940	93	1026009	31.152	ng/ul	97
29) 2,4-Dichlorophenol	10.169	162	686654	32.403	ng/ul	99
30) Naphthalene	10.563	128	2604232	31.884	ng/ul	99
32) 4-Chloroaniline	10.681	127	962507	29.183	ng/ul	98
33) Hexachlorobutadiene	10.851	225	428709	31.666	ng/ul	97
34) Caprolactam	11.469	113	200970	30.189	ng/ul	81
35) 4-Chloro-3-methylphenol	11.822	107	708842	30.863	ng/ul	92
36) 2-Methylnaphthalene	12.187	142	1600873	30.268	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062522\
 Data File : BP010834.D
 Acq On : 25 Jun 2022 18:25
 Operator : CG/JU
 Sample : PB145662BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS662

Quant Time: Jun 26 22:53:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 24 22:42:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.404	142	1595828	30.371	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.551	216	746611	32.468	ng/ul	98
40) Hexachlorocyclopentadiene	12.534	237	434183	29.797	ng/ul	96
41) 2,4,6-Trichlorophenol	12.804	196	466664	35.200	ng/ul	98
42) 2,4,5-Trichlorophenol	12.875	196	507706	35.687	ng/ul	99
43) 1,1'-Biphenyl	13.204	154	2001727	31.776	ng/ul#	98
44) 2-Chloronaphthalene	13.245	162	1542767	32.510	ng/ul	99
45) 2-Nitroaniline	13.457	65	404509	40.176	ng/ul#	78
47) Dimethylphthalate	13.845	163	1776391	32.360	ng/ul	99
48) 2,6-Dinitrotoluene	13.957	165	338258	41.126	ng/ul	92
50) Acenaphthylene	14.092	152	2359123	30.790	ng/ul	99
51) 3-Nitroaniline	14.287	138	385063	39.051	ng/ul	89
52) Acenaphthene	14.439	153	1609141	32.549	ng/ul	98
53) 2,4-Dinitrophenol	14.492	184	105495	33.410	ng/ul#	80
55) 4-Nitrophenol	14.610	109	240376	35.309	ng/ul#	75
56) Dibenzofuran	14.775	168	2136258	31.547	ng/ul	96
57) 2,4-Dinitrotoluene	14.745	165	469126	42.685	ng/ul#	83
58) 2,3,4,6-Tetrachlorophenol	15.004	232	373819	36.692	ng/ul	93
59) Diethylphthalate	15.216	149	1767084	32.379	ng/ul	99
61) Fluorene	15.428	166	1720878	31.870	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.428	204	799394	31.625	ng/ul	98
63) 4-Nitroaniline	15.457	138	412825	45.281	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.510	198	195305	37.202	ng/ul	95
67) N-Nitrosodiphenylamine	15.645	169	1440976	32.444	ng/ul	99
68) 4-Bromophenyl-phenylether	16.328	248	484479	32.967	ng/ul	98
69) Hexachlorobenzene	16.433	284	547812	32.183	ng/ul	97
70) Atrazine	16.616	200	477754	32.444	ng/ul	98
71) Pentachlorophenol	16.786	266	234977	26.809	ng/ul	96
72) Phenanthrene	17.181	178	2615634	33.495	ng/ul	99
74) Anthracene	17.275	178	2608427	33.212	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.163	216	766004	31.656	ng/ul	97
76) Pentachlorobenzene	14.692	250	676530	32.599	ng/ul	99
77) Carbazole	17.551	167	2427005	34.330	ng/ul	100
78) Di-n-butylphthalate	18.122	149	2842165	33.737	ng/ul	98
80) Fluoranthene	19.163	202	2937760	29.896	ng/ul#	90
82) Pyrene	19.516	202	3038638	30.620	ng/ul#	85
83) Butylbenzylphthalate	20.410	149	1320966	37.817	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.174	252	1150657	37.996	ng/ul#	99
85) Benzo(a)anthracene	21.233	228	3093725	33.730	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.180	149	1931799	34.924	ng/ul#	97
87) Chrysene	21.286	228	2907656	33.082	ng/ul	100
89) Di-n-octyl phthalate	22.098	149	3401624	35.564	ng/ul	100
90) Benzo(b)fluoranthene	22.898	252	3452290	34.451	ng/ul#	96
91) Benzo(k)fluoranthene	22.945	252	3285579	34.494	ng/ul#	95
93) Benzo(a)pyrene	23.515	252	3051708	31.306	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.080	276	4204188	35.742	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.104	278	3502337	34.713	ng/ul#	94
96) Benzo(g,h,i)perylene	26.833	276	3409792	34.079	ng/ul#	89





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.315 ng/uL
RT: 3.240 min Scan# 20
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion: 88 Resp: 139614
Ion Ratio Lower Upper
88 100
43 27.7 35.8 53.6#
58 75.5 68.1 102.1



#5
Pyridine
Concen: 30.374 ng/uL
RT: 3.634 min Scan# 93
Delta R.T. 0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion: 79 Resp: 870684
Ion Ratio Lower Upper
79 100
52 73.7 83.8 125.8#
51 34.1 41.0 61.4#









#6
Benzaldehyde
Concen: 43.031 ng/ul
RT: 6.863 min Scan# 643
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion: 77 Resp: 722761
Ion Ratio Lower Upper
77 100
105 100.2 76.0 114.0
106 94.6 73.2 109.8



#8
Phenol
Concen: 30.734 ng/ul
RT: 6.928 min Scan# 653
Delta R.T. 0.012 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion: 94 Resp: 1046396
Ion Ratio Lower Upper
94 100
65 27.0 27.4 41.2#
66 35.9 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 31.101 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion: 93 Resp: 850468
Ion Ratio Lower Upper
93 100
63 77.5 72.4 108.6
95 32.3 27.3 40.9



#12
2-Chlorophenol
Concen: 31.942 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:128 Resp: 870822
Ion Ratio Lower Upper
128 100
64 44.5 44.8 67.2#
130 32.3 25.8 38.6









#13
2-Methylphenol
Concen: 30.371 ng/ul
RT: 8.169 min Scan# 80
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:108 Resp: 769830
Ion Ratio Lower Upper
108 100
107 88.7 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 30.116 ng/ul
RT: 8.252 min Scan# 878
Delta R.T. -0.012 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion: 45 Resp: 1104920
Ion Ratio Lower Upper
45 100
77 14.7 11.2 16.8
79 10.8 6.9 10.3#









#16
Acetophenone
Concen: 30.959 ng/ul
RT: 8.546 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:105 Resp: 1196464
Ion Ratio Lower Upper
105 100
77 76.3 67.5 101.3
51 26.5 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 29.119 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion: 70 Resp: 569038
Ion Ratio Lower Upper
70 100
42 54.9 53.8 80.8
101 9.9 7.8 11.6
130 22.4 15.4 23.2









#18
4-Methylphenol
Concen: 30.804 ng/ul
RT: 8.505 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:108 Resp: 813121
Ion Ratio Lower Upper
108 100
107 109.8 88.3 132.5



#19
Hexachloroethane
Concen: 31.650 ng/ul
RT: 8.793 min Scan# 970
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:117 Resp: 349449
Ion Ratio Lower Upper
117 100
201 89.6 85.7 128.5
199 57.7 57.4 86.0









#22
Nitrobenzene
Concen: 35.458 ng/ul
RT: 8.922 min Scan# 991
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion: 77 Resp: 882772
Ion Ratio Lower Upper
77 100
123 48.1 31.5 47.3#
65 13.7 11.4 17.2



#23
Isophorone
Concen: 29.788 ng/ul
RT: 9.452 min Scan# 1082
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion: 82 Resp: 1555244
Ion Ratio Lower Upper
82 100
95 6.9 6.6 10.0
138 18.1 13.8 20.6









#25
2-Nitrophenol
Concen: 41.645 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:139 Resp: 416248
Ion Ratio Lower Upper
139 100
65 49.6 54.2 81.4#
109 28.5 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 30.041 ng/ul
RT: 9.704 min Scan# 1125
Delta R.T. 0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:107 Resp: 820353
Ion Ratio Lower Upper
107 100
121 52.7 39.1 58.7
122 88.8 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 31.152 ng/ul

RT: 9.940 min Scan# 1166

Delta R.T. -0.006 min

Lab File: BP010834.D

Acq: 25 Jun 2022 18:25

Instrument :





BNA_P

ClientSampleId :

SLCS662

Tgt Ion: 93 Resp: 1026009

Ion	Ratio	Lower	Upper
93	100		
95	31.8	26.6	39.8
123	11.3	7.8	11.8



#29

2,4-Dichlorophenol

Concen: 32.403 ng/ul

RT: 10.169 min Scan# 1204




Delta R.T. 0.006 min


Lab File: BP010834.D

Acq: 25 Jun 2022 18:25

Tgt Ion:162 Resp: 686654

Ion	Ratio	Lower	Upper
162	100		
164	65.5	52.5	78.7
98	37.6	31.4	47.0









#30
Naphthalene
Concen: 31.884 ng/ul
RT: 10.563 min Scan# 1272
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:128 Resp: 2604232
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 29.183 ng/ul
RT: 10.681 min Scan# 1291
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:127 Resp: 962507
Ion Ratio Lower Upper
127 100
129 31.7 26.2 39.2









#33
Hexachlorobutadiene
Concen: 31.666 ng/ul
RT: 10.851 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:225 Resp: 428709
Ion Ratio Lower Upper
225 100
223 64.1 53.7 80.5
227 64.2 49.4 74.2



#34
Caprolactam
Concen: 30.189 ng/ul
RT: 11.469 min Scan# 1425
Delta R.T. 0.023 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:113 Resp: 200970
Ion Ratio Lower Upper
113 100
55 160.5 153.8 230.8
56 118.5 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 30.863 ng/ul
RT: 11.822 min Scan# 1486
Delta R.T. 0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:107 Resp: 708842
Ion Ratio Lower Upper
107 100
144 25.5 18.6 27.8
142 81.9 59.6 89.4



#36
2-Methylnaphthalene
Concen: 30.268 ng/ul
RT: 12.187 min Scan# 1547
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:142 Resp: 1600873
Ion Ratio Lower Upper
142 100
141 89.0 70.1 105.1









#37
1-Methylnaphthalene
Concen: 30.371 ng/ul
RT: 12.404 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:142 Resp: 1595828
Ion Ratio Lower Upper
142 100
141 91.9 73.3 109.9
116 3.7 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.468 ng/ul
RT: 12.551 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:216 Resp: 746611
Ion Ratio Lower Upper
216 100
214 79.5 63.0 94.6
179 22.3 18.1 27.1
108 19.6 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 29.797 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:237 Resp: 434183
Ion Ratio Lower Upper
237 100
235 62.6 52.8 79.2
272 12.6 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 35.200 ng/ul
RT: 12.804 min Scan# 1652
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:196 Resp: 466664
Ion Ratio Lower Upper
196 100
198 95.8 78.2 117.4
200 31.3 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 35.687 ng/ul
RT: 12.875 min Scan# 1664
Delta R.T. 0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:196 Resp: 507706
Ion Ratio Lower Upper
196 100
198 97.3 77.0 115.4
200 31.1 24.8 37.2



#43
1,1'-Biphenyl
Concen: 31.776 ng/ul
RT: 13.204 min Scan# 1720
Delta R.T. -0.012 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:154 Resp: 2001727
Ion Ratio Lower Upper
154 100
153 40.5 31.9 47.9
76 14.4 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 32.510 ng/ul
RT: 13.245 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:162 Resp: 1542767
Ion Ratio Lower Upper
162 100
164 32.8 25.5 38.3
127 39.7 31.9 47.9



#45
2-Nitroaniline
Concen: 40.176 ng/ul
RT: 13.457 min Scan# 1763
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion: 65 Resp: 404509
Ion Ratio Lower Upper
65 100
92 69.9 49.0 73.6
138 109.4 66.4 99.6#









#47
Dimethylphthalate
Concen: 32.360 ng/ul
RT: 13.845 min Scan# 1829
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:163 Resp: 1776391
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 9.9 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 41.126 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:165 Resp: 338258
Ion Ratio Lower Upper
165 100
89 58.6 52.9 79.3
121 22.8 19.4 29.2









#50
Acenaphthylene
Concen: 30.790 ng/ul
RT: 14.092 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:152 Resp: 2359123
Ion Ratio Lower Upper
152 100
151 20.0 16.5 24.7
153 13.3 10.1 15.1



#51
3-Nitroaniline
Concen: 39.051 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:138 Resp: 385063
Ion Ratio Lower Upper
138 100
108 13.0 11.0 16.4
92 117.2 104.6 156.8









#52
Acenaphthene
Concen: 32.549 ng/ul
RT: 14.439 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:153 Resp: 1609141
Ion Ratio Lower Upper
153 100
152 47.8 36.2 54.2
154 89.0 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 33.410 ng/ul
RT: 14.492 min Scan# 1939
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:184 Resp: 105495
Ion Ratio Lower Upper
184 100
63 67.8 73.8 110.6#
154 60.6 57.2 85.8









#55
4-Nitrophenol
Concen: 35.309 ng/ul
RT: 14.610 min Scan# 1960
Delta R.T. 0.018 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:109 Resp: 240376
Ion Ratio Lower Upper
109 100
139 150.2 75.6 113.4#
65 137.7 107.6 161.4



#56
Dibenzofuran
Concen: 31.547 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:168 Resp: 2136258
Ion Ratio Lower Upper
168 100
139 38.6 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 42.685 ng/ul
RT: 14.745 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:165 Resp: 469126
Ion Ratio Lower Upper
165 100
63 45.4 47.0 70.6#
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 36.692 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:232 Resp: 373819
Ion Ratio Lower Upper
232 100
131 50.5 35.8 53.8
130 2.6 1.8 2.8
166 32.6 23.7 35.5









#59
Diethylphthalate
Concen: 32.379 ng/ul
RT: 15.216 min Scan# 2062
Delta R.T. -0.012 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:149 Resp: 1767084
Ion Ratio Lower Upper
149 100
177 21.4 17.6 26.4
150 12.2 10.3 15.5



#61
Fluorene
Concen: 31.870 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:166 Resp: 1720878
Ion Ratio Lower Upper
166 100
165 96.8 77.4 116.2
167 13.6 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 31.625 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.012 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:204 Resp: 799394
Ion Ratio Lower Upper
204 100
206 32.4 26.0 39.0
141 64.4 49.3 73.9



#63
4-Nitroaniline
Concen: 45.281 ng/ul
RT: 15.457 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:138 Resp: 412825
Ion Ratio Lower Upper
138 100
92 52.2 46.2 69.2
108 75.0 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 37.202 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.000 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:198 Resp: 195305
Ion Ratio Lower Upper
198 100
182 4.5 4.3 6.5
77 26.9 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 32.444 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:169 Resp: 1440976
Ion Ratio Lower Upper
169 100
168 66.5 52.6 78.8
167 36.0 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 32.967 ng/ul
RT: 16.328 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:248 Resp: 484479
Ion Ratio Lower Upper
248 100
250 98.7 80.8 121.2
141 82.0 63.9 95.9



#69
Hexachlorobenzene
Concen: 32.183 ng/ul
RT: 16.433 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:284 Resp: 547812
Ion Ratio Lower Upper
284 100
142 35.3 27.8 41.6
249 29.3 25.4 38.2





#70

Atrazine

Concen: 32.444 ng/ul

RT: 16.616 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP010834.D

Acq: 25 Jun 2022 18:25

Instrument :

BNA_P

ClientSampleId :

SLCS662





Tgt Ion:200 Resp: 477754

Ion Ratio Lower Upper

200 100

173 28.3 20.7 31.1

215 48.3 38.6 57.8



#71

Pentachlorophenol

Concen: 26.809 ng/ul

RT: 16.786 min Scan# 2329

Delta R.T. -0.000 min

Lab File: BP010834.D

Acq: 25 Jun 2022 18:25




Tgt Ion:266 Resp: 234977


Ion Ratio Lower Upper

266 100

264 61.8 51.4 77.2

268 63.3 53.5 80.3









#72
Phenanthrene
Concen: 33.495 ng/u1
RT: 17.181 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:178 Resp: 2615634
Ion Ratio Lower Upper
178 100
179 15.4 12.7 19.1
176 18.9 14.6 22.0



#74
Anthracene
Concen: 33.212 ng/u1
RT: 17.275 min Scan# 2412
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:178 Resp: 2608427
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.2
176 18.3 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.656 ng/uL
RT: 13.163 min Scan# 1713
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:216 Resp: 766004
Ion Ratio Lower Upper
216 100
214 77.9 65.0 97.4
218 47.1 39.1 58.7



#76
Pentachlorobenzene
Concen: 32.599 ng/uL
RT: 14.692 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:250 Resp: 676530
Ion Ratio Lower Upper
250 100
248 63.4 49.8 74.6
252 65.0 51.2 76.8





#77

Carbazole

Concen: 34.330 ng/ul

RT: 17.551 min Scan# 24

Delta R.T. -0.000 min

Lab File: BP010834.D

Acq: 25 Jun 2022 18:25

Instrument :

BNA_P

ClientSampleId :

SLCS662





Tgt Ion:167 Resp: 2427005

Ion Ratio Lower Upper

167 100

166 21.2 16.9 25.3

139 13.4 10.6 16.0



#78

Di-n-butylphthalate

Concen: 33.737 ng/ul

RT: 18.122 min Scan# 2556

Delta R.T. -0.012 min

Lab File: BP010834.D

Acq: 25 Jun 2022 18:25




Tgt Ion:149 Resp: 2842165


Ion Ratio Lower Upper

149 100

150 9.2 7.1 10.7

104 4.8 4.6 6.8









#80
Fluoranthene
Concen: 29.896 ng/u1
RT: 19.163 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:202 Resp: 2937760
Ion Ratio Lower Upper
202 100
101 13.2 7.1 10.7#
100 10.1 5.7 8.5#



#82
Pyrene
Concen: 30.620 ng/u1
RT: 19.516 min Scan# 2793
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:202 Resp: 3038638
Ion Ratio Lower Upper
202 100
101 15.7 7.9 11.9#
100 13.1 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 37.817 ng/ul
RT: 20.410 min Scan# 2945
Delta R.T. -0.012 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:149 Resp: 1320966
Ion Ratio Lower Upper
149 100
91 73.9 60.0 90.0
206 19.7 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 37.996 ng/ul
RT: 21.174 min Scan# 3075
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:252 Resp: 1150657
Ion Ratio Lower Upper
252 100
254 64.0 51.0 76.4
126 13.0 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 33.730 ng/ul
RT: 21.233 min Scan# 3086
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:228 Resp: 3093725
Ion Ratio Lower Upper
228 100
229 20.0 15.0 22.4
226 27.0 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 34.924 ng/ul
RT: 21.180 min Scan# 3076
Delta R.T. -0.018 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:149 Resp: 1931799
Ion Ratio Lower Upper
149 100
167 26.9 22.6 33.8
279 4.0 4.2 6.4





#87

Chrysene

Concen: 33.082 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.006 min

Lab File: BP010834.D

Acq: 25 Jun 2022 18:25

Instrument :

BNA_P

ClientSampleId :

SLCS662





Tgt Ion:228 Resp: 2907656

Ion Ratio Lower Upper

228 100

226 29.5 23.6 35.4

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 35.564 ng/ul




RT: 22.098 min Scan# 3232


Delta R.T. -0.024 min

Lab File: BP010834.D

Acq: 25 Jun 2022 18:25

Tgt Ion:149 Resp: 3401624









#90
Benzo(b)fluoranthene
Concen: 34.451 ng/ul
RT: 22.898 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:252 Resp: 3452290
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 12.2 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 34.494 ng/ul
RT: 22.945 min Scan# 3376
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:252 Resp: 3285579
Ion Ratio Lower Upper
252 100
253 22.0 17.0 25.4
125 12.9 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 31.306 ng/ul
RT: 23.515 min Scan# 34
Delta R.T. -0.006 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25




Instrument :
BNA_P
ClientSampleId :
SLCS662


Tgt Ion:252 Resp: 3051708
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 14.1 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 35.742 ng/ul
RT: 26.080 min Scan# 3909
Delta R.T. -0.012 min
Lab File: BP010834.D
Acq: 25 Jun 2022 18:25

Tgt Ion:276 Resp: 4204188
Ion Ratio Lower Upper
276 100
138 27.5 14.6 22.0#
277 24.6 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 34.713 ng/ul
 RT: 26.104 min Scan# 3915
 Delta R.T. -0.018 min
 Lab File: BP010834.D
 Acq: 25 Jun 2022 18:25

Instrument :
 BNA_P
 ClientSampleId :
 SLCS662

Tgt Ion:278 Resp: 3502337
 Ion Ratio Lower Upper
 278 100
 139 18.9 9.4 14.2#
 279 23.4 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 34.079 ng/ul
 RT: 26.833 min Scan# 4037
 Delta R.T. -0.012 min
 Lab File: BP010834.D
 Acq: 25 Jun 2022 18:25

Tgt Ion:276 Resp: 3409792
 Ion Ratio Lower Upper
 276 100
 138 26.2 12.7 19.1#
 277 23.6 18.2 27.4

