

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062821\
 Data File : BP006069.D
 Acq On : 28 Jun 2021 10:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 11:35:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 28 11:34:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	78	0.00
2	1,4-Dioxane	40.000	41.583	-4.0	86	0.00
3	Pyridine	40.000	44.364	-10.9	87	0.00
4	n-Nitrosodimethylamine	40.000	36.092	9.8	74	0.00
5 S	2-Fluorophenol	80.000	87.150	-8.9	89	0.00
6	Aniline	40.000	44.372	-10.9	91	0.00
7 S	Phenol-d6	80.000	91.383	-14.2	93	0.00
8	2-Chlorophenol	40.000	44.303	-10.8	92	0.00
9	Benzaldehyde	40.000	50.006	-25.0#	93	0.00
10 C	Phenol	40.000	45.168	-12.9	93	0.00
11	bis(2-Chloroethyl)ether	40.000	45.074	-12.7	94	0.00
12	1,3-Dichlorobenzene	40.000	43.147	-7.9	90	0.00
13 C	1,4-Dichlorobenzene	40.000	42.804	-7.0	89	0.00
14	1,2-Dichlorobenzene	40.000	42.714	-6.8	88	0.00
15	Benzyl Alcohol	40.000	42.881	-7.2	87	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.629	-6.6	87	0.00
17	2-Methylphenol	40.000	45.636	-14.1	93	0.00
18	Hexachloroethane	40.000	42.464	-6.2	87	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.315	-10.8	90	0.00
20	3+4-Methylphenols	40.000	45.947	-14.9	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	78	0.00
22	Acetophenone	40.000	44.723	-11.8	92	0.00
23 S	Nitrobenzene-d5	80.000	86.147	-7.7	89	0.00
24	Nitrobenzene	40.000	41.840	-4.6	88	0.00
25	Isophorone	40.000	42.290	-5.7	89	0.00
26 C	2-Nitrophenol	40.000	43.461	-8.7	91	0.00
27	2,4-Dimethylphenol	40.000	43.282	-8.2	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.630	-9.1	93	0.00
29 C	2,4-Dichlorophenol	40.000	42.305	-5.8	90	0.00
30	1,2,4-Trichlorobenzene	40.000	40.948	-2.4	88	0.00
31	Naphthalene	40.000	42.897	-7.2	92	0.00
32	Benzoic acid	40.000	42.821	-7.1	88	0.00
33	4-Chloroaniline	40.000	43.013	-7.5	91	0.00
34 C	Hexachlorobutadiene	40.000	40.432	-1.1	87	0.00
35	Caprolactam	40.000	47.537	-18.8	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	44.052	-10.1	94	0.00
37	2-Methylnaphthalene	40.000	42.163	-5.4	91	0.00
38	1-Methylnaphthalene	40.000	42.509	-6.3	91	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	77	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.870	-7.2	87	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.349	11.6	74	0.00
42 S	2,4,6-Tribromophenol	80.000	90.934	-13.7	92	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.268	-3.2	84	0.00
44	2,4,5-Trichlorophenol	40.000	40.630	-1.6	83	0.00
45 S	2-Fluorobiphenyl	80.000	86.635	-8.3	89	0.00
46	1,1'-Biphenyl	40.000	44.027	-10.1	89	0.00
47	2-Chloronaphthalene	40.000	42.334	-5.8	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062821\
 Data File : BP006069.D
 Acq On : 28 Jun 2021 10:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 28 11:35:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 28 11:34:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.417	-8.5	89	0.00
49	Acenaphthylene	40.000	43.197	-8.0	89	0.00
50	Dimethylphthalate	40.000	42.958	-7.4	91	0.00
51	2,6-Dinitrotoluene	40.000	43.759	-9.4	90	0.00
52 C	Acenaphthene	40.000	42.960	-7.4	91	0.00
53	3-Nitroaniline	40.000	44.049	-10.1	90	0.00
54 P	2,4-Dinitrophenol	40.000	47.825	-19.6	100	0.00
55	Dibenzofuran	40.000	42.443	-6.1	89	0.00
56 P	4-Nitrophenol	40.000	38.589	3.5	77	0.00
57	2,4-Dinitrotoluene	40.000	44.606	-11.5	90	0.00
58	Fluorene	40.000	43.660	-9.1	92	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.475	-3.7	85	0.00
60	Diethylphthalate	40.000	43.516	-8.8	91	0.00
61	4-Chlorophenyl-phenylether	40.000	42.689	-6.7	90	0.00
62	4-Nitroaniline	40.000	44.701	-11.8	92	0.00
63	Azobenzene	40.000	43.380	-8.5	89	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	75	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.711	-11.8	90	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.728	-9.3	90	0.00
67	4-Bromophenyl-phenylether	40.000	43.400	-8.5	89	0.00
68	Hexachlorobenzene	40.000	42.927	-7.3	89	0.00
69	Atrazine	40.000	41.943	-4.9	74	0.00
70 C	Pentachlorophenol	40.000	40.114	-0.3	75	0.00
71	Phenanthrene	40.000	43.334	-8.3	89	0.00
72	Anthracene	40.000	43.844	-9.6	90	0.00
73	Carbazole	40.000	44.436	-11.1	91	0.00
74	Di-n-butylphthalate	40.000	45.946	-14.9	93	0.00
75 C	Fluoranthene	40.000	44.267	-10.7	91	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	80	0.00
77	Benzydine	40.000	43.564	-8.9	77	0.00
78	Pyrene	40.000	41.930	-4.8	92	0.00
79 S	Terphenyl-d14	80.000	89.401	-11.8	97	0.00
80	Butylbenzylphthalate	40.000	45.153	-12.9	97	0.00
81	Benzo(a)anthracene	40.000	42.395	-6.0	94	0.00
82	3,3'-Dichlorobenzidine	40.000	42.752	-6.9	96	0.00
83	Chrysene	40.000	43.285	-8.2	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.174	-12.9	98	0.00
85 c	Di-n-octyl phthalate	40.000	45.156	-12.9	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	80	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.306	-5.8	94	0.00
88	Benzo(b)fluoranthene	40.000	42.977	-7.4	94	0.00
89	Benzo(k)fluoranthene	40.000	44.464	-11.2	98	0.00
90 C	Benzo(a)pyrene	40.000	43.116	-7.8	95	0.00
91	Dibenzo(a,h)anthracene	40.000	42.396	-6.0	93	0.00
92	Benzo(g,h,i)perylene	40.000	41.772	-4.4	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062821\
Data File : BP006069.D
Acq On : 28 Jun 2021 10:21
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 28 11:35:10 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 28 11:34:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0