

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062921\
 Data File : BP006097.D
 Acq On : 29 Jun 2021 18:37
 Operator : CG/JU
 Sample : M2856-18MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SB-09-COMPMS

Manual Integrations
 APPROVED

mohammad
 6/30/2021 2:14:16 PM

Quant Time: Jun 30 01:12:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 28 11:34:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	111275	20.000	ng	0.00
21) Naphthalene-d8	10.705	136	413333	20.000	ng	0.00
39) Acenaphthene-d10	14.534	164	235480	20.000	ng	0.00
64) Phenanthrene-d10	17.281	188	453185	20.000	ng	0.00
76) Chrysene-d12	21.357	240	423231	20.000	ng	# 0.00
86) Perylene-d12	23.757	264	401969	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	680471	98.127	ng	0.00
7) Phenol-d6	7.087	99	824077	91.683	ng	0.00
23) Nitrobenzene-d5	9.069	82	492168	58.378	ng	0.00
42) 2,4,6-Tribromophenol	16.028	330	368864	91.280	ng	0.00
45) 2-Fluorobiphenyl	13.157	172	1042502	58.134	ng	0.00
79) Terphenyl-d14	19.834	244	1573304	69.607	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.346	88	125700	40.122	ng	98
3) Pyridine	3.758	79	311660	42.326	ng	98
4) n-Nitrosodimethylamine	3.670	42	144945	42.348	ng	80
6) Aniline	7.234	93	434765	43.326	ng	99
8) 2-Chlorophenol	7.469	128	409873	51.550	ng	99
9) Benzaldehyde	7.046	77	227808	59.453	ng	98
10) Phenol	7.117	94	474702	52.868	ng	96
11) bis(2-Chloroethyl)ether	7.328	93	357759	52.580	ng	99
12) 1,3-Dichlorobenzene	7.787	146	443387	49.871	ng	99
13) 1,4-Dichlorobenzene	7.934	146	451764	49.826	ng	99
14) 1,2-Dichlorobenzene	8.252	146	427822	49.365	ng	99
15) Benzyl Alcohol	8.158	79	343392	50.722	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.428	45	313644	49.867	ng	100
17) 2-Methylphenol	8.364	107	314466	51.402	ng	98
18) Hexachloroethane	8.975	117	166221	50.683	ng	97
19) n-Nitroso-di-n-propyla...	8.722	70	262262	47.430	ng	96
20) 3+4-Methylphenols	8.693	107	414965	50.214	ng	98
22) Acetophenone	8.734	105	553779	50.902	ng	# 98
24) Nitrobenzene	9.116	77	401610	45.875	ng	99
25) Isophorone	9.646	82	691958	44.444	ng	99
26) 2-Nitrophenol	9.822	139	214091	51.254	ng	94
27) 2,4-Dimethylphenol	9.893	122	339731	54.593	ng	99
28) bis(2-Chloroethoxy)met...	10.116	93	433152	53.307	ng	99
29) 2,4-Dichlorophenol	10.369	162	356082	47.615	ng	100
30) 1,2,4-Trichlorobenzene	10.569	180	393525	46.831	ng	100
31) Naphthalene	10.758	128	1132472	47.354	ng	99
32) Benzoic acid	10.116	122	239406	50.906	ng	98
33) 4-Chloroaniline	10.881	127	145440	15.054	ng	99
34) Hexachlorobutadiene	11.034	225	262630	46.208	ng	99
35) Caprolactam	11.705	113	107069m	49.705	ng	
36) 4-Chloro-3-methylphenol	12.016	107	364743	48.034	ng	97
37) 2-Methylnaphthalene	12.363	142	790014	46.397	ng	99
38) 1-Methylnaphthalene	12.587	142	727348	45.350	ng	99
40) 1,2,4,5-Tetrachloroben...	12.734	216	418694	50.514	ng	99
41) Hexachlorocyclopentadiene	12.704	237	402820	95.574	ng	98
43) 2,4,6-Trichlorophenol	12.981	196	275506	48.459	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062921\
 Data File : BP006097.D
 Acq On : 29 Jun 2021 18:37
 Operator : CG/JU
 Sample : M2856-18MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SB-09-COMPMS

Manual Integrations
 APPROVED

mohammad
 6/30/2021 2:14:16 PM

Quant Time: Jun 30 01:12:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 28 11:34:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.063	196	296160	48.370	ng	97
46) 1,1'-Biphenyl	13.369	154	950522	50.063	ng	99
47) 2-Chloronaphthalene	13.416	162	754782	48.684	ng	99
48) 2-Nitroaniline	13.628	65	204892	50.263	ng	96
49) Acenaphthylene	14.257	152	1173627	49.342	ng	99
50) Dimethylphthalate	14.004	163	1027084	53.048	ng	99
51) 2,6-Dinitrotoluene	14.122	165	206968	47.031	ng	99
52) Acenaphthene	14.598	154	684485	47.387	ng	98
53) 3-Nitroaniline	14.457	138	154023	36.112	ng	98
54) 2,4-Dinitrophenol	14.675	184	222595	84.743	ng	96
55) Dibenzofuran	14.934	168	1081432	45.525	ng	98
56) 4-Nitrophenol	14.787	139	315777	91.341	ng	94
57) 2,4-Dinitrotoluene	14.916	165	280108	47.591	ng	98
58) Fluorene	15.581	166	867385	46.864	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.169	232	243116m	45.053	ng	
60) Diethylphthalate	15.357	149	927000	47.721	ng	99
61) 4-Chlorophenyl-phenyle...	15.575	204	445502	46.419	ng	99
62) 4-Nitroaniline	15.622	138	176119	40.118	ng	92
63) Azobenzene	15.869	77	806928	45.916	ng	95
65) 4,6-Dinitro-2-methylph...	15.681	198	147101	44.170	ng	98
66) n-Nitrosodiphenylamine	15.793	169	754078	50.177	ng	100
67) 4-Bromophenyl-phenylether	16.469	248	291268	50.074	ng	98
68) Hexachlorobenzene	16.592	284	343714	49.298	ng	95
69) Atrazine	16.751	200	276925	59.227	ng	99
70) Pentachlorophenol	16.940	266	383526	98.930	ng	99
71) Phenanthrene	17.328	178	1300304	47.778	ng	100
72) Anthracene	17.416	178	1342882	50.135	ng	99
73) Carbazole	17.692	167	1166306	45.352	ng	100
74) Di-n-butylphthalate	18.234	149	1562731	52.449	ng	99
75) Fluoranthene	19.286	202	1511407	45.729	ng	97
77) Benzidine	19.469	184	539074	61.213	ng	99
78) Pyrene	19.639	202	1548784	52.419	ng	99
80) Butylbenzylphthalate	20.504	149	679724	55.918	ng	97
81) Benzo(a)anthracene	21.345	228	1438029	47.070	ng	100
82) 3,3'-Dichlorobenzidine	21.275	252	447205	42.339	ng	98
83) Chrysene	21.398	228	1405818	47.509	ng	99
84) Bis(2-ethylhexyl)phtha...	21.257	149	1003533	56.291	ng	100
85) Di-n-octyl phthalate	22.175	149	1657815	51.982	ng	99
87) Indeno(1,2,3-cd)pyrene	26.292	276	1623902	47.242	ng	98
88) Benzo(b)fluoranthene	23.027	252	1409486	51.196	ng	100
89) Benzo(k)fluoranthene	23.075	252	1322076	49.522	ng	99
90) Benzo(a)pyrene	23.657	252	1315708	50.832	ng	98
91) Dibenzo(a,h)anthracene	26.304	278	1393009	47.083	ng	98
92) Benzo(g,h,i)perylene	27.068	276	1389438	47.540	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP062921\
Data File : BP006097.D
Acq On : 29 Jun 2021 18:37
Operator : CG/JU
Sample : M2856-18MS
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS

Manual Integrations
APPROVED

mohammad
6/30/2021 2:14:16 PM

Quant Time: Jun 30 01:12:11 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP060821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 28 11:34:30 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.899 min Scan# 818
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:152 Resp: 111275
Ion Ratio Lower Upper
152 100
150 149.8 125.6 188.4
115 60.6 48.5 72.7

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#2
1,4-Dioxane
Concen: 40.122 ng
RT: 3.346 min Scan# 44
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion: 88 Resp: 125700
Ion Ratio Lower Upper
88 100
58 58.9 46.7 70.1
43 23.4 20.3 30.5





#3
Pyridine
Concen: 42.326 ng
RT: 3.758 min Scan# 114
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion: 79 Resp: 311660
Ion Ratio Lower Upper
79 100
52 52.4 41.7 62.5
51 27.9 24.1 36.1

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#4
n-Nitrosodimethylamine
Concen: 42.348 ng
RT: 3.670 min Scan# 99
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion: 42 Resp: 144945
Ion Ratio Lower Upper
42 100
74 156.5 105.4 158.0
44 6.0 4.2 6.4





#5
2-Fluorophenol
Concen: 98.127 ng
RT: 5.481 min Scan# 407
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:112 Resp: 680471
Ion Ratio Lower Upper
112 100
64 52.2 41.4 62.2
63 27.2 23.8 35.8

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#6
Aniline
Concen: 43.326 ng
RT: 7.234 min Scan# 705
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37



Tgt Ion: 93 Resp: 434765
Ion Ratio Lower Upper
93 100
66 37.5 29.7 44.5
65 18.4 15.3 22.9









#7
Phenol-d6
Concen: 91.683 ng
RT: 7.087 min Scan# 680
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion: 99 Resp: 824077
Ion Ratio Lower Upper
99 100
42 14.7 13.4 20.0
71 36.0 30.6 46.0

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#8
2-Chlorophenol
Concen: 51.550 ng
RT: 7.469 min Scan# 745
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion: 128 Resp: 409873
Ion Ratio Lower Upper
128 100
130 31.6 12.4 52.4
64 40.4 20.3 60.3





#9
Benzaldehyde
Concen: 59.453 ng
RT: 7.046 min Scan# 673
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion: 77 Resp: 227808
Ion Ratio Lower Upper
77 100
105 99.6 76.7 116.7
106 96.2 76.2 116.2

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#10
Phenol
Concen: 52.868 ng
RT: 7.117 min Scan# 685
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion: 94 Resp: 474702
Ion Ratio Lower Upper
94 100
65 26.9 8.8 48.8
66 37.8 20.7 60.7





#11
bis(2-Chloroethyl)ether
Concen: 52.580 ng
RT: 7.328 min Scan# 721
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion: 93 Resp: 357759
Ion Ratio Lower Upper
93 100
63 65.8 46.2 86.2
95 32.7 12.2 52.2

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#12
1,3-Dichlorobenzene
Concen: 49.871 ng
RT: 7.787 min Scan# 799
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37



Tgt Ion: 146 Resp: 443387
Ion Ratio Lower Upper
146 100
148 63.7 51.5 77.3
75 30.7 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 49.826 ng
RT: 7.934 min Scan# 824
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:146 Resp: 451764
Ion Ratio Lower Upper
146 100
148 64.6 51.3 76.9
111 41.6 33.0 49.6

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#14
1,2-Dichlorobenzene
Concen: 49.365 ng
RT: 8.252 min Scan# 878
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:146 Resp: 427822
Ion Ratio Lower Upper
146 100
148 63.8 50.3 75.5
111 42.7 33.7 50.5





#15
Benzyl Alcohol
Concen: 50.722 ng
RT: 8.158 min Scan# 862
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion: 79 Resp: 343392
Ion Ratio Lower Upper
79 100
108 76.5 62.5 93.7
77 63.2 52.9 79.3

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 49.867 ng
RT: 8.428 min Scan# 908
Delta R.T. -0.012 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion: 45 Resp: 313644
Ion Ratio Lower Upper
45 100
77 22.2 2.0 42.0
79 16.8 0.0 36.7





#17
2-Methylphenol
Concen: 51.402 ng
RT: 8.364 min Scan# 897
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:107 Resp: 314466
Ion Ratio Lower Upper
107 100
108 110.9 87.7 131.5
77 47.3 40.2 60.2
79 46.5 39.0 58.4

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#18
Hexachloroethane
Concen: 50.683 ng
RT: 8.975 min Scan# 1001
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37



Tgt Ion:117 Resp: 166221
Ion Ratio Lower Upper
117 100
119 95.6 76.2 114.2
201 108.9 83.4 125.0






#19

n-Nitroso-di-n-propylamine
Concen: 47.430 ng
RT: 8.722 min Scan# 958
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:	70	Resp:	262262
Ion	Ratio	Lower	Upper
70	100		
42	46.3	40.8	61.2
101	10.7	8.8	13.2
130	24.1	18.9	28.3


Manual Integrations
APPROVED

mohammad
6/30/2021 2:14:16 PM






#20

3+4-Methylphenols
Concen: 50.214 ng
RT: 8.693 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:	107	Resp:	414965
Ion	Ratio	Lower	Upper
107	100		
108	90.5	71.0	111.0
77	33.8	16.6	56.6
79	29.9	12.7	52.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.705 min Scan# 1295
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:136 Resp: 413333
Ion Ratio Lower Upper
136 100
137 10.5 8.6 12.8
54 5.8 4.6 6.8
68 5.0 3.6 5.4




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#22
Acetophenone
Concen: 50.902 ng
RT: 8.734 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:105 Resp: 553779
Ion Ratio Lower Upper
105 100
71 3.1 1.9 2.9#
51 21.3 18.4 27.6
120 22.8 17.9 26.9









#23
Nitrobenzene-d5
Concen: 58.378 ng
RT: 9.069 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion: 82 Resp: 492168
Ion Ratio Lower Upper
82 100
128 44.9 35.7 53.5
54 40.8 35.0 52.4

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#24
Nitrobenzene
Concen: 45.875 ng
RT: 9.116 min Scan# 1025
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion: 77 Resp: 401610
Ion Ratio Lower Upper
77 100
123 48.0 38.2 57.2
65 12.7 11.4 17.2









#25
Isophorone
Concen: 44.444 ng
RT: 9.646 min Scan# 1115
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:	82	Resp:	691958
Ion	Ratio	Lower	Upper
82	100		
95	7.8	6.5	9.7
138	18.9	15.0	22.4

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#26
2-Nitrophenol
Concen: 51.254 ng
RT: 9.822 min Scan# 1145
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:	139	Resp:	214091
Ion	Ratio	Lower	Upper
139	100		
109	32.9	28.8	43.2
65	47.1	41.0	61.6









#27
2,4-Dimethylphenol
Concen: 54.593 ng
RT: 9.893 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion: 122 Resp: 339731
Ion Ratio Lower Upper
122 100
107 117.7 94.5 141.7
121 60.0 47.4 71.2

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#28
bis(2-Chloroethoxy)methane
Concen: 53.307 ng
RT: 10.116 min Scan# 1195
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion: 93 Resp: 433152
Ion Ratio Lower Upper
93 100
95 32.8 25.9 38.9
123 13.6 10.2 15.2





#29
2,4-Dichlorophenol
Concen: 47.615 ng
RT: 10.369 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:162 Resp: 356082
Ion Ratio Lower Upper
162 100
164 64.6 44.9 84.9
98 34.8 14.3 54.3

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#30
1,2,4-Trichlorobenzene
Concen: 46.831 ng
RT: 10.569 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:180 Resp: 393525
Ion Ratio Lower Upper
180 100
182 97.4 77.9 116.9
145 30.5 24.7 37.1





#31
Naphthalene
Concen: 47.354 ng
RT: 10.758 min Scan# 1304
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:128 Resp: 1132472
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.6
127 13.0 10.7 16.1




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#32
Benzoic acid
Concen: 50.906 ng
RT: 10.116 min Scan# 1195
Delta R.T. 0.035 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:122 Resp: 239406
Ion Ratio Lower Upper
122 100
105 124.5 103.2 143.2
77 92.9 76.2 116.2





#33
4-Chloroaniline
Concen: 15.054 ng
RT: 10.881 min Scan# 1325
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:127 Resp: 145440
Ion Ratio Lower Upper
127 100
129 33.3 26.1 39.1
65 27.7 22.5 33.7
92 19.9 15.0 22.6




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#34
Hexachlorobutadiene
Concen: 46.208 ng
RT: 11.034 min Scan# 1351
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:225 Resp: 262630
Ion Ratio Lower Upper
225 100
223 61.8 51.0 76.6
227 64.0 51.4 77.2





#35
Caprolactam
Concen: 49.705 ng m
RT: 11.705 min Scan# 1465
Delta R.T. 0.018 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:113 Resp: 107069
Ion Ratio Lower Upper
113 100
55 110.8 98.9 138.9
56 92.4 71.7 111.7

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#36
4-Chloro-3-methylphenol
Concen: 48.034 ng
RT: 12.016 min Scan# 1518
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37



Tgt Ion:107 Resp: 364743
Ion Ratio Lower Upper
107 100
144 26.1 21.1 31.7
142 79.0 66.0 99.0









#37
2-Methylnaphthalene
Concen: 46.397 ng
RT: 12.363 min Scan# 1577
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:142 Resp: 790014
Ion Ratio Lower Upper
142 100
141 89.3 71.8 107.6
115 37.3 31.6 47.4

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#38
1-Methylnaphthalene
Concen: 45.350 ng
RT: 12.587 min Scan# 1615
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:142 Resp: 727348
Ion Ratio Lower Upper
142 100
141 93.0 73.3 109.9
116 4.2 3.5 5.3





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.534 min Scan# 1946
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:164 Resp: 235480
Ion Ratio Lower Upper
164 100
162 103.9 80.2 120.2
160 45.8 34.5 51.7

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 50.514 ng
RT: 12.734 min Scan# 1640
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:216 Resp: 418694
Ion Ratio Lower Upper
216 100
214 77.8 61.9 92.9
179 21.1 16.8 25.2
108 19.1 14.9 22.3





#41
Hexachlorocyclopentadiene
Concen: 95.574 ng
RT: 12.704 min Scan# 1635
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion: 237 Resp: 402820
Ion Ratio Lower Upper
237 100
235 62.3 44.5 84.5
272 12.6 0.0 32.5




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#42
2,4,6-Tribromophenol
Concen: 91.280 ng
RT: 16.028 min Scan# 2200
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion: 330 Resp: 368864
Ion Ratio Lower Upper
330 100
332 97.9 77.8 116.8
141 28.9 25.4 38.2





#43
2,4,6-Trichlorophenol
Concen: 48.459 ng
RT: 12.981 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:196 Resp: 275506
Ion Ratio Lower Upper
196 100
198 94.7 79.5 119.3
200 30.2 25.0 37.6

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#44
2,4,5-Trichlorophenol
Concen: 48.370 ng
RT: 13.063 min Scan# 1696
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:196 Resp: 296160
Ion Ratio Lower Upper
196 100
198 96.6 74.7 112.1
97 48.2 37.4 56.0
132 27.9 20.0 30.0





#45
2-Fluorobiphenyl
Concen: 58.134 ng
RT: 13.157 min Scan# 1712
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:172 Resp: 1042502
Ion Ratio Lower Upper
172 100
171 35.4 27.9 41.9
170 22.9 18.6 27.8

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#46
1,1'-Biphenyl
Concen: 50.063 ng
RT: 13.369 min Scan# 1748
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:154 Resp: 950522
Ion Ratio Lower Upper
154 100
153 40.8 20.4 60.4
76 11.5 0.0 30.8





#47
2-Chloronaphthalene
Concen: 48.684 ng
RT: 13.416 min Scan# 1756
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:162 Resp: 754782
Ion Ratio Lower Upper
162 100
127 40.7 31.4 47.0
164 33.4 26.6 39.8

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#48
2-Nitroaniline
Concen: 50.263 ng
RT: 13.628 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion: 65 Resp: 204892
Ion Ratio Lower Upper
65 100
92 74.6 57.7 86.5
138 120.8 92.1 138.1





#49
Acenaphthylene
Concen: 49.342 ng
RT: 14.257 min Scan# 1899
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:152 Resp: 1173627
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.2
153 12.8 10.6 16.0




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#50
Dimethylphthalate
Concen: 53.048 ng
RT: 14.004 min Scan# 1856
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:163 Resp: 1027084
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.2
164 9.8 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 47.031 ng
RT: 14.122 min Scan# 1876
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:165 Resp: 206968
Ion Ratio Lower Upper
165 100
63 50.2 41.3 61.9
89 56.5 44.8 67.2




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#52
Acenaphthene
Concen: 47.387 ng
RT: 14.598 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:154 Resp: 684485
Ion Ratio Lower Upper
154 100
153 113.5 89.4 134.2
152 52.4 43.0 64.4





#53
3-Nitroaniline
Concen: 36.112 ng
RT: 14.457 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:138 Resp: 154023
Ion Ratio Lower Upper
138 100
108 10.1 8.4 12.6
92 111.7 87.8 131.8

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#54
2,4-Dinitrophenol
Concen: 84.743 ng
RT: 14.675 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:184 Resp: 222595
Ion Ratio Lower Upper
184 100
63 55.6 47.0 70.6
154 68.5 52.3 78.5





#55
Dibenzenofuran
Concen: 45.525 ng
RT: 14.934 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:168 Resp: 1081432
Ion Ratio Lower Upper
168 100
139 40.4 31.0 46.6
169 13.2 10.8 16.2




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#56
4-Nitrophenol
Concen: 91.341 ng
RT: 14.787 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:139 Resp: 315777
Ion Ratio Lower Upper
139 100
109 84.3 72.3 112.3
65 89.8 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 47.591 ng
RT: 14.916 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:	165	Resp:	280108
Ion Ratio	Lower	Upper	
165	100		
63	43.0	32.3	48.5
89	69.5	55.9	83.9
182	3.0	2.2	3.4

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#58
Fluorene
Concen: 46.864 ng
RT: 15.581 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:	166	Resp:	867385
Ion Ratio	Lower	Upper	
166	100		
165	98.6	78.0	117.0
167	13.6	10.5	15.7





#59
2,3,4,6-Tetrachlorophenol
Concen: 45.053 ng m
RT: 15.169 min Scan# 2054
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:232 Resp: 243116
Ion Ratio Lower Upper
232 100
131 39.8 32.2 48.4
130 2.5 2.0 3.0
166 25.2 19.4 29.2

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#60
Diethylphthalate
Concen: 47.721 ng
RT: 15.357 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:149 Resp: 927000
Ion Ratio Lower Upper
149 100
177 22.5 17.5 26.3
150 12.6 9.9 14.9





#61
4-Chlorophenyl-phenylether
Concen: 46.419 ng
RT: 15.575 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:204 Resp: 445502
Ion Ratio Lower Upper
204 100
206 32.7 26.6 40.0
141 60.1 48.6 73.0

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#62
4-Nitroaniline
Concen: 40.118 ng
RT: 15.622 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:138 Resp: 176119
Ion Ratio Lower Upper
138 100
92 50.3 31.2 71.2
108 87.4 78.6 118.6





#63
Azobenzene
Concen: 45.916 ng
RT: 15.869 min Scan# 2173
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion: 77 Resp: 806928
Ion Ratio Lower Upper
77 100
182 28.1 6.0 46.0
105 21.2 0.0 38.8
51 25.6 8.8 48.8




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.281 min Scan# 2413
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion: 188 Resp: 453185
Ion Ratio Lower Upper
188 100
94 7.9 5.6 8.4
80 8.8 6.3 9.5






#65
4,6-Dinitro-2-methylphenol
Concen: 44.170 ng
RT: 15.681 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS

Manual Integrations
APPROVED

mohammad
6/30/2021 2:14:16 PM







Tgt Ion:198 Resp: 147101
Ion Ratio Lower Upper
198 100
51 33.8 14.0 54.0
105 42.1 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 50.177 ng
RT: 15.793 min Scan# 2160
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:169 Resp: 754078
Ion Ratio Lower Upper
169 100
168 67.6 54.4 81.6
167 36.0 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 50.074 ng
RT: 16.469 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion: 248 Resp: 291268
Ion Ratio Lower Upper
248 100
250 96.8 79.8 119.8
141 70.9 57.8 86.6

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#68
Hexachlorobenzene
Concen: 49.298 ng
RT: 16.592 min Scan# 2296
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion: 284 Resp: 343714
Ion Ratio Lower Upper
284 100
142 28.7 26.9 40.3
249 30.0 24.5 36.7









#69
Atrazine
Concen: 59.227 ng
RT: 16.751 min Scan# 2323
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:200 Resp: 276925
Ion Ratio Lower Upper
200 100
173 25.8 4.1 44.1
215 47.6 27.7 67.7

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#70
Pentachlorophenol
Concen: 98.930 ng
RT: 16.940 min Scan# 2355
Delta R.T. -0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:266 Resp: 383526
Ion Ratio Lower Upper
266 100
268 64.8 51.6 77.4
264 63.5 50.3 75.5





#71
Phenanthrene
Concen: 47.778 ng
RT: 17.328 min Scan# 2421
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:178 Resp: 1300304
Ion Ratio Lower Upper
178 100
176 19.0 15.1 22.7
179 15.4 12.4 18.6

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#72
Anthracene
Concen: 50.135 ng
RT: 17.416 min Scan# 2436
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:178 Resp: 1342882
Ion Ratio Lower Upper
178 100
176 18.0 14.9 22.3
179 15.2 12.2 18.2





#73
 Carbazole
 Concen: 45.352 ng
 RT: 17.692 min Scan# 2483
 Delta R.T. 0.000 min
 Lab File: BP006097.D
 Acq: 29 Jun 2021 18:37

Instrument :
 BNA_P
 ClientSampleId :
 SB-09-COMPMS






Tgt Ion:167 Resp: 1166306

Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.1	25.7
139	13.0	10.7	16.1

Manual Integrations
 APPROVED




mohammad
 6/30/2021 2:14:16 PM




#74
 Di-n-butylphthalate
 Concen: 52.449 ng
 RT: 18.234 min Scan# 2575
 Delta R.T. 0.000 min
 Lab File: BP006097.D
 Acq: 29 Jun 2021 18:37

Tgt Ion:149 Resp: 1562731

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.3	10.9
104	5.2	4.5	6.7






#75
Fluoranthene
Concen: 45.729 ng
RT: 19.286 min Scan# 2754
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS

Manual Integrations
APPROVED

mohammad
6/30/2021 2:14:16 PM







Tgt Ion:202 Resp: 1511407
Ion Ratio Lower Upper
202 100
101 9.6 0.0 27.6
203 17.3 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.357 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:240 Resp: 423231
Ion Ratio Lower Upper
240 100
120 9.7 6.5 9.7#
236 25.1 20.6 31.0





#77
Benzidine
Concen: 61.213 ng
RT: 19.469 min Scan# 2785
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:184 Resp: 539074
Ion Ratio Lower Upper
184 100
185 14.0 11.0 16.4
183 11.7 9.5 14.3




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#78
Pyrene
Concen: 52.419 ng
RT: 19.639 min Scan# 2814
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:202 Resp: 1548784
Ion Ratio Lower Upper
202 100
200 20.9 16.3 24.5
203 17.4 14.2 21.2









#79
Terphenyl-d14
Concen: 69.607 ng
RT: 19.834 min Scan# 2847
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:244 Resp: 1573304
Ion Ratio Lower Upper
244 100
212 6.9 6.0 9.0
122 8.5 7.1 10.7

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#80
Butylbenzylphthalate
Concen: 55.918 ng
RT: 20.504 min Scan# 2961
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:149 Resp: 679724
Ion Ratio Lower Upper
149 100
91 66.0 54.6 81.8
206 22.6 16.7 25.1





#81
Benzo(a)anthracene
Concen: 47.070 ng
RT: 21.345 min Scan# 3104
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:228 Resp: 1438029
Ion Ratio Lower Upper
228 100
226 26.7 21.6 32.4
229 19.6 15.8 23.6




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#82
3,3'-Dichlorobenzidine
Concen: 42.339 ng
RT: 21.275 min Scan# 3092
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:252 Resp: 447205
Ion Ratio Lower Upper
252 100
254 65.0 50.6 76.0
126 10.1 7.0 10.6









#83
Chrysene
Concen: 47.509 ng
RT: 21.398 min Scan# 3113
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37

Tgt Ion:228 Resp: 1405818
Ion Ratio Lower Upper
228 100
226 30.0 23.8 35.6
229 19.5 15.4 23.2

Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#84
Bis(2-ethylhexyl)phthalate
Concen: 56.291 ng
RT: 21.257 min Scan# 3089
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:149 Resp: 1003533
Ion Ratio Lower Upper
149 100
167 27.5 21.9 32.9
279 5.4 4.3 6.5





#85
Di-n-octyl phthalate
Concen: 51.982 ng
RT: 22.175 min Scan# 3245
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:149 Resp: 1657815
Ion Ratio Lower Upper
149 100
167 1.7 1.4 2.0
43 6.5 5.6 8.4

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#86
Perylene-d12
Concen: 20.000 ng
RT: 23.757 min Scan# 3514
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:264 Resp: 401969
Ion Ratio Lower Upper
264 100
260 23.5 18.6 28.0
265 22.0 17.6 26.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 47.242 ng
RT: 26.292 min Scan# 3945
Delta R.T. 0.000 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS



Tgt Ion:276 Resp: 1623902
Ion Ratio Lower Upper
276 100
138 20.6 15.3 22.9
277 25.8 20.4 30.6




Manual Integrations
APPROVED


mohammad
6/30/2021 2:14:16 PM



#88
Benzo(b)fluoranthene
Concen: 51.196 ng
RT: 23.027 min Scan# 3390
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37


Tgt Ion:252 Resp: 1409486
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 8.1 6.2 9.4





#89
Benzo(k)fluoranthene
Concen: 49.522 ng
RT: 23.075 min Scan# 3398
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:252 Resp: 1322076
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 8.7 6.0 9.0

Manual Integrations
APPROVED



mohammad
6/30/2021 2:14:16 PM




#90
Benzo(a)pyrene
Concen: 50.832 ng
RT: 23.657 min Scan# 3497
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Tgt Ion:252 Resp: 1315708
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 9.4 6.5 9.7





#91
Dibenzo(a,h)anthracene
Concen: 47.083 ng
RT: 26.304 min Scan# 3947
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37




Instrument :
BNA_P
ClientSampleId :
SB-09-COMPMS




Tgt Ion:278 Resp: 1393009
Ion Ratio Lower Upper
278 100
139 13.1 9.3 13.9
279 23.7 18.8 28.2

Manual Integrations
APPROVED

mohammad
6/30/2021 2:14:16 PM



#92
Benzo(g,h,i)perylene
Concen: 47.540 ng
RT: 27.068 min Scan# 4077
Delta R.T. 0.006 min
Lab File: BP006097.D
Acq: 29 Jun 2021 18:37



Tgt Ion:276 Resp: 1389438
Ion Ratio Lower Upper
276 100
277 23.7 18.9 28.3
138 18.1 12.6 18.8

