

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP063022\
 Data File : BP010917.D
 Acq On : 30 Jun 2022 13:11
 Operator : CG/JU
 Sample : N3523-07
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0JX5

Quant Time: Jul 01 02:32:47 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 27 23:12:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	450615	20.000	ng/u1	0.00
20) Naphthalene-d8	10.498	136	1785376	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.363	164	892941	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.122	188	1585780	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.239	240	1527260	20.000	ng/u1	# 0.00
88) Perylene-d12	23.598	264	1701169	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	52647	4.420	ng/uL	0.00
4) Pyridine-d5	3.616	84	188163	5.934	ng/u1	0.00
7) Phenol-d5	6.881	99	219822	6.162	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	693020	29.707	ng/u1	0.00
11) 2-Chlorophenol-d4	7.240	132	678935	22.941	ng/u1	0.00
15) 4-Methylphenol-d8	8.422	113	403345	14.304	ng/u1	0.00
21) Nitrobenzene-d5	8.863	128	395875	34.831	ng/u1	0.00
24) 2-Nitrophenol-d4	9.587	143	362049	38.503	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.122	165	601289	25.880	ng/u1	0.00
31) 4-Chloroaniline-d4	10.646	131	620698	16.366	ng/u1	0.00
46) Dimethylphthalate-d6	13.781	166	1856609	30.353	ng/u1	0.00
49) Acenaphthylene-d8	14.051	160	2301158	29.477	ng/u1	0.00
54) 4-Nitrophenol-d4	14.575	143	47194	4.940	ng/u1	0.00
60) Fluorene-d10	15.357	176	1630505	31.063	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.481	200	149096	28.378	ng/u1	0.00
73) Anthracene-d10	17.222	188	2515586	35.655	ng/u1	-0.01
81) Pyrene-d10	19.474	212	2782434	33.440	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.445	264	3144415	37.405	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP063022\
Data File : BP010917.D
Acq On : 30 Jun 2022 13:11
Operator : CG/JU
Sample : N3523-07
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0JX5

Quant Time: Jul 01 02:32:47 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 27 23:12:49 2022
Response via : Initial Calibration

