

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP063022\
 Data File : BP010919.D
 Acq On : 30 Jun 2022 14:19
 Operator : CG/JU
 Sample : PB145923BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS923

Quant Time: Jul 01 02:33:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 27 23:12:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	490094	20.000	ng/ul	0.00
20) Naphthalene-d8	10.498	136	1963549	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.363	164	974334	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.128	188	1625734	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.239	240	1432315	20.000	ng/ul	# 0.00
88) Perylene-d12	23.598	264	1653342	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	80618	6.223	ng/uL	0.00
4) Pyridine-d5	3.605	84	1107319	32.108	ng/ul	0.00
7) Phenol-d5	6.887	99	1336145	34.440	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	837723	33.017	ng/ul	0.00
11) 2-Chlorophenol-d4	7.240	132	1115292	34.650	ng/ul	0.00
15) 4-Methylphenol-d8	8.428	113	1040954	33.942	ng/ul	0.00
21) Nitrobenzene-d5	8.869	128	501780	40.143	ng/ul	0.00
24) 2-Nitrophenol-d4	9.587	143	491079	47.486	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.122	165	889321	34.804	ng/ul	0.00
31) 4-Chloroaniline-d4	10.646	131	1226667	29.409	ng/ul	0.00
46) Dimethylphthalate-d6	13.781	166	2176147	32.605	ng/ul	0.00
49) Acenaphthylene-d8	14.051	160	2841318	33.356	ng/ul	0.00
54) 4-Nitrophenol-d4	14.581	143	370835	35.573	ng/ul	0.00
60) Fluorene-d10	15.357	176	1824273	31.851	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.481	200	236224	43.857	ng/ul	0.00
73) Anthracene-d10	17.228	188	2455450	33.947	ng/ul	0.00
81) Pyrene-d10	19.474	212	2545765	32.624	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.445	264	2934629	35.919	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	170868	12.367	ng/uL#	87
5) Pyridine	3.628	79	1105005	31.630	ng/ul#	73
6) Benzaldehyde	6.852	77	887876	43.374	ng/ul	95
8) Phenol	6.910	94	1370657	33.033	ng/ul#	86
10) Bis(2-Chloroethyl)ether	7.140	93	1098266	32.955	ng/ul	89
12) 2-Chlorophenol	7.269	128	1111298	33.447	ng/ul	92
13) 2-Methylphenol	8.157	108	1000332	32.382	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.246	45	1443173	32.276	ng/ul#	97
16) Acetophenone	8.534	105	1548807	32.884	ng/ul#	87
17) N-Nitroso-di-n-propyla...	8.528	70	742538	31.178	ng/ul	88
18) 4-Methylphenol	8.487	108	1055181	32.800	ng/ul	97
19) Hexachloroethane	8.781	117	436483	32.439	ng/ul#	83
22) Nitrobenzene	8.910	77	1151728	36.922	ng/ul#	90
23) Isophorone	9.440	82	2000004	30.573	ng/ul	98
25) 2-Nitrophenol	9.616	139	536561	42.845	ng/ul#	82
26) 2,4-Dimethylphenol	9.687	107	1081920	31.621	ng/ul	96
27) Bis(2-Chloroethoxy)met...	9.928	93	1335461	32.362	ng/ul	97
29) 2,4-Dichlorophenol	10.151	162	876362	33.006	ng/ul	99
30) Naphthalene	10.551	128	3312871	32.372	ng/ul	99
32) 4-Chloroaniline	10.669	127	1177643	28.497	ng/ul	99
33) Hexachlorobutadiene	10.834	225	522064	30.776	ng/ul	96
34) Caprolactam	11.463	113	246019	29.495	ng/ul	82
35) 4-Chloro-3-methylphenol	11.804	107	910672	31.646	ng/ul	95
36) 2-Methylnaphthalene	12.169	142	2044673	30.854	ng/ul	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP063022\
 Data File : BP010919.D
 Acq On : 30 Jun 2022 14:19
 Operator : CG/JU
 Sample : PB145923BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS923

Quant Time: Jul 01 02:33:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 27 23:12:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.392	142	2035793	30.922	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.540	216	930664	33.239	ng/ul	99
40) Hexachlorocyclopentadiene	12.516	237	530437	29.898	ng/ul	96
41) 2,4,6-Trichlorophenol	12.787	196	580046	35.933	ng/ul	98
42) 2,4,5-Trichlorophenol	12.857	196	621986	35.906	ng/ul	100
43) 1,1'-Biphenyl	13.192	154	2514068	32.777	ng/ul#	97
44) 2-Chloronaphthalene	13.228	162	1932482	33.445	ng/ul	97
45) 2-Nitroaniline	13.445	65	505442	41.230	ng/ul#	78
47) Dimethylphthalate	13.828	163	2163136	32.363	ng/ul	98
48) 2,6-Dinitrotoluene	13.945	165	412382	41.177	ng/ul	94
50) Acenaphthylene	14.081	152	2935000	31.461	ng/ul	99
51) 3-Nitroaniline	14.275	138	444490	37.021	ng/ul	89
52) Acenaphthene	14.428	153	1984792	32.973	ng/ul	99
53) 2,4-Dinitrophenol	14.481	184	141871	36.900	ng/ul#	79
55) 4-Nitrophenol	14.592	109	278142	33.555	ng/ul#	76
56) Dibenzofuran	14.763	168	2597615	31.504	ng/ul	97
57) 2,4-Dinitrotoluene	14.734	165	563921	42.141	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	14.992	232	447016	36.035	ng/ul	92
59) Diethylphthalate	15.204	149	2063359	31.051	ng/ul	98
61) Fluorene	15.416	166	2066193	31.427	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.416	204	955141	31.034	ng/ul	97
63) 4-Nitroaniline	15.445	138	461824	41.602	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.498	198	241526	41.497	ng/ul	95
67) N-Nitrosodiphenylamine	15.633	169	1705600	34.638	ng/ul	99
68) 4-Bromophenyl-phenylether	16.316	248	550761	33.803	ng/ul	97
69) Hexachlorobenzene	16.422	284	619101	32.806	ng/ul	97
70) Atrazine	16.604	200	523750	32.082	ng/ul	98
71) Pentachlorophenol	16.769	266	302569	31.137	ng/ul	96
72) Phenanthrene	17.169	178	2942175	33.984	ng/ul	99
74) Anthracene	17.263	178	2937347	33.734	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.151	216	944808	35.218	ng/uL	97
76) Pentachlorobenzene	14.675	250	815694	35.452	ng/uL	99
77) Carbazole	17.539	167	2646844	33.770	ng/ul	99
78) Di-n-butylphthalate	18.104	149	2904770	31.100	ng/ul	99
80) Fluoranthene	19.151	202	3026297	31.533	ng/ul#	89
82) Pyrene	19.504	202	3123552	32.228	ng/ul#	84
83) Butylbenzylphthalate	20.398	149	1180588	34.606	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.163	252	1035549	35.013	ng/ul#	97
85) Benzo(a)anthracene	21.221	228	3093087	34.529	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.163	149	1628722	30.149	ng/ul#	96
87) Chrysene	21.274	228	2905202	33.844	ng/ul	100
89) Di-n-octyl phthalate	22.080	149	2836355	28.902	ng/ul	100
90) Benzo(b)fluoranthene	22.880	252	3548541	34.513	ng/ul#	96
91) Benzo(k)fluoranthene	22.927	252	3392535	34.713	ng/ul#	95
93) Benzo(a)pyrene	23.498	252	3196782	31.962	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.051	276	4429512	36.702	ng/ul#	89
95) Dibenzo(a,h)anthracene	26.068	278	3698490	35.726	ng/ul#	93
96) Benzo(g,h,i)perylene	26.803	276	3611837	35.182	ng/ul#	89





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.367 ng/uL
RT: 3.234 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion: 88 Resp: 170868
Ion Ratio Lower Upper
88 100
43 29.2 35.8 53.6#
58 78.8 68.1 102.1



#5
Pyridine
Concen: 31.630 ng/uL
RT: 3.628 min Scan# 92
Delta R.T. -0.000 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion: 79 Resp: 1105005
Ion Ratio Lower Upper
79 100
52 74.4 83.8 125.8#
51 35.0 41.0 61.4#









#6
Benzaldehyde
Concen: 43.374 ng/ul
RT: 6.852 min Scan# 640
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion: 77 Resp: 887876
Ion Ratio Lower Upper
77 100
105 100.4 76.0 114.0
106 95.3 73.2 109.8



#8
Phenol
Concen: 33.033 ng/ul
RT: 6.910 min Scan# 650
Delta R.T. -0.000 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion: 94 Resp: 1370657
Ion Ratio Lower Upper
94 100
65 27.9 27.4 41.2
66 35.5 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 32.955 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Instrument :
BNA_P
ClientSampleId :
SLCS923

Tgt Ion: 93 Resp: 1098266




Ion	Ratio	Lower	Upper
93	100		
63	77.6	72.4	108.6
95	32.1	27.3	40.9




#12
2-Chlorophenol
Concen: 33.447 ng/ul
RT: 7.269 min Scan# 711
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:128 Resp: 1111298

Ion	Ratio	Lower	Upper
128	100		
64	47.4	44.8	67.2
130	31.7	25.8	38.6





#13

2-Methylphenol

Concen: 32.382 ng/ul

RT: 8.157 min Scan# 862

Delta R.T. 0.000 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :





SLCS923

Tgt Ion:108 Resp: 1000332

Ion Ratio Lower Upper

108 100

107 90.4 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 32.276 ng/ul

RT: 8.246 min Scan# 877

Delta R.T. -0.006 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19




Tgt Ion: 45 Resp: 1443173


Ion Ratio Lower Upper

45 100

77 14.0 11.2 16.8

79 11.2 6.9 10.3#









#16
Acetophenone
Concen: 32.884 ng/ul
RT: 8.534 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:105 Resp: 1548807
Ion Ratio Lower Upper
105 100
77 75.3 67.5 101.3
51 25.4 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 31.178 ng/ul
RT: 8.528 min Scan# 925
Delta R.T. 0.000 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion: 70 Resp: 742538
Ion Ratio Lower Upper
70 100
42 55.7 53.8 80.8
101 9.8 7.8 11.6
130 22.4 15.4 23.2









#18
4-Methylphenol
Concen: 32.800 ng/ul
RT: 8.487 min Scan# 918
Delta R.T. -0.000 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:108 Resp: 1055181
Ion Ratio Lower Upper
108 100
107 113.5 88.3 132.5



#19
Hexachloroethane
Concen: 32.439 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:117 Resp: 436483
Ion Ratio Lower Upper
117 100
201 90.0 85.7 128.5
199 56.6 57.4 86.0#









#22
Nitrobenzene
Concen: 36.922 ng/ul
RT: 8.910 min Scan# 990
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion: 77 Resp: 1151728
Ion Ratio Lower Upper
77 100
123 47.8 31.5 47.3#
65 13.6 11.4 17.2



#23
Isophorone
Concen: 30.573 ng/ul
RT: 9.440 min Scan# 1080
Delta R.T. 0.000 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion: 82 Resp: 2000004
Ion Ratio Lower Upper
82 100
95 6.9 6.6 10.0
138 18.0 13.8 20.6









#25
2-Nitrophenol
Concen: 42.845 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:139 Resp: 536561
Ion Ratio Lower Upper
139 100
65 51.6 54.2 81.4#
109 29.2 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 31.621 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:107 Resp: 1081920
Ion Ratio Lower Upper
107 100
121 51.6 39.1 58.7
122 86.5 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 32.362 ng/ul

RT: 9.928 min Scan# 1163

Delta R.T. -0.000 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :

SLCS923





Tgt Ion: 93 Resp: 1335461

Ion Ratio Lower Upper

93 100

95 32.0 26.6 39.8

123 11.5 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 33.006 ng/ul

RT: 10.151 min Scan# 1201

Delta R.T. -0.006 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19




Tgt Ion:162 Resp: 876362


Ion Ratio Lower Upper

162 100

164 65.0 52.5 78.7

98 38.0 31.4 47.0









#30
Naphthalene
Concen: 32.372 ng/ul
RT: 10.551 min Scan# 1269
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:128 Resp: 3312871
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 28.497 ng/ul
RT: 10.669 min Scan# 1289
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:127 Resp: 1177643
Ion Ratio Lower Upper
127 100
129 33.1 26.2 39.2









#33
Hexachlorobutadiene
Concen: 30.776 ng/ul
RT: 10.834 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:225 Resp: 522064
Ion Ratio Lower Upper
225 100
223 63.2 53.7 80.5
227 64.2 49.4 74.2



#34
Caprolactam
Concen: 29.495 ng/ul
RT: 11.463 min Scan# 1424
Delta R.T. 0.012 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:113 Resp: 246019
Ion Ratio Lower Upper
113 100
55 162.0 153.8 230.8
56 119.6 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 31.646 ng/ul
RT: 11.804 min Scan# 1482
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:107 Resp: 910672
Ion Ratio Lower Upper
107 100
144 25.4 18.6 27.8
142 79.3 59.6 89.4



#36
2-Methylnaphthalene
Concen: 30.854 ng/ul
RT: 12.169 min Scan# 1544
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:142 Resp: 2044673
Ion Ratio Lower Upper
142 100
141 89.5 70.1 105.1









#37
1-Methylnaphthalene
Concen: 30.922 ng/ul
RT: 12.392 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:142 Resp: 2035793
Ion Ratio Lower Upper
142 100
141 92.2 73.3 109.9
116 3.7 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.239 ng/ul
RT: 12.540 min Scan# 1607
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:216 Resp: 930664
Ion Ratio Lower Upper
216 100
214 78.4 63.0 94.6
179 21.6 18.1 27.1
108 18.9 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 29.898 ng/ul
RT: 12.516 min Scan# 1603
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:237 Resp: 530437
Ion Ratio Lower Upper
237 100
235 63.3 52.8 79.2
272 12.5 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 35.933 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:196 Resp: 580046
Ion Ratio Lower Upper
196 100
198 94.8 78.2 117.4
200 31.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 35.906 ng/ul
RT: 12.857 min Scan# 1661
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:196 Resp: 621986
Ion Ratio Lower Upper
196 100
198 96.7 77.0 115.4
200 31.1 24.8 37.2



#43
1,1'-Biphenyl
Concen: 32.777 ng/ul
RT: 13.192 min Scan# 1718
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:154 Resp: 2514068
Ion Ratio Lower Upper
154 100
153 40.9 31.9 47.9
76 14.0 9.2 13.8#









#44
2-Chloronaphthalene
Concen: 33.445 ng/ul
RT: 13.228 min Scan# 1724
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:162 Resp: 1932482
Ion Ratio Lower Upper
162 100
164 32.8 25.5 38.3
127 42.4 31.9 47.9



#45
2-Nitroaniline
Concen: 41.230 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion: 65 Resp: 505442
Ion Ratio Lower Upper
65 100
92 69.9 49.0 73.6
138 110.4 66.4 99.6#









#47
Dimethylphthalate
Concen: 32.363 ng/ul
RT: 13.828 min Scan# 1826
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:163 Resp: 2163136
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.6 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 41.177 ng/ul
RT: 13.945 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:165 Resp: 412382
Ion Ratio Lower Upper
165 100
89 60.9 52.9 79.3
121 22.7 19.4 29.2





#50

Acenaphthylene

Concen: 31.461 ng/ul

RT: 14.081 min Scan# 1869

Delta R.T. -0.006 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :

SLCS923





Tgt Ion:152 Resp: 2935000

Ion Ratio Lower Upper

152 100

151 20.2 16.5 24.7

153 13.4 10.1 15.1



#51

3-Nitroaniline

Concen: 37.021 ng/ul

RT: 14.275 min Scan# 1902

Delta R.T. -0.006 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19




Tgt Ion:138 Resp: 444490


Ion Ratio Lower Upper

138 100

108 11.6 11.0 16.4

92 117.3 104.6 156.8









#52
Acenaphthene
Concen: 32.973 ng/ul
RT: 14.428 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:153 Resp: 1984792
Ion Ratio Lower Upper
153 100
152 46.8 36.2 54.2
154 88.2 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 36.900 ng/ul
RT: 14.481 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:184 Resp: 141871
Ion Ratio Lower Upper
184 100
63 66.3 73.8 110.6#
154 60.7 57.2 85.8









#55
4-Nitrophenol
Concen: 33.555 ng/ul
RT: 14.592 min Scan# 1956
Delta R.T. -0.000 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:109 Resp: 278142
Ion Ratio Lower Upper
109 100
139 144.9 75.6 113.4#
65 139.1 107.6 161.4



#56
Dibenzofuran
Concen: 31.504 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:168 Resp: 2597615
Ion Ratio Lower Upper
168 100
139 39.3 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 42.141 ng/ul
RT: 14.734 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:165 Resp: 563921
Ion Ratio Lower Upper
165 100
63 45.9 47.0 70.6#
182 2.8 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 36.035 ng/ul
RT: 14.992 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:232 Resp: 447016
Ion Ratio Lower Upper
232 100
131 50.8 35.8 53.8
130 2.5 1.8 2.8
166 33.4 23.7 35.5









#59
Diethylphthalate
Concen: 31.051 ng/ul
RT: 15.204 min Scan# 2060
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:149 Resp: 2063359
Ion Ratio Lower Upper
149 100
177 21.0 17.6 26.4
150 12.3 10.3 15.5



#61
Fluorene
Concen: 31.427 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:166 Resp: 2066193
Ion Ratio Lower Upper
166 100
165 98.1 77.4 116.2
167 13.9 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 31.034 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:204 Resp: 955141
Ion Ratio Lower Upper
204 100
206 32.8 26.0 39.0
141 65.2 49.3 73.9



#63
4-Nitroaniline
Concen: 41.602 ng/ul
RT: 15.445 min Scan# 2101
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:138 Resp: 461824
Ion Ratio Lower Upper
138 100
92 52.9 46.2 69.2
108 73.9 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 41.497 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:198 Resp: 241526
Ion Ratio Lower Upper
198 100
182 4.3 4.3 6.5
77 26.4 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 34.638 ng/ul
RT: 15.633 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:169 Resp: 1705600
Ion Ratio Lower Upper
169 100
168 66.1 52.6 78.8
167 36.4 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 33.803 ng/ul
RT: 16.316 min Scan# 2249
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:248 Resp: 550761
Ion Ratio Lower Upper
248 100
250 98.4 80.8 121.2
141 84.0 63.9 95.9



#69
Hexachlorobenzene
Concen: 32.806 ng/ul
RT: 16.422 min Scan# 2267
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:284 Resp: 619101
Ion Ratio Lower Upper
284 100
142 37.6 27.8 41.6
249 30.8 25.4 38.2





#70

Atrazine

Concen: 32.082 ng/ul

RT: 16.604 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :

SLCS923





Tgt Ion:200 Resp: 523750

Ion Ratio Lower Upper

200 100

173 28.4 20.7 31.1

215 47.9 38.6 57.8



#71

Pentachlorophenol

Concen: 31.137 ng/ul

RT: 16.769 min Scan# 2326

Delta R.T. -0.012 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19




Tgt Ion:266 Resp: 302569


Ion Ratio Lower Upper

266 100

264 62.3 51.4 77.2

268 62.4 53.5 80.3









#72
Phenanthrene
Concen: 33.984 ng/ul
RT: 17.169 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:178 Resp: 2942175
Ion Ratio Lower Upper
178 100
179 15.3 12.7 19.1
176 18.9 14.6 22.0



#74
Anthracene
Concen: 33.734 ng/ul
RT: 17.263 min Scan# 2410
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:178 Resp: 2937347
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 18.4 14.5 21.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.218 ng/uL
RT: 13.151 min Scan# 1711
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:216 Resp: 944808
Ion Ratio Lower Upper
216 100
214 78.0 65.0 97.4
218 48.1 39.1 58.7



#76
Pentachlorobenzene
Concen: 35.452 ng/uL
RT: 14.675 min Scan# 1970
Delta R.T. -0.012 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:250 Resp: 815694
Ion Ratio Lower Upper
250 100
248 61.9 49.8 74.6
252 63.0 51.2 76.8





#77

Carbazole

Concen: 33.770 ng/ul

RT: 17.539 min Scan# 2457

Delta R.T. -0.006 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :

SLCS923





Tgt Ion:167 Resp: 2646844

Ion Ratio Lower Upper

167 100

166 21.4 16.9 25.3

139 13.2 10.6 16.0



#78

Di-n-butylphthalate

Concen: 31.100 ng/ul

RT: 18.104 min Scan# 2553

Delta R.T. -0.012 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19




Tgt Ion:149 Resp: 2904770


Ion Ratio Lower Upper

149 100

150 9.2 7.1 10.7

104 5.3 4.6 6.8









#80
Fluoranthene
Concen: 31.533 ng/u1
RT: 19.151 min Scan# 2731
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:202 Resp: 3026297
Ion Ratio Lower Upper
202 100
101 13.3 7.1 10.7#
100 10.1 5.7 8.5#



#82
Pyrene
Concen: 32.228 ng/u1
RT: 19.504 min Scan# 2791
Delta R.T. -0.006 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:202 Resp: 3123552
Ion Ratio Lower Upper
202 100
101 16.4 7.9 11.9#
100 13.2 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 34.606 ng/ul
RT: 20.398 min Scan# 2943
Delta R.T. -0.012 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19




Instrument :
BNA_P
ClientSampleId :
SLCS923


Tgt Ion:149 Resp: 1180588
Ion Ratio Lower Upper
149 100
91 70.9 60.0 90.0
206 18.9 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 35.013 ng/ul
RT: 21.163 min Scan# 3073
Delta R.T. -0.012 min
Lab File: BP010919.D
Acq: 30 Jun 2022 14:19

Tgt Ion:252 Resp: 1035549
Ion Ratio Lower Upper
252 100
254 65.1 51.0 76.4
126 13.9 8.6 12.8





#85

Benzo(a)anthracene

Concen: 34.529 ng/ul

RT: 21.221 min Scan# 3083

Delta R.T. -0.012 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :

SLCS923





Tgt Ion:228 Resp: 3093087

Ion Ratio Lower Upper

228 100

229 19.7 15.0 22.4

226 27.1 20.6 31.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 30.149 ng/ul

RT: 21.163 min Scan# 3073

Delta R.T. -0.012 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19




Tgt Ion:149 Resp: 1628722


Ion Ratio Lower Upper

149 100

167 26.2 22.6 33.8

279 3.6 4.2 6.4#





#87

Chrysene

Concen: 33.844 ng/ul

RT: 21.274 min Scan# 3092

Delta R.T. -0.006 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :

SLCS923





Tgt Ion:228 Resp: 2905202

Ion Ratio Lower Upper

228 100

226 29.7 23.6 35.4

229 19.8 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 28.902 ng/ul




RT: 22.080 min Scan# 3229


Delta R.T. -0.012 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Tgt Ion:149 Resp: 2836355





#90

Benzo(b)fluoranthene

Concen: 34.513 ng/ul

RT: 22.880 min Scan# 3365

Delta R.T. -0.012 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :

SLCS923





Tgt Ion:252 Resp: 3548541

Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.6

125 12.6 6.2 9.2#



#91

Benzo(k)fluoranthene

Concen: 34.713 ng/ul

RT: 22.927 min Scan# 3373

Delta R.T. -0.012 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19




Tgt Ion:252 Resp: 3392535


Ion Ratio Lower Upper

252 100

253 22.3 17.0 25.4

125 12.8 6.2 9.4#





#93

Benzo(a)pyrene

Concen: 31.962 ng/ul

RT: 23.498 min Scan# 3470

Delta R.T. -0.012 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19

Instrument :

BNA_P

ClientSampleId :

SLCS923





Tgt Ion:252 Resp: 3196782

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 14.3 6.6 10.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 36.702 ng/ul

RT: 26.051 min Scan# 3904

Delta R.T. -0.018 min

Lab File: BP010919.D

Acq: 30 Jun 2022 14:19




Tgt Ion:276 Resp: 4429512


Ion Ratio Lower Upper

276 100

138 29.2 14.6 22.0#

277 25.3 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 35.726 ng/ul
 RT: 26.068 min Scan# 3908
 Delta R.T. -0.024 min
 Lab File: BP010919.D
 Acq: 30 Jun 2022 14:19

Instrument :
 BNA_P
 ClientSampleId :
 SLCS923

Tgt Ion:278 Resp: 3698490
 Ion Ratio Lower Upper
 278 100
 139 20.3 9.4 14.2#
 279 23.6 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 35.182 ng/ul
 RT: 26.803 min Scan# 4032
 Delta R.T. -0.018 min
 Lab File: BP010919.D
 Acq: 30 Jun 2022 14:19

Tgt Ion:276 Resp: 3611837
 Ion Ratio Lower Upper
 276 100
 138 26.4 12.7 19.1#
 277 23.5 18.2 27.4

