

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP070120\
 Data File : BP002606.D
 Acq On : 01 Jul 2020 16:15
 Operator : CG/JU
 Sample : L3160-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MUDDY-DRUM

Quant Time: Jul 01 18:27:01 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP062920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 29 16:51:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	268210	20.00	ng	0.00
21) Naphthalene-d8	10.34	136	1055985	20.00	ng	0.00
39) Acenaphthene-d10	14.22	164	643066	20.00	ng	0.00
64) Phenanthrene-d10	16.97	188	1362059	20.00	ng	0.00
76) Chrysene-d12	21.09	240	1253052	20.00	ng	0.00
86) Perylene-d12	23.28	264	1137612	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	1710010	107.95	ng	0.00
7) Phenol-d6	6.76	99	2322838	105.06	ng	0.00
23) Nitrobenzene-d5	8.72	82	1610872	78.60	ng	0.00
42) 2,4,6-Tribromophenol	15.72	330	860612	110.03	ng	0.00
45) 2-Fluorobiphenyl	12.83	172	3276407	74.99	ng	0.00
79) Terphenyl-d14	19.57	244	4413673	76.03	ng	0.00
Target Compounds						
31) Naphthalene	10.39	128	1226104	21.933	ng	100
50) Dimethylphthalate	13.68	163	561811	11.123	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP070120\
Data File : BP002606.D
Acq On : 01 Jul 2020 16:15
Operator : CG/JU
Sample : L3160-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
MUDDY-DRUM

Quant Time: Jul 01 18:27:01 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP062920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 29 16:51:57 2020
Response via : Initial Calibration

