

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070124\
 Data File : BP020713.D
 Acq On : 01 Jul 2024 10:56
 Operator : MA/JU
 Sample : PB161602BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK602

Quant Time: Jul 02 00:53:48 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 25 18:30:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	204287	20.000	ng/u1	-0.03
20) Naphthalene-d8	10.499	136	893883	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.369	164	614693	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	1231273	20.000	ng/u1	0.00
79) Chrysene-d12	21.651	240	838199	20.000	ng/u1	0.01
88) Perylene-d12	25.021	264	918828	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	26436	5.557	ng/uL	-0.01
4) Pyridine-d5	3.670	84	390960	28.198	ng/u1	-0.02
7) Phenol-d5	6.905	99	498160	29.731	ng/u1	-0.02
9) Bis-(2-Chloroethyl)eth...	7.069	67	305786	28.787	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.263	132	407511	29.759	ng/u1	-0.02
15) 4-Methylphenol-d8	8.428	113	419404	30.820	ng/u1	-0.02
21) Nitrobenzene-d5	8.875	128	215684	32.682	ng/u1	-0.02
24) 2-Nitrophenol-d4	9.587	143	240994	36.490	ng/u1	-0.02
28) 2,4-Dichlorophenol-d3	10.122	165	420854	30.936	ng/u1	-0.02
31) 4-Chloroaniline-d4	10.634	131	612061	32.104	ng/u1	-0.02
46) Dimethylphthalate-d6	13.775	166	1417994	33.178	ng/u1	0.00
49) Acenaphthylene-d8	14.057	160	1569288	30.771	ng/u1	0.00
54) 4-Nitrophenol-d4	14.592	143	172262	26.380	ng/u1	0.00
60) Fluorene-d10	15.381	176	1166430	32.433	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.510	200	167339	26.969	ng/u1	0.00
73) Anthracene-d10	17.286	188	1771705	32.061	ng/u1	0.00
81) Pyrene-d10	19.675	212	1743913	34.632	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.792	264	1450892	32.462	ng/u1	0.03

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070124\
Data File : BP020713.D
Acq On : 01 Jul 2024 10:56
Operator : MA/JU
Sample : PB161602BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK602

Quant Time: Jul 02 00:53:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jun 25 18:30:06 2024
Response via : Initial Calibration

