

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070524\
 Data File : BP020785.D
 Acq On : 05 Jul 2024 12:13
 Operator : MA/JU
 Sample : PB161651BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK651

Quant Time: Jul 05 23:54:29 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 15:01:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	293999	20.000	ng/u1	0.00
20) Naphthalene-d8	10.499	136	1207990	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.357	164	791629	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.169	188	1699063	20.000	ng/u1	0.00
79) Chrysene-d12	21.621	240	1575028	20.000	ng/u1	0.00
88) Perylene-d12	24.962	264	1828606	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	36289	5.300	ng/uL	0.00
4) Pyridine-d5	3.675	84	540521	27.089	ng/u1	0.00
7) Phenol-d5	6.916	99	700902	29.067	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.075	67	414000	27.082	ng/u1	0.00
11) 2-Chlorophenol-d4	7.269	132	570150	28.931	ng/u1	0.00
15) 4-Methylphenol-d8	8.434	113	568482	29.027	ng/u1	0.00
21) Nitrobenzene-d5	8.875	128	283551	31.793	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.587	143	329402	36.908	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	562358	30.589	ng/u1	0.00
31) 4-Chloroaniline-d4	10.634	131	775653	30.106	ng/u1	0.00
46) Dimethylphthalate-d6	13.769	166	1738222	31.580	ng/u1	0.01
49) Acenaphthylene-d8	14.051	160	1941667	29.564	ng/u1	0.01
54) 4-Nitrophenol-d4	14.598	143	281977	33.530	ng/u1	0.02
60) Fluorene-d10	15.369	176	1443930	31.175	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.504	200	245090	28.624	ng/u1	0.00
73) Anthracene-d10	17.275	188	2357057	30.911	ng/u1	0.01
81) Pyrene-d10	19.657	212	2772323	29.299	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.745	264	2815838	31.656	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070524\
Data File : BP020785.D
Acq On : 05 Jul 2024 12:13
Operator : MA/JU
Sample : PB161651BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK651

Quant Time: Jul 05 23:54:29 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062524.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 03 15:01:56 2024
Response via : Initial Calibration

