

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070622\
 Data File : BP010925.D
 Acq On : 06 Jul 2022 11:58
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020692

Quant Time: Jul 06 23:15:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	448315	20.000	ng/ul	0.00
20) Naphthalene-d8	10.499	136	1776234	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.357	164	896965	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.122	188	1557224	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.239	240	1394121	20.000	ng/ul	# 0.00
88) Perylene-d12	23.598	264	1575967	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	98481	8.310	ng/uL	0.00
4) Pyridine-d5	3.611	84	625806	19.837	ng/ul	0.00
7) Phenol-d5	6.881	99	689343	19.424	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	461531	19.885	ng/ul	0.00
11) 2-Chlorophenol-d4	7.240	132	588337	19.982	ng/ul	0.00
15) 4-Methylphenol-d8	8.422	113	539475	19.230	ng/ul	0.00
21) Nitrobenzene-d5	8.863	128	253516	22.420	ng/ul	0.00
24) 2-Nitrophenol-d4	9.587	143	229909	24.576	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.122	165	455546	19.708	ng/ul	0.00
31) 4-Chloroaniline-d4	10.640	131	711050	18.845	ng/ul	0.00
46) Dimethylphthalate-d6	13.781	166	1147482	18.676	ng/ul	0.00
49) Acenaphthylene-d8	14.051	160	1483744	18.921	ng/ul	0.00
54) 4-Nitrophenol-d4	14.575	143	170700	17.787	ng/ul	0.00
60) Fluorene-d10	15.357	176	1001966	19.003	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.481	200	104352	20.226	ng/ul	0.00
73) Anthracene-d10	17.222	188	1339481	19.333	ng/ul	0.00
81) Pyrene-d10	19.474	212	1387638	18.270	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.445	264	1505241	19.328	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	101860	8.059	ng/uL#	87
5) Pyridine	3.628	79	641648	20.078	ng/ul#	74
6) Benzaldehyde	6.852	77	411591	21.981	ng/ul	96
8) Phenol	6.910	94	750228	19.765	ng/ul#	84
10) Bis(2-Chloroethyl)ether	7.140	93	603005	19.780	ng/ul	91
12) 2-Chlorophenol	7.269	128	608085	20.007	ng/ul	92
13) 2-Methylphenol	8.152	108	538206	19.046	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.246	45	811057	19.829	ng/ul#	97
16) Acetophenone	8.528	105	843166	19.570	ng/ul#	88
17) N-Nitroso-di-n-propyla...	8.522	70	406859	18.675	ng/ul	89
18) 4-Methylphenol	8.481	108	571378	19.416	ng/ul	96
19) Hexachloroethane	8.775	117	237181	19.269	ng/ul#	80
22) Nitrobenzene	8.904	77	609543	21.601	ng/ul	91
23) Isophorone	9.434	82	1053269	17.799	ng/ul	99
25) 2-Nitrophenol	9.616	139	269896	23.824	ng/ul#	81
26) 2,4-Dimethylphenol	9.681	107	603933	19.512	ng/ul	95
27) Bis(2-Chloroethoxy)met...	9.922	93	723602	19.384	ng/ul	98
29) 2,4-Dichlorophenol	10.146	162	479451	19.962	ng/ul	99
30) Naphthalene	10.546	128	1812977	19.584	ng/ul	99
32) 4-Chloroaniline	10.669	127	709297	18.974	ng/ul	100
33) Hexachlorobutadiene	10.834	225	293952	19.156	ng/ul	96
34) Caprolactam	11.451	113	119878	15.888	ng/ul	89
35) 4-Chloro-3-methylphenol	11.798	107	483635	18.578	ng/ul	94
36) 2-Methylnaphthalene	12.169	142	1129292	18.838	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070622\
 Data File : BP010925.D
 Acq On : 06 Jul 2022 11:58
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020692

Quant Time: Jul 06 23:15:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.387	142	1137283	19.096	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.534	216	518180	20.104	ng/ul#	98
40) Hexachlorocyclopentadiene	12.516	237	291961	17.876	ng/ul	94
41) 2,4,6-Trichlorophenol	12.787	196	300767	20.239	ng/ul	99
42) 2,4,5-Trichlorophenol	12.857	196	318570	19.977	ng/ul	99
44) 2-Chloronaphthalene	13.228	162	1063534	19.994	ng/ul	98
45) 2-Nitroaniline	13.440	65	257647	22.829	ng/ul#	80
47) Dimethylphthalate	13.828	163	1153109	18.740	ng/ul	99
48) 2,6-Dinitrotoluene	13.945	165	199901	21.682	ng/ul	94
50) Acenaphthylene	14.081	152	1669975	19.445	ng/ul	98
51) 3-Nitroaniline	14.275	138	219975	19.902	ng/ul	87
52) Acenaphthene	14.422	153	1094575	19.752	ng/ul	98
53) 2,4-Dinitrophenol	14.475	184	61775	17.453	ng/ul	88
55) 4-Nitrophenol	14.587	109	144008	18.871	ng/ul#	83
56) Dibenzofuran	14.757	168	1485718	19.573	ng/ul	97
57) 2,4-Dinitrotoluene	14.734	165	269424	21.870	ng/ul#	83
58) 2,3,4,6-Tetrachlorophenol	14.986	232	219076	19.184	ng/ul#	88
59) Diethylphthalate	15.198	149	1092588	17.860	ng/ul	97
61) Fluorene	15.416	166	1145090	18.919	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.416	204	538306	18.999	ng/ul	99
63) 4-Nitroaniline	15.439	138	211914	20.736	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.492	198	112428	20.166	ng/ul#	94
67) N-Nitrosodiphenylamine	15.634	169	953260	20.211	ng/ul	99
68) 4-Bromophenyl-phenylether	16.316	248	294177	18.850	ng/ul	98
69) Hexachlorobenzene	16.416	284	338389	18.720	ng/ul	94
70) Atrazine	16.598	200	272616	17.433	ng/ul	98
71) Pentachlorophenol	16.769	266	146881	15.781	ng/ul	96
72) Phenanthrene	17.163	178	1657425	19.987	ng/ul	98
74) Anthracene	17.257	178	1646161	19.737	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.145	216	526709	20.497	ng/uL	97
76) Pentachlorobenzene	14.675	250	448943	20.371	ng/uL	99
77) Carbazole	17.533	167	1492325	19.877	ng/ul	100
78) Di-n-butylphthalate	18.104	149	1487303	16.625	ng/ul	99
80) Fluoranthene	19.145	202	1691106	18.104	ng/ul#	87
82) Pyrene	19.504	202	1744605	18.494	ng/ul#	86
83) Butylbenzylphthalate	20.398	149	571029	17.197	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.163	252	545216	18.939	ng/ul#	99
85) Benzo(a)anthracene	21.221	228	1684727	19.322	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.163	149	829925	15.783	ng/ul#	96
87) Chrysene	21.274	228	1663281	19.907	ng/ul	100
89) Di-n-octyl phthalate	22.080	149	1409442	15.067	ng/ul	100
90) Benzo(b)fluoranthene	22.874	252	1843679	18.812	ng/ul#	96
91) Benzo(k)fluoranthene	22.927	252	1831997	19.666	ng/ul#	95
93) Benzo(a)pyrene	23.492	252	1880811	19.728	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.039	276	2272124	19.751	ng/ul#	90
95) Dibenzo(a,h)anthracene	26.062	278	1954581	19.808	ng/ul#	93
96) Benzo(g,h,i)perylene	26.792	276	1940856	19.834	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA P

BNA P

ClientSampleId :

SSTD020692


Quant Time: Jul 06 23:15:01 2022


Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M

Quant Title : SVOA CALIBRATION

QLast Update : Wed Jul 06 23:13:53 2022

Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.059 ng/uL
RT: 3.234 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion: 88 Resp: 101860
Ion Ratio Lower Upper
88 100
43 29.1 35.8 53.6#
58 78.4 68.1 102.1



#5
Pyridine
Concen: 20.078 ng/uL
RT: 3.628 min Scan# 92
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion: 79 Resp: 641648
Ion Ratio Lower Upper
79 100
52 76.3 83.8 125.8#
51 36.0 41.0 61.4#









#6
Benzaldehyde
Concen: 21.981 ng/ul
RT: 6.852 min Scan# 643
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion: 77 Resp: 411591
Ion Ratio Lower Upper
77 100
105 99.2 76.0 114.0
106 95.3 73.2 109.8



#8
Phenol
Concen: 19.765 ng/ul
RT: 6.910 min Scan# 650
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion: 94 Resp: 750228
Ion Ratio Lower Upper
94 100
65 26.8 27.4 41.2#
66 35.0 37.4 56.2#









#10
Bis(2-Chloroethyl)ether
Concen: 19.780 ng/ul
RT: 7.140 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion: 93 Resp: 603005
Ion Ratio Lower Upper
93 100
63 79.0 72.4 108.6
95 32.7 27.3 40.9



#12
2-Chlorophenol
Concen: 20.007 ng/ul
RT: 7.269 min Scan# 711
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion: 128 Resp: 608085
Ion Ratio Lower Upper
128 100
64 46.6 44.8 67.2
130 32.0 25.8 38.6









#13
2-Methylphenol
Concen: 19.046 ng/ul
RT: 8.152 min Scan# 80
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:108 Resp: 538206
Ion Ratio Lower Upper
108 100
107 89.5 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.829 ng/ul
RT: 8.246 min Scan# 877
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion: 45 Resp: 811057
Ion Ratio Lower Upper
45 100
77 13.8 11.2 16.8
79 11.2 6.9 10.3#









#16
Acetophenone
Concen: 19.570 ng/ul
RT: 8.528 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:105 Resp: 843166
Ion Ratio Lower Upper
105 100
77 75.9 67.5 101.3
51 27.1 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 18.675 ng/ul
RT: 8.522 min Scan# 924
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion: 70 Resp: 406859
Ion Ratio Lower Upper
70 100
42 56.4 53.8 80.8
101 9.7 7.8 11.6
130 22.2 15.4 23.2









#18
4-Methylphenol
Concen: 19.416 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:108 Resp: 571378
Ion Ratio Lower Upper
108 100
107 114.1 88.3 132.5



#19
Hexachloroethane
Concen: 19.269 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:117 Resp: 237181
Ion Ratio Lower Upper
117 100
201 86.1 85.7 128.5
199 55.3 57.4 86.0#









#22
Nitrobenzene
Concen: 21.601 ng/ul
RT: 8.904 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion: 77 Resp: 609543
Ion Ratio Lower Upper
77 100
123 46.4 31.5 47.3
65 13.1 11.4 17.2



#23
Isophorone
Concen: 17.799 ng/ul
RT: 9.434 min Scan# 1079
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion: 82 Resp: 1053269
Ion Ratio Lower Upper
82 100
95 7.1 6.6 10.0
138 17.3 13.8 20.6









#25
2-Nitrophenol
Concen: 23.824 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:139 Resp: 269896
Ion Ratio Lower Upper
139 100
65 50.9 54.2 81.4#
109 29.6 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 19.512 ng/ul
RT: 9.681 min Scan# 1121
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:107 Resp: 603933
Ion Ratio Lower Upper
107 100
121 51.9 39.1 58.7
122 87.0 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 19.384 ng/ul

RT: 9.922 min Scan# 1166

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020692





Tgt Ion: 93 Resp: 723602

Ion Ratio Lower Upper

93 100

95 32.0 26.6 39.8

123 11.2 7.8 11.8



#29

2,4-Dichlorophenol

Concen: 19.962 ng/ul

RT: 10.146 min Scan# 1200

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58




Tgt Ion: 162 Resp: 479451


Ion Ratio Lower Upper

162 100

164 64.2 52.5 78.7

98 39.5 31.4 47.0









#30
Naphthalene
Concen: 19.584 ng/ul
RT: 10.546 min Scan# 1268
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:128 Resp: 1812977
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 18.974 ng/ul
RT: 10.669 min Scan# 1289
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:127 Resp: 709297
Ion Ratio Lower Upper
127 100
129 32.6 26.2 39.2





#33

Hexachlorobutadiene

Concen: 19.156 ng/ul

RT: 10.834 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020692





Tgt Ion:225 Resp: 293952

Ion Ratio Lower Upper

225 100

223 61.6 53.7 80.5

227 61.6 49.4 74.2



#34

Caprolactam

Concen: 15.888 ng/ul

RT: 11.451 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58




Tgt Ion:113 Resp: 119878


Ion Ratio Lower Upper

113 100

55 172.4 153.8 230.8

56 128.6 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 18.578 ng/ul
RT: 11.798 min Scan# 1486
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:107 Resp: 483635
Ion Ratio Lower Upper
107 100
144 26.2 18.6 27.8
142 79.6 59.6 89.4



#36
2-Methylnaphthalene
Concen: 18.838 ng/ul
RT: 12.169 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:142 Resp: 1129292
Ion Ratio Lower Upper
142 100
141 87.6 70.1 105.1









#37
1-Methylnaphthalene
Concen: 19.096 ng/ul
RT: 12.387 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:142 Resp: 1137283
Ion Ratio Lower Upper
142 100
141 90.7 73.3 109.9
116 3.8 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.104 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:216 Resp: 518180
Ion Ratio Lower Upper
216 100
214 77.8 63.0 94.6
179 21.7 18.1 27.1
108 20.9 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 17.876 ng/ul
RT: 12.516 min Scan# 1603
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:237 Resp: 291961
Ion Ratio Lower Upper
237 100
235 60.6 52.8 79.2
272 12.5 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 20.239 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:196 Resp: 300767
Ion Ratio Lower Upper
196 100
198 96.7 78.2 117.4
200 31.0 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 19.977 ng/ul

RT: 12.857 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020692





Tgt Ion:196 Resp: 318570

Ion Ratio Lower Upper

196 100

198 95.3 77.0 115.4

200 30.3 24.8 37.2



#44

2-Chloronaphthalene

Concen: 19.994 ng/ul

RT: 13.228 min Scan# 1724

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58




Tgt Ion:162 Resp: 1063534


Ion Ratio Lower Upper

162 100

164 32.4 25.5 38.3

127 41.1 31.9 47.9









#45
2-Nitroaniline
Concen: 22.829 ng/ul
RT: 13.440 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion: 65 Resp: 257647
Ion Ratio Lower Upper
65 100
92 70.1 49.0 73.6
138 106.2 66.4 99.6#



#47
Dimethylphthalate
Concen: 18.740 ng/ul
RT: 13.828 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:163 Resp: 1153109
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 9.9 7.6 11.4









#48
2,6-Dinitrotoluene
Concen: 21.682 ng/ul
RT: 13.945 min Scan# 1846
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:165 Resp: 199901
Ion Ratio Lower Upper
165 100
89 59.7 52.9 79.3
121 24.4 19.4 29.2



#50
Acenaphthylene
Concen: 19.445 ng/ul
RT: 14.081 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:152 Resp: 1669975
Ion Ratio Lower Upper
152 100
151 19.5 16.5 24.7
153 12.9 10.1 15.1









#51
3-Nitroaniline
Concen: 19.902 ng/ul
RT: 14.275 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:138 Resp: 219975
Ion Ratio Lower Upper
138 100
108 13.2 11.0 16.4
92 114.6 104.6 156.8



#52
Acenaphthene
Concen: 19.752 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:153 Resp: 1094575
Ion Ratio Lower Upper
153 100
152 46.9 36.2 54.2
154 87.5 70.8 106.2









#53
2,4-Dinitrophenol
Concen: 17.453 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:184 Resp: 61775
Ion Ratio Lower Upper
184 100
63 77.3 73.8 110.6
154 66.0 57.2 85.8



#55
4-Nitrophenol
Concen: 18.871 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:109 Resp: 144008
Ion Ratio Lower Upper
109 100
139 134.0 75.6 113.4#
65 134.1 107.6 161.4









#56
Dibenzofuran
Concen: 19.573 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:168 Resp: 1485718
Ion Ratio Lower Upper
168 100
139 39.1 32.8 49.2



#57
2,4-Dinitrotoluene
Concen: 21.870 ng/ul
RT: 14.734 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:165 Resp: 269424
Ion Ratio Lower Upper
165 100
63 45.2 47.0 70.6#
182 2.4 2.2 3.2









#58
2,3,4,6-Tetrachlorophenol
Concen: 19.184 ng/ul
RT: 14.986 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:232	Resp:	219076
Ion Ratio	Lower	Upper
232	100	
131	54.4	35.8 53.8#
130	3.0	1.8 2.8#
166	34.8	23.7 35.5



#59
Diethylphthalate
Concen: 17.860 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:149	Resp:	1092588
Ion Ratio	Lower	Upper
149	100	
177	20.7	17.6 26.4
150	12.1	10.3 15.5









#61
Fluorene
Concen: 18.919 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:166 Resp: 1145090
Ion Ratio Lower Upper
166 100
165 97.5 77.4 116.2
167 13.2 10.8 16.2



#62
4-Chlorophenyl-phenylether
Concen: 18.999 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:204 Resp: 538306
Ion Ratio Lower Upper
204 100
206 32.6 26.0 39.0
141 63.1 49.3 73.9









#63
4-Nitroaniline
Concen: 20.736 ng/ul
RT: 15.439 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:138 Resp: 211914
Ion Ratio Lower Upper
138 100
92 54.2 46.2 69.2
108 78.0 85.8 128.8#



#66
4,6-Dinitro-2-methylphenol
Concen: 20.166 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:198 Resp: 112428
Ion Ratio Lower Upper
198 100
182 7.3 4.3 6.5#
77 32.7 23.7 35.5









#67
N-Nitrosodiphenylamine
Concen: 20.211 ng/ul
RT: 15.634 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:169 Resp: 953260
Ion Ratio Lower Upper
169 100
168 66.2 52.6 78.8
167 35.8 27.8 41.6



#68
4-Bromophenyl-phenylether
Concen: 18.850 ng/ul
RT: 16.316 min Scan# 2249
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:248 Resp: 294177
Ion Ratio Lower Upper
248 100
250 99.2 80.8 121.2
141 81.7 63.9 95.9





#69

Hexachlorobenzene

Concen: 18.720 ng/ul

RT: 16.416 min Scan# 2266

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020692





Tgt Ion:284 Resp: 338389

Ion Ratio Lower Upper

284 100

142 41.1 27.8 41.6

249 31.1 25.4 38.2



#70

Atrazine

Concen: 17.433 ng/ul

RT: 16.598 min Scan# 2297

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58




Tgt Ion:200 Resp: 272616


Ion Ratio Lower Upper

200 100

173 26.7 20.7 31.1

215 46.7 38.6 57.8





#71

Pentachlorophenol

Concen: 15.781 ng/ul

RT: 16.769 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020692





Tgt Ion:266 Resp: 146881

Ion Ratio Lower Upper

266 100

264 63.0 51.4 77.2

268 62.7 53.5 80.3



#72

Phenanthrene

Concen: 19.987 ng/ul

RT: 17.163 min Scan# 2393

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58




Tgt Ion:178 Resp: 1657425


Ion Ratio Lower Upper

178 100

179 15.1 12.7 19.1

176 19.1 14.6 22.0





#74

Anthracene

Concen: 19.737 ng/uI

RT: 17.257 min Scan# 2412

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020692





Tgt Ion:178 Resp: 1646161

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 18.7 14.5 21.7



#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.497 ng/uI

RT: 13.145 min Scan# 1710

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58




Tgt Ion:216 Resp: 526709


Ion Ratio Lower Upper

216 100

214 77.8 65.0 97.4

218 47.8 39.1 58.7









#76
Pentachlorobenzene
Concen: 20.371 ng/uL
RT: 14.675 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:250 Resp: 448943
Ion Ratio Lower Upper
250 100
248 63.2 49.8 74.6
252 63.8 51.2 76.8



#77
Carbazole
Concen: 19.877 ng/uL
RT: 17.533 min Scan# 2456
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:167 Resp: 1492325
Ion Ratio Lower Upper
167 100
166 21.2 16.9 25.3
139 13.5 10.6 16.0





#78

Di-n-butylphthalate

Concen: 16.625 ng/ul

RT: 18.104 min Scan# 2553

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020692





Tgt Ion:149 Resp: 1487303

Ion Ratio Lower Upper

149 100

150 8.9 7.1 10.7

104 5.2 4.6 6.8



#80

Fluoranthene

Concen: 18.104 ng/ul

RT: 19.145 min Scan# 2730

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58




Tgt Ion:202 Resp: 1691106


Ion Ratio Lower Upper

202 100

101 14.3 7.1 10.7#

100 10.8 5.7 8.5#









#82
Pyrene
Concen: 18.494 ng/ul
RT: 19.504 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:202 Resp: 1744605
Ion Ratio Lower Upper
202 100
101 15.6 7.9 11.9#
100 12.9 6.7 10.1#



#83
Butylbenzylphthalate
Concen: 17.197 ng/ul
RT: 20.398 min Scan# 2943
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:149 Resp: 571029
Ion Ratio Lower Upper
149 100
91 70.6 60.0 90.0
206 19.6 17.4 26.2









#84
3,3'-Dichlorobenzidine
Concen: 18.939 ng/ul
RT: 21.163 min Scan# 3073
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:252 Resp: 545216
Ion Ratio Lower Upper
252 100
254 63.6 51.0 76.4
126 13.4 8.6 12.8#



#85
Benzo(a)anthracene
Concen: 19.322 ng/ul
RT: 21.221 min Scan# 3083
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:228 Resp: 1684727
Ion Ratio Lower Upper
228 100
229 19.7 15.0 22.4
226 27.2 20.6 31.0









#86
Bis(2-ethylhexyl)phthalate
Concen: 15.783 ng/ul
RT: 21.163 min Scan# 3077
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:149 Resp: 829925
Ion Ratio Lower Upper
149 100
167 26.3 22.6 33.8
279 3.8 4.2 6.4#



#87
Chrysene
Concen: 19.907 ng/ul
RT: 21.274 min Scan# 3092
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:228 Resp: 1663281
Ion Ratio Lower Upper
228 100
226 29.4 23.6 35.4
229 19.8 15.6 23.4





#89

Di-n-octyl phthalate

Concen: 15.067 ng/ul

RT: 22.080 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58





Instrument :

BNA_P

ClientSampleId :

SSTD020692

Tgt Ion:149 Resp: 1409442



#90

Benzo(b)fluoranthene

Concen: 18.812 ng/ul

RT: 22.874 min Scan# 3364

Delta R.T. 0.000 min

Lab File: BP010925.D

Acq: 06 Jul 2022 11:58




Tgt Ion:252 Resp: 1843679


Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.6

125 12.7 6.2 9.2#









#91
Benzo(k)fluoranthene
Concen: 19.666 ng/u1
RT: 22.927 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020692


Tgt Ion:252 Resp: 1831997
Ion Ratio Lower Upper
252 100
253 21.8 17.0 25.4
125 12.5 6.2 9.4#



#93
Benzo(a)pyrene
Concen: 19.728 ng/u1
RT: 23.492 min Scan# 3469
Delta R.T. 0.000 min
Lab File: BP010925.D
Acq: 06 Jul 2022 11:58

Tgt Ion:252 Resp: 1880811
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 14.1 6.6 10.0#









#94
 Indeno(1,2,3-cd)pyrene
 Concen: 19.751 ng/ul
 RT: 26.039 min Scan# 3902
 Delta R.T. 0.000 min
 Lab File: BP010925.D
 Acq: 06 Jul 2022 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020692


Tgt Ion:276 Resp: 2272124
 Ion Ratio Lower Upper
 276 100
 138 27.6 14.6 22.0#
 277 23.8 19.6 29.4



#95
 Dibenzo(a,h)anthracene
 Concen: 19.808 ng/ul
 RT: 26.062 min Scan# 3906
 Delta R.T. 0.000 min
 Lab File: BP010925.D
 Acq: 06 Jul 2022 11:58

Tgt Ion:278 Resp: 1954581
 Ion Ratio Lower Upper
 278 100
 139 19.9 9.4 14.2#
 279 23.7 18.9 28.3





#96
 Benzo(g,h,i)perylene
 Concen: 19.834 ng/ul
 RT: 26.792 min Scan# 4030
 Delta R.T. 0.000 min
 Lab File: BP010925.D
 Acq: 06 Jul 2022 11:58

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020692

Tgt Ion:276 Resp: 1940856
 Ion Ratio Lower Upper
 276 100
 138 27.0 12.7 19.1#
 277 23.9 18.2 27.4

