

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070622\
 Data File : BP010927.D
 Acq On : 06 Jul 2022 13:14
 Operator : CG/JU
 Sample : PB145967BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS967

Quant Time: Jul 06 23:15:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	381304	20.000	ng/ul	0.00
20) Naphthalene-d8	10.499	136	1482391	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.357	164	750307	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.122	188	1439424	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.239	240	1564135	20.000	ng/ul	# 0.00
88) Perylene-d12	23.598	264	1738480	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	61460	6.097	ng/uL	0.00
4) Pyridine-d5	3.611	84	747829	27.871	ng/ul	0.00
7) Phenol-d5	6.881	99	862873	28.587	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	551017	27.913	ng/ul	0.00
11) 2-Chlorophenol-d4	7.240	132	723120	28.875	ng/ul	0.00
15) 4-Methylphenol-d8	8.422	113	659990	27.660	ng/ul	0.00
21) Nitrobenzene-d5	8.863	128	311020	32.958	ng/ul	0.00
24) 2-Nitrophenol-d4	9.587	143	279214	35.763	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.122	165	543280	28.162	ng/ul	0.00
31) 4-Chloroaniline-d4	10.640	131	471702	14.980	ng/ul	0.00
46) Dimethylphthalate-d6	13.781	166	1368767	26.632	ng/ul	0.00
49) Acenaphthylene-d8	14.051	160	1794164	27.352	ng/ul	0.00
54) 4-Nitrophenol-d4	14.575	143	268979	33.506	ng/ul	0.00
60) Fluorene-d10	15.357	176	1225521	27.786	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.481	200	159836	33.516	ng/ul	0.00
73) Anthracene-d10	17.222	188	1840965	28.746	ng/ul	0.00
81) Pyrene-d10	19.475	212	2163202	25.385	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.445	264	2505596	29.166	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	121919	11.342	ng/uL#	89
5) Pyridine	3.628	79	728216	26.792	ng/ul#	74
6) Benzaldehyde	6.852	77	534597	33.567	ng/ul	95
8) Phenol	6.911	94	860686	26.660	ng/ul#	85
10) Bis(2-Chloroethyl)ether	7.140	93	693322	26.740	ng/ul	90
12) 2-Chlorophenol	7.269	128	705870	27.306	ng/ul	93
13) 2-Methylphenol	8.158	108	612627	25.490	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.240	45	902500	25.943	ng/ul#	97
16) Acetophenone	8.534	105	952037	25.980	ng/ul#	87
17) N-Nitroso-di-n-propyla...	8.522	70	440539	23.775	ng/ul	90
18) 4-Methylphenol	8.487	108	655132	26.175	ng/ul	100
19) Hexachloroethane	8.775	117	275506	26.317	ng/ul#	79
22) Nitrobenzene	8.911	77	710991	30.191	ng/ul	91
23) Isophorone	9.434	82	1143291	23.150	ng/ul#	98
25) 2-Nitrophenol	9.616	139	307161	32.488	ng/ul#	82
26) 2,4-Dimethylphenol	9.687	107	661507	25.609	ng/ul	94
27) Bis(2-Chloroethoxy)met...	9.922	93	788236	25.301	ng/ul	99
29) 2,4-Dichlorophenol	10.152	162	533207	26.600	ng/ul	98
30) Naphthalene	10.546	128	2060773	26.673	ng/ul	99
32) 4-Chloroaniline	10.669	127	770138	24.685	ng/ul	99
33) Hexachlorobutadiene	10.834	225	315412	24.629	ng/ul	97
34) Caprolactam	11.452	113	146911	23.330	ng/ul	87
35) 4-Chloro-3-methylphenol	11.804	107	546468	25.153	ng/ul	94
36) 2-Methylnaphthalene	12.169	142	1240449	24.794	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070622\
 Data File : BP010927.D
 Acq On : 06 Jul 2022 13:14
 Operator : CG/JU
 Sample : PB145967BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS967


Quant Time: Jul 06 23:15:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 06 23:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.387	142	1231930	24.786	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.534	216	564342	26.174	ng/ul#	98
40) Hexachlorocyclopentadiene	12.516	237	324437	23.747	ng/ul	95
41) 2,4,6-Trichlorophenol	12.787	196	337303	27.135	ng/ul	98
42) 2,4,5-Trichlorophenol	12.857	196	369486	27.698	ng/ul	100
43) 1,1'-Biphenyl	13.193	154	1537303	26.027	ng/ul	98
44) 2-Chloronaphthalene	13.228	162	1191310	26.774	ng/ul	98
45) 2-Nitroaniline	13.440	65	315658	33.437	ng/ul	85
47) Dimethylphthalate	13.828	163	1316172	25.571	ng/ul	99
48) 2,6-Dinitrotoluene	13.946	165	236509	30.667	ng/ul	95
50) Acenaphthylene	14.081	152	1829546	25.467	ng/ul	98
51) 3-Nitroaniline	14.275	138	293736	31.770	ng/ul	89
52) Acenaphthene	14.422	153	1243548	26.827	ng/ul	98
53) 2,4-Dinitrophenol	14.475	184	89236	30.140	ng/ul	89
55) 4-Nitrophenol	14.587	109	201626	31.587	ng/ul#	76
56) Dibenzofuran	14.763	168	1686851	26.567	ng/ul	96
57) 2,4-Dinitrotoluene	14.734	165	348293	33.799	ng/ul#	82
58) 2,3,4,6-Tetrachlorophenol	14.987	232	270997	28.369	ng/ul#	86
59) Diethylphthalate	15.198	149	1265001	24.721	ng/ul	98
61) Fluorene	15.416	166	1348179	26.629	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.416	204	611554	25.803	ng/ul	98
63) 4-Nitroaniline	15.445	138	322977	37.782	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.493	198	158768	30.809	ng/ul	99
67) N-Nitrosodiphenylamine	15.634	169	1122372	25.744	ng/ul	98
68) 4-Bromophenyl-phenylether	16.316	248	350123	24.271	ng/ul	96
69) Hexachlorobenzene	16.416	284	411656	24.637	ng/ul	93
70) Atrazine	16.598	200	301187	20.837	ng/ul	97
71) Pentachlorophenol	16.769	266	197817	22.992	ng/ul	97
72) Phenanthrene	17.163	178	2088232	27.242	ng/ul	99
74) Anthracene	17.257	178	2133269	27.670	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.146	216	576182	24.257	ng/ul	96
76) Pentachlorobenzene	14.675	250	503607	24.721	ng/ul	98
77) Carbazole	17.534	167	2032999	29.295	ng/ul	99
78) Di-n-butylphthalate	18.104	149	2006679	24.266	ng/ul	99
80) Fluoranthene	19.145	202	2413723	23.031	ng/ul#	87
82) Pyrene	19.504	202	2566109	24.245	ng/ul#	86
83) Butylbenzylphthalate	20.398	149	961673	25.814	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.163	252	877995	27.184	ng/ul#	98
85) Benzo(a)anthracene	21.222	228	2669140	27.285	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.163	149	1444578	24.487	ng/ul#	97
87) Chrysene	21.275	228	2540095	27.097	ng/ul	99
89) Di-n-octyl phthalate	22.080	149	2537603	24.591	ng/ul	100
90) Benzo(b)fluoranthene	22.880	252	3008260	27.826	ng/ul#	95
91) Benzo(k)fluoranthene	22.927	252	2824104	27.482	ng/ul#	95
93) Benzo(a)pyrene	23.492	252	2654491	25.240	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.045	276	3566226	28.102	ng/ul#	90
95) Dibenzo(a,h)anthracene	26.068	278	2981754	27.392	ng/ul#	93
96) Benzo(g,h,i)perylene	26.798	276	2932258	27.164	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 06 23:15:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP062322.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 06 23:13:53 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 11.342 ng/uL
RT: 3.234 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion: 88 Resp: 121919
Ion Ratio Lower Upper
88 100
43 30.9 35.8 53.6#
58 80.4 68.1 102.1



#5
Pyridine
Concen: 26.792 ng/uL
RT: 3.628 min Scan# 92
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion: 79 Resp: 728216
Ion Ratio Lower Upper
79 100
52 76.3 83.8 125.8#
51 35.9 41.0 61.4#









#6
Benzaldehyde
Concen: 33.567 ng/ul
RT: 6.852 min Scan# 643
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion: 77 Resp: 534597
Ion Ratio Lower Upper
77 100
105 100.3 76.0 114.0
106 95.0 73.2 109.8



#8
Phenol
Concen: 26.660 ng/ul
RT: 6.911 min Scan# 650
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion: 94 Resp: 860686
Ion Ratio Lower Upper
94 100
65 26.6 27.4 41.2#
66 35.5 37.4 56.2#









#10
 Bis(2-Chloroethyl)ether
 Concen: 26.740 ng/ul
 RT: 7.140 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BP010927.D
 Acq: 06 Jul 2022 13:14




Instrument :
 BNA_P
 ClientSampleId :
 SLCS967


Tgt Ion: 93 Resp: 693322
 Ion Ratio Lower Upper
 93 100
 63 78.3 72.4 108.6
 95 33.0 27.3 40.9



#12
 2-Chlorophenol
 Concen: 27.306 ng/ul
 RT: 7.269 min Scan# 711
 Delta R.T. 0.000 min
 Lab File: BP010927.D
 Acq: 06 Jul 2022 13:14

Tgt Ion: 128 Resp: 705870
 Ion Ratio Lower Upper
 128 100
 64 47.6 44.8 67.2
 130 32.2 25.8 38.6









#13
2-Methylphenol
Concen: 25.490 ng/ul
RT: 8.158 min Scan# 80
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:108 Resp: 612627
Ion Ratio Lower Upper
108 100
107 87.6 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.943 ng/ul
RT: 8.240 min Scan# 876
Delta R.T. -0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion: 45 Resp: 902500
Ion Ratio Lower Upper
45 100
77 13.7 11.2 16.8
79 10.6 6.9 10.3#









#16
Acetophenone
Concen: 25.980 ng/ul
RT: 8.534 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:105 Resp: 952037
Ion Ratio Lower Upper
105 100
77 74.9 67.5 101.3
51 26.2 29.6 44.4#



#17
N-Nitroso-di-n-propylamine
Concen: 23.775 ng/ul
RT: 8.522 min Scan# 924
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion: 70 Resp: 440539
Ion Ratio Lower Upper
70 100
42 56.9 53.8 80.8
101 10.5 7.8 11.6
130 21.9 15.4 23.2









#18
4-Methylphenol
Concen: 26.175 ng/ul
RT: 8.487 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:108 Resp: 655132
Ion Ratio Lower Upper
108 100
107 110.8 88.3 132.5



#19
Hexachloroethane
Concen: 26.317 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:117 Resp: 275506
Ion Ratio Lower Upper
117 100
201 85.1 85.7 128.5#
199 55.2 57.4 86.0#









#22
Nitrobenzene
Concen: 30.191 ng/ul
RT: 8.911 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion: 77 Resp: 710991
Ion Ratio Lower Upper
77 100
123 46.1 31.5 47.3
65 13.1 11.4 17.2



#23
Isophorone
Concen: 23.150 ng/ul
RT: 9.434 min Scan# 1079
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion: 82 Resp: 1143291
Ion Ratio Lower Upper
82 100
95 6.6 6.6 10.0#
138 17.3 13.8 20.6









#25
2-Nitrophenol
Concen: 32.488 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:139 Resp: 307161
Ion Ratio Lower Upper
139 100
65 52.0 54.2 81.4#
109 28.9 30.6 45.8#



#26
2,4-Dimethylphenol
Concen: 25.609 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:107 Resp: 661507
Ion Ratio Lower Upper
107 100
121 51.9 39.1 58.7
122 88.3 66.1 99.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 25.301 ng/ul
 RT: 9.922 min Scan# 1166
 Delta R.T. 0.000 min
 Lab File: BP010927.D
 Acq: 06 Jul 2022 13:14




Instrument :
 BNA_P
 ClientSampleId :
 SLCS967


Tgt Ion: 93 Resp: 788236
 Ion Ratio Lower Upper
 93 100
 95 33.2 26.6 39.8
 123 11.5 7.8 11.8



#29
 2,4-Dichlorophenol
 Concen: 26.600 ng/ul
 RT: 10.152 min Scan# 1201
 Delta R.T. 0.006 min
 Lab File: BP010927.D
 Acq: 06 Jul 2022 13:14

Tgt Ion: 162 Resp: 533207
 Ion Ratio Lower Upper
 162 100
 164 64.5 52.5 78.7
 98 37.2 31.4 47.0









#30
Naphthalene
Concen: 26.673 ng/ul
RT: 10.546 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:128 Resp: 2060773
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.2 10.2 15.4



#32
4-Chloroaniline
Concen: 24.685 ng/ul
RT: 10.669 min Scan# 1289
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:127 Resp: 770138
Ion Ratio Lower Upper
127 100
129 32.1 26.2 39.2





#33

Hexachlorobutadiene

Concen: 24.629 ng/ul

RT: 10.834 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Instrument :

BNA_P

ClientSampleId :

SLCS967





Tgt Ion:225 Resp: 315412

Ion Ratio Lower Upper

225 100

223 65.4 53.7 80.5

227 65.4 49.4 74.2



#34

Caprolactam

Concen: 23.330 ng/ul

RT: 11.452 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14




Tgt Ion:113 Resp: 146911


Ion Ratio Lower Upper

113 100

55 172.8 153.8 230.8

56 123.8 110.2 165.4









#35
4-Chloro-3-methylphenol
Concen: 25.153 ng/ul
RT: 11.804 min Scan# 1482
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:107 Resp: 546468
Ion Ratio Lower Upper
107 100
144 26.1 18.6 27.8
142 79.5 59.6 89.4



#36
2-Methylnaphthalene
Concen: 24.794 ng/ul
RT: 12.169 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:142 Resp: 1240449
Ion Ratio Lower Upper
142 100
141 89.4 70.1 105.1









#37
1-Methylnaphthalene
Concen: 24.786 ng/ul
RT: 12.387 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:142 Resp: 1231930
Ion Ratio Lower Upper
142 100
141 92.5 73.3 109.9
116 3.7 4.0 6.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 26.174 ng/ul
RT: 12.534 min Scan# 1606
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:216 Resp: 564342
Ion Ratio Lower Upper
216 100
214 79.9 63.0 94.6
179 22.4 18.1 27.1
108 21.2 13.6 20.4#









#40
Hexachlorocyclopentadiene
Concen: 23.747 ng/ul
RT: 12.516 min Scan# 1603
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:237 Resp: 324437
Ion Ratio Lower Upper
237 100
235 62.4 52.8 79.2
272 12.3 11.5 17.3



#41
2,4,6-Trichlorophenol
Concen: 27.135 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:196 Resp: 337303
Ion Ratio Lower Upper
196 100
198 95.5 78.2 117.4
200 30.3 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 27.698 ng/ul

RT: 12.857 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Instrument :

BNA_P

ClientSampleId :

SLCS967





Tgt Ion:196 Resp: 369486

Ion Ratio Lower Upper

196 100

198 96.2 77.0 115.4

200 30.3 24.8 37.2



#43

1,1'-Biphenyl

Concen: 26.027 ng/ul

RT: 13.193 min Scan# 1718

Delta R.T. 0.006 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14




Tgt Ion:154 Resp: 1537303


Ion Ratio Lower Upper

154 100

153 40.8 31.9 47.9

76 13.6 9.2 13.8









#44
2-Chloronaphthalene
Concen: 26.774 ng/ul
RT: 13.228 min Scan# 1727
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:162 Resp: 1191310
Ion Ratio Lower Upper
162 100
164 33.0 25.5 38.3
127 41.6 31.9 47.9



#45
2-Nitroaniline
Concen: 33.437 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion: 65 Resp: 315658
Ion Ratio Lower Upper
65 100
92 68.8 49.0 73.6
138 99.5 66.4 99.6









#47
Dimethylphthalate
Concen: 25.571 ng/ul
RT: 13.828 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:163 Resp: 1316172
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 9.8 7.6 11.4



#48
2,6-Dinitrotoluene
Concen: 30.667 ng/ul
RT: 13.946 min Scan# 1846
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:165 Resp: 236509
Ion Ratio Lower Upper
165 100
89 61.4 52.9 79.3
121 23.1 19.4 29.2





#50

Acenaphthylene

Concen: 25.467 ng/ul

RT: 14.081 min Scan# 1869

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Instrument :

BNA_P

ClientSampleId :

SLCS967





Tgt Ion:152 Resp: 1829546

Ion Ratio Lower Upper

152 100

151 19.6 16.5 24.7

153 12.9 10.1 15.1



#51

3-Nitroaniline

Concen: 31.770 ng/ul

RT: 14.275 min Scan# 1902

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14




Tgt Ion:138 Resp: 293736


Ion Ratio Lower Upper

138 100

108 12.5 11.0 16.4

92 116.2 104.6 156.8









#52
Acenaphthene
Concen: 26.827 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:153 Resp: 1243548
Ion Ratio Lower Upper
153 100
152 48.4 36.2 54.2
154 89.5 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 30.140 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:184 Resp: 89236
Ion Ratio Lower Upper
184 100
63 76.6 73.8 110.6
154 67.4 57.2 85.8









#55
4-Nitrophenol
Concen: 31.587 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:109 Resp: 201626
Ion Ratio Lower Upper
109 100
139 143.8 75.6 113.4#
65 140.9 107.6 161.4



#56
Dibenzofuran
Concen: 26.567 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:168 Resp: 1686851
Ion Ratio Lower Upper
168 100
139 38.3 32.8 49.2









#57
2,4-Dinitrotoluene
Concen: 33.799 ng/ul
RT: 14.734 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:165 Resp: 348293
Ion Ratio Lower Upper
165 100
63 44.9 47.0 70.6#
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 28.369 ng/ul
RT: 14.987 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:232 Resp: 270997
Ion Ratio Lower Upper
232 100
131 55.4 35.8 53.8#
130 2.9 1.8 2.8#
166 35.8 23.7 35.5#









#59
Diethylphthalate
Concen: 24.721 ng/ul
RT: 15.198 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:149 Resp: 1265001
Ion Ratio Lower Upper
149 100
177 20.8 17.6 26.4
150 12.5 10.3 15.5



#61
Fluorene
Concen: 26.629 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:166 Resp: 1348179
Ion Ratio Lower Upper
166 100
165 98.3 77.4 116.2
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 25.803 ng/ul
RT: 15.416 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:204 Resp: 611554
Ion Ratio Lower Upper
204 100
206 33.1 26.0 39.0
141 64.0 49.3 73.9



#63
4-Nitroaniline
Concen: 37.782 ng/ul
RT: 15.445 min Scan# 2101
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:138 Resp: 322977
Ion Ratio Lower Upper
138 100
92 51.7 46.2 69.2
108 75.1 85.8 128.8#









#66
4,6-Dinitro-2-methylphenol
Concen: 30.809 ng/ul
RT: 15.493 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:198 Resp: 158768
Ion Ratio Lower Upper
198 100
182 6.0 4.3 6.5
77 30.3 23.7 35.5



#67
N-Nitrosodiphenylamine
Concen: 25.744 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:169 Resp: 1122372
Ion Ratio Lower Upper
169 100
168 66.2 52.6 78.8
167 36.8 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 24.271 ng/ul
RT: 16.316 min Scan# 2249
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:248 Resp: 350123
Ion Ratio Lower Upper
248 100
250 96.7 80.8 121.2
141 82.4 63.9 95.9



#69
Hexachlorobenzene
Concen: 24.637 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:284 Resp: 411656
Ion Ratio Lower Upper
284 100
142 41.0 27.8 41.6
249 30.7 25.4 38.2





#70

Atrazine

Concen: 20.837 ng/ul

RT: 16.598 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Instrument :

BNA_P

ClientSampleId :

SLCS967





Tgt Ion:200 Resp: 301187

Ion Ratio Lower Upper

200 100

173 28.1 20.7 31.1

215 47.1 38.6 57.8



#71

Pentachlorophenol

Concen: 22.992 ng/ul

RT: 16.769 min Scan# 2326

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14




Tgt Ion:266 Resp: 197817


Ion Ratio Lower Upper

266 100

264 62.9 51.4 77.2

268 63.6 53.5 80.3





#72

Phenanthrene

Concen: 27.242 ng/ul

RT: 17.163 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Instrument :

BNA_P

ClientSampleId :

SLCS967





Tgt Ion:178 Resp: 2088232

Ion Ratio Lower Upper

178 100

179 15.6 12.7 19.1

176 18.8 14.6 22.0



#74

Anthracene

Concen: 27.670 ng/ul

RT: 17.257 min Scan# 2409

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14




Tgt Ion:178 Resp: 2133269


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 18.6 14.5 21.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 24.257 ng/uL

RT: 13.146 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Instrument :

BNA_P

ClientSampleId :

SLCS967





Tgt Ion:216 Resp: 576182

Ion Ratio Lower Upper

216 100

214 76.9 65.0 97.4

218 46.8 39.1 58.7



#76

Pentachlorobenzene

Concen: 24.721 ng/uL

RT: 14.675 min Scan# 1970

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14




Tgt Ion:250 Resp: 503607


Ion Ratio Lower Upper

250 100

248 63.7 49.8 74.6

252 64.9 51.2 76.8









#77
 Carbazole
 Concen: 29.295 ng/ul
 RT: 17.534 min Scan# 2460
 Delta R.T. 0.000 min
 Lab File: BP010927.D
 Acq: 06 Jul 2022 13:14




Instrument :
 BNA_P
 ClientSampleId :
 SLCS967


Tgt Ion:167 Resp: 2032999
 Ion Ratio Lower Upper
 167 100
 166 21.3 16.9 25.3
 139 13.7 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 24.266 ng/ul
 RT: 18.104 min Scan# 2553
 Delta R.T. 0.000 min
 Lab File: BP010927.D
 Acq: 06 Jul 2022 13:14

Tgt Ion:149 Resp: 2006679
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.1 10.7
 104 5.2 4.6 6.8





#80

Fluoranthene

Concen: 23.031 ng/ul

RT: 19.145 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Instrument :

BNA_P

ClientSampleId :

SLCS967





Tgt Ion:202 Resp: 2413723

Ion Ratio Lower Upper

202 100

101 14.0 7.1 10.7#

100 10.8 5.7 8.5#



#82

Pyrene

Concen: 24.245 ng/ul

RT: 19.504 min Scan# 2791

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14




Tgt Ion:202 Resp: 2566109


Ion Ratio Lower Upper

202 100

101 15.7 7.9 11.9#

100 12.9 6.7 10.1#









#83
Butylbenzylphthalate
Concen: 25.814 ng/ul
RT: 20.398 min Scan# 2946
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:149 Resp: 961673
Ion Ratio Lower Upper
149 100
91 69.9 60.0 90.0
206 19.2 17.4 26.2



#84
3,3'-Dichlorobenzidine
Concen: 27.184 ng/ul
RT: 21.163 min Scan# 3073
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:252 Resp: 877995
Ion Ratio Lower Upper
252 100
254 64.3 51.0 76.4
126 13.6 8.6 12.8#









#85
Benzo(a)anthracene
Concen: 27.285 ng/ul
RT: 21.222 min Scan# 3086
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:228 Resp: 2669140
Ion Ratio Lower Upper
228 100
229 19.8 15.0 22.4
226 26.8 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 24.487 ng/ul
RT: 21.163 min Scan# 3073
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:149 Resp: 1444578
Ion Ratio Lower Upper
149 100
167 26.6 22.6 33.8
279 3.6 4.2 6.4#





#87

Chrysene

Concen: 27.097 ng/ul

RT: 21.275 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Instrument :

BNA_P

ClientSampleId :

SLCS967





Tgt Ion:228 Resp: 2540095

Ion Ratio Lower Upper

228 100

226 30.1 23.6 35.4

229 19.9 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 24.591 ng/ul




RT: 22.080 min Scan# 3229


Delta R.T. 0.000 min

Lab File: BP010927.D

Acq: 06 Jul 2022 13:14

Tgt Ion:149 Resp: 2537603









#90
Benzo(b)fluoranthene
Concen: 27.826 ng/ul
RT: 22.880 min Scan# 31
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:252 Resp: 3008260
Ion Ratio Lower Upper
252 100
253 21.3 17.8 26.6
125 12.7 6.2 9.2#



#91
Benzo(k)fluoranthene
Concen: 27.482 ng/ul
RT: 22.927 min Scan# 3373
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:252 Resp: 2824104
Ion Ratio Lower Upper
252 100
253 22.0 17.0 25.4
125 12.5 6.2 9.4#









#93
Benzo(a)pyrene
Concen: 25.240 ng/ul
RT: 23.492 min Scan# 34
Delta R.T. 0.000 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14




Instrument :
BNA_P
ClientSampleId :
SLCS967


Tgt Ion:252 Resp: 2654491
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 14.4 6.6 10.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 28.102 ng/ul
RT: 26.045 min Scan# 3903
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:276 Resp: 3566226
Ion Ratio Lower Upper
276 100
138 28.3 14.6 22.0#
277 24.7 19.6 29.4









#95
Dibenzo(a,h)anthracene
Concen: 27.392 ng/ul
RT: 26.068 min Scan# 3915
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Instrument :
BNA_P
ClientSampleId :
SLCS967

Tgt Ion:278 Resp: 2981754
Ion Ratio Lower Upper
278 100
139 19.9 9.4 14.2#
279 23.7 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 27.164 ng/ul
RT: 26.798 min Scan# 4031
Delta R.T. 0.006 min
Lab File: BP010927.D
Acq: 06 Jul 2022 13:14

Tgt Ion:276 Resp: 2932258
Ion Ratio Lower Upper
276 100
138 27.1 12.7 19.1#
277 24.0 18.2 27.4

