

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006127.D
 Acq On : 07 Jul 2021 11:20
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: Jul 07 13:49:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 13:48:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	165797	20.000	ng	0.00
21) Naphthalene-d8	10.681	136	661855	20.000	ng	0.00
39) Acenaphthene-d10	14.516	164	386359	20.000	ng	0.00
64) Phenanthrene-d10	17.263	188	738885	20.000	ng	# 0.00
76) Chrysene-d12	21.345	240	682395	20.000	ng	# 0.00
86) Perylene-d12	23.751	264	826745	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.464	112	99805	9.659	ng	0.00
7) Phenol-d6	7.069	99	126939	9.478	ng	0.00
23) Nitrobenzene-d5	9.052	82	97834	7.247	ng	0.00
42) 2,4,6-Tribromophenol	16.016	330	37556	5.664	ng	0.00
45) 2-Fluorobiphenyl	13.140	172	260576	8.856	ng	0.00
79) Terphenyl-d14	19.822	244	337081	9.249	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.323	88	20224	4.332	ng	# 86
3) Pyridine	3.787	79	42376	3.862	ng	# 78
4) n-Nitrosodimethylamine	3.664	42	18465	3.621	ng	# 95
6) Aniline	7.217	93	66751	4.464	ng	96
8) 2-Chlorophenol	7.452	128	56699	4.786	ng	98
9) Benzaldehyde	7.034	77	38008	6.657	ng	97
10) Phenol	7.099	94	58926	4.405	ng	99
11) bis(2-Chloroethyl)ether	7.305	93	44645	4.404	ng	96
12) 1,3-Dichlorobenzene	7.764	146	63589	4.800	ng	96
13) 1,4-Dichlorobenzene	7.917	146	66067	4.890	ng	99
14) 1,2-Dichlorobenzene	8.228	146	60041	4.650	ng	95
15) Benzyl Alcohol	8.152	79	40413	4.006	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.416	45	58285	6.219	ng	93
17) 2-Methylphenol	8.346	107	38515	4.225	ng	95
18) Hexachloroethane	8.952	117	21104	4.319	ng	91
19) n-Nitroso-di-n-propyla...	8.699	70	34195	4.150	ng	97
20) 3+4-Methylphenols	8.693	107	51091	4.149	ng	96
22) Acetophenone	8.716	105	71523	4.106	ng	# 94
24) Nitrobenzene	9.093	77	50648	3.613	ng	99
25) Isophorone	9.622	82	93853	3.765	ng	# 94
26) 2-Nitrophenol	9.811	139	25243	3.774	ng	99
27) 2,4-Dimethylphenol	9.887	122	39101	3.924	ng	94
28) bis(2-Chloroethoxy)met...	10.111	93	51295	3.942	ng	97
29) 2,4-Dichlorophenol	10.387	162	40578	3.389	ng	98
30) 1,2,4-Trichlorobenzene	10.552	180	48039	3.570	ng	99
31) Naphthalene	10.734	128	172673	4.509	ng	98
32) Benzoic acid	10.075	122	12481	1.952	ng	# 83
33) 4-Chloroaniline	10.875	127	58190	3.761	ng	97
34) Hexachlorobutadiene	11.010	225	26602	2.923	ng	98
35) Caprolactam	11.687	113	11494	3.332	ng	89
36) 4-Chloro-3-methylphenol	12.016	107	45643	3.754	ng	92
37) 2-Methylnaphthalene	12.346	142	116811	4.284	ng	95
38) 1-Methylnaphthalene	12.563	142	109774	4.274	ng	97
40) 1,2,4,5-Tetrachloroben...	12.722	216	44937	3.304	ng	96
41) Hexachlorocyclopentadiene	12.787	237	35	0.006	ng	83
43) 2,4,6-Trichlorophenol	12.975	196	30709	3.292	ng	94

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006127.D
 Acq On : 07 Jul 2021 11:20
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: Jul 07 13:49:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 13:48:57 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.069	196	31220	3.108	ng	# 96
46) 1,1'-Biphenyl	13.357	154	143558	4.608	ng	99
47) 2-Chloronaphthalene	13.399	162	107270	4.217	ng	97
48) 2-Nitroaniline	13.628	65	22870	3.419	ng	91
49) Acenaphthylene	14.240	152	173096	4.435	ng	100
50) Dimethylphthalate	13.981	163	138125	4.348	ng	99
51) 2,6-Dinitrotoluene	14.104	165	28699	3.975	ng	96
52) Acenaphthene	14.581	154	108519	4.579	ng	98
53) 3-Nitroaniline	14.457	138	24561	3.510	ng	96
54) 2,4-Dinitrophenol	14.734	184	1984	0.539	ng	90
55) Dibenzofuran	14.916	168	169433	4.347	ng	98
56) 4-Nitrophenol	14.863	139	3091	0.545	ng	84
57) 2,4-Dinitrotoluene	14.910	165	34559	3.579	ng	# 70
58) Fluorene	15.563	166	134061	4.415	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.163	232	24742	2.795	ng	# 70
60) Diethylphthalate	15.334	149	139218	4.368	ng	97
61) 4-Chlorophenyl-phenyle...	15.557	204	56634	3.597	ng	# 88
62) 4-Nitroaniline	15.628	138	21139	2.935	ng	# 84
63) Azobenzene	15.851	77	115232	3.996	ng	98
65) 4,6-Dinitro-2-methylph...	15.693	198	11713	2.157	ng	86
66) n-Nitrosodiphenylamine	15.775	169	112665	4.598	ng	98
67) 4-Bromophenyl-phenylether	16.457	248	32825	3.461	ng	# 86
68) Hexachlorobenzene	16.575	284	39794	3.501	ng	# 89
69) Atrazine	16.734	200	34987	4.590	ng	94
70) Pentachlorophenol	16.945	266	11477	1.816	ng	95
71) Phenanthrene	17.310	178	203489	4.586	ng	99
72) Anthracene	17.404	178	190879	4.371	ng	99
73) Carbazole	17.687	167	184802	4.408	ng	98
74) Di-n-butylphthalate	18.222	149	226617	4.665	ng	99
75) Fluoranthene	19.281	202	216424	4.016	ng	94
77) Benzidine	19.475	184	81684	5.753	ng	96
78) Pyrene	19.628	202	228254	4.791	ng	98
80) Butylbenzylphthalate	20.492	149	110875	5.657	ng	98
81) Benzo(a)anthracene	21.333	228	209555	4.254	ng	98
82) 3,3'-Dichlorobenzidine	21.275	252	57769	3.392	ng	# 94
83) Chrysene	21.386	228	212995	4.464	ng	99
84) Bis(2-ethylhexyl)phtha...	21.245	149	168960	5.878	ng	# 97
85) Di-n-octyl phthalate	22.163	149	298018	5.796	ng	# 95
87) Indeno(1,2,3-cd)pyrene	26.268	276	284902	4.030	ng	# 87
88) Benzo(b)fluoranthene	23.016	252	230985	4.079	ng	# 95
89) Benzo(k)fluoranthene	23.063	252	234515	4.271	ng	# 94
90) Benzo(a)pyrene	23.639	252	225991	4.245	ng	# 93
91) Dibenzo(a,h)anthracene	26.280	278	241404	3.967	ng	# 91
92) Benzo(g,h,i)perylene	27.051	276	246353	4.098	ng	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006127.D
 Acq On : 07 Jul 2021 11:20
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: Jul 07 13:49:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 13:48:57 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.875 min Scan# 814
 Delta R.T. 0.000 min
 Lab File: BP006127.D
 Acq: 07 Jul 2021 11:20




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:152 Resp: 165797
 Ion Ratio Lower Upper
 152 100
 150 158.1 125.6 188.4
 115 59.9 48.5 72.7



#2
 1,4-Dioxane
 Concen: 4.332 ng
 RT: 3.323 min Scan# 40
 Delta R.T. 0.006 min
 Lab File: BP006127.D
 Acq: 07 Jul 2021 11:20

Tgt Ion: 88 Resp: 20224
 Ion Ratio Lower Upper
 88 100
 58 71.7 46.7 70.1#
 43 28.9 20.3 30.5









#3
Pyridine
Concen: 3.862 ng
RT: 3.787 min Scan# 119
Delta R.T. 0.047 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 79 Resp: 42376
Ion Ratio Lower Upper
79 100
52 65.3 41.7 62.5#
51 44.4 24.1 36.1#



#4
n-Nitrosodimethylamine
Concen: 3.621 ng
RT: 3.664 min Scan# 98
Delta R.T. 0.018 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 42 Resp: 18465
Ion Ratio Lower Upper
42 100
74 135.7 105.4 158.0
44 18.4 4.2 6.4#









#5
2-Fluorophenol
Concen: 9.659 ng
RT: 5.464 min Scan# 404
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 112 Resp: 99805
Ion Ratio Lower Upper
112 100
64 49.7 41.4 62.2
63 25.0 23.8 35.8



#6
Aniline
Concen: 4.464 ng
RT: 7.217 min Scan# 702
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 93 Resp: 66751
Ion Ratio Lower Upper
93 100
66 34.1 29.7 44.5
65 19.5 15.3 22.9









#7
Phenol-d6
Concen: 9.478 ng
RT: 7.069 min Scan# 677
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 99 Resp: 126939
Ion Ratio Lower Upper
99 100
42 16.0 13.4 20.0
71 31.7 30.6 46.0



#8
2-Chlorophenol
Concen: 4.786 ng
RT: 7.452 min Scan# 742
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 128 Resp: 56699
Ion Ratio Lower Upper
128 100
130 33.2 12.4 52.4
64 38.6 20.3 60.3









#9
Benzaldehyde
Concen: 6.657 ng
RT: 7.034 min Scan# 671
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 77 Resp: 38008
Ion Ratio Lower Upper
77 100
105 99.7 76.7 116.7
106 94.2 76.2 116.2



#10
Phenol
Concen: 4.405 ng
RT: 7.099 min Scan# 682
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 94 Resp: 58926
Ion Ratio Lower Upper
94 100
65 28.1 8.8 48.8
66 41.5 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 4.404 ng
RT: 7.305 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 93 Resp: 44645
Ion Ratio Lower Upper
93 100
63 70.4 46.2 86.2
95 31.1 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 4.800 ng
RT: 7.764 min Scan# 795
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 146 Resp: 63589
Ion Ratio Lower Upper
146 100
148 65.2 51.5 77.3
75 25.9 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 4.890 ng
RT: 7.917 min Scan# 821
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:146 Resp: 66067
Ion Ratio Lower Upper
146 100
148 63.1 51.3 76.9
111 40.7 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 4.650 ng
RT: 8.228 min Scan# 874
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:146 Resp: 60041
Ion Ratio Lower Upper
146 100
148 66.6 50.3 75.5
111 45.9 33.7 50.5









#15
Benzyl Alcohol
Concen: 4.006 ng
RT: 8.152 min Scan# 861
Delta R.T. 0.018 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 79	Resp: 40413
Ion Ratio	Lower Upper
79 100	
108 82.6	62.5 93.7
77 65.8	52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 6.219 ng
RT: 8.416 min Scan# 906
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 45	Resp: 58285
Ion Ratio	Lower Upper
45 100	
77 18.4	2.0 42.0
79 14.1	0.0 36.7









#17
2-Methylphenol
Concen: 4.225 ng
RT: 8.346 min Scan# 894
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	107	Resp:	38515
Ion	Ratio	Lower	Upper
107	100		
108	110.2	87.7	131.5
77	42.2	40.2	60.2
79	42.1	39.0	58.4



#18
Hexachloroethane
Concen: 4.319 ng
RT: 8.952 min Scan# 997
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	117	Resp:	21104
Ion	Ratio	Lower	Upper
117	100		
119	99.2	76.2	114.2
201	90.4	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 4.150 ng
RT: 8.699 min Scan# 954
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	70	Resp:	34195
Ion	Ratio	Lower	Upper
70	100		
42	52.8	40.8	61.2
101	9.9	8.8	13.2
130	24.7	18.9	28.3



#20
3+4-Methylphenols
Concen: 4.149 ng
RT: 8.693 min Scan# 953
Delta R.T. 0.024 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	107	Resp:	51091
Ion	Ratio	Lower	Upper
107	100		
108	91.5	71.0	111.0
77	45.0	16.6	56.6
79	32.7	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.681 min Scan# 1291
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	136	Resp:	661855
Ion Ratio	Lower	Upper	
136	100		
137	10.3	8.6	12.8
54	6.7	4.6	6.8
68	4.7	3.6	5.4



#22
Acetophenone
Concen: 4.106 ng
RT: 8.716 min Scan# 957
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	105	Resp:	71523
Ion Ratio	Lower	Upper	
105	100		
71	2.4	1.9	2.9
51	27.8	18.4	27.6#
120	23.4	17.9	26.9









#23
Nitrobenzene-d5
Concen: 7.247 ng
RT: 9.052 min Scan# 1014
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 82 Resp: 97834
Ion Ratio Lower Upper
82 100
128 47.3 35.7 53.5
54 47.9 35.0 52.4



#24
Nitrobenzene
Concen: 3.613 ng
RT: 9.093 min Scan# 1021
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 77 Resp: 50648
Ion Ratio Lower Upper
77 100
123 48.6 38.2 57.2
65 14.1 11.4 17.2









#25
Isophorone
Concen: 3.765 ng
RT: 9.622 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	82	Resp:	93853
Ion Ratio	Lower	Upper	
82	100		
95	7.3	6.5	9.7
138	22.4	15.0	22.4#



#26
2-Nitrophenol
Concen: 3.774 ng
RT: 9.811 min Scan# 1143
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	139	Resp:	25243
Ion Ratio	Lower	Upper	
139	100		
109	34.8	28.8	43.2
65	50.7	41.0	61.6









#27
2,4-Dimethylphenol
Concen: 3.924 ng
RT: 9.887 min Scan# 1156
Delta R.T. 0.018 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:122 Resp: 39101
Ion Ratio Lower Upper
122 100
107 110.1 94.5 141.7
121 55.9 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 3.942 ng
RT: 10.111 min Scan# 1194
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 93 Resp: 51295
Ion Ratio Lower Upper
93 100
95 30.4 25.9 38.9
123 13.5 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 3.389 ng
RT: 10.387 min Scan# 1241
Delta R.T. 0.041 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:162 Resp: 40578
Ion Ratio Lower Upper
162 100
164 66.4 44.9 84.9
98 33.6 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 3.570 ng
RT: 10.552 min Scan# 1269
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:180 Resp: 48039
Ion Ratio Lower Upper
180 100
182 97.0 77.9 116.9
145 33.6 24.7 37.1









#31
Naphthalene
Concen: 4.509 ng
RT: 10.734 min Scan# 1300
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	128	Resp:	172673
Ion	Ratio	Lower	Upper
128	100		
129	10.8	9.0	13.6
127	12.3	10.7	16.1



#32
Benzoic acid
Concen: 1.952 ng
RT: 10.075 min Scan# 1188
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	122	Resp:	12481
Ion	Ratio	Lower	Upper
122	100		
105	99.8	103.2	143.2#
77	84.9	76.2	116.2









#33
4-Chloroaniline
Concen: 3.761 ng
RT: 10.875 min Scan# 1324
Delta R.T. 0.018 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	127	Resp:	58190
Ion	Ratio	Lower	Upper
127	100		
129	30.5	26.1	39.1
65	28.2	22.5	33.7
92	17.0	15.0	22.6



#34
Hexachlorobutadiene
Concen: 2.923 ng
RT: 11.010 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	225	Resp:	26602
Ion	Ratio	Lower	Upper
225	100		
223	62.0	51.0	76.6
227	62.8	51.4	77.2









#35
Caprolactam
Concen: 3.332 ng
RT: 11.687 min Scan# 1462
Delta R.T. 0.018 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	113	Resp:	11494
Ion	Ratio	Lower	Upper
113	100		
55	134.9	98.9	138.9
56	98.3	71.7	111.7



#36
4-Chloro-3-methylphenol
Concen: 3.754 ng
RT: 12.016 min Scan# 1518
Delta R.T. 0.024 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	107	Resp:	45643
Ion	Ratio	Lower	Upper
107	100		
144	24.9	21.1	31.7
142	73.7	66.0	99.0









#37
2-Methylnaphthalene
Concen: 4.284 ng
RT: 12.346 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:142 Resp: 116811
Ion Ratio Lower Upper
142 100
141 87.1 71.8 107.6
115 33.4 31.6 47.4



#38
1-Methylnaphthalene
Concen: 4.274 ng
RT: 12.563 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:142 Resp: 109774
Ion Ratio Lower Upper
142 100
141 94.9 73.3 109.9
116 3.8 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.516 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:164 Resp: 386359
Ion Ratio Lower Upper
164 100
162 104.8 80.2 120.2
160 46.5 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 3.304 ng
RT: 12.722 min Scan# 1638
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:216 Resp: 44937
Ion Ratio Lower Upper
216 100
214 79.8 61.9 92.9
179 24.3 16.8 25.2
108 21.7 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 0.006 ng
RT: 12.787 min Scan# 1649
Delta R.T. 0.106 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	237	Resp:	35
Ion	Ratio	Lower	Upper
237	100		
235	53.6	44.5	84.5
272	0.0	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 5.664 ng
RT: 16.016 min Scan# 2198
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	330	Resp:	37556
Ion	Ratio	Lower	Upper
330	100		
332	99.1	77.8	116.8
141	44.1	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 3.292 ng
RT: 12.975 min Scan# 1681
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:196 Resp: 30709
Ion Ratio Lower Upper
196 100
198 93.0 79.5 119.3
200 27.9 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 3.108 ng
RT: 13.069 min Scan# 1697
Delta R.T. 0.024 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:196 Resp: 31220
Ion Ratio Lower Upper
196 100
198 93.8 74.7 112.1
97 50.7 37.4 56.0
132 32.9 20.0 30.0#









#45
2-Fluorobiphenyl
Concen: 8.856 ng
RT: 13.140 min Scan# 1709
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	172	Resp:	260576
Ion	Ratio	Lower	Upper
172	100		
171	35.4	27.9	41.9
170	22.8	18.6	27.8



#46
1,1'-Biphenyl
Concen: 4.608 ng
RT: 13.357 min Scan# 1746
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	154	Resp:	143558
Ion	Ratio	Lower	Upper
154	100		
153	41.0	20.4	60.4
76	11.7	0.0	30.8









#47
2-Chloronaphthalene
Concen: 4.217 ng
RT: 13.399 min Scan# 1753
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:162 Resp: 107270
Ion Ratio Lower Upper
162 100
127 42.5 31.4 47.0
164 33.4 26.6 39.8



#48
2-Nitroaniline
Concen: 3.419 ng
RT: 13.628 min Scan# 1792
Delta R.T. 0.018 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion: 65 Resp: 22870
Ion Ratio Lower Upper
65 100
92 78.8 57.7 86.5
138 126.0 92.1 138.1









#49
Acenaphthylene
Concen: 4.435 ng
RT: 14.240 min Scan# 1896
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:152 Resp: 173096
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2
153 13.2 10.6 16.0



#50
Dimethylphthalate
Concen: 4.348 ng
RT: 13.981 min Scan# 1852
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:163 Resp: 138125
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.2
164 10.8 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 3.975 ng
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	165	Resp:	28699
Ion	Ratio	Lower	Upper
165	100		
63	56.5	41.3	61.9
89	56.3	44.8	67.2



#52
Acenaphthene
Concen: 4.579 ng
RT: 14.581 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	154	Resp:	108519
Ion	Ratio	Lower	Upper
154	100		
153	110.7	89.4	134.2
152	51.0	43.0	64.4









#53
3-Nitroaniline
Concen: 3.510 ng
RT: 14.457 min Scan# 1933
Delta R.T. 0.018 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:138 Resp: 24561
Ion Ratio Lower Upper
138 100
108 9.4 8.4 12.6
92 114.3 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 0.539 ng
RT: 14.734 min Scan# 1980
Delta R.T. 0.071 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:184 Resp: 1984
Ion Ratio Lower Upper
184 100
63 69.3 47.0 70.6
154 70.0 52.3 78.5









#55
Dibenzofuran
Concen: 4.347 ng
RT: 14.916 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:168 Resp: 169433
Ion Ratio Lower Upper
168 100
139 40.5 31.0 46.6
169 13.2 10.8 16.2



#56
4-Nitrophenol
Concen: 0.545 ng
RT: 14.863 min Scan# 2002
Delta R.T. 0.083 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:139 Resp: 3091
Ion Ratio Lower Upper
139 100
109 77.7 72.3 112.3
65 77.2 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 3.579 ng
RT: 14.910 min Scan# 2010
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:165 Resp: 34559
Ion Ratio Lower Upper
165 100
63 65.2 32.3 48.5#
89 90.2 55.9 83.9#
182 3.5 2.2 3.4#



#58
Fluorene
Concen: 4.415 ng
RT: 15.563 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:166 Resp: 134061
Ion Ratio Lower Upper
166 100
165 99.7 78.0 117.0
167 13.7 10.5 15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 2.795 ng
RT: 15.163 min Scan# 2053
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	232	Resp:	24742
Ion Ratio	Lower	Upper	
232	100		
131	60.9	32.2	48.4#
130	3.5	2.0	3.0#
166	38.6	19.4	29.2#



#60
Diethylphthalate
Concen: 4.368 ng
RT: 15.334 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:	149	Resp:	139218
Ion Ratio	Lower	Upper	
149	100		
177	20.3	17.5	26.3
150	13.2	9.9	14.9









#61
4-Chlorophenyl-phenylether
Concen: 3.597 ng
RT: 15.557 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:204 Resp: 56634
Ion Ratio Lower Upper
204 100
206 32.7 26.6 40.0
141 74.8 48.6 73.0#



#62
4-Nitroaniline
Concen: 2.935 ng
RT: 15.628 min Scan# 2132
Delta R.T. 0.024 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:138 Resp: 21139
Ion Ratio Lower Upper
138 100
92 53.0 31.2 71.2
108 76.3 78.6 118.6#









#63
Azobenzene
Concen: 3.996 ng
RT: 15.851 min Scan# 2170
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion	Ratio	Lower	Upper
77	100		
182	25.7	6.0	46.0
105	21.8	0.0	38.8
51	28.8	8.8	48.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.263 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion	Ratio	Lower	Upper
188	100		
94	10.3	5.6	8.4#
80	10.6	6.3	9.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 2.157 ng
RT: 15.693 min Scan# 2143
Delta R.T. 0.024 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:198 Resp: 11713
Ion Ratio Lower Upper
198 100
51 43.7 14.0 54.0
105 46.4 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 4.598 ng
RT: 15.775 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:169 Resp: 112665
Ion Ratio Lower Upper
169 100
168 69.8 54.4 81.6
167 36.7 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 3.461 ng
RT: 16.457 min Scan# 2273
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:248 Resp: 32825
Ion Ratio Lower Upper
248 100
250 105.1 79.8 119.8
141 94.6 57.8 86.6#



#68
Hexachlorobenzene
Concen: 3.501 ng
RT: 16.575 min Scan# 2293
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:284 Resp: 39794
Ion Ratio Lower Upper
284 100
142 44.5 26.9 40.3#
249 32.1 24.5 36.7





#69

Atrazine

Concen: 4.590 ng

RT: 16.734 min Scan# 2320

Delta R.T. 0.000 min

Lab File: BP006127.D

Acq: 07 Jul 2021 11:20





Instrument :

BNA_P

ClientSampleId :

SSTDIC005

Tgt Ion:	200	Resp:	34987
Ion	Ratio	Lower	Upper
200	100		
173	27.6	4.1	44.1
215	44.3	27.7	67.7



#70

Pentachlorophenol

Concen: 1.816 ng




RT: 16.945 min Scan# 2356


Delta R.T. 0.018 min

Lab File: BP006127.D

Acq: 07 Jul 2021 11:20

Tgt Ion:	266	Resp:	11477
Ion	Ratio	Lower	Upper
266	100		
268	58.0	51.6	77.4
264	63.8	50.3	75.5









#71
Phenanthrene
Concen: 4.586 ng
RT: 17.310 min Scan# 2418
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:178 Resp: 203489
Ion Ratio Lower Upper
178 100
176 19.4 15.1 22.7
179 15.5 12.4 18.6



#72
Anthracene
Concen: 4.371 ng
RT: 17.404 min Scan# 2434
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:178 Resp: 190879
Ion Ratio Lower Upper
178 100
176 18.3 14.9 22.3
179 15.9 12.2 18.2









#73
 Carbazole
 Concen: 4.408 ng
 RT: 17.687 min Scan# 2482
 Delta R.T. 0.012 min
 Lab File: BP006127.D
 Acq: 07 Jul 2021 11:20




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:167 Resp: 184802
 Ion Ratio Lower Upper
 167 100
 166 22.5 17.1 25.7
 139 14.2 10.7 16.1



#74
 Di-n-butylphthalate
 Concen: 4.665 ng
 RT: 18.222 min Scan# 2573
 Delta R.T. 0.006 min
 Lab File: BP006127.D
 Acq: 07 Jul 2021 11:20

Tgt Ion:149 Resp: 226617
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 4.8 4.5 6.7









#75
Fluoranthene
Concen: 4.016 ng
RT: 19.281 min Scan# 2753
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:202 Resp: 216424
Ion Ratio Lower Upper
202 100
101 12.1 0.0 27.6
203 16.4 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.345 min Scan# 3104
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:240 Resp: 682395
Ion Ratio Lower Upper
240 100
120 15.2 6.5 9.7#
236 25.6 20.6 31.0





#77

Benzidine

Concen: 5.753 ng

RT: 19.475 min Scan# 2786

Delta R.T. 0.018 min

Lab File: BP006127.D

Acq: 07 Jul 2021 11:20





Instrument :

BNA_P

ClientSampleId :

SSTDICC005

Tgt Ion:184	Resp:	81684
Ion Ratio	Lower	Upper
184	100	
185	15.8	11.0 16.4
183	12.9	9.5 14.3



#78

Pyrene

Concen: 4.791 ng




RT: 19.628 min Scan# 2812


Delta R.T. 0.000 min

Lab File: BP006127.D

Acq: 07 Jul 2021 11:20

Tgt Ion:202	Resp:	228254
Ion Ratio	Lower	Upper
202	100	
200	21.0	16.3 24.5
203	16.9	14.2 21.2









#79
Terphenyl-d14
Concen: 9.249 ng
RT: 19.822 min Scan# 2845
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:244 Resp: 337081
Ion Ratio Lower Upper
244 100
212 7.2 6.0 9.0
122 12.8 7.1 10.7#



#80
Butylbenzylphthalate
Concen: 5.657 ng
RT: 20.492 min Scan# 2959
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:149 Resp: 110875
Ion Ratio Lower Upper
149 100
91 67.5 54.6 81.8
206 17.5 16.7 25.1





#81

Benzo(a)anthracene

Concen: 4.254 ng

RT: 21.333 min Scan# 3102

Delta R.T. 0.006 min

Lab File: BP006127.D

Acq: 07 Jul 2021 11:20





Tgt Ion:228 Resp: 209555

Ion Ratio Lower Upper

228 100

226 28.3 21.6 32.4

229 19.8 15.8 23.6



#82

3,3'-Dichlorobenzidine

Concen: 3.392 ng

RT: 21.275 min Scan# 3092

Delta R.T. 0.012 min

Lab File: BP006127.D

Acq: 07 Jul 2021 11:20




Tgt Ion:252 Resp: 57769


Ion Ratio Lower Upper

252 100

254 67.1 50.6 76.0

126 13.5 7.0 10.6#









#83
Chrysene
Concen: 4.464 ng
RT: 21.386 min Scan# 3111
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:228 Resp: 212995
Ion Ratio Lower Upper
228 100
226 29.9 23.8 35.6
229 20.1 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 5.878 ng
RT: 21.245 min Scan# 3087
Delta R.T. 0.000 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:149 Resp: 168960
Ion Ratio Lower Upper
149 100
167 26.3 21.9 32.9
279 3.5 4.3 6.5#









#85
Di-n-octyl phthalate
Concen: 5.796 ng
RT: 22.163 min Scan# 3243
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:149 Resp: 298018
Ion Ratio Lower Upper
149 100
167 1.7 1.4 2.0
43 9.0 5.6 8.4#



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.751 min Scan# 3513
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:264 Resp: 826745
Ion Ratio Lower Upper
264 100
260 23.9 18.6 28.0
265 21.9 17.6 26.4





#87

Indeno(1,2,3-cd)pyrene

Concen: 4.030 ng

RT: 26.268 min Scan# 3941

Delta R.T. 0.006 min

Lab File: BP006127.D

Acq: 07 Jul 2021 11:20

Instrument :

BNA_P

ClientSampleId :

SSTDIC005





Tgt Ion:276 Resp: 284902

Ion Ratio Lower Upper

276 100

138 32.6 15.3 22.9#

277 25.2 20.4 30.6



#88

Benzo(b)fluoranthene

Concen: 4.079 ng

RT: 23.016 min Scan# 3388

Delta R.T. 0.012 min

Lab File: BP006127.D

Acq: 07 Jul 2021 11:20




Tgt Ion:252 Resp: 230985


Ion Ratio Lower Upper

252 100

253 22.3 17.5 26.3

125 13.9 6.2 9.4#









#89
Benzo(k)fluoranthene
Concen: 4.271 ng
RT: 23.063 min Scan# 3396
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:252 Resp: 234515
Ion Ratio Lower Upper
252 100
253 22.5 17.6 26.4
125 14.4 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 4.245 ng
RT: 23.639 min Scan# 3494
Delta R.T. 0.006 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:252 Resp: 225991
Ion Ratio Lower Upper
252 100
253 22.7 17.9 26.9
125 16.3 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 3.967 ng
RT: 26.280 min Scan# 3943
Delta R.T. 0.012 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:278 Resp: 241404
Ion Ratio Lower Upper
278 100
139 21.0 9.3 13.9#
279 24.3 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 4.098 ng
RT: 27.051 min Scan# 4074
Delta R.T. 0.018 min
Lab File: BP006127.D
Acq: 07 Jul 2021 11:20

Tgt Ion:276 Resp: 246353
Ion Ratio Lower Upper
276 100
277 23.2 18.9 28.3
138 28.4 12.6 18.8#

