


Benzoic acid

Response Ratio




$$\text{Response} = 2.331\text{e-}001 * \text{Amt} - 8.044\text{e-}002$$

Coef of Det (r^2) = 0.996292 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070721.M


Calibration Table Last Updated: Wed Jul 07 16:08:13 2021

Hexachlorocyclopentadiene



$R = 2.212e-002 A^2 + 4.534e-002 A - 2.390e-002$
Coef of Det (r^2) = 0.998761 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070721.M
Calibration Table Last Updated: Wed Jul 07 16:08:13 2021


2,4-Dinitrophenol



$\text{Response} = 1.916\text{e-}001 * \text{Amt} - 8.507\text{e-}002$
 Coef of Det (r^2) = 0.995154 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070721.M
 Calibration Table Last Updated: Wed Jul 07 16:08:13 2021

4-Nitrophenol

Response Ratio



$$\text{Response} = 2.779\text{e-}001 * \text{Amt} - 9.900\text{e-}002$$


Coef of Det (r^2) = 0.996525 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070721.M

Calibration Table Last Updated: Wed Jul 07 16:08:13 2021

4-Nitroaniline

Response Ratio




$$\text{Response} = 3.576\text{e-}001 * \text{Amt} - 5.290\text{e-}002$$

Coef of Det (r^2) = 0.999099 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070721.M

Calibration Table Last Updated: Wed Jul 07 16:08:13 2021

4,6-Dinitro-2-methylphenol



$$\text{Response} = 1.438\text{e-}001 * \text{Amt} - 3.178\text{e-}002$$

Coef of Det (r^2) = 0.999485 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070721.M

Calibration Table Last Updated: Wed Jul 07 16:08:13 2021