

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006134.D
 Acq On : 07 Jul 2021 16:00
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP070721

Quant Time: Jul 07 16:33:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 16:08:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	85	0.00
2	1,4-Dioxane	40.000	42.655	-6.6	93	0.00
3	Pyridine	40.000	45.028	-12.6	94	0.00
4	n-Nitrosodimethylamine	40.000	44.316	-10.8	97	0.00
5 S	2-Fluorophenol	80.000	84.900	-6.1	90	-0.01
6	Aniline	40.000	41.793	-4.5	89	0.00
7 S	Phenol-d6	80.000	83.441	-4.3	88	0.00
8	2-Chlorophenol	40.000	41.110	-2.8	87	0.00
9	Benzaldehyde	40.000	41.925	-4.8	89	0.00
10 C	Phenol	40.000	42.117	-5.3	90	-0.01
11	bis(2-Chloroethyl)ether	40.000	41.818	-4.5	88	0.00
12	1,3-Dichlorobenzene	40.000	40.362	-0.9	86	0.00
13 C	1,4-Dichlorobenzene	40.000	39.688	0.8	84	0.00
14	1,2-Dichlorobenzene	40.000	39.642	0.9	84	0.00
15	Benzyl Alcohol	40.000	43.187	-8.0	90	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.757	-1.9	89	0.00
17	2-Methylphenol	40.000	40.845	-2.1	86	0.00
18	Hexachloroethane	40.000	38.911	2.7	83	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.648	-1.6	87	0.00
20	3+4-Methylphenols	40.000	40.182	-0.5	84	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	82	0.00
22	Acetophenone	40.000	41.765	-4.4	85	0.00
23 S	Nitrobenzene-d5	80.000	84.277	-5.3	86	0.00
24	Nitrobenzene	40.000	42.312	-5.8	87	0.00
25	Isophorone	40.000	41.978	-4.9	85	0.00
26 C	2-Nitrophenol	40.000	41.532	-3.8	80	0.00
27	2,4-Dimethylphenol	40.000	41.511	-3.8	82	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.218	-3.0	83	0.00
29 C	2,4-Dichlorophenol	40.000	41.226	-3.1	81	0.00
30	1,2,4-Trichlorobenzene	40.000	39.777	0.6	79	0.00
31	Naphthalene	40.000	40.398	-1.0	83	0.00
32	Benzoic acid	40.000	40.838	-2.1	87	0.00
33	4-Chloroaniline	40.000	41.680	-4.2	82	0.00
34 C	Hexachlorobutadiene	40.000	39.363	1.6	77	0.00
35	Caprolactam	40.000	43.271	-8.2	85	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.332	-5.8	85	0.00
37	2-Methylnaphthalene	40.000	40.649	-1.6	82	0.00
38	1-Methylnaphthalene	40.000	40.759	-1.9	82	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	82	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.643	0.9	78	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.560	-1.4	84	0.00
42 S	2,4,6-Tribromophenol	80.000	76.570	4.3	72	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.420	-3.6	82	0.00
44	2,4,5-Trichlorophenol	40.000	41.752	-4.4	82	0.00
45 S	2-Fluorobiphenyl	80.000	80.772	-1.0	82	0.00
46	1,1'-Biphenyl	40.000	40.834	-2.1	84	0.00
47	2-Chloronaphthalene	40.000	41.048	-2.6	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
 Data File : BP006134.D
 Acq On : 07 Jul 2021 16:00
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP070721

Quant Time: Jul 07 16:33:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 07 16:08:13 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.044	-10.1	89	0.00
49	Acenaphthylene	40.000	41.054	-2.6	83	0.00
50	Dimethylphthalate	40.000	40.878	-2.2	83	0.00
51	2,6-Dinitrotoluene	40.000	42.068	-5.2	84	0.00
52 C	Acenaphthene	40.000	40.284	-0.7	83	0.00
53	3-Nitroaniline	40.000	43.243	-8.1	85	0.00
54 P	2,4-Dinitrophenol	40.000	40.263	-0.7	87	0.00
55	Dibenzofuran	40.000	40.524	-1.3	83	0.00
56 P	4-Nitrophenol	40.000	40.786	-2.0	89	0.00
57	2,4-Dinitrotoluene	40.000	42.406	-6.0	83	0.00
58	Fluorene	40.000	39.973	0.1	82	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.839	-2.1	79	0.00
60	Diethylphthalate	40.000	40.366	-0.9	83	0.00
61	4-Chlorophenyl-phenylether	40.000	39.354	1.6	79	0.00
62	4-Nitroaniline	40.000	39.972	0.1	84	0.00
63	Azobenzene	40.000	41.663	-4.2	87	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	79	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.738	-1.8	81	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.135	-2.8	80	0.00
67	4-Bromophenyl-phenylether	40.000	39.185	2.0	74	0.00
68	Hexachlorobenzene	40.000	38.648	3.4	72	0.00
69	Atrazine	40.000	41.288	-3.2	80	0.00
70 C	Pentachlorophenol	40.000	40.739	-1.8	76	0.00
71	Phenanthrene	40.000	40.455	-1.1	80	0.00
72	Anthracene	40.000	40.577	-1.4	79	0.00
73	Carbazole	40.000	41.342	-3.4	81	0.00
74	Di-n-butylphthalate	40.000	41.931	-4.8	82	0.00
75 C	Fluoranthene	40.000	39.983	0.0	78	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	77	0.00
77	Benzydine	40.000	45.535	-13.8	81	0.00
78	Pyrene	40.000	40.440	-1.1	77	0.00
79 S	Terphenyl-d14	80.000	80.913	-1.1	75	0.01
80	Butylbenzylphthalate	40.000	42.326	-5.8	82	0.01
81	Benzo(a)anthracene	40.000	40.765	-1.9	78	0.01
82	3,3'-Dichlorobenzidine	40.000	40.528	-1.3	75	0.00
83	Chrysene	40.000	40.608	-1.5	78	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	42.806	-7.0	85	0.00
85 c	Di-n-octyl phthalate	40.000	43.470	-8.7	86	0.01
86 I	Perylene-d12	20.000	20.000	0.0	83	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.749	-1.9	82	0.02
88	Benzo(b)fluoranthene	40.000	40.662	-1.7	80	0.02
89	Benzo(k)fluoranthene	40.000	40.701	-1.8	85	0.01
90 C	Benzo(a)pyrene	40.000	40.528	-1.3	83	0.02
91	Dibenzo(a,h)anthracene	40.000	40.877	-2.2	82	0.03
92	Benzo(g,h,i)perylene	40.000	40.533	-1.3	83	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070721\
Data File : BP006134.D
Acq On : 07 Jul 2021 16:00
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP070721

Quant Time: Jul 07 16:33:02 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 07 16:08:13 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0